

Electronic Peak Flow Meter

(Product Model: DL-DF01)

Instruction For Use

Effective Date: 2022-03-20

Doc. No.: [FDA-003](#)

Version: A/0



Content

Package Contents.....	1
1 Overview	2
2 Product composition	4
3 How to use	7
4 Cleaning and disinfection.....	14
5 Maintenance	15
6 Common troubleshooting.....	16
7 Product features	17
8 Technical Specifications	17
9 Symbol used in labeling	18
10 Guidance and manufacturer's declaration.....	19
11 Information of Manufacturer	26



Package Contents

No.	Item	Quantity
1	Main unit	1
2	Turbine	1
3	Blowing Mouthpiece	1
4	AAA Battery(optional)	2
5	Instruction for use	1
6	Certificate of conformity	1



Instruction For Use

- ◆ Please read this manual carefully and follow the instructions before you use.
- ◆ Please keep this manual properly so that you can refer to it when you need it in the future.
- ◆◆ We recommend user to use the product exclusively and avoid sharing with others.

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

This device complies with Industry Canada license-exempt RSS standard(s). Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device.

CONVENTIONS USED IN THIS MANUAL

	Indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.
	Indicates a hazardous situation which, if not avoided, may result in minor or moderate injury.
NOTICE	Indicates an operating tip, a maintenance suggestion, or a hazard that may result in device damage

1 Overview

1.1 Working principle:

The device is an impeller type flowmeter. When the gas flows through the pipe of the flowmeter, a spiral airflow is generated, which pushes the blades to rotate. During this period, the light beam emitted by the infrared generator will be interrupted by the rotating blade. The pulse signal formed by each interruption will be converted into an electrical signal, and be output and transmitted to the microprocessor. The number of revolutions can therefore be calculated, and the gas flow and volume can also be calculated based on it.



1.2 Function introduction:

The electronic peak flow meter is a portable device that can perform lung function tests. It can be used to measure the maximum expiratory flow rate (PEF).

The maximum expiratory flow rate (PEF) refers to the highest flow rate during forced exhalation, also known as the highest expiratory flow rate, the maximum expiratory flow rate, and the best expiratory flow rate.

PEF can better reflect the patency of the airway, help diagnose and evaluate asthma, and diagnose and classify the degree of airway obstruction.

1.3 Intended use:

This device is intended for monitoring PEF (Peak Expiratory Flow Rate) for patient home use. The device is designed for adults and children over 5 years of age with caregiver supervision.

1.4 Contraindications:

There are no known Contraindications.

1.5 Safety Precautions



:

- Please stop using this meter immediately if it is damaged. Do Not touch any internal parts to avoid electric shock.
- In order to avoid electric shock or detector failure, do not allow any liquid to permeate into the meter.
- Do Not use the meter in the presence of flammable materials, solutions or gases, or in an oxygen enriched environment.
- Any unauthorized personnel should not tamper with the device.
- Only original accessories as specified by the manufacturer can be used with the device.
- In the event of an accident of any kind arising from use of the device, you are strongly recommended to inform the manufacturer or your doctor so that he/she can notify the authorities as required by local legislation.
- Any Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the



equipment.

- This equipment should be installed and operated with a minimum distance of 0mm between the radiator and your extremity, 60mm for your head.



:

- The patient is an intended operator. The patient can use and maintain the meter and its accessories according to this manual.
- Put the meter into the specified storage case during transportation.
- Keep the device out of the reach of children/pets to avoid inhalation or swallowing of small parts.
- The supervision of an adult is required for monitoring elderly subjects, children over the age of five and disabled persons.
- Periodically check that no impurities or foreign bodies, such as skin, hairs have deposited inside the turbine. This may cause errors in measurement or compromise the correct functioning of the device.
- The device is not designed to be used in direct air currents (e.g. wind), sources of heat or cold, direct sun rays or other sources of light or energy, dust, sand or chemical substances.
- Use and store the device in compliance with the environmental conditions specified in the Technical Specifications. If the device is subjected to environmental conditions other than those specified, it may malfunction and/or display incorrect results.
- At least 30min required for ME equipment to warm from the minimum storage temperature to the measuring environment temperature until it is ready for intended use.
- At least 30min required for ME equipment to cool from the maximum storage temperature to the measuring environment temperature until it is ready for intended use.

2 Product composition

Electronic Peak Flow Meter consists of main unit, turbine and blowing mouthpiece. Product structure diagram is shown below.







Figure 1. Side view of product




Figure 2. Front view of product






Figure 3. Screen interface picture

Materials of components is shown below.

Component name	Material
Housing	Acrylonitrile Butadiene Styrene (ABS)
Button	SEBS Thermoplastic Elastomer Compounding Rubber (SEBS TPR)
Screen panel	Polycarbonate (PC)
Turbine	Polycarbonate (PC)
	Polycarbonate+Acrylonitrile Butadiene Styrene (PC+ABS)
Blowing mouthpiece	Polypropylene (PP)

Meaning of display icon in the screen is shown below.



	Two-line displays the number ● Setting phase: the first line does not display, the second line displays weight, age, height, time (year, month, day, hour, minute) ● Testing phase: the first line shows the PEF value, the second line shows the PEF%
	History record icon
	Sound Icon
	Battery icon
	Blowing preparation prompt
	PEF Value

♀	Male icon
♂	Female icon
kg	Weight icon
Height icon	Height icon
▶	Indicator to traffic colour outside surface of main unit, the position of arrow means the degree of PEF normal or declined.
kg	Kilogram (kg) icon
%	PEF%
Age	Age icon
Set	Setting icon


3 How to use

3.1. Installation

- Open the battery cover and insert two AAA batteries into the battery compartment.
- Install the turbine into the main unit, clamp the groove of the turbine to the slot in the main unit, and tighten it to the right according to the direction shown in the figure on the right.




- Insert the mouthpiece into the turbine according to the direction of arrow marked on the mouthpiece, and press it gently to ensure that there is no air leakage at the connection.
 - Remove the accessories and batteries: remove it
- NOTICE:**
- Install the batteries correctly according to the sign terminals in the battery compartment.



3.2. Boot

- Long press the “ON/OFF” button for 2 seconds until you hear the “DI” sound, prompting the device has been turned on.



- When using for the first time, the screen interface displays the “SET”. Follow the interface prompts to enter age, gender, height, weight and other information in sequence, and then follow the interface prompts to set the year, month, date, hour, and minute in sequence.
- Short press the “ON/OFF” button to adjust the setting item, and short press “RIGHT” or “LEFT” button can increase or decrease the number.
- After the setting of basic information is completed, click the “ON/OFF”

button. You can see the “” icon is flashing, and when you hear the continuous “DIDIDI” sound, indicating that the device is preparation for measurement. You can Long press the “LEFT” button for 2 seconds to reset the basic information.




Prompt:

1. After the setting of basic information is completed, the device will calculate an estimated value based on your input parameters, which is used to calculate PEF%.
2. The predicted value is not displayed, but PEF% will be displayed after each measurement is completed.

3.3. Testing

- Deeply inhale and use your lips to wrap the mouthpiece firmly, and make sure that there is no air leakage between the mouth and the mouthpiece. At the same time, you exhale the gas forcefully in the shortest time (one second or more).
- When you hear the "DI" sound, it indicates that the test has been completed. The screen interface displays the PEF test value and its percentage of the predicted value. The icon "▶" displayed on the right can indicate the area (green, yellow or red) that represents the test result.

**NOTICE:**

- Green area: indicates that your PEF is above 80% of the predicted value and is within the normal range, which means that your condition is under control. Please follow the doctor's recommendation to decide whether you require the medication or do not require a higher dose of medication.
- Yellow area: indicates that your PEF is in the range of 60-80% of the predicted value, which means that you must pay attention to it. Please follow the doctor's recommendation to take appropriate treatment measures, such as increasing the dose of medication.
- Red area: indicates that your PEF is below 60% of the predicted value, which means that your situation is dangerous. Please follow the doctor's



recommendation to take appropriate treatment measures, such as increasing the dose of medication or seeking emergency treatment.



:


- Please clean mouthpiece according to the Manual Chapter 4 to prevent from dust because of transport or storage when you use the device for the first time.
- Do not block the air outlet with your tongue or teeth, otherwise the measurement result will be affected.
- Mouthpiece must be used by one person, don't cross-use.

3.4 Remeasure

- Short press the “ON/OFF” button. When you hear the continuous “DIDIDI” sound, repeat step 3.3 to repeat the measurement testing.
- The device will automatically store the highest PEF value measured as the final measurement result. Performing the measurement testing 3 times is recommended. Doctors generally recommend taking one measurement in the morning and one in the evening every day.

3.5 History Record

- Click the “LEFT” or “RIGHT” button to check the corresponding history measurement record, and the corresponding history record number and test time will be displayed at the same time.



- After selecting the required history record number, the upper line displays the history record number, the lower line displays the corresponding test



time (display in order: month, day, hour, minute). Waiting for about 2 seconds, the interface will automatically display the PEF value in the corresponding history record number (upper line) and PEF% (lower line).

The device can store 500 sets of data. When the memory data exceeds 500 sets of data, the device will automatically overwrite the earlier data in the order of the memory data.

3.6 Data Transmission

The device can also be connected to communication devices (such as the phone) through Bluetooth. The patient's PEF and PEFR data changes within one week can be monitored by APP, so that it is more convenient to monitor your respiratory health status.

3.6.1 Register and log in APP


- Open the APP and follow the prompts to register and log in to the APP.
- Fill in personal information on the APP, such as name, gender, age, height, weight and so on.

NOTICE:


- Please fill in your information truthfully, do not fill in false information. The APP will calculate your PEF% according to the information you fill in, and the measurement results will guide doctors to use medicine.

3.6.2 Intelligent detecion

1) On the initial interface of the APP, select the Bluetooth symbol "📶" in the upper right corner to enter the Bluetooth connection interface.



2) Click "Bond" and then "This APP needs to use Bluetooth" going to pop up. Click "Allow", the APP will search automatically. When "Connected to: JDY-08" is displayed, the connection is successful. Bluetooth symbol "📶" will then appear on the screen of the device.



3) The test can be performed according to 3.3 test steps for measurement, and the measured values will be uploaded to the APP synchronously.

4) Click "Got it" to enter the test interface.

5) Repeat the measurements as per steps 1-4.


NOTICE:

- The measurement should be taken three times in the morning and three in the afternoon. If the measurement is less than three times, the symbol "⚠️" on test interface is red, prompting you to continue to complete the measurement. After the measurement has been completed for three times, it turns blue to indicate that the measurement has been completed.

3.6.4 Report

- Click "📄+" in the test interface and select "Data Report" to view the data graph of PEF value and PEFr within a week.





- Click "E+" in the test interface and select "Personal information" to modify your name, gender, age, height, weight, etc.

3.6 Shutdown

- Manual shutdown: long press the "ON/OFF" button for 2 seconds until you hear the "DIDI" sound and the screen interface displays "OFF" flashing twice, and the device will shut down.
- Automatic shutdown: If there is no any operation for 3 minutes in the standby state of the device, you will hear the "DIDI" sound and the screen interface will display "OFF" flashing twice, the device will automatically shut down.




4 Cleaning and disinfection

4.1 Cleaning of the turbine

NOTICE:

- Do not immerse the turbine in hot or boiling water.
- To avoid irreparable damage to the turbine, never try to clean the turbine under a direct jet of water or other liquids.



Correct operation of the turbine is guaranteed only if it is "clean" and free of foreign objects that affect its movement. The presence of dust or foreign bodies (such as hairs, sputum etc.) could slow or block the moving parts of the turbine and make the result less accurate, or damage the turbine itself. After each use, check the cleanliness of the turbine.

Weekly cleaning is recommended. To clean the turbine, remove it from the main unit and separate it from the mouthpiece. Immerse the turbine in warmy soapy water and agitate the turbine for 2-3 minutes. Rinse in purified water and shake gently to remove any excess water. If the device is not visually clean, the user should either repeat the relevant previous cleaning steps untill it is clean. Then wipe it dry with a clean soft cloth or allow to air dry on a towel. Finally store it in a clean, ventilated, and dry place.

4.2 Cleaning and disinfection of the mouthpiece

Make sure to clean and disinfect the mouthpiece after each use.

To clean the mouthpiece, simply pull it apart from the turbine. Immerse the mouthpiece in warmy soapy water and agitate the mouthpiece for 2-3 minutes. Rinse in purified water and shake gently to remove any excess water. If the device is not visually clean, the user should either repeat the relevant previous



cleaning steps until it is clean. Then wipe it dry with a clean soft cloth.

To disinfect the mouthpiece, wipe the clean mouthpiece with an alcohol cotton pad or a cotton cloth moistened with 75% alcohol for 2 minutes. Allow to air dry on a towel. Finally store it in a clean, ventilated and dry place.

4.3 Cleaning of the main unit

Weekly cleaning is recommended. To clean, wipe the device's surfaces with a soft damp cloth. Dry with a soft cloth, or allow to air dry. Ensure that all surfaces are completely dry. Finally store it in a clean, ventilated, and dry place.

NOTICE:

- Do not clean the main unit in water or other liquids.

5 Maintenance

5.1 Battery

When the battery is too low, the “” symbol will flash 2 times, and the “OFF” will flash 2 times immediately accompanied by "DIDI" sound. The device will shut down at this time, and you should replace the battery with a new one in time.

NOTICE:

- Please take out batteries from battery case to avoid battery leakage if you do not use the device for a long time.
- Please select batteries with regular factory instead of fake and shoddy batteries
- If the battery volume is not sufficient, the accuracy of parameter of PEF may be affected

5.2 Repair

The repair of the device is limited to qualified maintenance personnel designated by the manufacturer. Users are not allowed to disassemble and repair it. If necessary, the relevant technical documents such as circuit diagrams and component lists required for maintenance can be provided.

Important:

Changes or modifications to this unit not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.



5.3 Storage

Please store the electronic peak flow meter in a clean and dry place under the storage environment condition as specified in the manual. Do not place it under the sun or extreme high or low temperature to avoid severe impact. Otherwise, these may cause the device to fail to work normally or even be damaged.

5.4 Waste disposal

The meter, accessories and consumables are to be disposed of according to local law or regulation after they have exceeded their service life. Alternatively, they can be returned to the local distributor or the manufacturer for recycling or proper disposal.

For the detailed disposal information, consult the manufacturer or the local distributor.

6 Common troubleshooting

Abnormal phenomenon	Possible reason	Resolution method
Screen shows nothing	The battery level is too low or the battery is installed incorrectly.	Replace the battery with a new one and pay attention to the direction in which the positive and negative electrodes are installed.
Screen shows "Err 1"	1. No blowing in 30 seconds 2. System failure	1. Follow operation step 3.3 again to perform the testing 2. System failure, please contact the manufacturer or dealer
Screen shows "Err 2"	Out of measurement range (PEF Value < 50L/min)	Follow operation step 3.3 again to perform the testing



Screen shows “Err 3”	Out of measurement range (PEF Value > 999L/min)	Remeasure or replace other equipment to measure
The display suddenly disappears.	The battery is loose.	Reinstall the battery

NOTICE:

- If you take the above measures, the electronic peak flow meter still can not be used normally, it is recommended to contact the manufacturer or dealer.

7 Product features

Classified by type of electric shock	Internal power supply
Classified by the degree of protection against electric shock	Type BF application part
Degree of IP protection	IP22
Classification of safety levels when used in the presence of flammable anesthetic gases mixed with flammable anesthetic gases or nitrous oxide mixed with air	Non-AP/APG Device
Classified by operating mode	Continuous
Classification of sterilization method	Non-Sterile Medical Device

8 Technical Specifications

Product name	Electronic peak flow meter
Model	DL-DF01
Measurement principle	Infrared interruption
Wight	108.0±1.0g
Size	61.49mm×88.86mm×126.70mm (H×W×D)
Power supply	d.c.3.0V (2pcs AAA LR03 batteries)
Measurement range	50~999L/min
Accurancy	±10L/minor ±10%, whichever is greater



Resistance to flow	<0.35kPa/L/s
Measuring resolution	1L/min
BTPS	PEF results displayed at BTPS conditions
Expected Service Life	3 years
Operating environment condition	Ambient temperature range: 5°C~40°C Relative humidity range: 15~90%RH Atmospheric pressure range: 70KPa~106KPa
Storage/transport environment condition	Ambient temperature range: -25°C~55°C Relative humidity range: 15~93%RH Atmospheric pressure range: 86KPa ~ 106KPa (under uncondensed state)

9 Symbol used in labeling

Symbol	Meaning
	Manufacturer
	Date of manufacture
	Use-By-Date
	Type BF applied part
	Batch code
	Refer to instruction manual/ booklet
	Caution



	This symbol indicates that waste electrical and electronic equipment must not be disposed of as unsorted municipal waste and must be collected separately. Please contact an authorized representative of the manufacturer for information concerning the decommissioning of your equipment.
IP22	Degree of IP protection
	Temperature limitation
	Humidity limitation
	Atmospheric pressure limitation
	RF

10 Guidance and manufacturer's declaration

Electronic Peak Flow Meter complies with IEC 60601-1-2:2020 on electromagnetic compatibility (EMC for medical devices) for both immunity and emissions.

10.1 Electromagnetic emission

1	Guidance and manufacturer's declaration – electromagnetic emission		
2	The model DL-DF01 is intended for use in the electromagnetic environment specified below. The customer or the user of the model DL-DF01 should assure that it is used in such an environment.		
3	Emissions test	Compliance	Electromagnetic environment – guidance
4	RF emissions CISPR 11	Group 1	The Model DL-DF01 uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. The Model DL-DF01 is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
5	RF emissions CISPR 11	Class B	
6	Harmonic emissions	Not applicable	



	IEC 61000-3-2		
7	Voltage fluctuations / flicker emissions IEC 61000-3-3	Not applicable	

NOTICE:

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation.

10.2 Electromagnetic immunity

Guidance and manufacturer's declaration – electromagnetic immunity			
The Model DL-DF01 are intended for use in the electromagnetic environment specified below. The customer or the user of the Model DL-DF01 should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Electrostatic discharge (ESD) IEC 61000-4-2	± 8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	± 8 kV contact ±2 kV, ±4 kV, ±8 kV, ±15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrostatic transient / burst IEC 61000-4-4	± 2 kV for power supply lines 100 kHz repetition frequency ± 1 kV for input/output lines	N/A	N/A
Surge IEC 61000-4-5	± 0.5 kV, ± 1 kV differential mode line-line	N/A	N/A
Voltage dips, short	0 % UT (100 % dip in UT)	N/A	N/A



interruptions and voltage variations on power supply input lines IEC 61000-4-11	for 0.5 cycle at 0° , 45° , 90° , 135° , 180° , 225° , 270° , and 315° 0 % UT (100 % dip in UT) for 1 cycle at 0° 70 % UT (30 % dip in UT) for 25/30 cycles at 0° 0 % UT (100 % dip in UT) for 250/300 cycle at 0°		
Power frequency (50/60 Hz) magnetic field IEC 61000-4-8	30 A/m, 50/60Hz	30 A/m, 50/60Hz	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE: UT is the a. c. mains voltage prior to application of the test level.			

Guidance and manufacturer's declaration – electromagnetic immunity			
The DL-DF01 is intended for use in the electromagnetic environment specified below. The customer or the user of the DL-DF01 should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance



Conducted RF IEC 61000-4-6	3 Vrms 150 kHz to 80 MHz 6 Vrms150 kHz to 80 MHz outside ISM bands ^a	N/A	Portable and mobile RF communications equipment should be used no closer to any part of the Models DL-DF01, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Radiated RF IEC 61000-4-3	10 V/m 80 MHz to 2.7 GHz	10 V/m	Recommended separation distance $d = \left[\frac{3.5}{V_1} \right] \sqrt{P}$ $d = \left[\frac{3.5}{E_1} \right] \sqrt{P} \quad 80\text{MHz to } 800\text{MHz}$ $d = \left[\frac{7}{E_1} \right] \sqrt{P} \quad 800\text{MHz to } 2.7\text{GHz}$ where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres(m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,^a should be less than the compliance level in each frequency range^b Interference may occur in the



			vicinity of equipment marked with the following symbol: 
NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies. NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			



- a The ISM (industrial, scientific and medical) bands between 0,15 MHz and 80 MHz are 6,765 MHz to 6,795 MHz; 13,553 MHz to 13,567 MHz; 26,957 MHz to 27,283 MHz; and 40,66 MHz to 40,70 MHz. The amateur radio bands between 0,15 MHz and 80 MHz are 1,8 MHz to 2,0 MHz, 3,5 MHz to 4,0 MHz, 5,3 MHz to 5,4 MHz, 7 MHz to 7,3 MHz, 10,1 MHz to 10,15 MHz, 14 MHz to 14,2 MHz, 18,07 MHz to 18,17 MHz, 21,0 MHz to 21,4 MHz, 24,89 MHz to 24,99 MHz, 28,0 MHz to 29,7 MHz and 50,0 MHz to 54,0 MHz.
- b The compliance levels in the ISM frequency bands between 150 kHz and 80 MHz and in the frequency range 80 MHz to 2,7 GHz are intended to decrease the likelihood that mobile/portable communications equipment could cause interference if it is inadvertently brought into patient areas. For this reason, an additional factor of 10/3 has been incorporated into the formulae used in calculating the recommended separation distance for transmitters in these frequency ranges.
- c Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the DL-DF01 is used exceeds the applicable RF compliance level above, the DL-DF01 should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the DL-DF01.
- d Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

10.3 Recommended separation distances

Recommended separation distances between portable and mobile RF communications equipment and the model DL-DF01

The Model DL-DF01 is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Model DL-DF01 can help prevent electromagnetic interference by



maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Model DL-DF01 as recommended below, according to the maximum output power of the communications equipment.

Rated maximum output of transmitter W	Separation distance according to frequency of transmitter m		
	150 kHz to 80 MHz $d = [\frac{3.5}{V_1}] \sqrt{P}$	80 MHz to 800 MHz $d = [\frac{3.5}{E_1}] \sqrt{P}$	800 MHz to 2.7 GHz $d = [\frac{7}{E_1}] \sqrt{P}$
0.01	0.12	0.04	0.07
0.1	0.37	0.12	0.23
1	1.17	0.35	0.7
10	3.7	1.11	2.22
100	11.7	3.5	7.0

For transmitters rated at a maximum output power not listed above the recommended separation distance d in metres (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.



:

- Do not use the meter near other devices (computers, cordless phones, cell phones, etc.) that generate strong magnetic fields. Keep these devices at a minimum distance of 30 centimeters. If it is necessary to use it at shorter distances, the meter and the other devices must be kept under observation to verify that they work normally.
- Due to the increasing number of electronic devices (computers, cordless phones, cell phones, etc.) medical devices may be susceptible to electromagnetic interference from other equipment. Such electromagnetic interference could cause the medical device to malfunction, as a



measurement accuracy lower than stated, and create a potentially unsafe situation.

- The use of accessories other than those specified, may result in increased emissions or decreased immunity of the equipment or system
- To maintain compliance with the RF exposure requirement, a separation distance of 20 cm between the device and the human should be maintained.

11 Information of Manufacturer

	Name: Taian Dalu Medical Instrument Co., Ltd. Address: West Part of Yitianmen Street, Hi-tech Zone, Taian 271000, Shandong, P.R.China Tel: +86 538-8931968 Fax: +86 538-8931969
---	---

