




POTMASTER

USER MANUAL

Multi Channel Systems MCS GmbH	Phone	+49 7121 90925 26
Aspenhastrasse 21	Fax	+49 7121 90925 11
D-72770 Reutlingen	Website	www.heka.com / www.elproscan.com
Germany	Email	eimeaorderscmt@harvardbioscience.com support@heka.com

Harvard Bioscience, Inc.	Phone	+1 800 597 0580
84 October Hill Road	Fax	+1 508 429 5732
01746 Holliston	Website	www.heka.com / www.elproscan.com
Massachusetts	Email	us-sales@smart-ephys.com support@heka.com
United States		





Table of Contents

Table of Contents	x
List of Figures	xi
List of Tables	xiii
1 Introduction	1
1.1 Disclaimer	1
1.2 Scope of the Program	1
1.3 Supported System Software	1
1.4 Naming Conventions	1
1.4.1 Windows Versions	1
1.4.2 Syntax	2
1.5 Windows and Mac Key Conventions	2
1.6 Installing POTMASTER	2
1.7 Starting POTMASTER	3
1.8 Get Online Help on Keys	3
1.9 Get Online Help Controls	3
1.10 Closing POTMASTER	3
1.11 Support Hotline	4
2 User Interface	5
2.1 Dialog Controls	5
2.1.1 Numerical Input	6
2.2 Modifying Dialogs and Controls	6
2.2.1 Dialog Control Window	7
2.2.2 Hiding Controls	9
2.2.3 Deleting Controls	10
2.3 Saving Dialogs and Controls	10
2.4 Toggling between Windows	11
2.5 Copy and Paste Functions	12



3	Keys	13
3.1	The Key Listing	13
3.2	The POTMASTER.key File	14
4	Menus	17
4.1	File Menu	17
4.2	Edit Menu	21
4.3	Windows Menu	22
4.4	Replay Menu	23
4.5	Display Menu	29
4.5.1	3D-Mode	30
4.5.2	Trace Properties Dialog	31
4.6	Buffer Menu	33
4.6.1	Handling of Parameters by the Buffer	34
4.7	Notebook Menu	34
4.8	Protocols Menu	35
4.9	PG 300 / PG 600 Menu	36
4.10	Help Menu	39
5	Configuration Window	41
5.1	Individualize POTMASTER	42
5.2	Save	43
5.3	General	43
5.3.1	Max Shown Traces and Values	43
5.3.2	Memory Allocation	43
5.3.3	Minimum Wait Time	44
5.3.4	Batch Communication	44
5.3.5	User Name	45
5.4	Hardware	45
5.4.1	Amplifier and Digitizer Selection	46
5.4.2	Output and Input Channels	47
5.4.3	LockIn	47
5.4.4	Spectroscopy	48
5.4.5	SmartLUX	48
5.4.6	DocuLUX	48
5.4.7	Photometry	48
5.4.8	ElProScan	49
5.4.9	Serial Out	50



5.5	Files	51
5.5.1	General Advice on Naming Folders and Files	52
5.5.2	Miscellaneous Settings	52
5.6	Display	54
5.6.1	Fonts and Colors	54
5.6.2	Notebook, Display, and Analysis	54
5.6.3	Show Options	55
5.7	I/O Control	56
5.7.1	AD/DA Input/Output Scaling	57
5.7.2	List of I/O Parameters	59
5.7.3	Show Digital In / Out	60
5.8	Trace Assign	61
5.9	Miscellaneous	64
6	Oscilloscope Window	67
6.1	Display	67
6.2	Navigation	68
6.3	Measurements on Traces	68
6.4	Display Refreshing	68
6.5	Overlay Options	68
6.6	Display Scaling	69
7	Potentiostat Window	71
7.1	Controlling the PG 310/390 USB - Main Functions	71
7.1.1	Monitor	71
7.1.2	Tuning	73
7.1.3	Filter Settings	73
7.1.4	Show All	75
7.1.5	Charge Controls	77
7.1.6	Current Density Control	78
7.2	Controlling the PG 610/690 USB - Main Functions	79
7.2.1	Monitor	79
7.2.2	Tuning	80
7.2.3	Filter Settings	81
7.2.4	Show All	81
7.2.5	Charge Controls	83
7.2.6	Current Density Control	83
7.3	Controlling the PG 340 USB - Special Functions	85



7.3.1	Electrode Conditions	85
7.3.2	Potential/Current Settings	86
7.3.3	Additional Settings	86
7.3.4	Electrode Mode	86
7.4	Controlling the PG 618 USB - Special Functions	87
7.4.1	Electrode Conditions	87
7.4.2	Settings of Amplifier-1	87
7.4.3	Settings of Amplifier-2	88
7.4.4	Filter Settings of Amplifier-2	88
7.4.5	Additional Features of Amplifier-2	89
8	Control Window	91
8.1	Information about the Experiment	91
8.2	Controlling Data Acquisition	92
8.3	Starting Pulse Generator Sequence	93
8.4	Starting a Protocol	93
8.5	Fixed Control Protocols	93
9	Replay Window	95
9.1	Main window functions	95
9.1.1	Tree Handling	96
10	Pulse Generator Window	99
10.1	Different Views of the Pulse Generator	100
10.2	Sequence Selector	101
10.3	Sequence Pool Handling	101
10.4	Acquisition Modes	102
10.5	Auto Sample	102
10.6	Timing	102
10.7	Check and Execute	103
10.8	Sweep and Channel Length	104
10.9	Wave Parameters	104
10.9.1	Sine Wave Parameters	104
10.9.1.1	Use as Simple Sinewave: Common Frequency	105
10.9.1.2	Use as Simple Sinewave: Separate Frequency	105
10.9.1.3	Use as LockIn Sinewave	106
10.9.2	Square Wave Parameters	107
10.9.2.1	Use as Simple Squarewave: Common Frequency	107
10.9.2.2	Use as Simple Squarewave: Separate Frequency	108



10.9.2.3 Use as LockIn Squarewave	108
10.9.3 Chirp Wave Parameters	109
10.9.3.1 Linear Chirp	110
10.9.3.2 Exponential Chirp	110
10.9.3.3 Spectroscopy Chirp	111
10.9.4 Photometry Wave Parameters	111
10.9.5 Imaging Wave Parameters	112
10.10 Channel Settings for DA Output and AD Input	114
10.10.1 DA output channel settings	114
10.10.2 AD input channel settings	117
10.11 Segments	121
10.11.1 Segment Classes	121
10.11.2 Scan Rates	123
10.11.3 Increment Modes	123
10.11.3.1 Logarithmic increment modes	123
10.12 Miscellaneous	125
10.13 Stimulus Template Preview	126
10.14 E-initial	127
10.15 P/n Leak Correction	127
10.16 PGF Parameters	130
10.17 Error Handling	130
11 Protocol Editor Window	131
11.1 Protocol Handling	131
11.2 Event Handling	132
11.3 Recurring Functions	133
11.4 Events - Overview	134
11.4.1 Protocol Sequence	135
11.4.1.1 BREAK	135
11.4.1.2 Chain Protocol	135
11.4.1.3 Clear Key	135
11.4.1.4 GOTO	136
11.4.1.5 GOTO_MARK	136
11.4.1.6 IF... THEN	136
11.4.1.7 ELSEIF... THEN	138
11.4.1.8 ELSE	138
11.4.1.9 Launch	139
11.4.1.10 Macro Command	139



11.4.1.11 REPEAT	139
11.4.1.12 Switch Window	140
11.4.1.13 WAIT	140
11.4.2 Acquisition	141
11.4.2.1 Acquire Series	141
11.4.2.2 Acquire Each Sweep	141
11.4.2.3 Acquire Properties	142
11.4.2.4 Set Sweep Label	142
11.4.3 Hardware	142
11.4.3.1 Amplifier	142
11.4.3.2 Serial Output	143
11.4.3.3 Set DAC	143
11.4.3.4 Set Digital Bit	144
11.4.3.5 Set Digital Word	144
11.4.3.6 Set Solution Changer	144
11.4.4 Data/Display	145
11.4.4.1 Analysis	145
11.4.4.2 Digital Filters	145
11.4.4.3 Display Properties	146
11.4.4.4 Export	147
11.4.4.5 File Operation	147
11.4.4.6 Replay	148
11.4.4.7 Trace Buffer	149
11.4.5 Value/Parameter	150
11.4.5.1 PGF Parameters	150
11.4.5.2 Set Solutions	150
11.4.5.3 Set Value	151
11.4.6 Messages	152
11.4.6.1 Annotation	152
11.4.6.2 Beep	153
11.4.6.3 Write Icon Value	153
11.4.7 Extensions	153
11.4.7.1 ElProScan	153
11.4.7.2 Photometry/Imaging	155
11.4.8 Echem-Methods	155
12 Protocol Editor Window - Electrochemical Methods	157
12.1 Cyclic Voltammetry (CV)	161



12.1.1	CV Parameters	161
12.1.2	Acquiring a Cyclic Voltammogram	163
12.1.3	Customizing the Analysis Method	164
12.1.4	Customizing the CV Acquisition Sequence	166
12.2	Electrochemical Impedance Spectroscopy (EIS)	168
12.2.1	Parameters for EIS	168
12.2.2	Acquiring an Electrochemical Impedance Spectrum	169
12.2.3	Customizing the Analysis Method	171
12.3	Open Circuit Potential (OCP)	172
12.3.1	OCP Parameters	172
12.3.2	Acquiring the OCP	172
12.4	Square Wave Voltammetry (SWV)	174
12.4.1	SWV Parameters	174
12.4.2	Acquiring a Square Wave Voltammogram	176
12.5	Optical Spectroscopy	177
13	Analysis Window	179
13.1	Stimulus Control	180
13.2	Analysis Methods	180
13.3	Analysis Functions	181
13.3.1	Analysis Function Settings	181
13.3.2	Analysis Function Types	185
13.3.2.1	Timing	185
13.3.2.2	Stim. Properties	185
13.3.2.3	Measurements	186
13.3.2.4	Lock-In	187
13.3.2.5	Trace Parameters	187
13.3.2.6	Sweep Parameters	187
13.3.2.7	Math	187
13.3.2.8	Trace	188
13.3.2.9	Power Spectra	189
13.3.2.10	Histogram	189
13.3.2.11	ElProScan	189
13.4	Analysis Graphs	190
13.4.1	Scale Axis	191
13.4.2	Graph Entries	194
13.5	Graph Positions	195



14 Parameters Window	197
14.1 Parameter Tabs	198
14.1.1 Root Items	198
14.1.2 Group Items	199
14.1.3 Series Items	200
14.1.4 Sweep Items	201
14.1.5 Trace Items	202
14.1.6 Amplifier Items	204
14.1.7 ElProScan Items	206
14.1.8 Marked Items	208
14.2 Flagging	208
15 Notebook Window	211
16 I/O Control Window	213
16.1 Dig-In / Out	213
16.2 DA / AD-Channels	214
16.3 Values	214
16.4 Serial Output	214
16.5 Solutions	214
16.6 Parameters	214
17 Zero Potential (E-zero)	217
18 Solution Base	221
18.1 Activating the Solution Base	221
18.2 Using Solution Indices	222
18.3 Solution Data Base	222
19 Solution Changer	225
20 Markers	227
20.1 New Marker	227
20.2 Edit Marker	229
21 Calculator Window and Equations	231
21.1 Equation Syntax	232
21.2 Equations - Where to Use	233
21.2.1 Analysis	233
21.2.2 Trace Buffer	235



21.2.3	DA Scaling	235
21.2.4	Values	236
22	Photometry Extension	237
22.1	Photometry Configuration TILL Photonics/Cairn OptoScan	238
22.1.1	Wavelength Calibration	239
22.1.1.1	Calibrating the different TILL monochrometers	240
22.2	Photometry Configuration DG-4/DG-5	240
22.2.0.1	PGF Setting of DG-4/DG-5	241
22.3	Photometry Configuration Lambda-10	242
22.3.0.1	PGF Setting of Lambda-10	242
22.3.1	Lambda 10-3	243
22.4	Photometry Configuration pti DeltaRAM	245
23	SmartLUX (Imaging Extension)	247
24	Software LockIn Extension	249
24.1	LockIn Configuration Window	249
24.2	The LockIn Traces	250
24.3	Capacitance Measurements - Step by Step	252
24.3.1	Activating the LockIn Extension	252
24.3.2	LockIn Configuration	252
24.3.3	Creating a PGF sequence	253
24.3.4	Amplifier Settings	256
24.3.5	Running the PGF sequence	256
24.3.6	Analysis	257
25	Spectroscopy Extension	261
25.1	Introduction	261
25.1.1	Chirp Wave Forms	261
25.1.2	Transfer Function Ratio	262
25.2	Activating the Spectroscopy Extension	262
25.2.1	The Spectroscopy Traces	264
25.3	Setting up a Spectroscopy Acquisition	265
25.3.1	The Chirp Wave Dialog	265
25.3.2	Parametrization of the Chirp Stimuli	265
25.3.3	Adding Chirp Analysis Traces	266
25.3.4	Minimizing Swing-In Effects	266
25.4	Display of Spectra in the Oscilloscope Window	269



26 Appendix I: File Overview	271
26.1 File Types	271
27 Appendix II: Data Format	273
27.1 Data Files	273
27.1.1 Raw Data File	273
27.1.2 Markers File	274
27.1.3 Protocol Methods File	274
27.1.4 Analysis Methods File	274
27.1.5 Stimulation Template File	274
27.1.6 Acquisition Parameters File	275
27.1.7 Solutions File	275
27.1.8 Notebook File	275
27.2 The Tree Format	275
27.3 File Template	277
27.3.1 Filename	277
27.3.2 Data Format of the File Template	277
27.3.2.1 Potentiostatic Stimulation or plain DA output	277



List of Figures

2.1 Icon Configuration Dialog 8

4.1 3D Mode of the Oscilloscope Window 30

4.2 Pseudo Color Data Representation 31

4.3 Trace Properties Dialog 31

26.1 The different file types of PATCHMASTER, POTMASTER AND CHARTMASTER 271





List of Tables

2.1 Numerical Input Options 6

2.2 Icon Configuration Options 7

17.1 pH dependence of the potential of the Relative Hydrogen Electrode at 298 K (1013 hPa) 219

27.1 Markers File Description 274

27.2 Analysis Methods File Description 274

27.3 Stimulus Template File Description 274

27.4 Acquisition Parameters File Description 275

27.5 Solutions File Description 275





1 Introduction

1.1 Disclaimer

This *Product* relies on the tools of Microsoft Windows OS (Windows Vista, Windows 7, Windows 8, Windows 10) or Mac OS (10.6 - 10.14). HEKA is not responsible for: i) the contents of these third party products, ii) any links contained in these third party products, iii) changes or upgrades to these third party products and iv) for any consequential damages resulting from the use of these products.

This *Product* may not be reverse engineered, decompiled or disassembled without the express written consent of HEKA.

In no event shall HEKA be responsible for any incidental, punitive, indirect, or consequential damages whatsoever, (including but not limited to loss of data, privacy of data or other pecuniary loss), arising from or relating to the use, or the inability to use, this *Product* or the provision, or lack of provision, of support services.

In all cases where HEKA is liable, the extend of HEKA's liability shall be limited to the actual cost of the *Product* or to the provision of a replacement version of the *Product*.

1.2 Scope of the Program

Extended analysis functions for POTMASTER data are provided by the FITMASTER software. FITMASTER features analysis and fitting routines designed to fully support the extended POTMASTER data structure. Analysis can be performed on the levels of *Traces*, *Sweeps* and *Series*. There are standard fit functions such as "Polynomial, Exponential, Gaussian, and Boltzmann" tailored functions to fit e.g. whole-cell current traces according to the Hodgkin & Huxley gating formalism, current-voltage relationships and dose-response curves.

1.3 Supported System Software



POTMASTER is supported on all recent MS Windows Versions: Windows Vista, Windows 7, Windows 8, Windows 10.



POTMASTER is supported on Mac OS (10.6 - 10.14). Older Mac OS X versions and Mac OS 9 are not supported.

1.4 Naming Conventions

1.4.1 Windows Versions

Throughout this manual we will address all the supported Microsoft Windows versions as "MS Windows". We will explicitly mention the particular MS Windows versions if required.



1.4.2 Syntax

- **POTMASTER:** Small Capitals denote a HEKA program / hardware or keys on the keyboard.
- **Replay:** Typewriter font denotes menu entries and chapters or filenames
- **Protocol Editor:** Sans serif font denotes window names or dialogs.
- **Sinewave:** Roman font denotes options or buttons
- **Italic:** Denotes general emphasis.
- **Bullet list:** Lists points.
- **Numbered list:** Lists actions to be performed sequentially.



&



or (MS Windows) & (Macintosh) denotes system-specific keys and actions.

1.5 Windows and Mac Key Conventions

In POTMASTER, all key commands are saved in the file `Potmaster.key` and will be read at the program start (see chapter 3 on page 13).

Note that all key commands specified in this manual refer to the default setting in `Potmaster.key`. Since all commands can be customized by the user, the settings in your working version of POTMASTER might differ from these default settings.

Please also be aware that in case the file `Potmaster.key` is not available at the program start, no key commands are available!

The basic conventions for the system keys on MS Windows and Mac are as follows:



- **CTRL:** Stands for the CONTROL key.
- **ALT:** Stands for the ALT key.



- **CTRL:** Stands for the CONTROL key.
- **CMD:** Stands for the COMMAND key (apple or cloverleaf symbol).
- **ALT:** Stands for the OPTION key.

1.6 Installing POTMASTER

Please install the hardware and the software according to the `Installation Guide` which is supplied with your order. It can also be obtained from the HEKA homepage.



1.7 Starting POTMASTER

Upon clicking on POTMASTER the software starts and various controls and windows become available:

- The drop-down menus File, Edit, Windows, Replay, Display, Buffer, Notebook, Protocols, EPC10, Help.
- Frequently used POTMASTER windows such as Amplifier, Control Window, Analysis Window 1, Oscilloscope, Replay.
- A scrolling text window called Notebook.

Most of these windows can be minimized; they can be re-opened by clicking on the window bar or by selecting them in the drop-down menu Windows.

1.8 Get Online Help on Keys

The option Show Keys in the drop-down menu Help (compare chapter 4.10 on page 39) displays the shortcut keys that are assigned to various controls of the active windows.

By selecting List Keys in the drop-down menu Help you can list all keys in the Notebook window. This is useful e.g., if you want to print the complete list.

For further information on the assigned keys, see chapter 3 on page 13.

1.9 Get Online Help Controls

The option Show Tooltips in the drop-down menu Help (chapter 4.10 on page 39) displays a short description of the control under the mouse tip.

1.10 Closing POTMASTER

To exit from POTMASTER, choose Quit from the drop-down menu File or



press CTRL + Q.



press CMD + Q.

You have three possibilities:

- **Save + Exit:** Saves configuration and data files and quits the program.
- **Exit:** Saves data files and quits the program.
- **Cancel:** Exit process will be aborted, you return to the program.

Note: If you changed the protocols, the Pulse Generator pool, or the Analysis, then you will be asked independently if you want to save your changes.

Note: Data files are always saved.



1.11 Support Hotline

If you have any questions, suggestions, or improvements, please contact HEKA's support team. The best way is to send an e-mail to "support@heka.com" specifying as much information as possible:

- Your contact information
- The program name: e.g. PATCHMASTER, POTMASTER or CHARTMASTER software
- The program version number: e.g. v2.65, v2.73.1
- Your operating system and its version: e.g. Mac OS 10.8, Windows 7 64-bit
- Your type of computer: e.g. Intel Core i3 3.2 GHz with 2 GB RAM
- Your acquisition hardware, if applicable: e.g. EPC 10 USB, EPC 800 + LIH 8+8 or PG 340 USB
- The serial number and version of your amplifier, if applicable: e.g. EPC 9 single, version "920552 D"
- The questions, problems, or suggestions you have
- Under which conditions and how often the problem occurs

We will address the problem as soon as possible.

Contact Address:

HEKA Elektronik GmbH
a division of Harvard Bioscience
Aspenhastrasse 21
D-72770 Reutlingen
Germany
Phone: +49 7121 90925 26
Fax: +49 7121 90925 11
E-mail: support@heka.com
Web: www.heka.com
www.elproscan.com



2 User Interface

The following chapter describes the user interface of POTMASTER. It consists of four sections:

Dialog controls explains all types of control controls available in the POTMASTER user interface.

Modifying the Dialogs and Controls describes how size, color, and arrangement can be changed.



Saving modified Dialogs and Controls describes how to save the dialog settings.


Toggling between Windows describes how to switch very fast in between different POTMASTER dialog windows.


2.1 Dialog Controls

Controls are defined as parts of the dialogs that allow user input (data, options or execution of actions).

The general rules for dialog controls in this program are:

-  : Box items with a drop shadow enclose changeable values, either as
 - list item or
 - pop-up menu list or
 - drag item
-  : Rounded rectangles are items that perform some action.
-  : Simple rectangles (without drop shadow) display a measured value.
-  : Plain text is for titles only.

 **Drag:** A number in a box with a drop shadow. The parameter value in a drag item can be changed by clicking on it and dragging the mouse up and down. Alternatively, you can double-click on it, or SHIFT-click, or right-click (MS Windows), and then type in a new value. Terminate input with RETURN or ENTER. Using TAB will cycle through all *Drag* items of the active dialog window.

 **List:** Similar in appearance to a *Drag* item. Clicking on it will pop up a menu list from which one can choose a setting.

 **Edit Text:** A text string in a box with a drop shadow. Clicking on it will allow editing the displayed string.

 **Button:** Rounded corner rectangle. Clicking on it will cause the respective action to occur.

  **Switch:** Rounded corner rectangle. Clicking on it toggles the parameter value. The switch is "On" or activated if the item is highlighted. A switch can optionally also execute some action.

 **Radio Button / Checkbox:** Identical to the standard dialog items. Clicking on them will toggle the respective parameters.



-70 mV **Framed Text / Number / Boolean:** Simple box with optionally some text. The Boolean value is indicated by its color, inactive controls are gray **1.00 | 0.00**.

Enter: Pressing ENTER on the extended keyboard always brings you back to edit the control that was edited last. The feature is very useful when one often edits the same control (e.g., a duration of a specific segment in the Pulse Generator or the *Display Gain* in the Oscilloscope window).

Background Color: The color that appears while the user is dragging or entering a value is set by the *Highlight Color* in the Mac OS control panel.



Be careful if changing – the user will not be able to read the edited number if the highlight color is set to a very dark color. The MS Windows version displays highlighted controls with white text on a black background.

2.1.1 Numerical Input

The numerical values can be entered in scientific notation (e.g., "2.3e-3", "2.3E-3") or in engineering format (e.g., "2.3m"). Numbers outside the range for engineering numbers (see table) are always displayed in scientific notation. The old value is erased as soon as the user starts to type. To preserve the old string, move the LEFT or RIGHT cursor first. To leave the previous value unchanged although a new one has been entered already, just clear the input by pressing ESC, then RETURN or ENTER.

Table 2.1: Numerical Input Options

Name	eng.	sci.
Tera	T	E12
Giga	G	E9
Mega	M	E6
kilo	k	E3
milli	m	E-3
micro	μ / u	E-6
nano	n	E-9
pico	p	E-12
femto	f	E-15

SI Units: POTMASTER expects most units to be SI units, i.e., meters, seconds, amperes, or derived units like Hertz, etc. However, for convenience there are exceptions to that rule. In such cases the item title contains an identifier for what unit is to be used, e.g., "mV" if a voltage is to be entered in millivolts rather than in volts.

String Buffer: Whenever an edit process is finished with RETURN the edited string is entered into a cyclic buffer of edit strings consisting of 10 entries. These strings can be accessed during editing using CURSOR UP and CURSOR DOWN. This feature is useful when identical or similar strings have to be typed into various string items.

2.2 Modifying Dialogs and Controls

Dialog items can be modified in many different ways, e.g. background or item color, text font, position of one item, position of all items in the window.



To modify dialogs and control items in the POTMASTER user interface, you have to select `Enable Icon Configuration` from the `Windows` menu.



Select `Enable Icon Configuration`. Now you can customize the windows:

- To drag and resize an item, right-click on the item and drag.
- To drag a group of items, press `CTRL` and right-click on the group and drag. A group of items is indicated by a grey background field.
- To bring up the `Dialog Control` window, press `CTRL` and left-click on the item. Here you can modify the item settings e.g. such as color, text font, dragging speed (see below).



Select `Enable Icon Configuration`. Now you can customize the windows:

- To drag and resize an item, press `ALT` and left-click on the item and drag. The new item position will be ignored if `ALT` is up when the mouse button is released.
- To drag a group of items, press `CMD` and `ALT` and left-click on the group and drag.
- To bring up the `Dialog Control` window, press `CMD` and click on the item. Here you can modify the item settings e.g. such as color, text font, dragging speed (see below).



Be careful when e.g. changing positions, sizes or colors of buttons. If you deselect the option visible in the `Dialog Control` window the button will disappear. Be reminded that it is not possible to get the button back after that action.

The following table summarizes all actions (`Enable Icon Configuration` has to be selected!):

Table 2.2: Icon Configuration Options

Action	Mac OS	MS Windows
Open a configuration dialog	<code>CMD + click</code>	<code>CTRL + left-click</code>
Move one item	<code>ALT + left-click + drag</code>	<code>ALT + right-click + drag</code>
Move group of items	<code>CMD + ALT + left-click+drag</code>	<code>CTRL + right-click + drag</code>

All windows except the `Configuration` window can be iconized i.e. reduced to a minimal size window. Such a window can easily be expanded to the original size (and shrunk again) by clicking in its zoom box.

2.2.1 Dialog Control Window

In the `Dialog Control` window, you can control the properties of an item.






Figure 2.1: Icon Configuration Dialog

Here you can see the Dialog Control window for a *Drag* button, in this case the amplifier *R-series* button. For other buttons, e.g., with numbers or execution commands, there may be more or less controls available.

Here is the complete list of the controls:

Rectangle: Gives the position of the rectangle in the dialog window (left **l**, bottom **b**) and the size of the item (width, height) in pixel.

Position: Gives the position of the text in the item (left **l**, bottom **b**) in pixel. You can select a position from the list: left/center/right/last. "Last" is useful if you have text that may be longer than the specified button and where the end of the text is more important than the beginning, e.g. a complete file path and name. In this case, only the end of the text will be displayed, cutting off the beginning.

Text: Text label of the item.

Note: The label of a button can only be changed when there is already a name in the text field inserted. Buttons with the default entry "—" will not be effected by a change of the label.

Unit: Unit of numeric values.

inverted: Changes the text color from black to white or vice versa.

Length: Maximum number of characters for this field. If the number is too low no number/value is displayed.

Digits: Number of decimal places.

Font: Select a font type from the list of available fonts on the system.

Size: Change the font size of the text.

Style: Define the style of the text: normal, bold, italic, underline, outline, shadow, condense, extend.

Key: Define the character which is assigned to that item. This enables the item to be executed from the keyboard. When a key is applied to a **list** item, there are two keys assignments. E.g. one for increment and one for decrement.

front only: If marked, the key will only work if this window is in front, i.e. the active one. This prevents key command collision in case you want to use the same key settings in several windows.

any modifier: If marked, any of the modifier keys can be used.

Note: In newer MS Windows versions this function is limited.

executable: The button starts action by clicking. If not executable, the button will stay in its original color and is inactive.

visible: The button is visible.

Note: If not activated, the button will be invisible and cannot be edited anymore! To restore the original setting you have to delete the file `Potmaster.set`.

enabled: The button can be activated or edited. If not enabled, the button will be gray and inactive.

Back Col: Color for a switch button that is *not* active (default: pink) or for any other button.


High Col: Color for a switch button that is active (default: red). For other buttons this feature is disabled.

Format of values:

- integer: Value has to be entered as integer, e.g., "5".
- fixed: Value has to be entered as floating point with maximum as many decimal places as given in **Digits**, e.g., "0.001".
- scientific: Value has to be entered in scientific format, e.g., "e-9" for "nano".
- engineering: Value has to be entered in engineering format, e.g., "n" for "nano".

WMF: Saves the window as Windows Meta File (WMF, MS Windows) or PICT (Mac).

Goto Item: Index number of the button. Please do not change this number, it is for internal handling only.

 In the caption of the dialog, you can see, if the item can be called from a *Macro Command* in a protocol ('macro').

To complete the input, you have the following possibilities:

- Click on Update to see the change in the item of the respective window.
- Click on Cancel to leave the Dialog Control window without changes.
- Click on Done to leave the Dialog Control window and save all changes.

2.2.2 Hiding Controls

If you want to hide controls that are usually not necessary for your tasks, the easiest way to hide them is to pull them to the side of the window.

Proceed as follows:

1. Select `Enable Icon Configuration` from the **Windows** menu.
2. Increase the size of a window. For that pull with the mouse on the lower end of the window.
3. Press CTRL and click with the right mouse button (MS Windows) or while holding the CMD key down click (Mac OS) on a control and drag it to the edge of the window. While moving, you will see a gray rectangle underneath it.
4. Re-size the window again to the original size. The controls are now hidden in the not visible area.
5. Now you can save it using the `Save Front Dialog` function in the **Windows** menu. The window setting will be saved as `*.dia` file.
6. To restore the old default settings, delete the `*.dia` files from the home path as set in the Configuration window.



2.2.3 Deleting Controls

If you want to customize POTMASTER for special purposes, e.g., for not allowing the user access to controls, the secure version is to delete the controls. For this, you have to edit the button properties in the Dialog Control window.

Proceed as follows:

1. Select `Enable Icon Configuration` from the `Windows` menu.
2. Press `CTRL` and click with the left mouse button (MS Windows) or while holding the `CMD` key down click (Mac OS) on a control. The Dialog Control window opens.
3. Deselect `enable` or `executable` to simply deactivate the control. It will still be visible then.
4. Alternatively deselect `visible` - the control disappears completely.
5. Click on `Update` to see the effect, or on `Done` to save the control setting.
6. Now you can save it using the `Save Front Dialog` function in the `Windows` menu. The window setting will be saved as `*.dia` file.

To prevent users from changing your settings, place the `.dia` files into access restricted folders and change the home path in the `Configuration` window accordingly.

To restore the old default settings, delete the `*.dia` files from the home path as set in the `Configuration` window.

2.3 Saving Dialogs and Controls

Position and size of all POTMASTER windows are saved in the configuration file (e.g., `Potmaster.set`).

However, position, size, and state (iconized/not iconized) of each window can also be stored using the menu option `Save Front Dialog` in the `Windows` menu (e.g., `IO_Control.dia`). If a dialog file for a window exists, then the settings from that dialog file overrule the settings from the configuration file.


Note: When installing a new version of POTMASTER these customized dialogs are likely to become incompatible, because additional items may have been introduced in the new version. It is, therefore, best to discard these custom dialogs when upgrading.

Note: When you have saved a wrongly configured window by chance, delete the corresponding file from the POTMASTER directory, the naming convention is `[windowname].dia`, e.g., `Oscilloscope.dia`.



2.4 Toggling between Windows

Hitting the ESC key will close whatever window is in front. Systematically pressing the ESC key will close all windows one at a time (except the Notebook window).



The SPACE key can be used to quickly toggle between windows. The sequence of switching can be set in the *Switch to control*. This control is extremely useful in several respects, firstly, it enables you to switch from one window to another, depending upon what *Switch to* window option is selected.

Secondly, switching between windows can be performed very quickly and easily if you assign a keystroke to perform the function. Simply select the desired *Switch to* window option and then assign a key to this window in the previously described Icon Configuration window. To achieve this press the CTRL and select the *Switch to* button with your mouse. The SPACE key is the default key but this can be changed. Thirdly, having the *Switch to* control, enables window switching to be included within macro recordings in the Protocol Editor.

Note: Since this feature is not needed normally, it is available only if either *Enable Icon Configuration* is selected or you are in the Macro Recording mode. You can then find the control in the invisible window area.

Furthermore, this control is also available in the Protocol Editor, event *Switch Window* where it allows the inclusion of window switching in a protocol sequence.



2.5 Copy and Paste Functions

In this section we give an overview of the Cut and Paste functions within the different POTMASTER windows and dialogs.

While editing an item text in a dialog the Cut and Paste functions will operate on the edited text of the active, highlighted item.

The menu function `Edit → Copy` is enabled for the following windows:

- Notebook Window: Standard copy functionality
- Equation Editor Dialog: Complete equation is copied.
- Calculator Window: Result of the equation if the SHIFT is not pressed and equation plus result if the SHIFT is pressed.
- Amplifier Window: Amplifier state as defined in the Parameter window (Amplifier tab).
- Analysis Window 1 & 2: Respective *Analysis Graphs* are copied
- Analysis Window: The function list of the active *Analysis Method*, equivalent to the *List* function.
- Oscilloscope Window: Oscilloscope graph is copied.
- PGF Editor Window: Text description of selected stimulus is copied. The stimulus cartoon is copied instead of the text listing, if the `Cartoon View` tab is active.
- Protocol Editor Window: Parameters as defined in the active tab are copied.
- Replay Window: Text listing of the data tree is copied.
- Protocol Methods: The protocol method is copied.
- Markers Window: Copies *Series* and *Markers* to the clipboard, equivalent to the function "Write to Notebook".
- In most other functions or when the `OPTION` is down during the Copy function in the above functions the window content is copied to the clipboard as PICT and WMF for Mac OS and MS Windows, respectively. The graph is stored to disk, if the `SHIFT` is hold down while selecting the Copy function. Be aware that *Trace* representations will be a bit map (not vectors) when exported as part of the dialog window.

The menu function `Edit → Paste` is enabled for the following windows:

- Notebook Window: Standard paste functionality
- Equation Editor Dialog: Insert a complete equation
- Calculator Window: Insert an equation
- Protocol Editor Window: Insert a protocol. This includes a method copied in the Protocol Methods window. Events are inserted at the active position, if the pasted text contains events without a leading protocol definition. New protocols are appended to the present protocol pool.

Thus, Copy and Paste operations are target specific. E.g., pasting an equation while a sub-equation is being edited in FITMASTER, will paste the new text into the selected sub-equation string, while pasting when no item is selected, will replace the complete equation.



3 Keys

Controls within windows can be accessed from the keyboard. The key assignments are saved in the `Potmaster.key` file and will be read upon starting the program. All commands can be customized by the user.

Please be aware that in case the `Potmaster.key` file is not available when the program is started, no key commands are assigned!

The option `Show Keys` in the `Help` menu displays the key assignments in the various windows. For further information see chapter 1.8 on page 3.

To list the keys, choose `List Keys` in the `Help` menu. The complete key list is displayed in the `Notebook` window.

To save the keys, choose `Save Keys` in the `Help` menu. The keys are saved in the `Potmaster.key` file. Old keyboard assignments will be automatically saved with an incrementing extension, e.g., `*.k00`, `*.k01`, `*.k02...`

You can freely customize the key commands by:

- changing and saving key assignments via the dialog control (chapter 2.2 on page 6) or by
- directly modifying the `Potmaster.key` file, e.g., in a text editor.

3.1 The Key Listing

All keys can be listed in the `Notebook` by choosing `List Keys` in the `Help` menu. Alternatively open the `Potmaster.key` file with a text editor.

In the following the notation of the key listing is described:

`;` denotes remark – will not be executed

`@` denotes menu command.

B, P, N... denotes the window (B = Protocol Editor, P = Pulse Generator, N = Notebook...)

```
; B -> Protocol Editor
B WindowSwitch          Key SPACE any
```

char denotes a letter – simply press it.

```
R UnmarkIt              Key Char U
```

any denotes that any modifier key may be pressed and the key will function.

```
R WindowSwitch          Key SPACE any
```

back denotes that this key will also work when the window is in the background (reverse to the "front only" function in the Dialog Configuration window).



E HoldDec2

Key CursorLeft option back

shift, option, caps, command, control requires that the respective modifier key must be pressed.

HOME, END denotes the corresponding keys on your keyboard.

R ScrollHome

Key HOME any

NUM denotes that you have to use the numeric keypad for these numbers.

E Macro3

Key Numeric3

3.2 The POTMASTER.key File

MACRO-FILE

900

```
; @ -> Menu Keys
@ File      "New..."           Key "N" [Press "Key" + CTRL]
@ File      "Open Read Only..." Key "O" [Press "Key" + CTRL]
@ File      "Open Modify..."   Key "M" [Press "Key" + CTRL]
@ File      "Update File"        Key "U" [Press "Key" + CTRL]
@ File      "File Status"        Key "I" [Press "Key" + CTRL]
@ File      "Quit"               Key "Q" [Press "Key" + CTRL]

; the standard keys should not be changed:
; @ Edit     "Undo"               Key "Z" [Press "Key" + CTRL]
; @ Edit     "Cut"                Key "X" [Press "Key" + CTRL]
; @ Edit     "Copy"               Key "C" [Press "Key" + CTRL]
; @ Edit     "Paste"              Key "V" [Press "Key" + CTRL]
@ Edit     "Select All"           Key "A" [Press "Key" + CTRL]
@ Edit     "Find..."            Key "F" [Press "Key" + CTRL]
@ Edit     "Find Same"           Key "G" [Press "Key" + CTRL]
@ Edit     "Find Selection..."  Key "H" [Press "Key" + CTRL]
@ Edit     "Replace..."         Key "R" [Press "Key" + CTRL]
@ Edit     "Replace Same"        Key "T" [Press "Key" + CTRL]

@ Windows   "Oscilloscope"       Key F12
@ Windows   "Potentiostat"       Key F11
@ Windows   "Replay"             Key F10
@ Windows   "Protocol Editor"    Key F9
@ Windows   "Pulse Generator"    Key F8
@ Windows   "Analysis"           Key F7
@ Windows   "Notebook"           Key F5
@ Windows   "Close Front Window" Key "W" [Press "Key" + CTRL]

@ Notebook  "Save"               Key "S" [Press "Key" + CTRL]
@ Notebook  "Print..."          Key "P" [Press "Key" + CTRL]
@ Notebook  "Clear"              Key "B" [Press "Key" + CTRL]
@ Notebook  "Zoom In"            Key "K" [Press "Key" + CTRL]
@ Notebook  "Zoom Out"           Key "L" [Press "Key" + CTRL]

; A -> Analysis
A WindowSwitch      Key SPACE any

; B -> Protocol Editor
B WindowSwitch      Key SPACE any
B End               Key END any
B Home              Key HOME any
B DownPage          Key PageUp any
B UpPage            Key PageDown any
B Down              Key CursorUp any
B Up                Key CursorDown any
```



3.2. The POTMASTER.key File

```
; C -> Solution Base
C   Done                Key RETURN any

; D -> Parameters
D   WindowSwitch        Key SPACE any

; E -> Amplifier
E   WindowSwitch        Key SPACE any
E   StopStandby         Key Numeric0 any back
E   IRange              Key CursorDown back
E   IRange              Key2 CursorUp back

; H -> Photometry
H   WindowSwitch        Key SPACE any

; I -> I/O Control
I   WindowSwitch        Key SPACE any

; J -> Calculator

; K -> Markers
K   WindowSwitch        Key SPACE any

; L -> LockIn
L   WindowSwitch        Key SPACE any
L   LockInDone          Key RETURN any

; N -> Control Window
N   WindowSwitch        Key SPACE any
N   TimerSet            Key Char T any back
N   Resume              Key Char R control back
N   Wait                Key Char W control back
N   Next                Key Char N control back
N   Break               Key Char B control back
N   Stop                Key Char S control back
N   PG6                 Key Char 6
N   PG5                 Key Char 5
N   PG4                 Key Char 4
N   PG3                 Key Char 3
N   PG2                 Key Char 2
N   PG1                 Key Char 1

; O -> Oscilloscope
O   WindowSwitch        Key SPACE any
O   YOffsetDec          Key Numeric- shift back
O   YOffsetInc          Key Numeric+ shift back
O   YScaleDec           Key Numeric- back
O   YScaleInc           Key Numeric+ back
O   YCenter             Key Numeric. back
O   ResetY              Key Numeric* back
O   DispTrace           Key Numeric4 back
O   DispTrace           Key2 Numeric5 back
O   Wipe                Key DeleteLeft any back
O   MoveRRight          Key Char > any [for "Scan" function, jumps 10 points]
O   MoveRLeft           Key Char < any [for "Scan" function, jumps 10 points]
O   MoveLRight          Key Char . [for "Scan" function, point by point]
O   MoveLLeft           Key Char , [for "Scan" function, point by point]

; P -> PGF-Editor
P   WindowSwitch        Key SPACE any

; R -> Replay
R   WindowSwitch        Key SPACE any
R   ScrollEnd           Key END any
R   ScrollPageDown      Key PageDown any
```



3.2. The POTMASTER.key File

```
R ScrollDown      Key CursorDown any
R ScrollUp        Key CursorUp  any
R ScrollPageUp    Key PageUp    any
R ScrollHome      Key HOME      any
R ScrollRight     Key CursorRight any
R ScrollLeft      Key CursorLeft any
R UnmarkIt        Key Char U
R MarkIt          Key Char M
R ShowIt          Key RETURN  any

; S -> Configuration
S WindowSwitch    Key SPACE  any

; T -> Trace Properties
T Done            Key RETURN  any

; U -> Online_1
U WindowSwitch    Key SPACE  any

; V -> Online_2
V WindowSwitch    Key SPACE  any

; Z -> Ezero
Z WindowSwitch    Key SPACE  any
Z Cancel          Key ESC    any
Z Done            Key RETURN  any
```



4 Menus

The following section describes the various drop-down menus in POTMASTER.

4.1 File Menu

The **File** menu has all options to handle POTMASTER experiment files. A single POTMASTER "Experiment", that can hold a variable number of single electrochemical experiments, can consist of several types of files:

- The *.dat file contains only the actual raw data without any timing or scaling information (=data).
- The *.mac file contains macros (=Macro). [Disposed since version 2x52]
- The *.mrk file contains the used marker information (=Marker).
- The *.meth file contains the used methods (=Method).
- The *.onl file contains Analysis information (=Online Analysis).
- The *.pgf file contains the used stimulus templates (= Pulse Generator File).
- The *.pro file contains Protocol Editor information (=Protocol Editor).
- The *.pul file contains the complete data tree (=Pulsed Tree).
- If the solution database is used, then a file with the extension *.sol is stored with the data files (=Solution). It contains information on the solutions used in the stored experiments.
- In the *.txt file (=Text), the contents of the Notebook are stored (If Auto Store is activated in the Notebook menu).

If the option *Make Bundle File* is checked in the Configuration window (see chapter 5.5.2 on page 52), then the files *.pul, *.pgf, *.onl, *.mrk, *.meth and *.dat are merged into one single *.dat file.

Raw data acquired by POTMASTER is only written to disk in the so called *Store* mode (i.e. if the *Store* button in the Control window is switched on).

The data are written to disk either:

- upon completion of a *Sweep* or *Series* or
- during the acquisition, in case of a continuous *Sweep*.

The structural information is kept in RAM; they are stored to disk only when:

- a new file (File → New), a new group (File → New Group) or a new experiment (File → New Experiment) is created or
- the Update File function is executed (File → Update File) or
- the program is terminated (File → Quit)



New...: Creates a new, empty data file that is ready for data acquisition. The file has read and write permission.

Open Read Only...: Loads an existing file in *Read Only* mode. Modification of the file is not allowed. Use this option when you want to analyze data and to make sure not to change or delete anything.

Open Modify...: Loads an existing file with read and write permission. Modification of the file such as appending or deleting data is allowed.



Deleting entries in the data file is not reversible. Make sure to always have a backup of the original files when modifying an experiment. The exception is, of course, when you really want to delete a part of the stored data.

Merge...: Appends the data tree from an existing data file to the data tree of the actually opened file.

Update File: Updates the whole experiment to disk. This includes all files involved (see above). If you encounter computer crashes leading to data loss, use this option frequently or enable one of the settings to be found via `File → Write Tree Files`.

Close: Closes the actual file.

File Status: Prints information about the status of the currently opened file such as the path, length, etc. to the Notebook window. A typical output could look as follows:

```
Storing to disk ENABLED,
read-write file: "C:\HEKA\Data\Exp001.dat"
length: 284 kb; free disk space: 44.33 GB.
```

New Group: Generates a new *Group* in the output data tree of the *Replay* window if the file is opened without write protection. After the addition of a new *Group* a file update is automatically performed.

New Experiment: Generates a new experiment and a new *Group* (see above) and increments the experiment number.

Write Tree File: Defines when the data are flushed to disk. Choose one of these options:

- After Update File: Files are updated after an *Update File* command, and after creating a new experiment or *Group*.
- After Series: Writes raw data to disk after each acquired *Series*.
- After Sweeps: Writes raw data to disk after each *Sweep*.

The second option will make sure that the system file cache gets written to disk (i.e., "flushed") after acquiring a *Sweep*, and the third option performs a cache flush after acquiring a complete *Series*. Deselecting both options will suppress file cache flushing. In that case, the operating system will flush the file cache when it overflows.

The flushing of the file cache may take some seconds, depending on its size. Thus, if one lets the system decide when to flush, it may occur at an inappropriate moment, such as in the middle of a series. On the other hand, if POTMASTER would always force file cache flushing (as it does when the option *Write After Sweep* is active), one could not take advantage of the fact that writing to the file cache in RAM is faster than physically writing to disk.

It is safest to select the *Write After Sweep* option. This ensures that the data are immediately written to disk. The timing between *Sweeps* is also not interrupted by the system when a possibly large file cache is written to disk. If one must get the fastest disk performance possible, one can deselect the options. In that case, data are written to

New...	Ctrl+N
Open Read Only...	Ctrl+O
Open Modify...	Ctrl+M
Merge...	
Update File	Ctrl+U
Close	
File Status	Ctrl+I
New Group	
New Experiment	
Write Tree Files	▶
Convert To PPC Format	
Convert To Intel Format	
Page Setup...	
Page Properties Notebook...	
Page Properties Traces...	
Quit	Ctrl+Q



RAM, not directly to disk. However, this can only work as long as fewer data are acquired than there is space in the file cache.


***Note:** POTMASTER writes only that part of the tree files which were added or moved, thereby drastically reducing the time required for updating the files to disk.*

Convert to PPC: Converts the raw data to PowerPC format ("big endian"). This is required if a third party program is used that can not distinguish between Intel ("little endian") and PPC data format but requires data to be read in PPC format.


Convert to Intel Format: Converts the raw data to Intel ("little endian") format. This is required if a third party program is used that can not distinguish between Intel and PPC data format but requires data to be read in Intel format.

Page Setup...: Calls the Printer/Page Setup dialog of the operating system.

Page Properties Notebook...: Calls a dialog to set the page margins (left, right, top and bottom) and the font for printing. These settings apply for printing the Notebook.



Page Properties Traces...: Calls a dialog to set the number of columns and rows, the page margins (left, right, top and bottom) and the font for printing. These settings apply for printing data graphs.



Quit: Exits POTMASTER.

The default quit options are:

- **Save + Exit:** Saves configuration and data files and then quits the program.
- **Exit:** Saves data files and quits program.
- **Cancel:** Exit process will be aborted, you return to the program.

These default behavior can be influenced by the options *Save Settings* and *Verify Quit* in the configuration dialog of POTMASTER.

Note: *If you changed the Protocol Editor, the Pulse Generator pool or the Analysis you will be asked independently if you want to save your changes.*



Data files are always saved.



4.2 Edit Menu

The Edit menu applies to text manipulation in the Notebook window. The *Copy* function can also be used to copy graphs from the Oscilloscope or from an Online window into the clipboard.

Please note that most entries of this menu are disabled unless the Notebook window is in front!

The menu entries conform to the typical functions of the actual operating system (Mac OS or MS Windows).

Undo: Cancels the last action performed in the Notebook window.

Cut: Cuts the text selection from the Notebook window.

Copy: Copies a text selection from the Notebook. The *Copy* function can also be used to copy graphs from the Oscilloscope or from an Online Window into the clipboard. The active window (Notebook, Oscilloscope or Analysis) determines what is copied.

Paste: Pastes text from or to the clipboard.

Clear: Removes any content of the Notebook window.

Select All: Selects the whole Notebook window content.

Find...: Finds the entered search string.

Note: The Find function is case-sensitive and only searches from the cursor position to the end of the string.

Find Same: Finds the next appearance of the entered search string.

Find Selection...: Finds the search string that was marked (highlighted) in the Notebook window.

Replace...: Replaces the entered search string by some new string.

Replace Same: Finds and replaces the next appearance of the entered search string.

Undo	Ctrl+Z
Cut	Ctrl+X
Copy	Ctrl+C
Paste	Ctrl+V
Clear	
Select All	Ctrl+A
Find...	Ctrl+F
Find Same	Ctrl+G
Find Selection...	Ctrl+H
Replace...	Ctrl+R
Replace Same	Ctrl+T



4.3 Windows Menu

The `Windows` menu applies to the windows in POTMASTER. Clicking on a menu entry either opens the respective window or brings the already open window to the front. For most windows there are key commands assigned which will be displayed in the menu (see image beneath and also chapter 3 on page 13).

Control Window: Opens the Control Window or brings it to the front (see chapter 8 on page 91).

Configuration: Opens the Configuration window or brings it to the front (see chapter 5 on page 41).

Oscilloscope: Opens the Oscilloscope window or brings it to the front (see chapter 6 on page 67).

Potentiostat: Opens the Potentiostat window or brings it to the front (see chapter 7 on page 71).

Replay: Opens the Replay window or brings it to the front (see chapter 9 on page 95).

Protocol Editor: Opens the Protocol Editor window or brings it to the front (see chapter 11 on page 131).

Pulse Generator: Opens the Pulse Generator window or brings it to the front (see chapter 10 on page 99).

Analysis: Opens the Analysis window or brings it to the front (see chapter 13 on page 179).

Analysis Window 1, 2: Opens the specified Analysis Window window or brings it to the front (see chapter 13.4 on page 190).

Parameters: Opens the Parameters window or brings it to the front (see chapter 14 on page 197).

I/O Control: Opens the I/O Control window or brings it to the front (see chapter 16 on page 213).

Photometry/Imaging: Opens the Photometry Configuration or Imaging Configuration window or brings it to the front if the `Photometry` or `Imaging Extension` is activated (see chapter 5.4.7 on page 48 or chapter 5.4.5 on page 48).

Zero Potential (E-zero): Opens the Zero Potential window or brings it to the front (see chapter 17 on page 217).

LockIn: Opens the LockIn Control window or brings it to the front if the `LockIn Extension` is activated (see chapter 5.4.3 on page 47).

Spectroscopy: Opens the Spectroscopy Control window or brings it to the front if the `Spectroscopy Extension` is activated (see chapter 5.4.4 on page 48).

Solution Base: Opens the Solution Base window (see chapter 18 on page 221). If no solution base file is loaded you will be prompted to create a `*.sol` file.

Solution Changer: Opens the Solution Changer dialog (see chapter 19 on page 225).

Markers: The Markers window allows to set sweep markers and/or trace markers during a continuous recording (see chapter 20 on page 227).

Protocol Methods: Opens a dialog which provides information about the used *Protocol Method*.

Calculator: The Calculator window allows to perform numeric calculations in POTMASTER (see chapter 21 on

Control Window	
Configuration	
Oscilloscope	F12
Potentiostat	F11
Replay	F10
Protocol Editor	F9
Pulse Generator	F8
Analysis	F7
Analysis Window 1	
Analysis Window 2	
Parameters	
I/O Control	
Photometry/Imaging	
Zero Potential (E-zero)	
LockIn	
Spectroscopy	
Solution Base	
Solution Changer	
Markers	
Protocol Methods	
Calculator	
Notebook	F5
Close Front Window	Ctrl+W
Save Front Dialog	
Reset Front Window Position	
Enable Icon Configuration	



page 231).

Notebook: Opens the Notebook window or brings it to the front (see chapter 15 on page 211).

Close Front Window: Closes the window that is in front display.

Save Front Dialog: Dialogs modified by the user can be saved via this menu entry (see chapter 2 on page 5).

Reset Front Window Position: The position of the active (front) window is reset to a predefined "home" position. This function is essential when accidentally the window position has a huge offset and cannot be brought back into the field of view.

Enable Icon Configuration: If selected, the dialogs and controls in the POTMASTER user interface can be modified (see chapter 2.2 on page 6).

4.4 Replay Menu

The **Replay** menu functions are active when the **Replay** window is selected.

Show: Displays the content of the selected target. *Traces* are displayed according to the settings specified in the **Display** menu and the **Oscilloscope** window. If the **Root** is selected for **Show**, all *Traces* are displayed one after another. This can be aborted by mouse click on the **Break** or **Stop** buttons in the **Oscilloscope** window.

Show PGF Template: This option opens the stimulation file (PGF-Template) of the corresponding selected *Series* in the **Replay** tree (or the first *Series*, if a *Group* is the target). Leave the window open and scroll through your acquired data in the **Replay** tree to get an overview of the stimulation settings of your recordings (read-only).

Copy PGF to PGF-Pool: Copies the stimulus protocol of a selected *Trace*, *Sweep* or *Series* into the current **Pulse Generator** file.

Show Method: Displays the name and the used settings of the *Protocol Method* of the selected target (read-only). There are two additional options:

- **Copy To Analysis:** Copies the used *Analysis Method* of the *Protocol Method* to the *Analysis Methods* pool of the **Analysis**.
- **Copy To Protocol:** Copies the protocol event lines of the used *Protocol Method* into the active protocol at the selected position of the **Protocol Editor**.

Note: By using the Copy function from the Edit menu one can copy the the method text to the clipboard. Then one can paste the method at any position in the Protocol Editor by the Paste option.

Reference Series: Selects a target as reference with the **Set** function. The *Reference Series* is subtracted from all replayed data, if *Subtract Ref. Series* in the **Display** menu is active. The subtraction of the *Reference Series* not only affects the display of the data in the **Oscilloscope** window, but also data analysis (see chapter 4.5.2 on page 31) and export. A previously selected *Reference Series* can be turned off with the **Off** function.

Delete: Deletes the selected target.

Show
Show PGF-Template
Copy PGF to PGF-Pool
Show Protocol Method
Reference Series ▶
Delete
Print
Export
Export Full Sweep
Export Format: ASCII ▶
Export Mode ▶
ASCII Text Format ▶
Import Trace ▶
Export Trace ▶
With Marked Target and Children
Show Target
Print Target
Export Target
Export Full Target
Recompute Zero Offset
Average Sweep
Average Series
Duplicate Target
Compress Target
Collapse Target
Delete Target



Print: Prints the selected target and the *Analysis Graphs* if an *Analysis Method* was active. The *Analysis Graph* gets only printed when the *Series* level is the selected target.

Export: Exports the selected target. Display gain, leak subtraction or zero line subtraction will be applied (special case for *Igor Binary* export see below).

Note: The Export option will try to keep a "what-you-see-is-what-you-get" behavior. This means that the display options define the export options; e.g., only the displayed data are exported; or when Show Leak Traces is on, the leak Traces are also exported. For full Sweep export please use Export Full Sweep or make sure that the display is set to full time range, i.e. Start Time = 0%, End Time = 100% .

Export Full Sweep: Exports the full Sweep. This function works independent from the time range settings in the Oscilloscope window. Display gain, leak subtraction or zero line subtraction will be applied.

Export Format: This determines the output device and the type of output to be created. Output is generated in the way the data are displayed in the Oscilloscope window, e.g. when the digital filter is on, filtered data are output.

The following export formats are implemented:

- **ASCII:** Sweeps are output as columns of ASCII numbers representing the data *Traces* (all in scientific format). Each *Sweep* and *Series* starts with an identifier. The separator can be modified (space, comma, or tab separators) by using the ASCII-Text format option in the *Replay* menu.

Please note, that this may create huge ASCII files when the output target is a *Group*, for example.

- **Igor Pro:** Exports Igor Pro format. For further format options, see the lower part of the menu.
- **MatLab:** Sweeps are exported as MatLab file.
-  **PICT:** Sweeps are exported as Mac OS PICT file. Each file contains a single *Sweep*. When a *Series* is output, the *Sweep* files are generated automatically with the same name convention as waves for Igor Pro files: indices of "Group_Series_Sweep" are appended to the name.
-  **WMF:** Sweeps are exported as Windows Meta Files. Each file only contains a single *Sweep*. When a *Series* is output, the *Sweep* files are generated automatically with the same name convention as waves for Igor Pro files: indices of "Group_Series_Sweep" are appended to the name.

Note: Instead of exporting graphs in PICT or WMF format, you might want to copy (CTRL+C) the traces and to paste (CTRL+V) them into another application.

- **Printer:** Defines a printer for the export of the selected data. In contrast to the direct *Print* option where you always print the data and the *Analysis Graphs*, if available, the *Printer* allows to set either *Traces* and/or *Analysis - Graphs* in the *Export Mode*. If one wants to have individual *Sweeps* rather than a complete *Series* plotted in the page sections, one has to turn off *Overl. Ser.* in the *Oscilloscope* window or use the export option *Export Full Sweep*.

Further export parameters can be set here:

- **Trace Time relative to Sweep:** Individual *Sweeps* will be exported in an overlaid fashion. Time of the first data point of each *Sweep* starts with 0.
- **Trace Time relative to Series:** Individual *Sweeps* will be exported in an overlaid fashion. Time of the first data point of each *Sweep* starts with the difference between the start time of the *Series* and the start time of the 1. *Sweep* of that *Series*.



- **Trace Time relative to Timer:** Individual *Sweeps* will be exported in a concatenated fashion. The *Trace* time is offset by the *TimerTime* at *Sweep* start.
- **Relative Trace Time:** Individual *Sweeps* will be exported in a concatenated fashion. Same as above, but the time is subtracted by the *TimerTime* at the start time of the 1. *Sweep*. Resulting that the time of the first *Sweep* starts with 0.

The following format options refer to Igor Pro format only:

- **Igor - Allow Raw Data Access:** Whenever possible an Igor Info file is generated allowing access to raw POTMASTER data which should be stored in the same folder. Using this output option assures fastest output and the smallest output files.
- **Igor - Create Binary Wave:** Export of *Sweeps* to Igor Pro as binary data. This function generates an Igor macro which contains the instructions for Igor on how the data are to be loaded, scaled, and displayed. It has the extension *.itx. A double-click on it will make Igor Pro load that macro file and execute the instructions in it, importing, scaling, and displaying the data. The actual data are not really exported when using the *Allow Raw Data Access* option. That option will make use of the "GBLoadWave" Igor extension to read the data directly from the POTMASTER raw data file, i.e., the *.dat file. The data are converted to Igor binary waves when the simple Export option is used. However, even in that case a macro file is generated and you should load the data via that file. When you want to import data from within Igor Pro, use the option *Load...Igor Text* to load the macro file. Use the option *Load...Igor Binary* only when you want to explicitly load one of the generated Igor binary waves (file extensions *.ibw or *.wav).

Note: It is much faster to work with "Create Binary Wave" than with "Create Text Wave", and the created files are considerably smaller.

- **Igor - Create Text Wave:** Export of *Sweeps* as ASCII waves in "Igor Text" format for the analysis and display program Igor. In Igor Pro, each wave is identified by indices "Group_Series_Sweep_Trace" (e.g., "Name2_4_3_1"). If the file name starts with a number, a "W" is placed in front of it, because in Igor Pro, waves are not allowed to start with a number. The created file has the extension *.itx and is recognized by Igor Pro, i.e., double-click on this file will start Igor Pro, load and display the file content (not for *Sweeps*). The waves will immediately be displayed in Igor Pro only when the *Igor: Make Graphs* option was selected. Otherwise, the *Sweeps* will be loaded, but must be displayed by Igor Pro's "Display Wave" or "Make Graph" command. When loading Igor Text output files, do not use the "General Text" import option in Igor; always use the option "Load...Igor Text".
- **Igor - Make Graphs:** The *Sweeps* are exported as Igor graphs as they appear in POTMASTER in the Oscilloscope window. The created files have the extension *.itx and are recognized by Igor Pro, i.e., double-click on this file will start Igor Pro, load and display the file content.
- **Igor - Make Layouts:** This option is useful if several *Series* shall be exported to Igor Pro. The exported *Traces* are arranged on an Igor Pro layout page. The created files have the extension *.itx and are recognized by Igor Pro, i.e., double-click on this file will start Igor Pro, load and display the file content.
- **Igor - One Graph per Full Sweep:** With this option, a *Trace* that consists of several pages (e.g. continuous data) is exported as one single Igor Pro wave. Without this option, one wave per page is exported.

Export Mode: Shows a sub-menu that allows determining what will be exported:

- **Traces:** Exports the selected *Traces*.
- **Stimulus:** Exports information on the used stimulation pattern (segment duration and amplitude).

Note: This option can only be used when either Traces or Analysis - Graphs is selected.



- **Analysis - Graphs:** The graphs of the **Analysis** get exported.
- **Analysis - Notebook:** The **Analysis** function results which are displayed in the **Notebook** get exported. Please note that the option **Notebook** in the **Analysis** window has to be selected.

***Note:** This option is not available when using the Igor Pro Export Format.*

ASCII-Text Format: Shows a sub-menu that allows to specify the type of separator used when generating ASCII tables, and the format of the exported text:

- **Space/Comma/Tab Separator:** Specifies how values are separated.
- **Include Headers:** If checked, a header that specifies various parameters of the exported data will precede the actual values.

Import Trace: Imports data into a selected *Trace*.

When importing data you first have to generate a corresponding data structure (*Tree*) in POTMASTER. Make sure that the *Gain* and *Sampling Interval* of the data structure match the data you intend to import:

- *Gain* mismatch might result in loss of resolution or saturation of the data.
- The length of the imported *Trace* must be equal or greater than the *Trace* selected in the **Replay** window. Points exceeding the length of the *Trace* in the **Replay** window are discarded.

You can import the following three formats:

- from ASCII file: Imports data which were saved as CSV (=Comma-Separated Values).
- from Binary file: Imports data which were saved in the Igor binary format.
- from Igor Wave: Imports data which were saved in the Igor Wave format.

***Note:** Since the Traces will be imported into the open file, the import function does not work when your file is read-only.*


Export Trace: Exports data. You have four possibilities:

- as ASCII file: Exports data in ASCII file format.
- as Binary file: Exports data in Binary file format.
- as Igor Wave: Exports data in Igor wave format.
- as Stimulus Template file: Exports data as a stimulus template file.

With marked Target and Children...


Actions executed from this section of the **Replay** menu are performed on all marked children of the selected *Target* in the data tree:





In example "A" the `With Marked Target and Children → Export` command would export only Sweep 3 to 8 of the first IV, while in example "B" all marked Sweeps would be exported.

- **Show:** Displays marked *Targets* in the selected tree branch.
- **Print:** Prints the marked *Targets* and children of the selected tree branch.
- **Export Target:** Exports marked targets and children of the selected tree branch. The display options in the Oscilloscope window define the export options.
- **Export Full Target:** Exports the full marked targets and children independent from the time range settings in the Oscilloscope window.
- **Recompute Zero Offset:** The zero offset value of all *Traces* in the selected target is recomputed. The zero offset value is used by the *Subtract Zero Offset* function (see chapter 4.5 on page 29).



- **target trace (1...32):** Selects the target *Trace* for the zero offset recomputing.
- **left bound [%](0...99) & right bound [%](0...100):** Defines the range in between POTMASTER should recompute the zero offset.
- **segment (0=use zero seg.):** Defines the Segment (of the PGF) for the zero offset calculation.

Note: Only visible Traces are recomputed.

Note: Only Traces originating from an AD-channel are recomputed; virtual Traces are not affected.



Note: The zero offset of leak Traces are only recomputed if the "Show Leak Traces" is selected from the Display menu.

- Average Sweep: Averages all Sweeps of a Series and stores them as Series with one Sweep. The Sweeps have to have the same length.
- Average Series: Generates one new Series in which each Sweep is the average of all marked Series with the same Sweep index and each Trace with the same Trace index.
- Duplicate: Creates a new Group containing all marked targets. This is e.g. convenient when one needs to compress a Trace for export.
- Compress: Compresses all marked Traces by a given compression factor. Optionally one can supply a maximal compression factor. This can be used e.g., if one has Traces which are already compressed and having much fewer points per second, e.g., a FURA trace. The Traces cannot reside in a read-only file. To compress a Trace in a read-only file, duplicate the required targets.
- Collapse: The Sweeps of all marked Series are copied into a new Series. This is typically used when one acquired many Series with one single Sweep and one wants to combine them into one Series for easier Analysis.
- Delete: Deletes the marked targets and children of the selected tree branch.

Note: Delete and Delete Marked Target Children are disabled when a data file is opened with File → Open Read Only.



4.5 Display Menu

This menu sets some parameters for the display of data in the Oscilloscope window.

Auto Show: Automatically displays the data selected in the Replay window. When this option is disabled, a double-click on the targets in the tree is required for displaying the data.


Trace Properties: In the Trace Properties dialog the user can set *Show*, *Vector* (connect individual data points) and *Center* flags, select the *Marker* type and the *Reference Series* properties. All settings are explained in detail in chapter 4.5.2 on page 31.

Also, the *Trace Properties* can be set via macro commands to allow automatic display adjustment depending on type of acquisition. Have a look for the macro commands starting with T (Help menu → List All Macro Items).

The color of *Traces* can be set in the Line Colors dialog of the Configuration window (see chapter 5.6.1 on page 54).

Show Zero Line: Draws a reference zero line in the Oscilloscope window.

Subtract Zero Offset: Subtracts the zero offset from the measured signals. The zero offset is the mean value, calculated from the *Zero* segment. The *Zero* segment is defined in the PGF (see chapter 10.10.2 on page 120). Usually, the zero offset value is determined online, but it can also be computed offline by use of the *Recompute Zero Offset* function from the *Replay* menu (see chapter 4.4 on page 27).



Note: Subtract Zero in the Display menu is kept “On” even if the conditions for Zero Subtraction are not met. If one wants to see the subtracted Trace offline, one has to switch to another Trace in the Replay Tree and then back to the just acquired Trace. A simple “Wipe” or “Redraw” of the displayed data is not sufficient.

Subtract Ref. Series: If this item is checked, then the *Reference Series* is subtracted from all replayed data (see chapter 4.5.2 on page 31).

Subtract Trace Buffers: If this item is checked, then the *Trace Buffer* is subtracted from all replayed *Traces* (see chapter 4.5.2 on page 31).

Dimmed Overlay: Turns the *Dimmed Overlay* mode on or off. In *Dimmed Overlay* mode, the last displayed *Trace* is drawn in another color than the previously plotted *Traces*.

Overlay Pages: During continuous acquisition the data are displayed in pages. The pages can be overlaid.

Overlay Traces: Overlay of all *Traces* of a selected *Sweep*.

Overlay Sweeps: Overlay of all *Sweeps* of a *Series*.

Overlay Series: Overlay of all *Series*, i.e. overlay of all.

Auto Y-scale: The Y-scaling is automatically adapted to the replayed data to fit minimal and maximal data of a *Series* onto the screen. The range of the Y-scaling depends on the *Gain* settings.

Fixed Y-scale: The Y-scale is fixed. In the Oscilloscope window the controls *Y-scale* and *Y-offs.* turn into *Y-max* and *Y-min*.

Note: Auto Y-scale has priority over Fixed Y-scale.

Fixed X-width: The X-scale is fixed. In the Oscilloscope window the controls *Start Time* and *End* turn into *X-min* and *X-max*.



Freeze Zero Line: This checkbox determines the behavior of the display scaling in the Oscilloscope window. When *Freeze Zero Line* is selected, the Oscilloscope behaves like an oscilloscope, i.e. when the display gain is changed the zero line stays at its position. When deselected, display scaling is performed with reference to the center of the display. I.e., when *Y-offs* is not zero, the zero line may alter its position on the screen.

Labeling: Determines the labels in the main display of the Oscilloscope window and the *PGF-Editor Grid* independently:

- Labels Only: Draws calibration bars.
- Grids + Labels: Draws a grid and units/division.
- Grids + Values: Draws a labeled grid.
- No Labels: No labels, no grids.

Show Tree Info: When this option is selected, information about the replayed target is shown in the Notebook window.

4.5.1 3D-Mode

Shows a sub-menu for selection of the pseudo three-dimensional graphing.




Figure 4.1: 3D Mode of the Oscilloscope Window

The following settings can be made:

- 3D-Graph - On: The results are displayed in black and white in pseudo three-dimensional mode by displaying subsequent *Sweeps* with a horizontal and vertical offset. The *dX* and *dY* values for the 3D feature can be entered in the Configuration window. The default value for both is 0.010.
- 3D-Graph - Color: The results are displayed in pseudo colors leading to a two-dimensional height profile. You can set the color selection and the contrast by changing the values in *Y-scale* and *Y-offs*.
- 3D-Graph - On and 3D-Graph - Color: The results are displayed in pseudo colors and 3D-mode to give the impression of contour lines, thus leading to a 3D height profile.




Figure 4.2: Pseudo Color Data Representation

4.5.2 Trace Properties Dialog




Figure 4.3: Trace Properties Dialog

The Trace Properties dialog provides access to certain display settings of individual *Traces* and *Buffers*.

Extend: Extends the setting of the first *Trace* to all other *Traces*.

Trace 1...16, Buffer 1...4: Up to 16 *Traces* and 4 buffers can be individually modified.

Note: The amount of traces (default: 16) is defined in the *Configuration* (General pane).

Show: Here, one can decide if a *Trace* shall be displayed (*Show*) in the Oscilloscope window or not.

Vectors: When option *Vectors* is checked then the markers are connected by lines.

Marker: Markers, such as

- Point
- Plus
- Star
- Diamond
- Cross
- Square

can be used for plotting data points. This is particularly useful, if LockIn or FURA data are acquired.

Ref. Series Corr.: If *Ref. Series Corr.* is checked for a *Trace*, then the *Reference Series* is subtracted from that *Trace*. A *Reference Series* can be defined by use of the *Reference Series - Set* item from the *Replay* menu. The *Ref. Series Corr.* can be turned 'On' or 'Off' globally in the *Display* menu.

Center: If this option is set the traces get centered during acquisition. Therefore the left 10% of the first page are brought to the center of the display.

Trace Buffer Corr.: If *Trace Buffer Corr.* is set to one of the four available *Trace Buffers* then the contents of the chosen *Buffer* are subtracted from that *Trace*. This can be used for subtracting a control *Trace* from other *Traces*. The *Trace Buffer* subtraction can be turned 'On' or 'Off' globally in the *Display* menu.

Use Trace Scaling: With this option, the display scaling of a *Trace Buffer* can be coupled to the display scaling of a certain *Trace*.

Repaint: Repaints the last displayed *Sweep* in the Oscilloscope with the new *Trace Properties*.

Done: Closes the *Trace Properties* dialog.

Note: Like other menu settings, the Trace Properties are stored in the POTMASTER configuration file.



4.6 Buffer Menu

POTMASTER offers four independent *Trace Buffers* for storing and processing of intermediate *Trace* data.

Buffer 1...Buffer 4: Select (make active) the buffer to work on.

Use Full Trace: If this option is selected the functions *Add Trace*, *Subtract Trace*, and *Accumulate Trace* work on the complete *Trace*. If this option is not selected the above mentioned functions work on the part of the *Trace* which is displayed in the *Oscilloscope* window only.

Clear All: Clears all *Buffers* (*Buffer 1...4*).

Clear: Clears the selected *Buffer*.

Show: Shows the selected *Buffer* in the *Oscilloscope*.

Scale: Scales the selected *Buffer* with *Scale* and *Offset*.

Equation: Allows to perform a *Buffer* calculation based on an equation. For details see chapter 21 on page 231 and chapter 21.2.2 on page 235.

Import: Imports data into a selected *Buffer*.

Please make sure that the key parameters *Gain* and *Sampling Interval* of the imported data match the parameters of the *Buffer*. In case a *Buffer* is filled with data for the first time, the parameters are retrieved from the *Trace* currently selected in the *Replay* window.

You can import the following three formats:

- from ASCII file: Imports data which were saved as CSV (=Comma-Separated Values).
- from Binary file: Imports data which were saved in the IGOR binary format.
- from Igor Wave: Imports data which were saved in the IGOR Wave format.

Export: Exports data from a selected *Buffer*. You have four possibilities:

- as ASCII file: Exports data in ASCII file format.
- as Binary file: Exports data in Binary file format.
- as Igor Wave: Exports data in Igor wave format.
- as Stimulus Template file: Exports data as a stimulus template file (see Using a Recorded Waveform as Stimulus in the PATCHMASTER Tutorial).


Add Trace: Adds the *Trace* selected in the *Replay* window to the active *Buffer*.

Subtract Trace: Subtracts the *Trace* selected in the *Replay* window from the active *Buffer* (*Trace*).

Accumulate Trace: Accumulates the *Trace* selected in the *Replay* window to the active *Buffer*:

$$\text{Buffer}(n+1) = (\text{Buffer}(n) * n + \text{Trace})/n+1$$

with n = number of *Traces* in the *Buffer*.



Deaccumulate Trace: Deaccumulates the *Trace* selected in the *Replay* window from the active *Buffer*:

$$\text{Buffer}(n-1) = (\text{Buffer}(n) * n - \text{Trace}) / n-1$$

with n = number of *Traces* in the *Buffer*.

With Marked Target and Children: Multiple *Traces* can be marked in the *Replay* window and processed simultaneously.

- Use: Specify which *Trace* of the marked *Sweeps* should be processed.
- Add, Subtract, Accumulate, Deaccumulate: Work on all marked targets. For further explanations of the functions refer to individual functions explained above (see chapter 4.4 on page 26).

Replace Target Trace: The *Trace* selected in the *Replay* window (target *Trace*) will be replaced with the active *Buffer*.

4.6.1 Handling of Parameters by the Buffer

When a *Buffer* is filled with data for the first time (after a *Clear* command) all parameters of the target *Trace* are transferred to the *Buffer* data structure. When adding further *Traces* to the *Buffer* these parameters are not changed. It is assumed that all *Traces* that are added to the *Buffer* are of same type. When replacing a target *Trace* with the data of a *Buffer*, data only are transferred. The parameters of the target *Trace* stay unchanged.



It is the responsibility of the user that the key parameters such as Gain and Sampling Interval are chosen adequately when adding importing data to a Buffer or processing different Traces in a Buffer. Once a Trace is replaced by a Buffer, most parameters i.e. Zero Offset or C-slow become meaningless.

4.7 Notebook Menu

The Notebook is used to display experiment information. The ASCII-table separator setting of the *Replay* menu is used for the Notebook as well. This enables e.g. to transfer data directly to spread sheets requiring a TAB separator, such as MS Excel, via "Cut and Paste". The options in the Notebook menu are:

Save: Saves the Notebook under its default name: "Notebook_Date".

Save As...: Asks for a file name before saving.

Merge...: Merges a text file to the content of the Notebook.

Print...: Output the content of the Notebook to a printer.

Clear when Saved: Automatically clears the Notebook after the present content is saved to disk.

Clear: Clears the Notebook.

Set Length...: Specifies the maximal number of text lines in the Notebook. The maximal number of lines is given in parentheses.

Line Numbers: Shows line and column numbers when moving with the cursors inside the Notebook.

Font Size...: Allows to select the font size of the Notebook text.

Zoom In: Shrinks the Notebook window to default (i.e. small) size.

Save	Ctrl+S
Save As...	
Merge...	
Print...	Ctrl+P
☒ Clear when Saved	
Clear	Ctrl+B
Set Length... (20000)	
Line Numbers	
Font Size...	
Zoom In	Ctrl+K
Zoom Out	Ctrl+L
Scientific Notation	
Auto Store	



Zoom Out: Expands the Notebook window to full program screen size.

Scientific Notation: If set, the results of the Analysis are written to the Notebook in scientific notation (e.g., 1.23e-12). The default is engineering format (e.g., 1.23p). The scientific notation is mostly used when the user wants to copy results from the Notebook to a spread sheet program by copying to the clipboard.

Note: This setting also applies to the ASCII-export format entry in the Replay menu.

Auto Store: This option will automatically store the Notebook together with the data file ([data file name].txt). Upon opening a data file, its Notebook file will automatically be loaded as well.

4.8 Protocols Menu

The **Macros** menu of older POTMASTER versions has been replaced by the new **Protocols** menu (since version 2x52). This drop-down menu can be used for the conversion of macro files from older POTMASTER versions to protocols.


Append Macro File: Converts an existing macro file (*.mac) to a protocol and appends this new protocol to the actually opened protocol file (*.pro).

Execute while Recording: If selected, every button you press during macro/protocol recording will be executed. If not selected, this option allows recording a protocol without executing the buttons. Thus, during protocol recording, commands and values are recorded, listed in the Notebook, and then the parameters are immediately set back to their previous value.

Note: Macro/protocol recording can be started from the Protocol Editor dialog (see chapter 11.1 on page 131).

Execute: Executes the selected, user defined protocol. Alternatively, protocols can be executed from:

- the Control window, if the option *Hide Protocol-Bar* in the Control window is not activated on the Display tab (see chapter 5.6.3 on page 55).



4.9 PG 300 / PG 600 Menu

The PG 300/PG 600 menu contains some user specific settings which have effects on the data acquisition or on the replay of recorded data.

The options **E vs. Reference**, **E vs. HE**, **E vs. OCP**, **Use Present Value**, and **Use Value in Data File** are display options for data which are originating from voltage input. In many cases these options are needed for the display of recorded current potential curves. Please note that the effects of these options are only visible in the Analysis window where you can use potential traces as x axis.

E vs Reference: The displayed cell potential is in accordance to the connected reference electrode. Using this option the recorded potentials are not corrected against any constants and relate directly to the potentials measured between the reference and the working electrode. This is the default setting for data acquisition in combination with **Use value in data file**. For data replay and potential corrections please chose the options **E vs. HE** and **E vs. OCP**.

***Note:** Please pay attention, that the connected reference electrode has measured and stable potential. When the potential of your reference electrode is not stable during the experiments this shift will affect the recorded potential data.*

E vs HE: The displayed cell potential is in accordance to the potential of the normal hydrogen electrode (NHE). The potential difference between the Normal Hydrogen Electrode and other reference electrodes is selected in the window **Zero potential**. To relate recorded potentials versus a reference potential from the window **Zero potential** please chose **E vs. HE**, activate the option **Use present value** and open the window **Zero potential**. In this window you can select a reference potential or a user specific value. There are various options but the only interesting field is the first blue panel beside **Manual Input**. Only here stands the value of the selected reference potential which is applied for a correction of your recorded potential values! Click **Done** and select in the **Replay** window your potential data.



E vs OCP: The displayed cell potential is in accordance to the measured **Open Cell Potential (OCP)**. In corrosion experiments the OCP is sometimes called "Corrosion Potential" or "Polarisation Potential" and is used as **Initial Potential** for Tafel Analysis or linear polarization plots. When some potential data have been recorded with the option **E vs. Reference**, the potentials can be related to a recorded **Open Cell Potential (=OCP)**. Please open the window **Configuration** and type an OCP value in the section **I/O Parameters**. Opening the **I/O Control** window this value appears and can also be changed here. If no OCP has been determined, the OCP value is set to 0 V.

The following options are important for replaying recorded potentials, they take no effect in the parameters **E-initial** and **Charge** of the **Potentiostat** window.

Use Present Value: Replayng recorded traces are shown with the actual selected potential correction.

Use Value in Data File: Replayng recorded potential traces are shown with the recorded potential correction. This is only visible if you have selected **E vs. HE** and 0 mV in the window **Zero potential**.

Update Delay: This option defines the refresh interval time of **I-cell** and **E-cell** depicted in the **Monitor** tab of the



Potentiostat window (see chapter 7 on page 71). The default value is 0.25 s.

Stimulus Ranging: If this option is activated, then the minimal required voltage range of the potentiostat is set automatically when acquiring data via a **Pulse Generator** sequence.

Initialize PG300/PG600: This is used to restart the AD/DA interface; e.g., in case POTMASTER was started with the interface being turned off.

Calibrate PG300/PG600: Calibrate the PG 300. Note that this may take several minutes. The calibration procedure creates a new scale file.

EIS Calibration: Selecting this option start the calibration routine for EIS (Electrochemical Impedance Spectroscopy) measurements. The calibration of the impedance spectroscopy is done with help of the **Protocol Editor** in a semi automatic way. In case of a multi-potentiostat each amplifier must be calibrated separately. First the select the amplifier in the **Amplifier** window and then start the EIS Calibration process from the PG300/PG600 Menu. The following amplifiers will be calibrated in a fully automated way. The calibration resistors are built-in to the amplifier and can be switched automatically:

- PG 610 USB
- PG 612 USB
- PG 690 USB
- PG 618 USB Amplifier 1



The calibration process of those amplifiers takes several hours and should be run overnight.


In case a pre-amplifier is used (PG 611 USB and PG 618 USB Amplifier 2), then the model cell MC 61 EIS (Order number: 895336) is required. After starting the calibration process you will be prompted connect the model cell MC 61 EIS to the headstage of the amplifier. Each feedback gain range will be calibrated separately, and you will be asked to switch the model cell to the appropriate resistor for each range.



This calibration process takes about 30 minutes per gain range. Before starting the calibration of a gain range your interaction is required.

Follow the instructions of the software to finalize the calibration process.





Once the calibration process has finished, a lookup table is saved (e.g. PG618_USB_332355_1.EIS), and one can now use the *EIS Method* in the Protocol Editor (see 12.2 on page 168). One calibration file (a text file with extension *.EIS) is stored per amplifier and contains the calibration parameters for each frequency and each gain range.



The EIS calibration is absolutely necessary to do before starting the EIS experiments.



The PG 300 Series of Potentiostats do not support Electrochemical Impedance Spectroscopy.




Serial No.: Displays the serial number (e.g. '560141') and board version (e.g. 'E').

4.10 Help Menu


With the additional tools listed in the Help menu you can get more familiar with the available functions of POTMASTER.

Show Tooltips
Show Keys
Hide Keys
List Keys
Save Keys
List All Items
List All Macro Items
About PotMaster

Show Tooltips: When activated information about the button that is located under the mouse tip is displayed.



Show Keys: Displays the key commands with the corresponding buttons.



Hide Keys: Hides the key commands.

List Keys: Lists all keys to the Notebook window (see chapter 3 on page 13).

Save Keys: Saves the actual key command settings to the default key file. The old `Potmaster.key` file will be saved with the extension `*.kXX` (X = consecutive index number).


List all Items: Lists all items and their values in the Notebook window.

List all Macro Items: Lists all macro items and their values in the Notebook window.

About POTMASTER: Displays the POTMASTER software version and contact information.



5 Configuration Window



Settings like sources for external parameters, default values, display settings, colors, fonts, default files, ... can be edited in the Configuration window. To access the Configuration window select the drop-down menu Windows → Configuration. These and other settings can be stored in *.set files; by default this file is Potmaster.set. By means of different *.set files every user can define their individual program layout to meet specific requirements.



5.1 Individualize POTMASTER



The name of a POTMASTER setting file can be used as a command line parameter upon starting the program.

That way, one can start POTMASTER with various configurations which is very useful when several users work on one setup and have a need for different POTMASTER settings.


The procedure for an individualized POTMASTER starting procedure is as follows:

- Rename the *.set file in your HEKA folder (e.g. "Potmaster_User.set").

Note: The term "User" is just a place holder for individual names.

- Create a shortcut of Potmaster.exe on the desktop.
- Rename the shortcut (e.g. "Potmaster_User").
- Right-click on the shortcut: Properties - Shortcut - Target.
- The following command line should be written in the "Target" field, including the inverted comas and empty spaces:

```
"C:\Program Files (x86)\HEKA\Potmaster\Potmaster.exe" "C:\Program Files (x86)\HEKA\Potmaster\Potmaster_User.set".
```



Note: Please be aware that your installation path of your POTMASTER may vary on your PC. The "Patchmaster" definition on the screenshot corresponds to POTMASTER or CHARTMASTER, respectively.



5.2 Save



SAVE: Saves the configuration file with its current name.


SAVE AS...: Saves the configuration file under a new name.

Default Windows: Resets the position of most windows. This helps to restore the windows position once one gets dragged off screen (for example after reducing the screen resolution).

Default Settings: Sets all settings back to the original settings. Note that this needs a restart of the program!

Note: There is no Load option because POTMASTER has to be restarted upon substantial changes of the settings. Thus, in order to use another setting, quit POTMASTER and restart. Before you do so, make certain that there is no Potmaster.set file available. In this case, upon start-up, POTMASTER will prompt you to select a new setup file (this can have any name).

5.3 General



5.3.1 Max Shown Traces and Values

Max. Shown Traces: This number defines how many *Traces* are used by POTMASTER. The user interface is adapted to show only the defined number of *Traces*. A small number makes the dialogs and lists within POTMASTER shorter. The new setting only takes effect after saving the configuration file and upon restarting the program. When using the *Probe Selector* it may be necessary to increase *Max. Shown Traces*.

Max. Shown Traces	16
Max. Shown Values	16

Max. Shown Values: This number defines how many values are used by POTMASTER. The user interface is adapted to show only the defined number of values. A small number makes the dialogs and lists within POTMASTER shorter. The new setting only takes effect after saving the configuration file and upon restarting the program.

5.3.2 Memory Allocation



Max. Sample Points: Sets the maximum number of points that can be acquired in one *Trace* (see chapter 10.8 on page 104). This parameter is limited by the amount of available memory.

Max. Sample Points	500000
Max. Stim. Traces x Points	5 2000000

Max. Stim. Traces x Points: Maximum number of stimulation *Traces* and their respective number of points. This parameter is limited by the amount of available memory.

Note: In case you want to drive the number of Max. Sample Points to its limit you might reduce the number of Max. Stim. Traces to make more memory available for allocating sample points.

Spectrum Buffer Size: Maximum number of samples used for spectrum analysis calculation (Fast Fourier Transformation). If more data points were acquired than defined an average of the FFT calculation will be performed.

- 1024
- 4096
- 16384
- 65536
- 262144

Note: It is recommended to adapt the number of Spectrum Buffer Size in respect to the sampling frequency/sample interval and the frequency of interest of your power spectrum.

Log Compression: In a power spectrum analysis (FFT calculation) one gets an exponential increase of data points with increasing frequencies. To balance that unequal distribution the *Log Compression* can be used. Here, the logarithmic frequencies are divided into equal sized bins and for each bin the average value is calculated. This allows a smooth representation of the power spectrum in the analysis.

5.3.3 Minimum Wait Time

Min. Wait Time: Sets the time POTMASTER reserves to wait for the correct time to start when executing individual repeat loops such as acquisition of a *Series of Sweeps* or an *Repeat* loop in the Protocol Editor.

Min. Wait Time	100. ms
----------------	---------

For example, to ensure an *Acquire Each Sweep* repeat loop keeping the same repeat time as set by the *Duration* in the *Repeat* event, the program will wait *Min. Wait Time* before executing the next acquisition to hit the scheduled starting time of the next *Sweep* as exactly as possible.

Note: In case you are observing incorrect and unstable repeat times, please try to increase the Min. Wait Time.

This parameter can also be set via the *Acquire Properties* event in the Protocol Editor.

5.3.4 Batch Communication

Batch Communication: The *Batch Communication* with other applications can be disabled or enabled (see *Controlling PATCHMASTER* in the PATCHMASTER Tutorial).

Batch Communication	Disable	☒ Logging	1 Connection
---------------------	---------	---	--------------

- **Disable:** If this option is checked, the *Batch Communication* is turned off.



- **Enable as Receiver:** POTMASTER is configured to serve as *Receiver* and will wait for command inputs of other applications.
- **Enable as Sender:** POTMASTER is configured to serve as *Sender*, thereby controlling another program, e.g. FITMASTER.
- **Synchronize Files:** With this option the *Sender* application will send the "FileOpened" message, when a *.dat file, *.pgf file, *.pro file or *.onl file is opened, and the "FileUpdated" message, when the data file is updated. The "FileClosed" message is send, when the data file is closed. The *Receiver* will then open, update, or close the same data file as the *Sender* application.
- **Polling:** In the *Polling* mode, the master application polls the slave application for commands. The slave then answers to the query and appends all pending commands, i.e., it no longer sends commands independently. In doing so, commands do no longer get lost, if heavy bi-directional communication is ongoing.


Logging: If the checkbox is activated the communication via the *Batch Communication* interface is written to the Notebook.

Connections: Specify the number of connections to other software instances which can communicate via batch communication.

5.3.5 User Name

User Name: A user name that is entered here will be stored with the raw data.

5.4 Hardware



5.4.1 Amplifier and Digitizer Selection

Amplifier: POTMASTER supports the following amplifiers:




POTMASTER supports the following potentiostats/galvanostats:

- PG 600 USB Amplifier: A PG 610/690 USB, PG 611 USB PG 612 USB or PG 618 USB amplifier.
- PG 330 USB Amplifier: A PG 330 USB amplifier.
- PG 300 USB Amplifier: A PG 310/390 USB, PG 390 USB or PG 340 USB amplifier.
- PG 300 Amplifier: A PG 310, PG 390 or PG 340 amplifier.
- PG 400 Remote: A PG 400 or PG 500 amplifier (PG 410, PG 490, PG 510 or PG 590) is connected with the data acquisition interface LIH 1600 (by three BNC cables at Monitor Output Voltage and Current, External Input). The serial communication port of the potentiostat is connected with a serial port of the computer. In this case the amplifier settings are controlled by the software.
- PG 400 Local: A PG 400 or PG 500 amplifier (PG 410, PG 490, PG 510 or PG 590) is connected only with the data acquisition interface LIH 1600 and the serial communication port of the potentiostat is connected with a serial port of the computer. In this case the amplifier settings must be done by the knobs on the front panel of the potentiostat. POTMASTER monitors gain, amplifier modes and filter settings.
- PG 285 Amplifier: To connect any analog potentiostat with the data acquisition interface LIH 1600 with 3 BNC cables (Voltage Output, Current Output and External Input). All amplifier settings must be done manually and must be selected manually also in POTMASTER.
- PG 284 Amplifier: To connect any analog potentiostat with the data acquisition interface LIH 1600 with 3 BNC cables (Voltage Output, Current Output and External Input). All amplifier settings must be done manually and must be selected manually also in POTMASTER.
- Telegraphing Amplifier: For connection of an amplifier with telegraphing capabilities.

Digitizer Selection: Select the appropriate AD/DA-converter, if it is not automatically assigned via the amplifier selection.

POTMASTER supports the following AD/DA-converters:

- ITC-16
- ITC-18
- LIH 1600 / ITC-1600
- LIH 8+8



Board Type: Select the used *Board Type*. There are two options:



- 1...3 PCI: Additional set-up information, if there is more than one possible card slot to connect to the digitizer.
- USB: InstruTech's USB-16 or USB-18 adapters allow for connecting an ITC-16, EPC 9, ITC-18 to the computer via an USB 2.0 port. In case, such an adapter is used, it should be enabled here.

PCI
PCI
PCI
☒ USB
USB
USB

Demo Mode: The functionality of POTMASTER is restricted in the *Demo Mode*. You can only read your own data files or read and analyze the demo data files.

If a *Telegraphing Amplifier* is used, the *Gain* and/or *Bandwidth* settings for the amplifier are read from an AD-channel and encoded by means of a *Lookup Table*. Upon selection of this amplifier type, the user must specify the AD-channel for the *Current Gain* and *Bandwidth* and load the appropriate *Gain* and *Bandwidth* tables. Both settings are made in the *I/O Parameter* section (see chapter 5.7 on page 56).

Lookup Tables can be created easily as they are ASCII files with a simple structure with a series of text lines each containing:

Voltage | Gain (mV/pA)

or

Voltage | Bandwidth (Hz)

The voltage is the threshold above which the following *Gain* or *Bandwidth* setting applies. The voltage values are scanned beginning with the first value. Thus, voltages must be in *descending* order! The voltage thresholds for discrimination between adjacent settings should be halfway between the nominal telegraph voltages.

Note: The *DacLookup* file in the *LookupTables* folder (inside the *POTMASTER* folder) depicts exemplary how to create your own lookup table.

5.4.2 Output and Input Channels

The following settings define the default output and input channels (see chapter 10 on page 99).

E-Initial Out: Assigns the output DA-channel of the stimulus.

Stim. Scale: Scaling factor of the stimulus signal in voltage clamp mode. The voltage amplitude that is set in the PGF sequence is *divided* by this factor before output to the DA channel.

Current In: Assigns the input AD channel for the current *Trace*. This input channel is scaled by the *Current Gain* (see chapter 5.7.2 on page 59).

Voltage In: Assigns the input AD for the voltage *Trace*.

Stim. Scale	1.000
E-initial Out	Stim-out
Current In	I-mon
Voltage In	E-mon

5.4.3 LockIn

Here, the LockIn amplifier of POTMASTER can be activated. For a detailed description see chapter *Software LockIn Extension*, 24 on page 249, and the PATCHMASTER Tutorial "Capacitance Measurements using the LockIn Extension".




5.4.4 Spectroscopy

Here, the `Spectroscopy` Extension of POTMASTER can be activated. For a detailed description see chapter `Spectroscopy`, 25 on page 261, and the PATCHMASTER Tutorial "Using the `Spectroscopy` Extension".

5.4.5 SmartLUX

Here, the `Imaging` Extension (SMARTLUX) of POTMASTER can be activated. Depending on the fluorescence excitation light source or wavelength switcher you have installed in your setup, you can choose between different kinds of `Imaging` Extensions:

- TILL Photonics/Cairn OptoScan: "TILL Photonics" or "Cairn OptoScan", can be controlled via an analog output of the amplifier or any AD/DA converter supported by POTMASTER. The exposure of the camera should then be triggered via a digital output channel.
- DG-4/DG-5: The "DG-4/DG-5" can be controlled via the digital output of the EPC 10, ITC-16 or ITC-18. The exposure of the camera should then be triggered via an analog output channel.
- Lamba-10: The "Sutter filter wheel Lambda-10" can be controlled via the digital output of the EPC 10, ITC-16 or ITC-18. The exposure of the camera should then be triggered via an analog output channel.
- pti DeltaRAM: The "High-Speed Random-Wavelength Illuminator" can be controlled via the digital output of the EPC 10, ITC-16 or ITC-18. The exposure of the camera should then be triggered via a digital output channel.
- no device: No light source control.



For details on the `Imaging` Extension aka "SMARTLUX" please refer to the SMARTLUX manual.

5.4.6 DocuLUX


Enable `DocuLUX` features for POTMASTER here. Please be aware, that you need to buy the `DocuLUX` bundle (order number: 895286 or 895287) first.

5.4.7 Photometry

This will activate stimulation of a monochromator and acquisition via a photo-multiplier or photo diode.

You can choose four kinds of `Photometry` Extensions:

- TILL Photonics/Cairn OptoScan: "TILL Photonics" or "Cairn OptoScan", can be controlled via an analog output of the amplifier or any AD/DA converter supported by POTMASTER.
- DG-4/DG-5: The "DG-4/DG-5" can be controlled via the digital output of the EPC 10, ITC-16 or ITC-18.
- Lamba-10: The "Sutter filter wheel Lambda-10" can be controlled via the digital output of the EPC 10, ITC-16 or ITC-18.




- pti DeltaRAM: The “High-Speed Random-Wavelength Illuminator” can be controlled via the digital output of the EPC 10, ITC-16 or ITC-18.

For details on the Photometry Extension please refer to the chapter Photometry, 22 on page 237.


5.4.8 ElProScan

The ElProScan Extension can only be used with additional hardware. Further, our ELPROSCAN customers will get a special hardware configuration file. If you activate the extension without having the configuration file you will get an error message indicating that the *ElProScan Hardware file* is missing.

After the activation of the ElProScan Extension two additional windows will pop up:



Further, there is an additional menu entry in POTMASTER named ElProScan:



The functions and options of the new windows and the new menu are explained in detail in our ELPROSCAN manual.

5.4.9 Serial Out

This control is used to open and set up the serial port communication mode. If a so-called *Serial Communication* has been established between the computer and another device, POTMASTER can send strings over a serial port, but it will not receive instructions.

The options for the serial communication are:

- Off: No connected device.
- Comm 1...4: Any device that can receive strings through the serial port. Up to 4 ports can be defined and triggered.

When opening a serial communication, POTMASTER will allow you to configure the serial device. Make sure that the settings match on both communicating machines.

- Serial Port: No Port, Comm 1... Comm 8.
- Baud Rate: 300, 600, 1200, 1800, 2400, 3600, 4800, 7200, 9600, 19200, 57600, 115200, 128000 or 256000 bps.
- Stop Bit: 1.0, 1.5 or 2.0.
- Parity: No Parity, Even Parity or Odd Parity.
- Data Bit: 5, 6, 7 or 8 data bits.
- XOn/XOff: On or Off.
- Rts/Cts: On or Off.

To change all settings or leave the dialog, do one of the following:


- Edit Start String: Allows to enter a specific start string.
The *Start String* (Initialization String) is send once after opening the COM port.
- Open: Opens the configured port (icon text changes to "Close")
- To Notebook: Commands sent and received are written to the Notebook.
- Get Defaults: Sets the default settings given in POTMASTER.
- Undo: Cancels all changes without closing the dialog.
- Cancel: Cancels all changes and closes the dialog.
- Done: Saves all changes and closes the dialog.

If a serial out is active then command strings can be entered and sent via the I/O Control window (see chapter 16 on page 213). There is also a *Serial Out* event available in the Protocol Editor, 11.4.3.2 on page 143.



5.5 Files

In this section, the user has to specify the paths for several files. POTMASTER will use these paths when saving or retrieving files.



To rename a path, click on the button and set a new path.

Home Path: Sets the path to the folder `LookupTables` and to the key assignment file (`Potmaster.key`). It is also the path where the default files `DefPgf_v9.pgf`, `DefAnal.onl` and `DefProt.pro` are searched automatically as long as no configuration file was saved.

The path to the `Configuration` file is shown in the title of the `Configuration` window. This path can be changed by saving the `Configuration` under another path using the `SAVE...AS` function.

Data File: Sets the path to the data file (file extension: `*.dat`).

Protocol File: Sets the path to the protocol files (file extension: `*.pro`).

PGF Pool File: Sets the path to the Pulse Generator file (file extension: `*.pgf`).

Analysis File: Sets the path to the Analysis file (file extension: `*.onl`).

Solution Base: Sets the path to the solution data base files (file extension: `*.sol`).

Equations File: Sets the path to the equation file, which can contain equation formulas. This path can be cleared by use of the  button.

Batch Path: Path for batch file control protocols, e.g., `E9BatchIn` and `E9BatchOut`. This path can be cleared by use of the  button. For further information we refer to the `PATCHMASTER` Tutorial: "Controlling `PATCHMASTER`".

Dialog Files: Sets the path to the dialog files, in which button placements of the different windows can be stored in case they should deviate from the default. This path can be cleared by use of the  button.

Temp. Path: Optionally a path different from the home path can be specified for storing temporary data. This



path can be cleared by use of the  button.

I-gain Lookup: Path to the *Current Gain Lookup Table* if selected in the configuration (see chapter 5.7.2 on page 59). Only necessary for telegraphing amplifiers.

5.5.1 General Advice on Naming Folders and Files

Here is some advice to keep in mind when naming folders and files:

- The use of invisible characters and spaces is not recommended (a blank in a file name is very often overlooked).
- Names should not begin with a number, because some other applications, e.g., Igor, do not allow names with a number as the first character (exported Igor waves inherit characters of the data file).
- POTMASTER only shows the first 14 letters of the file name in the title of the Oscilloscope window.
- The first few letters of a name are the more important ones, because they ease file selection with the file selector. The file selector continuously selects the files while the user types.
- Experiment data should not be stored within the HEKA software folder. This will ease upgrading the software without moving the data.

5.5.2 Miscellaneous Settings

Save Settings File: The *Configuration* settings will be saved automatically on exit, thus overwriting the previous settings.

Ask for Data File: When this option is enabled, POTMASTER will ask for a data file upon start-up.

Make Bundle Files: When this option is selected, all files generated for a data set (e.g. the raw data *.dat, *.pgf, *.pul, ...) are combined into one *.dat file.


Verify Quit: When this option is enabled, POTMASTER will ask to save modified setting files (*.pgf, *.pro, *.onl) when you quit the program. When it is not enabled, the *Configuration* settings will not be saved on quit!

Save after Break: When this option is enabled, POTMASTER will save the data in a *Sweep* that have been acquired when the acquisition was terminated via the *Break* button. With this option disabled the entire *Sweep* will be discarded if during its acquisition the *Break* button was pressed.

Solution Base: The *Solution Base* is activated by default, therefore the default solution base (DefSolutionBase.sol) is loaded (see chapter 18 on page 221).

Auto Filename: Allows an automatic file name generation. The file names can be generated out of a set of special components.

- [DATE=<descriptor>] = descriptor compounded by "YYYY" or "YY", "MMM" or "MM", and "DD". Characters separating the blocks are copied to the name, e.g. [DATE=DD-MMM-YYYY] = 31-Dec-2015
- [DATE] or [DATE1] = DD-MMM-YYYY (e.g. "31-DEC-2015")
- [DATE2] = YYYY-MM-DD, international custom data format, directly sorts according to dates e.g. "2015-12-31"



☒ Save Settings File	☒ Verify Quit
☒ Ask for Data File	☐ Save after Break
☒ Make Bundle Files	☐ Solution Base
☐ Auto Filename	
Startup Protocol	
Experiment Number	1
Max. File Size Alert	500. MB



- [DATE3] = YYYYMMDD, analog to [DATE2], but without hyphens e.g. "20151231"
- [DATE4] = YYMMDD, e.g. "151231"
- [TIME] or [TIME1] = HH-MM-SS (e.g. "24-59-59")
- [TIME2] = HHMMSS, analog to [TIME1], but without hyphens (e.g. "245959")
- [USERxx] = username, "xx" is a number giving the maximal number of characters. No "xx" means maximal allowed characters, presently 32
- [000] = an incrementing counter, each digit being a placeholder. The given number is the starting value of the counter. The counter will increment until a unique filename is created
- [NAME] = a placeholder for the data file name

The components are composed to a file name by the insertion of an underscore "_".

A typical entry could look like: HERG_[DATE]_[001] = "HERG_31-Dec-2015_001.dat"

Do not use the following characters:

- path separators: "/", "\", ":" (unacceptable)
- extension separator: "." (unacceptable)
- blank: " " (compatibility problems, danger of oversight)

Startup Protocol: Here, you can enter the name of a protocol that is executed after the end of program initialization (i.e. after the configuration file is loaded, the amplifier initialized and the other settings activated). E.g., this allows for additional settings like the immediate activation of the second amplifier of an EPC 10 Double at start-up.


Experiment Number: This number can be used to identify experiments. It is automatically incremented when a new Experiment is generated with `File → New Experiment`.

Max. File Size Alert: This feature will not limit the size of any file written to disk, but can be used as a reminder so that no files are created that e.g., do not fit onto a CD. A warning is written to the Notebook if the size of the raw data file exceeds this value. Entering a very large number will - in practice - suppress any warnings.

The maximum file size is limited to 2 GB. If you exceed this number during data acquisition you have to consider data loss.



5.6 Display



5.6.1 Fonts and Colors

Fonts / Button Colors / Line Colors:

Colors and text fonts for the program layout can be selected here. These are global settings for all window dialogs and they are installed upon restart of POTMASTER. The colors and fonts are stored in the configuration file (*.set), i.e., they are independent of the dialog files (see chapter 2 on page 5).

Note: If dialog files are present, they will overwrite the Configuration settings.

Note: You can make the background colors of the windows dark (useful when doing light-sensitive experiments) by selecting the option **Button Colors**. In this case, you may also have to change the color of lines, like the **Trace Colors**, for example, using the option **Line Colors**.

5.6.2 Notebook, Display, and Analysis

Wipe at Start: If this option is activated the Analysis and Oscilloscope windows will be wiped before data acquisition.

Digits:

☒ Wipe at Start		
Digits: Notebook Time	6	8
Grid Digits: X Y	6	6
Graph-3D: dX dY	0.010	0.010

- **Notebook:** Numbers of digits for data display in the Notebook window (range 3 – 24).
- **Time:** Numbers of digits for the *Time* and *Timer* function in the Oscilloscope window (range 8 – 11). The maximal number is 11, this results in "hh:mm:ss.ms".

Grid Digits X|Y: Number of digits to be shown in grid annotation of the Oscilloscope and the stimulation template preview of the Pulse Generator (range 6 – 10).

Graph 3D dX|dY: Allows to specify the horizontal and vertical offset of subsequent Sweeps to be displayed in the 3D graphing mode (range 0.001 – 1.000). You can also enter 0 for one or both. Note that the *3D-Graph* feature has to be on for this option to take effect (see chapter 4.5 on page 29).

5.6.3 Show Options

The appearance of some windows can be configured using the following options:

Hide PGF-Bar in Control Window: The complete PGF pool in the Control window can be hidden.

Hide Protocol-Bar in Control Window: The complete protocol pool in the Control window can be hidden.

Show Fixed Control Protocols: In case the user does not want to scroll through the protocol pool, a specified number of protocols (0, 7, 15, 23, 31 or 39) can be shown as buttons in the Control window.

Show Amplifier Protocols: Protocols can be started via a button click from the Amplifier window. Here, the user can specify how many protocol buttons (0, 4, 8, 12, 16, 20 or 24) should be created in the Amplifier window. To see the new protocol buttons you have to enlarge the Amplifier window by activating the *Show All* tab.


Note: When using Amplifier Protocols please be aware of some restrictions (see chapter 11.3 on page 133).

Show Leak Pulses in PGF Editor: Enables the leak pulse settings in the PGF Editor.

Hide Charge Controls: Disables the *anodic charge* and *cathodic charge* controls in the Potentiostat window.

Hide Current Density Controls: Disables the *I-Initial(Density)* and *I-Range(Density)* controls in the Potentiostat window.

Hide PG611/PG618 Compensation Controls: Disables the compensation options in the Potentiostat window if the PG 611 or PG 618 is used.




5.7 I/O Control

Selected options in this section of the Configuration window will appear in the I/O Control window (see chapter 16 on page 213) and can be edited there.

Parameters such as e.g. *Temperature*, *Gain*, *R-series*, are used to customize the recording environment of a given experimental setup.

The checkbox on the left determines whether the parameter is displayed (and updated) in the I/O Control window.

For each I/O parameter you have to enter a default value and a source.



Default: The default values can be edited in this column.

Source: In this row one can determine how the value is obtained. For most of the parameters there are three alternatives:


- Default: The specified default value is taken.
- PG300 USB: Parameters are obtained directly from the given PG 300 USB amplifier, or PG 600 if using a PG 600 USB amplifier.
- AD-Channels: Parameters are sampled via a specified AD-channel.

list : Pressing this button writes all scaling and offset values for the I/O parameter in the Notebook window. Use this feature for getting a quick overview on the scaling of all your AD- and DA-channels.


5.7.1 AD/DA Input/Output Scaling

In the I/O Control pane of the Configuration window you can define the scaling methods for the DA- and AD-channels. Please note that the I/O Control window always displays the raw (unscaled) voltage values of the channels. The scaling of DA- and AD-channels can be used for stimulus output and for data acquisition via the Pulse Generator.


In case of an AD-channel you can choose between a *Scaling Factor* and a *Lookup table*:




For direct input of scaling factors, enter a *Scale Factor* and a *Scale Offset*:




Example: A temperature controller has a measure range from 10 °C to 50 °C and a corresponding voltage output range of +/- 10 V. The correlation between measured temperature and voltage output is linear (see figure below).




In this example, *Scaling Factor* should be set to "2" and *Scale Offset* should be set to "30". If the correlation between measured signal and voltage output is not linear then a *Lookup table* should be used. For input via a *Lookup table*, you have to choose the file from which the program will read the data.



In case of a DA-channel you have in addition the the possibility to apply a scaling based on an equation. The equation has to be entered here in the Control window, whereas the scaling has to be activated in the Pulse Generator (see chapter 10.10.1 on page 115).



The options *Lookup* and *Scaling* work similar to the methods described for the AD-channel scaling. In case you choose *Equation* the following dialog appears:



In the command field an equation can be entered, the calculated result is displayed in the *Results* field beneath.

Equation **User Defined** : A label for the equation can be entered in this field.

Select Equation : Already saved equations can be selected in this pop-up menu.

Save Equation : After defining a name for the entered equation it can be saved. After doing this the formula is available in the pop-up menu of *Select Equation*.

Note: The new equations is saved in a default *Equations.txt* file. If you want to change the file or the file name for your equations you have to this via the *Save* button.

? : Prints a help text on the equation syntax to the Notebook.

LOAD : Brings up a file selector to open an equation file.

SAVE : Defines the place and the name of the file in which your equations are stored.

For a detailed description on the equation syntax please refer to the chapter *Calculator and Equations*, 21 on page 231.

As was mentioned earlier, *Lookup tables* for commonly used amplifiers are supplied. If your amplifier is not one of those provided, a *Lookup Table* can be created as an ASCII file.

Note: Have a look for our video tutorial "Using Telegraphing Inputs" to learn how to create your own *Lookup table*.

Note: It takes CPU time to read AD-channels and to convert the data to the desired values. In order to minimize handling overhead, only activate those parameter fields that are actually used.

5.7.2 List of I/O Parameters

☐ Current Gain	100.0 mA/V	PG300_USE	list
☐ scale	AD-0	1.000	scale
	DA-0	1.000	scale
☐ User 1	V	0.000 V	Default
☐ User 2	V	0.000 V	Default
☐ User 3	V	0.000 V	Default
☐ User 4	V	0.000 V	Default
☐ User 5	V	0.000 V	Default
☐ User 6	V	0.000 V	Default
☐ User 7	V	0.000 V	Default
☐ User 8	V	0.000 V	Default
☐ User 9	V	0.000 V	Default
☐ User 10	V	0.000 V	Default
☐ Temperature	20.00 C	Default	
☐ Curr. Bandwidth	10.01 kHz	PG300_USE	
☒ OCP	0.000 V	Default	
☐ Electrode Area	1.000 cm2	Default	
☐ Show Solution			

scale:

- AD-Channel: If data *Traces* from AD-channels are acquired in a PGF sequence, the voltages can be scaled to convert them to appropriate units. For scaling you can use a fixed value or *Lookup tables* (see chapter 5.7.1 on page 57).



- **DA-Channel:** Free DA output channels that can be used for stimulation during a PGF sequence can be scaled with a factor entered in the *Value* field.

User 1, 2: There are two parameter fields reserved for user-specific assignments. These fields act like other parameter fields with the addition that parameter name and unit can be specified (e.g., "pH" and "U"). Name and unit are also stored in the output *.pul file on the level of each *Sweep*, so they can be changed during an experiment, if required.

Temperature: Temperature (from a recording device).

Curr. Bandwidth: Recording bandwidth.

OCP: If selected this setting displays the specified or actually measured Open Cell Potential in the I/O Control window.

Electrode Area: If selected, this setting displays the specified electroactive area of the working electrode in the I/O Control window. The value must be in cm^2 (!). This value is used to calculate the current densities.

Show Solutions: If the checkbox is active you can enter/change the number of the external and internal solutions in the I/O Control window.

5.7.3 Show Digital In / Out

Show Digital In: If the checkbox is deselected the Digital Inputs in the I/O Control window will be hidden.

Show Digital Out: If the checkbox is deselected the Digital Outputs in the I/O Control window will be hidden.


Show DA Controls: If the checkbox is deselected the drag-able DA-channel fields in the I/O Control window will be hidden.

Show AD Values: If the checkbox is deselected the AD-channel display fields in the I/O Control window will be hidden.


Show AD Channels: Via these checkboxes the user can deactivate any *AD Value* field in the I/O Control window.

Show Values: You can specify how many values (4, 8, 12, 16, 20, 24, 28 or 32) should be listed in the I/O Control window.

Name of Value: A user defined name can be specified for each value (right text field) in order to make the parameter input and output more meaningful. A list of all value names pops up if you select the left box.



5.8 Trace Assign



The *Trace Assignment* function is very useful if your number of input and output channels of a PGF is increasing. This function allows you a defined labeling of each acquired *Trace*. E.g. if you define acquisition channels for current, voltage, ADC input, LockIn and Leak, they get labeled according to their origin, i.e. *Imon-1*, *Vmon-1*, *Adc-0*, *Adc-1*, *LockIn_CM*, *LockIn_GM*, *LockIn_GS* and *Leak-8*, instead of simple *Trace* counts (e.g. *Trace 1...8*). It is also possible to assign several *Traces* with the same input signal, so you can define e.g. three times *Imon-1* but the compression rate for each channel can be set differently. The benefit comes with the consistent labeling of these *Traces* within POTMASTER (e.g. Analysis, Trace Properties, Oscilloscope etc.) which helps the user to keep the overview of the data.

In the *Trace Assign* dialog it is possible by default to assign labels for up to 16 *Traces* (*Trace 1...16*). If you increase the number of *Max. Shown Traces* in the *General* tab of the Configuration window this number will also increase.

For each *Trace Assignment* you have to select the type of channel:


Trace Count: Define a *Trace* number for the selected *Trace*:

- *Trace 1...32*

Amplifier: Define a current or voltage monitor for the selected *Trace*:

- *Imon*
- *Emon*

AD-Channel: Define an AD-channel for the selected *Trace*:



- *Adc-0...2*
- *Mux*
- *Adc-8...15*

Leak: Define a leak channel for the selected *Trace*:

- *Leak-1...32*

Digital: Define a digital channel for the selected *Trace*:

- *Dig-in(word)*
- *Dig-in 0...15*

Virtual: Define a virtual channel for the selected *Trace*:

- *Virtual-1...32*

LockIn: Define a *LockIn* channel for the selected *Trace*. The following channels are available:

- *LockIn_CM*
- *LockIn_GM*
- *LockIn_GS*
- *LockIn_Real(Y)*
- *LockIn_Imag(Y)*
- *LockIn_DC*
- *LockIn_Avg*
- *LockIn_CV*
- *LockIn_GP*
- *LockIn_Admit(Y)*
- *LockIn_Phase*
- *LockIn_Real(Z)*
- *LockIn_Imag(Z)*
- *LockIn_IMP|Z|*

Spectroscopy: Define a spectroscopy channel for the selected *Trace*. The following channels are available:

- *Chirp_Avg*
- *Chirp_Phase*
- *Chirp_Admit(Y)*
- *Chirp_Real(Y)*



- *Chirp_Imag(Y)*
- *Chirp_Real(Z)*
- *Chirp_Imag(Z)*
- *Chirp_Imp|Z|*

Photometry: Define a photometry channel for the selected *Trace*. The following channels are available:

- *Photo_W1*
- *Photo_W2*
- *Photo_W3*
- *Photo_R*
- *Photo_Ca*

Imaging(SmartLUX): Define an imaging channel for the selected *Trace*. The following set of channels is available for six times:

- *Image_W1*
- *Image_W2*
- *Image_W3*
- *Image_R*
- *Image_Ca*

ElProScan: Define an ELPROSCAN channel for the selected *Trace*. The following channels are available:

- *SF-ampl*
- *SF-phase*
- *z_pos*
- *z_slope*

S.W.Voltammetry: Define a square wave voltammetry channel for the selected *Trace*. The following channels are available:

- *I-forward*
- *I-reverse*
- *I-diff*
- *E-mean*

Undo


: The last change will be reset.

Reset

: Resets all *Trace* assignments to the default value.



5.9 Miscellaneous



Analysis: Sweeps | Results: The Analysis needs to allocate memory for storage of its results. The *Analysis Result* buffer is a 2-dimensional buffer (table).

- Analysis: Sweeps: Maximum Analysis results per Analysis function (default = 1024)
- Analysis: Results: Maximum Analysis functions (default = 36)

For example if you use the Analysis for chart like recording during your experiment. The experiment lasts for about 10 minutes and you analyze one set of *Analysis: Results* per second then the online buffer has space for 600 *Analysis: Sweeps*.

***Note:** The Analysis buffer is cleared after each Wipe.*

In case you select more Analysis functions in your Analysis Method than specified here, then all results exceeding this number will not be computed, hence not displayed!

Analysis Mode: Handling of *Analysis: Sweeps | Results*:

- Fixed: The number of *Analysis: Sweeps | Results* given above is fixed, i.e. limited to that number. For long chart like recordings and many Analysis results the number of Sweeps and results might have to be set to such high values that available memory become a problem.



- **Extend:** The number of *Analysis: Sweeps | Results* given above will be extended, if more data are acquired. With this mode selected the system might run into trouble when no further free memory is available.

Analysis | Max.Swp.Redraw: Defines the maximum number of *Sweeps* which can be displayed in an *Analysis* window.

Replay Scroll Rate: Defines the time interval when scrolling through the *Replay* tree.

Analysis: Compress Vectors: If this checkbox is selected, the data of the *Analysis* results get compressed to increase the drawing speed of the graphs in the *Analysis Windows*.

Analysis: Include Headers: If this checkbox is selected, the column headers of the *Analysis* results written into the *Notebook* is displayed. If you deactivate the checkbox the headers will not be shown.


Analysis: Write Sweep Index: If this checkbox is active, the sweep index for each *Sweep* is printed into the *Notebook* and exported.

Clear Comment: Clears the *Comment* field in the *Control* window after the *Comment* is stored with the data.





6 Oscilloscope Window



The Oscilloscope window is mainly used for monitoring the data. The *Traces* of a *Sweep* are displayed versus the time. Controls for display scaling and data handling are provided. The title of the window contains the currently active data file name.

6.1 Display


Zooming – Lassoing: It is possible to set the display scaling in the Oscilloscope window by "lassoing" a screen region while pressing the left mouse button. When you release the mouse button, the marked area will be set to fill the Oscilloscope screen. The scaling has to be done for each *Trace* separately.

If you want to display data that is outside the active screen area, you can either enter some values into *Y-scale* or *Start/End Time* or drag the mouse *outside* the active screen area. The display gains gets reduced by 20% as long as the mouse is outside, and stops when it moves back on the active display area. When the mouse is below or above the screen, the *Y-scale* is changed, and when the mouse is to the left or right of the screen, the *X-scale* is changed.

Cursor shapes: To accommodate different user preferences there are multiple cursor shapes available. Press the CTRL key while the "cross" cursor is displayed and you can cycle through all available cursor shapes.



6.2 Navigation



Group_Series_Sweep of Sweep: Currently active *Group*, *Series* and *Sweep* number within the data tree. The total number of *Sweeps* per *Series* is given after "of". The numbers are updated with every start of a new *Sweep* acquisition. During data replay these information are replaced by the information of the replayed data. Note that this can interfere with the above mentioned updating process if data are replayed in the wait time between data acquisition cycles.

6.3 Measurements on Traces

Measure : Allows to measure amplitudes and duration. With this option the user can measure the displayed data by pointing and dragging the mouse or by clicking on and . Two mouse-driven horizontal lines are provided to measure signal differences. By clicking on the button , the measurement data is written to the Notebook window.

Scan : Allows to measure signal amplitudes by scanning the individual data points. When selected, a marker is displayed on the *Trace*. This marker can be moved forward and backward by dragging with the mouse or by clicking on and . By clicking on the button , the measurement data are written to the Notebook window.

*Note: The functions "Measure" and "Scan" work on the selected Trace in the **Replay** window. If anything other than a Trace is selected in the **Replay** window, then the functions work on the Trace that is selected in the **Oscilloscope** window (right panel).*

6.4 Display Refreshing

Freeze : Freezes the present content of the Oscilloscope. Any change in display parameters, such as scaling etc. will unfreeze the display again, as does a **Wipe** or switching to the **Amplifier** window.

Wipe : Clears the Oscilloscope completely.

Repaint : Clears the Oscilloscope first, then redraws the last displayed *Trace*. This is useful e.g., if you had the *Overlay Sweep* option activated, but then decided that you only want to see the last *Sweep* of the *Series*.

6.5 Overlay Options

Overl.Swp : Displays all *Sweeps* of a *Series* without erasing the screen in between *Sweeps*. The next *Series*, however, will erase the screen.

Overl.Ser : Displays all incoming *Sweeps* without erasing the screen. This allows *Sweeps* of different *Series* to be computed. The screen can be wiped by pressing BACKSPACE.



6.6 Display Scaling

Trace Color: Displays the color of the selected *Trace*.

Trace 1...16 (default): Selects the *Trace* to which the scaling setting applies. The color of the selected *Trace* is displayed above the *Trace* control.

Note: When using the Trace Assignment function (see chapter 5.8 on page 61) the names of the Traces may be different.

Dig. Filter: Shows the currently selected bandwidth of a digital non-lagging Gaussian low-pass filter (i.e. software filter). The -3dB cutoff frequency is specified in Hertz. This allows to reduce the original bandwidth by the factor 0.0025 (1/400). It is for display purposes only; no changes to the data are performed.

Y-scale: For each *Trace* you can control the display scaling. The value of 1 corresponds to full scale of ± 10.24 V. This scaling does not affect the display of *Test Pulses*. Pressing + from the numeric keypad (take care not to have NUMLOCK activated!) increases the display scaling by a factor of 2 and - decreases it by a factor of 2. It is also possible to auto-adjust (*Auto Y-scale*) or fix the Y-scale (*Fixed Y-scale*) of the Oscilloscope. The reset button  sets the Y-scale to "1". For further information we refer to chapter 4.5 on page 29.

Note: This display scaling does not affect the display of the Test Pulse Current Trace. If you want to be able to scale the Test Pulse as well, activate the option Scale Test Pulse in the Configuration window and use I-scale in the Amplifier dialog. We advise against activating that option because one can easily overlook that the amplifier Gain is not correctly set, when the current Trace is scaled by the display scaling.

R: Resets the Y-scale to the default values.

Y-offs.: For each *Trace* you can control the offset of the zero line. Offsets of *Traces* may be between -1 and 1 relative to full scale of display (default = 0). SHIFT + and SHIFT - (from the numeric keypad) increase and decrease the display offset by 0.1.

: Clicking on the *Center* button automatically centers the active *Trace* on the screen. When pressing the SHIFT key together with this button all *Traces* will be centered.

Auto Swp: Pressing the *Auto Swp* button calculates the maximum Y-scaling factors (*Y-scale*, *Y-offs.*) to display the selected *Trace* of a *Sweep* as large as possible in the Oscilloscope. The basis for the calculation are the *Y-min* and *Y-max* values of the selected *Trace* of the *Sweep* (see chapter 14.1.5 on page 202).

Auto Ser: Pressing the *Auto Ser* button calculates the maximum Y-scaling factors (*Y-scale*, *Y-offs.*) to display the selected *Traces* of a *Series* as large as possible in the Oscilloscope. The basis for the calculation are the *Y-min* and *Y-max* values of the *Sweep Traces* of the *Series* (see chapter 14.1.5 on page 202).


Start Time / End: Section of the *Sweep* to be shown on the screen in % (*Start - End*).

R: The *Reset* button sets the full *Sweep* length (0 to 100 %).

Note: The full time scale provided for Sweep display is based on the longest Sweep within a Series. Alternatively, one can use the Fixed Scale option (see below).


Page: Page of display during replay of continuous data *Sweeps* or when time axis is chosen to be less than 100%. Clicking on the right/left arrow control will display the next/previous page of the current *Sweep*. Dragging the page number scrolls the data forward or backward; entering a page number will display that particular page. Page is highlighted whenever there is more than one page available.

R: The *Reset* button sets the page number to "1.0".





7 Potentiostat Window



The Potentiostat window is used for controlling, adjusting and displaying the amplifier settings. The window might differ depending on the type of amplifier you are controlling with POTMASTER. We will describe the window of the PG 310 USB and PG 610 USB in most detail and focus to special details only when describing the windows of other amplifiers.

7.1 Controlling the PG 310/390 USB - Main Functions


The PG310 USB Potentiostat window is used for controlling, adjusting and displaying the Potentiostat/Galvanostat operating modes.

The Potentiostat window has three different *Panes* (*Monitor*, *Tuning*, and *Show All*) to allow switching between the display of different sets of controls.

7.1.1 Monitor

On the Monitor pane the most important parameters and controls of the potentiostat are shown.





Mode Selection: Via these buttons the connection of the Potentiostat/Galvanostat to the cell can be controlled.



- **Cell:** Connects all the electrodes to the respective inputs of the Potentiostat/Galvanostat. In the potentiostatic mode the cell potential is controlled by the PG 310 USB, and the current flowing through the cell is displayed as *I-cell*. In the galvanostatic mode the current flow through the cell is controlled by the PG 310 USB, and the respective cell potential is shown in the *E-cell* display.
- **Standby:** In the *Standby* mode all the connections are switched off, and the potential can neither be read nor set.
- **OCP:** In the *OCP* (Open Cell Potential) mode the counter-electrode is disconnected. The cell is in the zero current state, and the open cell potential is exhibited in the *E-cell* display. The small *Set* insert button is used to set the initial potential value to the actual *OCP*.

Note: It is also possible to switch to the Standby mode by pressing the '0' key of the numeric keypad.

OCP-Set: The button OCP-Set is used to set the Initial Potential value to the actual Open Cell Potential. A new Open Cell Potential can be measured in the OCP Mode and is stored internally upon pressing the OCP-Set button.

E-cell: Cell potential monitor. The cell potential is displayed in relation to the chosen Zero Potential (E-zero).

I-cell: Direct current monitor. The current flowing through the cell is displayed.

E-initial/I-initial: Sets the desired initial value of the cell potential in the potentiostatic mode or the cell current in the galvanostatic mode. The range is +/-10 V and can be set by dragging the mouse or entering via the keyboard.

- Dragging with the mouse will change the potential in 1 mV steps.
- Pressing ← and → changes E-initial in steps of 10 mV.
- Pressing Option ← / → (MacOS) or Ctrl ← / → (Windows) will change E-initial in 1 mV steps.

In the galvanostatic mode, the item will display the I-initial (Initial current). The current is calculated by multiplying the current density and the electrode area, which can be specified either in the Configuration or Parameter window. It is also possible to set I-initial (Density) instead of I-initial. In the potentiostatic mode, the I-initial (Density) field is not activated.

Current Range: Sets the scaling of the current monitor output (*I-cell*). The ranges are 1 μ A to 2 A incremented by decades.

To set the *Current Range*, click with the mouse on the *Current Range* field and selecting the desired range in the pop-up window by tagging.

If the high gain pre-amplifier is connected to the PG 310 USB, the current ranges from 100 pA to 100 nA are selectable. The *Auto Range* mode is also available (see below).


If the HIGH CURRENT BOOSTER, HCB 50 is connected, the 50 A range is selectable.

If there is a saturation of amplifiers in the current monitor circuitry, a blinking box labeled "Clip" on the *Current Range* button is displayed. This is a warning that excess artifacts or noise may occur as a result of the saturation of amplifiers.

Note: The option I-Range (Density) is only available if you deactivate Hide Current Density in the Configuration window (see chapter 5.6.3 on page 55).

7.1.2 Tuning

Shows additionally the settings of Filters and Amplifier Mode.



Break+Standby: Corresponds to the buttons *Break* and *Standby*, thus stopping all data acquisition immediately and turning the cell to *Standby* mode.

Auto Filter: The *Auto Filter* function automatically adjusts the *Stimulus Filter*, *Voltage Filter* and the *Current Filter* in relation to the set *Sample Interval* and *Filter Factor* of your PGF.

Amplifier Mode: Allows a selection between potentiostatic and galvanostatic mode.

7.1.3 Filter Settings

Notch Filter: The PG 310 USB provides a notch filter for both the voltage and the current pathway. The notch frequency (50 Hz or 60 Hz) is preset at the factory as desired by the customer, and cannot be set by the user. If the notch filter is 'On', incoming signals at the power supply frequency will be effectively filtered out.

Control Amplifier Bandwidth	100 kHz
Notch Filters	Off
Stimulus Filter	Bypass
Voltage Filter	Max
Curr. Filter	Bessel 10.0 kHz

Control Amplifier Bandwidth, Stimulus Filter, Voltage Filter, Current Filter: These filters can be set either manually or automatically (The *Notch Filter* has to be switched manually by the user in any case).

For automatic filter mode, please activate the *Auto Filter* option in the Configuration window. The automatic settings dependence is that way:

- The automatic settings of the *Control Amplifier Bandwidth*, the *Voltage Filter*, and the *Current Filter* depend on the user defined *Sample Interval* and *Filter Factor* in the Pulse Generator (Free waveform) window.

- The automatic setting of the *Stimulus Filter* depends on the *Sample Interval*.

If the *Auto Filter* mode is selected, the filter settings are unable for manual settings and displayed with grayish numbers.

Note: *If you are using an electrochemical cell with a high capacity and low Ohmic drop, a problem may arise with the automatic setting of the Stimulus Filter. In combination with a high frequency Stimulus Filter a pulse-like stimulus may cause oscillations at the potential steps with the use of such cells. In this case, try to select the manual filter mode and reduce the frequency of the Stimulus Filter. The other three filter settings proposed in the Auto Filter mode can be left unchanged.*

Control Amplifier Bandwidth: The potential control amplifier can be filtered to avoid oscillations of the control circuit. Eight of the nine possible settings of the bandwidth (plus Max) are available:

10 Hz - 100 Hz - 300 Hz - 1 kHz - 3 kHz - 10 kHz - 30 kHz - 100 kHz - 300 kHz - Max

The possible settings depend on the Potentiostat version:

- If you are using a PG 310/390 with a 'A' or 'B' labeled serial number, the low frequency bandwidth of 10 Hz will not be available.
- If you are using a PG 310/390 with a 'C' or 'D' labeled serial number, the high frequency bandwidth of 300 kHz will not be available.

Note: *A transient problem may arise if the potential change is very fast in combination with a low bandwidth setting. There will be a rise time lag in the cell potential with respect to the given potential pulse. The time constant as the inverse of the bandwidth defines the shortest time domain over which the cell will accept a significant perturbation.*

Stimulus Filter: The stimulus can be filtered (2-pole Bessel) to reduce the amplitude of fast capacitance transients when the speed of potential changes is not critical. The following settings are available:

PG 310 USB: Bypass, 10 kHz - 7 kHz - 5 kHz - 3 kHz - 2 kHz - 1 kHz

Note: *In a ramp segment the stimulus change is step-like rather than line-like. The number of steps needed for representing the ramp is defined by the quotient of the segment length and the Sample Interval (see also chapter Pulse Generator, 10.6 on page 102). To improve the smoothness of the stimulus change, these steps are filtered to yield exponential, which reach about 63 percent of the step height during a term of one time constant. Hence, a filtered stimulus will become a chain of exponential, which exhibits a phase shift relative to the unfiltered stimulus. For instance, if you choose a filter time constant equal to the sample interval, the filtered stimulus will be a nearly smooth line with a phase shift of about 90 percent of the step length (i.e. of the Sample Interval).*

Voltage Filter: The voltage output (U-cell monitor) can be filtered by a 3-pole Bessel filter to reduce high frequency noise. The bandwidth can be set in four steps:

10 Hz - 100 Hz - 1000 Hz - Max

With the selection of Max no filtering is active and all frequencies are allowed to pass the AD channel.

Current Filter: Controls a switchable analog Bessel/Butterworth filter (4-pole) in the current monitor pathway. The menu provides the following settings:



- Bessel
- Butterworth
- Bypass (e.g., no filter active)

Dragging the mouse or entry on the keyboard allows bandwidth fine adjustment from 0.1-16 kHz in 0.1 kHz steps (Bessel) or 0.1 - 25.5 kHz (Butterworth). They differ in the following way:


- The *Bessel* setting is the best characteristic for general use. As such it is always used in automatic mode.
- The *Butterworth* response rolls off more rapidly with frequency and is useful mainly for power spectral analysis.

Under most conditions a 10 kHz bandwidth is more than ample, and the filtering reduces the high-frequency noise substantially.

Note: In automatic mode always the Bessel filter is used.

7.1.4 Show All

Shows all controls.



IR-Comp.: The series resistance compensation corrects for the voltage drop between the reference and working electrodes under conditions of high access resistance or high current flow between counter- and working electrode. In the 4 *Electrode Mode* the serial resistance between both reference electrodes is corrected. The compensation is based on the value of R-series, which can be changed by dragging the mouse or entering on the keyboard (range: 0 to 1 M) and will be effective only when *IR-Comp* is 'On' (see also PG310/390 manual, IR

Compensation).

Ext. Input: The `External Input` (front panel of the PG 310/390) is scaled by a selectable factor (range: 0.1x and 1x), to allow for different external stimulators. It is strongly recommended to set *Ext. Input* to 'Off' (i.e., equal to zero), if no external stimulator is connected to *Ext. Input*. This will prevent pick-up of external noise.

***Note:** The *E-Initial* (the initial potential) is not affected by changing the external scale factor. The scaling will only affect the initial potential, if the user sets the initial potential externally (e.g., with a stimulator or another computer).*

Reset: Selecting this button will reset the PG 310/390 to its initial default configuration. Reset is very useful for defining the initial state of the PG 310/390. It will reset the DA channels to zero. *E-Initial* (the initial potential) will be unchanged by resetting the PG 310/390.

Electrode Mode: Two separate cell setups are supported by the PG 310/390:

- 3 Electrode Mode
- 4 Electrode Mode

The *3 Electrode Mode* consists of a working, a counter and a reference electrode. In this mode the second reference electrode input (Reference II) on the front panel of the PG 310/390 is short-circuited to the working electrode, so that the potentiostat controls the voltage between the working and reference electrodes. This commonly used mode is the default setting.

The *4 Electrode Mode* provides two distinct reference electrodes (Reference I and Reference II on the front panel of PG 310/390), between which the potentiostat controls the voltage.

Auto Range: Before starting the acquisition or if the potentiostat is set to *Cell*, POTMASTER can adjust the current range to higher or lower ranges if the flowing current is out of bounds.

Specifies the lower threshold of current within the given current range in percentages of full range. If the current reaches the *Min* percentage of the actual current range, the next lower current range is selected and the current resolution increases. The minimum value for *Min* is 1%, the maximum value for *Max* is 100%.

Specifies the upper threshold of current within the given current range in percentages of full range. If the current reaches the *Max* percentage of the actual current range, the amplifier switches to the next higher current range. The current resolution decreases. The values for *Max* range between 1% and 100%.

***Note:** Auto Range is also possible with the extended current ranges of the external preamplifier and the High Current Booster.*

The desired thresholds of current within a given current range are defined in the *Min* / *Max* boxes in percentages of full range. For instance, a *Max* setting with 90% will force POTMASTER to switch to the next higher current range if the current exceeds 90% of the actual range.

Ampl. Mode switch: set E/I-init: This feature ensures that by switching the amplifier mode between potentiostatic and galvanostatic, no current or voltage leap will occur. Thus, in the potentiostatic mode the actual current is measured and applied to the cell after the amplifier mode is switched to galvanostatic. This is the default setting.



If the function Ampl. Mode switch: keep E/I-init is chosen, current or voltage leaps may occur and affect the electrochemical processes at the working electrode.


E-initial range: Via this drop-down menu you define the voltage range (working range) of your potentiostat/galvanostat. When reducing the voltage range the resolution of your signal gets better.

The following options are available:


- $\pm 10\text{V}$
- $\pm 5\text{V}$
- $\pm 2\text{V}$

Amplifier Protocols: If you activate the option Show Amplifier Protocol in the Configuration window (Display tab) a defined number of Protocol buttons appear in the Potentiostat window (see 5.6.3 on page 55). In this example we show 4 so-called *Amplifier Protocols*. Use the Assign button to assign a specific protocol to the button.

7.1.5 Charge Controls




Two more fields for anodic and cathodic charge will be shown in the potentiostat window under I-cell displays (see upper red box in the screenshot below). The anodic and cathodic charge will be displayed continuously until the Charge Reset (see lower red box) is pressed. If pressed, the charge will be zeroed, and the charge measurement will then continue.




7.1.6 Current Density Control

In case you want to address the current readings as density rather than raw current values, you can activate the "Current Density Controls".

The Current Density Control items can be activated by unchecking the option "Hide Current Density Controls" on the pane "Display" of the POTMASTER Configuration window.



The fields I-initial (Density) and I-Range (Density) will be added below the standard Initial and Current Range controls. The I-Range (Density) corresponds to the selected Current Range and is scaled in accordance to the Electrode Area, which is shown below.



The *Current Range* is calculated by multiplying the *I-range (Density)* and the *Electrode area*. Alternatively, the user can choose the range of *I-Range (Density)* instead of the *Current Range*. The unit of the *I-Range (Density)* is labeled with a prime, e.g., 'A'.

In the galvanostatic mode, the item will display the *I-initial* (Initial current). The current is calculated by multiplying the *Current Density* and the *Electrode Area*. It is also possible to set *I-initial (Density)* instead of *I-initial*. In the potentiostatic mode, the *I-initial (Density)* field is *not* activated.

Electrode Area: Showing the *Electrode Area* can be activated by checking its box on the pane "I/O Control" of the POTMASTER Configuration window. Here you can enter the size of the electrode area. The entered value influences the *I-Range Density*.

Note: This option is only available if *Hide Current Density Controls* in the *Configuration* is deactivated (see chapter 5.6.3 on page 55).

Note: When entering an electrode area larger than "1 cm²" you need to clear that value with your current size defined in the PGF sequence. E.g. if you want to inject 100 mA current and your electrode area is "3 cm²" you need to enter "300 mA" in the PGF sequence.


7.2 Controlling the PG 610/690 USB - Main Functions

The PG610 USB Potentiostat window is used for controlling, adjusting and displaying the Potentiostat/Galvanostat operating modes.

The Potentiostat window has three different *Panes* (*Monitor*, *Tuning*, and *Show All*) to allow switching between the display of different sets of controls.

7.2.1 Monitor

On the Monitor pane the most important parameters and controls of the potentiostat are shown.



E-cell: Cell potential monitor. The cell potential is displayed in relation to the chosen Zero Potential (E-zero).

I-cell: Direct current monitor. The current flowing through the cell is displayed.

Mode Selection: A drop down menu allows to select the operation mode.

- **StandBy:** All electrodes are disconnected and the potential can neither be read nor set.
- **OCP:** The counter-electrode is disconnected and the cell is in the zero current state. The potential between working electrode and reference electrode is measured and shown as open cell potential in the E-Cell display.
- **Potentiostatic:** Connects all the electrodes to the respective inputs of the PG 610 USB. In the potentiostatic mode the cell potential is controlled by the PG 610 USB, and the current flowing through the cell is displayed as I-Cell.
- **Galvanostatic:** Connects all the electrodes to the respective inputs of the PG 610 USB. In the galvanostatic mode the current flow through the cell is controlled by the PG 610 USB, and the respective cell potential is shown in the E-Cell display.

Note: It is also possible to switch to the Standby mode by pressing the '0' key of the numeric keypad.

OCP-Set: The button OCP-Set is used to set the Initial Potential value to the actual Open Cell Potential. A new Open Cell Potential can be measured in the OCP Mode and is stored internally upon pressing the OCP-Set button.

Mode Switching Behavior: In the field to the right of the OCP-Set button you find a selection for the mode switching behavior. You can choose between:

- **Last Init:** When switching between potentiostatic and galvanostatic mode, the amplifier will use as Initial value (E-Init or I-Init) the value that was set previously in the respective mode.
- **Set Zero:** When switching between potentiostatic and galvanostatic mode, the amplifier will use zero current or voltage as initial value in the new mode.



- **Gentle Switch:** When switching from potentiostatic to galvanostatic mode, the system reads the actual current monitor signal and uses this current as initial current (I-Init). When switching from galvanostatic to potentiostatic mode, the system reads the actual voltage monitor signal and uses this voltage as initial potential (E-Init).

E-initial/I-initial: Sets the desired initial value of the cell potential in the potentiostatic mode or the cell current in the galvanostatic mode. The range is ± 10 V and can be set by dragging the mouse or entering via the keyboard.

- Dragging with the mouse will change the potential in 1 mV steps.
- Pressing $\leftarrow$ and $\rightarrow$ changes E-initial in steps of 10 mV.
- Pressing Option $\leftarrow$ / $\rightarrow$ (MacOS) or Ctrl $\leftarrow$ / $\rightarrow$ (Windows) will change E-initial in 1 mV steps.

In the galvanostatic mode, the item will display the I-initial(Initial current). The current is calculated by multiplying the current density and the electrode area, which can be specified either in the Configuration or Parameter window. It is also possible to set I-initial (Density) instead of I-initial. In the potentiostatic mode, the I-initial (Density) field is not activated.


Current Range: Sets the gain of the current measuring path. The ranges span 20nA to 100mA. To set the current range, click with the mouse on the Current Range button and selecting the desired range in the pop-up window by tagging.

If the High Current Booster HCB 50 is connected, the 50 A range is selectable.

If there is a saturation of amplifiers in the current monitor circuitry, a blinking box labelled Clip on the Current Range button is displayed. This is a warning that excess artifacts or noise may occur as a result of the saturation of amplifiers.

7.2.2 Tuning

Shows additionally the settings of Filters and Amplifier Mode.



Voltage Range: 4 different voltage ranges are provided: ± 10 V, ± 2 V, ± 200 mV, ± 20 mV. In case your voltage does not exceed a certain range, the resolution of the voltage measurement can be improved by choosing the smallest voltage range that just includes the expected signals.

***Note:** The POTMASTER software provides a feature that automatically adjusts the voltage range in accordance to a FreeWaveForm Generator sequence (*.pgf sequence) at the time of execution*

of the sequence. This option can be activated by checking “Stimulus Ranging” from the menu “PG610_USB”.

DC Subtraction: The total current gain is given by the “feedback gain” and a “secondary gain”. The PG 610 USB provides 3 feedback gain ranges. The basic feedback current ranges, which have a secondary gain of 1, are: 100 mA, 1 mA, and 10 μ A. The additional current ranges in between have secondary gains of 5, 50, and 500. A DC current/voltage can be subtracted from the signal in a current range with secondary gain of 1, allowing to use a current range with larger secondary gain subtraction. If the Subtract button is not active, the I-dc field shows the DC offset value. After activating the Subtract button (dark red, if activated), automatically the actual DC offset will be read and permanently subtracted from the current signal path. The subtracted value is shown in the field on the right to the Subtract button. The remaining offset is still shown in the I-dc field.

The I-Cell/E-cell displays as well as the recordings always show the sum of subtracted and remaining signals.

Break + Standby: Corresponds to the buttons Break and Standby, thus stopping all data acquisition immediately and turning the PG 610 USB to standby mode.

7.2.3 Filter Settings

Auto Filter: If activated, the Stimulus Filter and Output Filters are set in dependence of the Sample Frequency and the Filter Factor defined in the Pulse Generator of POTMASTER. The fields for Stimulus Filter and Output Filter are blocked for entries (greyed out) but show the current filter setting.

Control Amplifier Bandwidth: Sets the bandwidth of the main control loop. Typically, the bandwidth should be set to a value larger than the expected fastest change in the signal. Eight possible settings are available:

100 Hz, 300 Hz, 1 kHz, 3 kHz, 10 kHz, 30 kHz, 100 kHz and 300 kHz.

The Control Amplifier Bandwidth is not changed by the Auto Filter function.

***Note:** A transient problem may arise if the potential change is very fast in combination with a low bandwidth setting. There will be a rise time lag in the cell potential with respect to the given potential pulse. The time constant as the inverse of the bandwidth defines the shortest time domain over which the cell will accept a significant perturbation.*

Stimulus Filter: The output of the digital scan generator is filtered by a 2-pole Sallen-Key Filter. It is used to e.g. smooth a ramp type stimulus or limit the rise time in a voltage step stimulus. Eight cut-off frequencies are available:

20 Hz, 600 Hz, 2 kHz, 6 kHz, 20 kHz, 60 kHz, 200 kHz and bypass

Output Filter: Filters the current and voltage signals with a 4-pole Bessel Filter before digitization. Seven cut-off frequencies are available:


20 Hz, 600 Hz, 2 kHz, 6 kHz, 20 kHz, 60 kHz and 200 kHz

If the Auto Filter mode is selected, the Stimulus Filter and Output Filter settings are disabled for manual change but display the actual setting with grayish numbers.

7.2.4 Show All

Shows all controls.





IR-Comp.: The series resistance compensation corrects for the voltage drop between the reference and working electrodes under conditions of high access resistance or high current flow between counter- and working electrode. In the 4-Electrode Mode the serial resistance between both reference electrodes is corrected. The compensation is based on the value of R-series, which can be changed by dragging the mouse or entering on the keyboard (range: 0 to 1 M) and will be effective only when IR-Comp is On.

Electrode (Cell) Mode: Two separate cell setups are supported:

- 3-Electrode Mode
- 4-Electrode Mode

The 3-Electrode Mode consists of a working, a counter and a reference electrode. In this mode the second reference electrode input (Reference II) on the front panel of the PG 610 USB is short-circuited internally to the working electrode, so that the potentiostat controls the voltage between the working and reference electrodes. This commonly used mode is the default setting. The 4-Electrode Mode provides two distinct reference electrodes (Reference I and Reference II on the front panel of PG 610 USB), between which the potentiostat controls the voltage.

Reset: Selecting this button will reset the PG 610 USB to its initial default configuration. Reset is very useful for defining the initial state of the PG 610 USB when recording a macro. It will reset the DA channels to zero. E-Initial (the Initial Potential) will be unchanged by resetting the PG 610 USB.

Auto Range: The PG 610 USB can automatically find an appropriate current range by analyzing the DC current and the noise in the digitized signal.

Min: Specifies the lower threshold of current within the given current range in percentages of full range. If the current reaches the Min percentage of the actual current range, the next lower current range is selected and the current resolution increases. The minimum value for Min is 1%, the maximum value for Max is 100%.

Max: Specifies the upper threshold of current within the given current range in percentages of full range. If the current reaches the Max percentage of the actual current range, the amplifier switches to the next higher current range. The current resolution decreases. The values for Max range between 1% and 100%.




Rms: The noise in a signal is typically a good measurement of the capacitive load connected to the working electrode. With a large noise (capacitive load value) a larger current transient will be expected upon a potential change. Hence with a larger Rms value a larger current range will be automatically selected.


In order to find a good starting current range, activate AutoRange until it has selected a range and then deactivate it again. During the acquisition of a data sweep the current range stays constant.

Amplifier Protocols: If you activate the option `Show Amplifier Protocol` in the Configuration window (Display tab) a defined number of Protocol buttons appear in the Potentiostat window (see 5.6.3 on page 55). In this example we show 4 so-called *Amplifier Protocols*. Use the `Assign` button to assign a specific protocol to the button.

7.2.5 Charge Controls



Two more fields for anodic and cathodic charge will be shown in the potentiostat window under I-cell displays (see upper red box in the screenshot below). The anodic and cathodic charge will be displayed continuously until the Charge Reset (see lower red box) is pressed. If pressed, the charge will be zeroed, and the charge measurement will then continue.



7.2.6 Current Density Control

In case you want to address the current readings as density rather than raw current values, you can activate the "Current Density Controls".

The Current Density Control items can be activated by unchecking the option "Hide Current Density Controls" on the pane "Display" of the POTMASTER Configuration window.



- ☒ Hide Charge Controls
- ☒ Hide Current Density Controls

The fields *I-initial (Density)* and *I-Range (Density)* will be added below the standard Initial and Current Range controls. The *I-Range (Density)* corresponds to the selected Current Range and is scaled in accordance to the Electrode Area, which is shown below.

The screenshot shows the PG610_USB software interface with the 'Tuning' tab selected. The 'I-initial (Density)' and 'I-Range (Density)' fields are highlighted with a red box. The 'Electrode Area' field is also highlighted with a red box. The 'I-Range (Density)' field is set to 200.0 mA, and the 'Electrode Area' is set to 0.50 cm².

The *Current Range* is calculated by multiplying the *I-range (Density)* and the *Electrode area*. Alternatively, the user can choose the range of *I-Range (Density)* instead of the *Current Range*. The unit of the *I-Range (Density)* is labeled with a prime, e.g., 'A'.

In the galvanostatic mode, the item will display the *I-initial* (Initial current). The current is calculated by multiplying the *Current Density* and the *Electrode Area*. It is also possible to set *I-initial (Density)* instead of *I-initial*. In the potentiostatic mode, the *I-initial (Density)* field is *not* activated.

Electrode Area: Showing the *Electrode Area* can be activated by checking its box on the pane "I/O Control" of the POTMASTER Configuration window. Here you can enter the size of the electrode area. The entered value influences the *I-Range Density*.

Note: This option is only available if *Hide Current Density Controls* in the *Configuration* is deactivated (see chapter 5.6.3 on page 55).

Note: When entering an electrode area larger than "1 cm²" you need to clear that value with your current size defined in the PGF sequence. E.g. if you want to inject 100 mA current and your electrode area is "3 cm²" you need to enter "300 mA" in the PGF sequence.


7.3 Controlling the PG 340 USB - Special Functions

With the POTMASTER software you can also control the PG 340 USB Bipotentiostat/Galvanostat.

The main feature of the PG 340 USB Ring/Disc Potentiostat/Galvanostat is its design as a double potentiostat which allows to control two independent working-electrodes. This not only provides the tools to control rotating ring-disc electrodes, which are almost essential for the thorough study of mechanisms and kinetics of electrochemical reactions. It also allows the control of microelectrodes and ultra-microelectrodes connecting two external preamplifiers and makes innovative microelectrochemical technologies possible as generator- and collector-electrode mode for a Scanning Electrochemical Microscope (SECM).

The PG 340 USB allows to set a sample substrate to a well defined potential versus a reference electrode while recording currents in the lower pA range at an ultramicroelectrode in the same solution. Using two current channels strongly different currents can be simultaneously recorded. The bipotentiostat PG 340 is therefore the heart of the HEKA ELPROSCAN, the Electrochemical Probe Scanner which is much more than a SECM.

The Potentiostat window of PG 340 USB is amended with some buttons and functions for the two working electrodes, in comparison to the normal Potentiostat window (see chapter PG 310/390 Potentiostat/Galvanostat, 7.1 on page 71).



7.3.1 Electrode Conditions

The control between the two channels of the potentiostat can be switched by the buttons Amplifier-1 and Amplifier-2. The active channel is red.

- Amplifier-1 – all settings in the Potentiostat window apply to the *DISK* electrode
- Amplifier-2 – all settings in the Potentiostat window apply to the *RING* electrode

7.3.2 Potential/Current Settings

Almost all functions are the same as for the standard PG 310/390, see chapter *Potentiostat Window*, 7.1 on page 71. However, there are some small differences to be taken into account.

All settings apply to the selected electrode, e.g., *DISK* or *RING*.

Exceptions are:

- *Standby*, *Cell* and *OCP*, which act on the *DISK* and the *RING* electrode.
- The galvanostatic mode can only be set with the *DISK* electrode.
- *Control Amplifier Bandwidth* filtering, which acts on the *DISK* and the *RING* electrode.
- *Stimulus Filter*, which is active only with the *DISK* electrode.

7.3.3 Additional Settings

The IR-compensation in the *3 Electrode Mode* as well as in the *4 Electrode Mode* applies only to the *DISK* electrode.

7.3.4 Electrode Mode

Cell Mode:

Two separate cell setups are supported by the PG 340:

- 3 Electrode Mode
- 4 Electrode Mode

The *3 Electrode Mode* consists of a working (*DISK* and *RING*), a counter and a reference electrode. In this mode the second reference electrode input (Reference II) on the front panel of the PG 340 is short-circuited to the working electrode (*DISK*), so that the potentiostat controls the voltage between the working (*DISK* and *RING*) and reference electrodes. This commonly used mode is the default setting.

The *4 Electrode Mode* provides two distinct reference electrodes (Reference I and Reference II on the front panel of PG 340), between which the potentiostat controls the voltage.




7.4 Controlling the PG 618 USB - Special Functions

With the POTMASTER software you can also control the PG 618 USB Bipotentiostat/Galvanostat.

The main feature of the PG 618 USB Bipotentiostat/Galvanostat is its design as a bipotentiostat which allows to control two independent working-electrodes. Two different potentiostat boards make the PG 618 USB the first asymmetric bipotentiostat in the market. It is optimized for the use of microelectrodes and ultra-microelectrodes using an external preamplifier on one channel and controlling larger electrodes using the second channel.

The PG 618 USB allows to set a sample substrate to a well defined potential versus a reference electrode while recording currents in the lower pA range at an ultramicroelectrode in the same solution. Using two current channels strongly different currents can be simultaneously recorded. The bipotentiostat PG 618 USB is therefore the heart of the HEKA ELPROSCAN, the Electrochemical Probe Scanner which is much more than a SECM.

The Potentiostat window of PG 618 USB is amended with some buttons and functions for the two working electrodes, in comparison to the normal Potentiostat window (see chapter PG 610/690 USB Potentiostat/Galvanostat, 7.2 on page 79).



7.4.1 Electrode Conditions


The control between the two channels of the potentiostat can be switched by the buttons Amplifier-1 and Amplifier-2. The active channel is shown in red.

- Amplifier-1 – all settings in the Potentiostat window apply to the *Amplifier-1* electrode
- Amplifier-2 – all settings in the Potentiostat window apply to the *Amplifier-2* electrode

7.4.2 Settings of Amplifier-1

All functions of Amplifier-1 are the same as for the standard PG 610/690 USB, see chapter Potentiostat Window, 7.2 on page 79.

7.4.3 Settings of Amplifier-2



Voltage Range: The Amplifier-2 provides the voltage ranges ± 1 V (default) and ± 5 V.

Note: In the voltage range ± 5 V the current ranges 200 μ A and smaller are disabled.

Break + Standby: Corresponds to the buttons Break and Standby, thus stopping all data acquisition immediately and turning the Amplifier 1 to Standby mode.

Auto Filter: If activated, the Stim.Filter as well as Filter1 and Filter2 are set in dependence of the Sample Frequency and the Filter Factor defined in the Pulse Generator (Free Waveform Generator) of POTMASTER. The fields for Stimulus Filter, Filter1 and the cut-off frequency of Filter2 are blocked for entries (greyed out), but show the current filter setting.

V-zero: Allows to adjust the offset potential in the stimulus pathway.

Auto: Executes an automatic adjustment of the stimulus offset potential to yield zero current output.

7.4.4 Filter Settings of Amplifier-2

The Amplifier 2 provides two analog filters. Filter 1 is a 5-pole Bessel filter which acts as prefilter in the signal pathway. Filter 2 is a tunable 4-pole filter with selectable Bessel or Butterworth characteristic. The Filter 2 can be applied either to the current or to the voltage signal.

Filter 2: Selection of the Filter pathway and characteristic.

- I_Bessel
- I_Butterworth
- Bypass
- V_Bessel
- V_Butterworth

In the number field to the right you can enter the cut-off frequency in units of kHz.

Filter 1: You may select between the following filter settings:

- Bessel 100kHz
- Bessel 30kHz
- Bessel 10kHz
- Bessel 15Hz

Stimulus Filter: The stimulus signal which is generated by the DA converter can be filtered with a 10kHz or 100kHz filter with Sallen-Key characteristics.

7.4.5 Additional Features of Amplifier-2

On the pane “Show All” of the potentiostat, the following items will be shown if the Amplifier 2 is active:

2 Electrode Mode	Reset
Probe Offset	2048
CC-Stim. Offset	2048
Ref. Gain Trim	2550

Electrode (cell) Mode: Two separate cell setups are supported by the Amplifier 2 of PG 618 USB Potentiostat:

- 2 Electrode Mode
- 3 Electrode Mode

The 2 Electrode Mode consists of a working and a combined reference/counter electrode. The working electrode is connected to the input of the preamplifier (S-Probe) and the reference/counter electrode is connected to the GND jack on the Preamplifier (or at the Signal Ground of the main amplifier unit). The potential is applied to the working electrode with reference to the reference/counter electrode. The 2-electrode mode can be used, if the currents are small and they won't affect the potential of the reference electrode too much. In 3 Electrode Mode, the counter electrode is separated from the reference electrode. Counter is connected to GND jack and the reference electrode to the Bath Sense jack at the side panel of the Preamplifier. The potential drop between reference electrode and counter electrode is measured and added to the potential, which is applied to the working electrode.

Reset: Selecting this button will reset the PG 618 USB to its initial default configuration. Reset is very useful for defining the initial state of the PG 618 USB, when recording a macro. It will reset the DA channels to zero. E-Initial (the Initial Potential) will be unchanged by resetting the PG 618 USB.

The following three parameters are calibration parameters, which can be accessed for fine tuning. Normally, such parameters don't need to be changed.

Note: The parameters will be reset to the last calibration values after startup of POTMASTER. Hence you can't permanently change them. If you want to set those values routinely, you may write your custom protocols to do this automatically.

***Note:** Changing these values will affect the stimulus and reading of current and voltage signals. Hence, you may falsify your results!*

Probe Offset: Control for adjustment of the offset potential at the output of the current to voltage converter of the preamplifier. Each feedback range has its own Probe Offset value, which is stored in the calibration settings of the amplifier. After startup of the POTMASTER software the calibration value for the probe offset will be read




and set. You may fine tune the offset for certain applications by changing the DAC value to the right. Typically, the probe offset is adjusted with shielded preamplifier input (no current flowing) for zero reading at the current monitor (I-cell) of amplifier 2.

CC-Stim.Offset: In the galvanostatic mode the CC-Stim.Offset is activated. CC-Stim.Offset stands for “current clamp stimulus offset” whereas current clamp is equal to galvanostatic mode. This control is comparable to the V-zero adjustment in potentiostatic mode.

Ref. Gain Trim: In 3-Electrode Mode the reference potential of Amplifier 1 is measured and added to the stimulus potential of Amplifier 2. That means, with respect to the potential the Amplifier 2 is floating on the potential of Amplifier 1. That is, because both amplifiers share the same Ground/Counter electrode. The pathway of adding the reference potential to the stimulus of Amplifier 2 contains a gain adjustment amplifier. This gain is adjusted for unity gain to prevent any crosstalk between both amplifier stimuli.



8 Control Window



From this window, protocols or Pulse Generator sequences are started or stopped. It allows comments to be entered and it provides timing information. The number of buttons for executing protocols can be set in the `Display` section of the Configuration dialog (see chapter 5.6.3 on page 55).

8.1 Information about the Experiment

State: Current state of the experiment. All possible states are listed below:

- "aq: waiting": POTMASTER is waiting for the user to click on the *Resume* button during an acquisition (because the *Wait* button was activated by the user).
- "acquiring": Acquiring data in the *Interactive* mode.
- "ex: waiting": POTMASTER is waiting for the user to click on the *Resume* button during a protocol (because of a *Wait* event with the wait type "Resume Icon" or "wait for key").
- "gap-free": Acquiring data in the *Gap-free* mode.
- "idle": All else.
- "locked": User is locked out (PATCHMASTER PRO only).
- "noise": Measures noise.
- "recording": Recording a macro.
- "run macro": Runs a macro (may be part of a protocol).
- "run protocol": Runs a protocol.
- "test": *Test Pulse* is running.
- "test (demo)": *Test Pulse* in *Demo* mode.



Time: Actual system time measured since midnight.

Timer: This item functions as a stopwatch or timer called *TimerTime*. It can be reset at any time by clicking on the *Set* item or by pressing T. It may be useful to keep track of the experiment (e.g., to monitor the time spent in a specific experiment mode). The *TimerTime* is updated also during the *Series* execution. The timer value is stored at the beginning of the *Sweep* acquisition. The internal *Timer* tick corresponds to 1 ms. When you click on *Set*, the *Timer* will be reset to zero and started again.

Comment: Comment to the currently active *Series*. This field can be edited. It will result in a modification of the text of the present *Series* if a file was opened with write permission. This means that one has to enter a text to a *Series* *after* it has been acquired. The new comment will be copied into all incoming new *Series* until a new text is entered.

8.2 Controlling Data Acquisition



Store: Store is used to save the data that will be acquired from now on. A dialog window opens and you have to enter the name of the new data file.

In case *Store* is not activated but the file is opened with write permission, the last *Sweep* will be stored temporarily. This *Sweep* is marked with a "#" in its label. It will be deleted when:

- the next *Sweep* is acquired
- or a new *Group* is created or
- the file is closed.

Break: Break is used to stop *Series* acquisition or replay. The *Break* flag is reset, when the next target is displayed or executed. If *Break* is pressed during acquisition of a *Sweep*, this particular *Sweep* will not be completed and the data acquired for this *Sweep* thus far will be discarded. All previous *Sweeps* of the *Series* will be saved; thus, there can be a *Series* with fewer *Sweeps* than specified in the Pulse Generator. The *Break* button can also be used in *Replay* mode when performing a lengthy operation.

Stop: This button is used similarly to *Break*. However, during acquisition of a *Sweep*, this particular *Sweep* will be completed. The *Stop* button can also be used in *Replay* mode when performing a lengthy operation.

Next: At the end of the actually acquired *Sweep* the *Series* is stopped and the next protocol event will be executed.

Wait: This button is used to pause *Series* execution. Its command key is W. *Series* execution is resumed by a click on *Resume* or by R. *Wait* will become effective after completion of the currently acquired *Sweep*. The *Wait* button can also be used in *Replay* mode when performing a lengthy operation; e.g., it allows inspecting a particular *Sweep*, when replaying a collection of *Sweeps*.

Resume: Opposite of *Wait*. Press *Resume* to continue data acquisition or replay.


Average: By use of this function, a PGF sequence can be repeated several times. Only the average is stored to disk. Please enter the number of averages acquired for one *Sweep*.

Note: It is not mandatory but advised to use even numbers of averages in order to avoid possible problems with the Alternating Leak Averaging feature (see chapter 10.15 on page 128).

Note: The Average function uses the Trace buffers for temporary storage. Thus, the average is limited to 4 Traces per Sweep. Also, it will overwrite whatever is in the Trace buffers at time of acquisition.



8.3 Starting Pulse Generator Sequence



If you have not activated the option *Hide PGF-Bar in Control Window* on the **Display** tab of the **Configuration** (see chapter 5.6.3 on page 55) then you will see a row with buttons listing the various sequences of the loaded Pulse Generator file (*.pgf). Six buttons are shown at a time. But the row of buttons can be scrolled using the arrow buttons on the left and right side.

- A single click on one of those buttons will start the respective Pulse Generator sequence.
- By entering a number into the field to the right of the button row, the sequence with this index is started.

8.4 Starting a Protocol



If you have not activated the option *Hide Protocol-Bar in Control Window* on the **Display** tab of the **Configuration** (see chapter 5.6.3 on page 55) then you will see a scroll-able row with buttons listing the various protocols of the loaded protocol file (*.pro). Six buttons are shown at a time. But the row of buttons can be scrolled using the arrow buttons on the left and right side. If you have activated the option *Show fixed Protocols in Control Window* on the **Display** tab of the **Configuration** (see chapter 5.6.3 on page 55) then you will see a buttons listing the various protocols of the loaded protocol file (*.pro). A single click on one of those buttons will start the respective *Protocol*.

- A single click on one button in the scroll bar will start the respective protocol.
- By entering a number into the field to the right of the button row, the protocol with this index is started.
- A single click on one button in the row of the fixed protocols will start the respective protocol.

8.5 Fixed Control Protocols


If you have activated the option *Show X Fixed Control Protocols* in the **Configuration** window (see chapter 5.6.3 on page 55) there will be a certain number of buttons available below the protocol bar.



The position of the protocols is fixed. The order can only be changed when using the *Assign* button. In the upcoming dialog the user is asked to enter the name of an existing protocol and the position to which button it should be moved within the *Fixed Control Protocols*.




9 Replay Window



If data has been acquired in the *Store* mode or if an old data file has been opened, they can be reviewed and edited in the *Replay* window. To open it select *Windows* → *Replay*. Up to five levels of the data tree are displayed.

9.1 Main window functions



Text: Displays the comment to the target (either *Root*, *Group* or *Series*). Use this function to edit the comment. The entered text can be seen in the field *Text* of the tab for *Root* and *Group* or in the field *Comment* of the tab for *Series* in the *Parameter* window.

Note: A comment for a *Series* can be written directly in the field *Comment* of the *Oscilloscope* window.

Label: Displays the target label (*Groups/Series/Sweep/Trace*). All objects have default labels assigned:

- Group: E-1
- Series: Name of Pulse Generator sequence



- Sweep: No label
- Trace: Name of acquired AD channel or assigned *Trace* name of the Pulse Generator

You may use the *Label* function to edit the label.

When creating a new *Experiment* or a new *Group* with the *File Operation* event or acquiring a *Series* with the *Acquire* event in the Protocol Editor, then you can specify a custom label.

show / to: Defines which part of the data tree will be displayed. The highlighted part of the data tree is referred to as target throughout this manual. Available parts of the data tree are *Root*, *Group*, *Series*, *Sweep*, and *Trace*.

tracing: Specify the part of the data tree that will be traced during replay. This feature might be of importance if the *Replay* window is very narrow (allowing to display three levels of the tree only) and many tree entries exit. Setting the *Tracing* item to *Series* then the active items during replay are the series objects and still *Root*, *Group* and *Series* are shown in the *Replay* window. If you would set the *Tracing* item to *Trace* keeping the same narrow *Replay* window, then the whole tree would be scrolled to the right and only *Series*, *Sweep* and *Trace* are shown in the *Replay* window.

visible/Apply: This feature can be used to expand or minimize the whole tree. If we specify *Series* and click *Apply* then the tree is reduced to the level of the *Series*, all *Sweep* and *Trace* objects are hidden.



Show: Replay the selected target. Alternatively, you can perform a double-click on the target.

Mark: Mark the selected target. You can mark more than one entry. Several functions in the *Replay* menu apply for the marked entries (see chapter 4.4 on page 23). Note that set marks will be written into the *.pul file of the data when you save the data. That means that marks will also be reloaded when you open such a marked data file.

Unmark: Unmark the selected marked target.

Note: You can also quickly mark or unmark your targets by selecting them and pressing 'M' or 'U', respectively.

Show/Hide Arrows: These buttons can be used to expand and close the child tree.

Mark All: Mark all targets in the selected data range with the named label.

repeat rate: Enter a time interval for controlling the scroll speed when keeping a scroll button (arrows) pressed. Every time value between 0.01 and 1 s is possible.

(◀▶ | ⏏, ⏏) **Window Scroll Arrows:** The arrow buttons on the window frame can be used to scroll/jump through the data tree.

9.1.1 Tree Handling


Interpreting the Tree objects: Besides a running index, text is displayed in the icons of the tree entries. The displayed text is the *Label* of the object. Usually the *Root* icon contains the file name, the *Group* icon contains the experiment number, and the *Series* icon shows the name of the *Stimulation Sequence*. The *Sweep* icon holds the *Sweep* index and the *Trace* icon the *Trace* name.

Example: In the screenshot of the *Replay* window below you can see the following entries in the first line:

- File name: Demo
- Group name: HEK; The group "HEK" is the first group of the file Demo.



- Series name: `Lock 1`; The *Series Lock* is the first *Series* of the *Group*.
- Sweep name: 1; the *Sweep* carries no name or label. It is the first *Sweep* of the *Series*.
- Trace names: *Imon-1*, *CM 2*



Labels starting with a "#" (*Series "6"*) denote a currently acquired object that will not be stored after acquisition.

Maneuvering through the Tree: The cursor keys (LEFT, RIGHT, UP and DOWN) of the keyboard allow you to scroll through the data structure of the tree. PAGE UP and PAGE DOWN can be used to scroll one window up or down. HOME and END will move to the start or end of the tree.

Selecting/Marking multiple objects: An example: For analysis purposes you would like to export all data obtained with `Lock` protocols for a given cell. Thus, enter `Lock` in the name field, select the *Group* for the cell under consideration, and click on *Mark All*. Then all *Series* named `Lock` of this group can be exported with *Replay* → *Export Marked Target Children*.

Mark a Target and its Children:



SHIFT+CLICK = Marks Target and Children between two selections.




STRG+CLICK = Unmark a specific Target Children





10 Pulse Generator Window



The Pulse Generator window defines the layout of stimulation pulse patterns. On the level of stimulation, they are called "sequence" throughout this manual. The result of a sequence stimulation, i.e. the families of sampled data *Traces*, are called a *Series*. Entries that are not default (e.g. adjustable parameters) are highlighted in bright background colors (*active* parameters), the rest is shown in dim background colors (*inactive* parameters). Because of this automatic feature, the color of individual editable controls cannot be changed in this dialog. All other controls can be modified.


10.1 Different Views of the Pulse Generator

There are three different views available for the Pulse Generator window.

Full View: Shows the complete Pulse Generator window (see figure above).


Note: You can use the "Copy to Clipboard" function to copy the text listing of the PGF setting somewhere else.

Condensed View: The condensed view provides access to the most important functions of a simple experiment. In case segment parameters have to be changed frequently, then the Pulse Generator window may stay open during the entire experiment with minimal space requirement.




Cartoon View: Shows only the graphical representation of the selected Pulse Generator sequence.

Note: If you use the "Copy to Clipboard" function when the *Cartoon View* pane is active, the stimulus cartoon is copied instead of the text listing of the PGF settings.



10.2 Sequence Selector



Sequence Buttons: The first row displays a section of the pool of available sequences. It is a paging bar in units of six sequences. Two arrows at each side allow scrolling through the available pulse protocols (the innermost arrows move in increments of one page, i.e., six sequences; the outermost arrows move to the start/end of the sequence list). A sequence is selected by clicking on it.

The pool of sequences is read from a file, usually the default file `DefPgf_v9.pgf`.

Note: If no Pulse Generator file is available, POTMASTER creates the default file. This file only contains one stimulation sequence, called "Test". This sequence can also be edited.

More sequences are added to the pool by copying an already existing sequence (**COPY** button, see below) or by creating a new sequence (**NEW** button, see below).

10.3 Sequence Pool Handling



LOAD: Loads the pool of available stimulation sequences (*.pgf file). The present file name is indicated in the title bar of the Pulse Generator window, e.g., "Pulse Generator File: DefPgf_v9".

MERGE: Loads individual stimulation sequences from another (.pgf file) into the currently opened file. In case a sequence name is not unique, you are prompted to either skip this sequence, mark it invalid, or give it a new name.

SAVE: Saves the pool of available stimulation sequences. After modification of the existing pool of sequences, the entire Pulse Generator File (PGF) should be saved to disk (Save button). It can be saved under any name. POTMASTER automatically appends the extension *.pgf to the file name. If this new PGF file should be loaded into the Pulse Generator as a default, the new name of the PGF file has to be specified in the Configuration window and the configuration file has to be saved.

Name: Editable name of the present sequence.

NEW: Creates a new stimulation sequences in the loaded PGF pool.

COPY: Duplicates the actual sequence into the first free position. A new name has to be entered.

MOVE: Moves the sequence to a new position. A number for the new position has to be entered. Use this option to move the most commonly used sequences to one of the handy first six positions, or to rearrange your pool.

DELETE: Removes the present sequence from the pool.

LIST: Writes the settings of the actual Series into the Notebook window. Use this option, if you want to create a listing of your sequence to be able to recreate it on another machine. A PGF listing could look as follows:

```
PGF-stimulus: pulse, Sequence: 4 - polarogr
TIMING:
NumberSweeps: 1, Averages: 0, SweepInterval: 0.000 s,
SampleInterval: 10.00 us
Wait before 1. sweep: FALSE
```




```

CHANNEL: 1, channel 1 time.
AD-Channel [A]: AD-6, Comp.Factor: 1, Comp.Mode: 00000000,
CompSkip: 0
Rel.DacChannel: 1, Rel.Segments - X: 2, Y: 2, Write: TRUE,
Holding: 0.000 V
LEAK SUBTRACTION:
Rel.DacChannel: 1, Rel.Segments - X: 2, Y: 2, Write: TRUE,
Holding: 0.000 V no leak pulses
DA-Channel [V]: DA-3, use StimScalerrelative to Vmemb
Amplifier mode: VoltageClamp, Set last Seg. Ampl.: FALSE
SEGMENTS: Voltage Duration VFact VIncr TFact Tincr

```

10.4 Acquisition Modes


Interactive Mode: In the *Interactive Mode* there is a minimum waiting period between the recording of single *Sweeps*, the so-called "Sweep Gap". The size of the gap depends on the speed of your computer system and the complexity and duration of the stimulus to be calculated and output. If you need to do any modification (e.g. *C-slow*, recording modes...) during the acquisition which effects the *Sweep* recording you have to select the *Interactive Mode*. During the "Sweep Gap" the set *Holding Potential* will be applied.



Gap Free Mode: The *Gap Free Mode* allows a timely precise acquisition of *Sweeps* without any gap in between. If you do not want to modify the recording settings during acquisition we highly recommend to select the *Gap Free Mode*.

10.5 Auto Sample


Auto Sample: This function helps to adjust *Sample Interval* and the *Compression* of a *Sweep* in relation to the used *Segment* type.



10.6 Timing

Timing:

- Wait before 1. Sweep: Forces POTMASTER to wait the time indicated by *Sweep Interval* before executing the sequence after activating it.
- No Wait before 1. Sweep: If you want the sequence to start immediately without a delay.
- Allow Continuous Redraw: Check this option to allow continuous display of the data in the Oscilloscope during acquisition. Minimum required *Sweep* length is 1 second to activate this function.



No. of Sweeps: Determines the number of *Sweeps* within one sequence, e.g., how many times the specified pulse pattern is run.

Sweep Interval/Sweep Delay: One can select either a *Sweep Interval* time or a *Sweep Delay* time. The *Sweep Interval* defines a waiting time from the start of one *Sweep* to the start of the subsequent *Sweep*. The *Sweep Delay* defines a waiting time from the end of the last *Sweep* to the beginning of the subsequent *Sweep*.



Note: Only one selection, the active one, is taken into consideration for the recording.

Sample Interval: The timing of data acquisition is given as *Sample Interval* (in seconds) and as sampling frequency (in Hz). The shortest *Sample Interval* (in case only one channel is used) is 5 μ s (200 kHz); the longest interval is 1 s (1 Hz). Only the *Sample Interval* has to be entered, the sampling frequency is just displayed as a reference.

Note: If the *Sample Interval* is so long as to yield only one point per segment, POTMASTER asks whether the segment duration should be fixed, e.g., set to a higher value.

Trigger Mode: Determines if and how data acquisition is triggered by an external TTL pulse. The default is *Not Triggered*, which means that stimulation is immediately elicited by the user within POTMASTER. Otherwise you have to activate the sequence and then one or more external triggers have to be applied.

- Not Triggered: No external triggering.
- Trigger Series: One trigger at the start of a sequence.
- Trigger Sweeps, Main + Leaks: A trigger is required for the main pulse and for each single leak pulse.
- Trigger Sweep, Main only: One trigger is required to start the complete Sweep, main as well as leak pulses.

Use Durations / Use Scan Rates:

Segments	1	2	3
Segment Class	Constant	Ramp	Ramp
Potential [V]	val 0.000	val 500.0m	val -200.0m
Duration [s]	val 2.000	val 5.000m	val 7.000m
E-incr. Mode	Increase	Increase	Increase
E-fact./incr. [V]	1.00 0.000	1.00 0.000	1.00 0.000
t-incr. Mode	Increase	Increase	Increase
Scan-fact./incr.	1.00 0.000	1.00 10.00	1.00 10.00
Scan Rate [V/s]	---	100.0	-100.0

Ramp segments can be specified either by their duration or by their *Scan Rate*. If the *Scan Rate* is given by the user, the duration will be calculated, and vice versa. The desired mode can be selected from the list:

- Use Durations: The time between the ending and beginning of a *Ramp* is set in seconds in the *Duration* field of the sequence.
- Use Scan Rates: The *Scan Rate* is set in 'V/s' in the *Scan Rate* field of the sequence and is calculated as *Voltage/Duration*.

Start Seg: Sets the start segment.

Start Time: Sets the start time within the start segment.



Start Seg and Start Time determine the start from where acquired data are stored. The full stimulation template is always output.

10.7 Check and Execute

Not Checking / Checking: This option determines whether a check is performed for any inconsistencies that might occur when entering values.



NOT Checking : When *Not Checking* is enabled, the validity checking of the sequence editing is suspended. This is convenient when one wants to perform multiple changes, especially when some intermediate steps would result in a (temporarily) faulty sequence.

Checking : When *Checking* is enabled, the active sequence is checked after any modification of segments and when storing, switching, or leaving the PGF editor. If the input is faulty, the user is notified and the last operation is canceled, until the sequence is valid. The checking should be done at least at the end of the sequence input, before executing or storing a sequence.

EXECUTE : Allows to output the presently active stimulation sequence. Upon termination of data acquisition, the program returns to the Pulse Generator window. In this way, pulse patterns can be adjusted interactively without changing windows until they yield the desired responses.

10.8 Sweep and Channel Length

Sweep Length: Is the duration of the Sweep. It is given by the longest Trace (stored or not stored) in a Sweep.

Sweep Length	Total	30.00 ms	1500 pts
	Stored	25.00 ms	5000 bytes
Channel Length	Stimulus	30.00 ms	1500 pts

Shows the length of the executed Sweep and stored data as determined by the timing, compression and segments settings.

- Total: Denotes the total duration of the longest Trace of the Sweep in 'ms' and 'points'.

Note: If the total Sweep length exceeds Max. Sample Points (see chapter Memory Allocation, 5.3.2 on page 43) a warning appears. If you would like to acquire longer Sweeps or a Sweep at higher time resolution, you might have to increase the Max. Sample Points parameter in the Configuration window.

- Stored: Denotes the total duration stored of one Sweep of the given sequence in 'ms'. The Sweep size is given in 'bytes'.

Note: All Traces contribute to the Sweep size but not to the Sweep Length. Total and Stored duration may be different when a Start Seg and Start Time are set.

Channel Length: Length of the actual DA stimulation. This can be shorter than the Sweep length, e.g., a short trigger impulse.

- Stimulus: Denotes the duration of the selected stimulus signal in 'ms' and 'points'.

10.9 Wave Parameters

Depending on the Segment type (see chapter Segment Classes, 10.11.1 on page 121) and the Stim → DA setting (see chapter 10.10.1 on page 115) different wave parameter buttons will appear.

10.9.1 Sine Wave Parameters

Sine Wave For segments of the Sine type. This button will appear between the buttons *Checking* and *Execute* and allows the specification of the wave characteristics.



10.9.1.1 Use as Simple Sinewave: Common Frequency

Use this setting if all wave segments within a *Sweep* should have the same frequency.

Peak Ampl. [V]: Defines the size of the *Peak Amplitude*. The amplitude is half of the peak-to-peak amplitude.


Value or parameter: If *Value* is selected the amplitude of the sine wave can be entered in the field *Peak Ampl.*. Alternatively, the *Peak Ampl.* can be determined by a *PGF parameter*. This allows to automatically change the *Peak Ampl.* of the sine wave between different series acquisitions.

Requested Freq.: Desired frequency of the sine wave in Hertz (Hz). Once the *Wave Parameters* window is closed, the *Requested Freq.* is set to the *Actual Frequency*.

Actual Frequency: Only certain frequencies are possible because they are generated by dividing a fundamental clock frequency by an integer. Nevertheless, the *Actual Frequency* is usually within a couple of percent of the *Requested Freq.*. The *Actual Frequency* is a function of the *Requested Freq.*.

Points / Cycle: Number of sample points per full wave length. This number is calculated and can not be edited. It is a display value only and gives the number of points per sinus wave which is the cycle length divided by the sample interval.


Below a sine wave created with a *Peak Ampl.* of 10 mV and a *Actual Frequency* of 100 Hz is shown.



10.9.1.2 Use as Simple Sinewave: Separate Frequency

Use this setting if the frequency of the wave is determined by the segment duration. Each segment will contain one wave cycle only. This option allows to create a *Sweep* containing wave cycles of different frequencies.






All available settings have been explained before (see chapter 10.9.1.1 on the previous page).

10.9.1.3 Use as LockIn Sinewave

Use this setting if the sine wave should be analyzed with the software LockIn.



Note: For the LockIn Sinewave to be an active option, the LockIn Extension must first be turned on in the Configuration window (see chapter 5.4.3 on page 47).

Use Theoretical Stimulus: Specify whether to use the theoretical stimulus or a measured stimulus Trace for computing real and imaginary traces.

Note: This option is needed when using the Protocol Method "Impedance Spectroscopy", also known as "EIS".

Peak Ampl. [V]: See chapter 10.9.1.1 on the previous page.

Requested Freq.: See chapter 10.9.1.1 on the preceding page.

Actual Frequency: See chapter 10.9.1.1 on the previous page.

Points / Cycle: See chapter 10.9.1.1 on the preceding page.

Cycles to Skip: When a Sinewave segment begins, there is a "capacitive" transient current response just as there is a transient response when a voltage step is given. In order to prevent this from causing an artifact in the C_m



Trace, sine wave cycles are skipped at the beginning of each *Sinewave* segment. If the frequency of the sinusoidal wave is chosen appropriately, then the transient should decay within a single cycle.

Note: A value of 1 Cycle to Skip is recommended.

Cycles to Average: If a value larger than '1' is entered, the specified number of cycles will be averaged. This reduces the noise and the time resolution (number of data points) in the results.

Total Cycles: Number of all cycles.

Checking: Allows to deactivate the internal checking procedures during the editing process.

For more information concerning the LockIn extension, we refer to the chapter Software LockIn Extension, 24 on page 249, and the PATCHMASTER Tutorial Capacitance Measurements using the LockIn Extension.

10.9.2 Square Wave Parameters

Square Wave For segments of the *Square* type. This button will appear above the *Execute* button and allows the specification of the wave characteristics.

10.9.2.1 Use as Simple Squarewave: Common Frequency

Use this setting if all wave segments within a *Sweep* should have the same frequency.

Peak Ampl. [V]: Defines the size of the *Peak Amplitude*. The amplitude is half of the peak-to-peak amplitude.

Value or parameter: If *Value* is selected the amplitude of the square wave can be entered in the field *Peak Ampl.*. Alternatively, the *Peak Ampl.* can be determined by a *PGF parameter*. This allows to automatically change the *Peak Ampl.* of the square wave between different *Series* acquisitions.

Requested Freq.: Desired frequency of the square wave in Hertz (Hz). Once the *Wave Parameters* window is closed, the *Requested Freq.* is set to the *Actual Frequency*.

Actual Frequency: Only certain frequencies are possible because they are generated by dividing a fundamental clock frequency by an integer. Nevertheless, the *Actual Frequency* is usually within a couple of percent of the *Requested Freq.*. The *Actual Frequency* is a function of the *Requested Freq.*.


In contrast to the sine wave the square wave allows to use different positive and negative amplitudes as well as different half cycle times. By the use of the following two parameters, e.g. stimulation trains can be easily parameterized.



Neg. Ampl. [V]: If a *Pos. Dur. Factor* unequal zero is used, the negative amplitude of the square wave can be entered.


Base Incr. [V]: Defines an increment amplitude which causes a baseline shift of the square waves ("Staircase" effect).

Below a square wave created with a *Peak Amplitude* of 10 mV and an *Ampl. Incr.* of 5 mV) is shown.



Pos. Dur. Factor: Factors between 0 and 1 are allowed. A factor of '0.5' means that both half cycles have the same duration. A factor smaller than '0.5' shortens the positive half cycle of the square wave.

Below a square wave created with a *Pos. Dur. Factor* of '0.2' and a *Peak Ampl.* of 10 mV and a *Neg. Ampl.* of -5 mV is shown.



10.9.2.2 Use as Simple Squarewave: Separate Frequency

Use this setting if the frequency of the wave is determined by the segment duration. Each segment will contain one wave cycle only. This option allows to create a *Sweep* containing wave cycles of different frequencies.

All available settings have been explained before (see chapter 10.9.2.1 on the preceding page).

10.9.2.3 Use as LockIn Squarewave

Use this setting if the square wave should be analyzed with the *LockIn Extension*.

Squarewave Parameters

Use as LockIn Squarewave

Use Theoretical Stimulus

Peak Ampl. [V] 10.00m value

Requested Freq. 100.0 Hz

Actual Frequency 100.0 Hz

Points / Cycle 200

Cycles to Skip 0

Cycles to Average 1

Total Cycles 1

Checking

Cancel

Done

Note: For the LockIn Squarewave to be an active option, the LockIn Extension must first be turned on in the Configuration window (see chapter 5.4.3 on page 47).

Use Theoretical Stimulus: Specify whether to use the theoretical stimulus or a measured stimulus Trace for computing real and imaginary traces.

Note: This option is needed when using the Protocol Method "Electrochemical Impedance Spectroscopy", also known as "EIS".

Peak Ampl. [V]: See chapter 10.9.2.1 on page 107.

Requested Freq.: See chapter 10.9.2.1 on page 107.

Actual Frequency: See chapter 10.9.2.1 on page 107.

Points / Cycle: Number of sample points per full wave length. This number is calculated and can not be edited. It is a display value only and gives the number of points per square wave which is the cycle length divided by the sample interval.

Cycles to Skip: When a Squarewave segment begins, there is a "capacitive" transient current response just as there is a transient response when a voltage step is given. In order to prevent this from causing an artifact in the C_m Trace, square wave cycles are skipped at the beginning of each Squarewave segment. If the frequency of the square wave is chosen appropriately, then the transient should decay within a single cycle.

Note: A value of '1' Cycle to Skip is recommended.

Cycles to Average: If a value larger than '1' is entered, the specified number of cycles will be averaged. This reduces the noise and the time resolution (number of data points) in the results.

Total Cycles: Number of all cycles.

Checking: Allows to deactivate the internal checking procedures during the editing process.


For more information concerning the LockIn extension, we refer to the chapter Software LockIn Extension, 24 on page 249, and the PATCHMASTER Tutorial Capacitance Measurements using the LockIn Extension.

10.9.3 Chirp Wave Parameters


Chirp Wave For segments of the Chirp type. This button will appear between the buttons Checking and Execute, allowing specification of the wave parameters. The Chirp wave's characteristic is either an increasing or decreasing frequency over time. This frequency modulation can be in a linear, exponential or spectroscopic manner. An example for a Linear Chirp and an Exponential Chirp is given below:



Linear Chirp



Exponential Chirp



50pA
1ms

10.9.3.1 Linear Chirp

Use this setting if the chirp frequency should change in a linear manner.

Amplitude [mV]: Defines the size of the *Peak Amplitude*. The amplitude is half of the peak-to-peak amplitude.

Value or parameter: If *Value* is selected the amplitude of the chirp wave can be entered in the field *Peak Ampl.*. Alternatively, the *Peak Ampl.* can be determined by a *PGF parameter*. This allows to automatically change the *Peak Ampl.* of the chirp wave between different *Series* acquisitions.

Start Frequency: Defines the frequency where the chirp wave starts.

End Frequency: Defines the frequency where the chirp wave ends.

Min. Points / Cycle: Indicates the minimum number of sample points per full wave length for the *End Frequency*.

Segment Points: Total amount of points for the *Chirp* segment. It is a display value only.

10.9.3.2 Exponential Chirp

Use this setting if the chirp frequency should change in an exponential manner.

Chirpwave Parameters

Exponential Chirp

Amplitude [V]	10.00m	value
Start Frequency	10.00 kHz	
End Frequency	100.0 Hz	
Min. Points / Cycle	2.0	
Segment Points	200	

Checking
Cancel
Done

All available settings have been explained before (see chapter 10.9.3.1 on the facing page).

10.9.3.3 Spectroscopy Chirp

Use this setting if the *Chirp* wave should be analyzed with the Spectroscopy Extension.

The screenshot shows a dialog box titled "Chirpwave Parameters" with a sub-tab "Spectroscopy Chirp". It contains several input fields and buttons:

- Amplitude [V]**: 10.00m, with a "value" button next to it.
- Start Frequency**: 78.13 Hz, with a checkbox for "Pre-Chirp" to its right.
- End Frequency**: 10.00 kHz.
- Min. Points / Cycle**: 2.0.
- Chirps to Skip**: 0.
- Segment Points**: 200.
- Three buttons on the right: "Checking" (red), "Cancel" (yellow), and "Done" (yellow).

Note: For the Spectroscopy Chirp to be an active option, the Spectroscopy Extension must first be turned on in the Configuration window of Potmaster (see chapter 5.4.4 on page 48).

Amplitude [V]: See chapter 10.9.3.1 on the facing page.

Value or parameter: See chapter 10.9.3.1 on the preceding page.

Start Frequency: See chapter 10.9.3.1 on the facing page.

End Frequency: See chapter 10.9.3.1 on the preceding page.

Min. Points / Cycle: See chapter 10.9.3.1 on the facing page.

Segment Points: See chapter 10.9.3.1 on the preceding page.

*Note: End Freq. = Min. Points / Cycle * Sampling Freq.*

Pre-Chirp: Attaches a small chirp fragment to the beginning of the chirp wave to avoid

Checking: Allows to deactivate the internal checking procedures during the editing process.

For details on the parametrization of *Chirp* wave form we refer to the chapter Spectroscopy Extension, 25 on page 261 and the POTMASTER tutorial Using the Spectroscopy Extension.

10.9.4 Photometry Wave Parameters

Photometry Wave If the option *Use for Wavelength* is active for a DA output channel, then the *Photometry Wave* button will appear above the *Checking* button. It assists you in creating repetitive stimulation pattern for controlling the excitation wavelengths.

The screenshot shows a dialog box titled "Photometry Parameters". It contains the following elements:

- Number of Segments**: 3
- ☒ **Adapt to Maximal Sweep Length**
- Number of Cycles**: 1
- Two buttons: "Trunc" and "Expand".
- Two buttons at the bottom: "Checking" (red) and "Done" (yellow).



Number of Segments: The number of creating a single cycle for a measurement. Typically there are three segments used for a ratio-metric measurement. Segment '1' and '2' for the wavelength setting and a third segment for a waiting time.


Adapt to Maximal Sweep Length: If checked, then the number of cycles is calculated by the software in respect to the duration of the longest output channel.

Number of Cycles: As input determines how many times the first *Number of Segments* are duplicated with the *Expand* command. In case *Adapt to Maximal Sweep Length* is checked, then it shows how many cycles have been fit into the stimulus.

Expand: Creates the complete photometry stimulus. The first *Number of Segments* are duplicated until the *Maximal Sweep Length* or the *Number of Cycles* is reached.

Trunc: Removes all segment except the first *Number of Segments*. This option should be used to clean the stimulus before changing any parameters in the photometry stimulus. After parameter adjustment an *Expand* creates the new stimulus.

Checking: Allows to turn off the internal consistency checks. Turning this transiently off might be necessary when multiple changes of parameter would lead to intermediate invalid configurations.




In this example a single cycle consists of 3 segments. You have just to edit the first three segments. A click on *Expand* adds 15 additional segments to the PGF sequences. The resulting PGF sequence then consists of $6 * 3 = 18$ segments. *Trunc.* removes all segments except the initial 3 segments.

10.9.5 Imaging Wave Parameters

In case you want to perform image acquisition during the execution of a Pulse Generator sequence, please we recommend reading first the chapter *Imaging Extension*, 23 on page 247.

The *Imaging Wave* parameters are usually set in the *PGF Primer* of SmartLUX and automatically integrated into the Pulse Generator sequence by the *Prime* function of SmartLUX. Hence, usually **it is not necessary to modify the parameters in this dialog.**



The image shows a software dialog box titled "Imaging Parameters". It contains a list of parameters, each with a corresponding input field. The parameters are: "Adapt to Maximal Sweep Length" (checkbox), "Number of Cycles" (1), "Cycle Time" (0.00 s), "Readout Time" (100. ms), "Start Wait Time" (10.0 ms), "Exposure Time 1" (0.00 s), "Wavelength 1" (0.00), "Exposure Time 2" (0.00 s), "Wavelength 2" (0.00), "Exposure Time 3" (0.00 s), "Wavelength 3" (0.00), "Resting Wavelength" (280.), "Dead Time" (2.00 ms), "Regions of Interest" (0), "Rows to Read" (0), and "Binning" (1x1). At the bottom, there are four buttons: "Trunc" (yellow), "Prime" (yellow), "Checking" (red), and "Cancel" (yellow).

Parameter	Value
☐ Adapt to Maximal Sweep Length	
Number of Cycles	1
Cycle Time	0.00 s
Readout Time	100. ms
Start Wait Time	10.0 ms
Exposure Time 1	0.00 s
Wavelength 1	0.00
Exposure Time 2	0.00 s
Wavelength 2	0.00
Exposure Time 3	0.00 s
Wavelength 3	0.00
Resting Wavelength	280.
Dead Time	2.00 ms
Regions of Interest	0
Rows to Read	0
Binning	1x1

Buttons: Trunc, Prime, Checking, Cancel

For a detailed description of the imaging parameters we refer to the SMARTLUX manual.



10.10 Channel Settings for DA Output and AD Input

1	DA	Unit	Stimulus -> DA	Leak	AD	Unit	Link	Compr.	Points	Store	Zero	Leak	
 Ch-1	Stim-out	V	StimScale		I-mon	A	1	64	C	4690		1	Store Avg
 Ch-2	off	V	absolute voltage		E-mon	V	1	64	C	4690		0	No Leak
 ---	off		absolute voltage		off		---	---	C	---		---	No Leak
 ---	off		absolute voltage		off		---	---	C	---		---	No Leak

Ch-1...16: Settings for each channel. The default is "1", the other channels (Channels = 2...) may be used to simultaneously record the potential or an amperometric signal, for example.

The DA section allows you to set the properties of the DA output, e.g., the stimulus signal.

The AD section allows you to set the properties of the AD input, e.g., the acquired data. DA and AD settings are basically independent from each other. Their reference is only given by the variable *Link* in the AD settings! This allows you to associate several AD inputs to the same DA stimulation.

10.10.1 DA output channel settings

DA: The DA channel for stimulation has to be specified. The figure below shows the options for the EPC 10 Single amplifier.

- DA-0...2: Analog output.
- Stim-out: Standard output channel.
- off: No output of stimulus, disables this channel.
- Dig-out (word): Output as digital word.
- Dig-0...15: Digital output channel.

☒

DA-0
DA-1
DA-2
Stim-out
DA-4
DA-5
DA-6
DA-7

off

Dig-out (word)

Dig-0
Dig-1
Dig-2
Dig-3
Dig-4
Dig-5
Dig-6
Dig-7
Dig-8
Dig-9
Dig-10
Dig-11
Dig-12
Dig-13
Dig-14
Dig-15



Unit: Units for output channels. Defaults are "V" for voltage, and "A" for current.

Stimulus → DA: Conversion between the stimulus and the actual analog output.

- apply StimScale: Apply standard conversion.
- relative to E-init: Calculates the signal relative to *E-init*.
- load from file template: Load a stimulus template from file. A detailed description how to use a recorded waveform as a stimulus template can be found in the chapter [Using a Recorded Waveform as Stimulus in the PATCHMASTER Tutorial](#).
- use for LockIn: Has to be activated if the stimulus should be used for the software LockIn. For details please refer to the chapter [Software LockIn Extension, 24 on page 249](#).
- use for Wavelength: Has to be activated if the stimulus should be used to control a wavelength. Amplitudes of segments can then be entered in "nm". This option is available if the [Photometry or Imaging Extension](#) is activated.
- use DA-Scaling: Has to be activated if the stimulus should be scaled depending on the settings made in the *I/O Control* tab of the *Configuration*. For further information we refer to the chapter [AD/DA Input/Output Scaling, 5.7.1 on page 57](#).
- use for Spectroscopy: Has to be activated if a chirp wave form should be used for impedance or admittance analysis with the [Spectroscopy Extension](#). For details please refer to the chapter [Spectroscopy Extension, 25 on page 261](#).
- use for Camera Trigger: Has to be activated if the stimulus should be used to control the exposure of the camera. This option is available if the [Imaging Extension](#) is activated. For further information we refer to the chapter [Imaging Extension, 23 on page 247](#).

Stim to DA

☐ apply StimScale
☐ relative to E-init
☐ load from file template
☐ use for LockIn
☐ use for Wavelength
☒ use DA-Scaling
☐ use for Spectroscopy
☐ use for Camera Trigger

If nothing is selected, "absolute voltage" will be applied.

Use of digital Outputs: Digital outputs are mainly used for controlling (triggering) other devices, such as perfusion systems, filter wheels, flash lamps, or others. The EPC 10 amplifier (and the LIH 8+8) are equipped with 16 digital output lines. These outputs can be accessed from the rear panel of the amplifier. The first three of the 16 available digital outputs can be accessed also from the front panel of the amplifier (interface). The BNC outputs ("Trigger-Outputs") "Out-0", "Out-1", "Out-2" correspond with the out channels "Dig-0", "Dig-1", and "Dig-2" in the Pulse Generator dialog of POTMASTER.

There are two ways to control the digital lines:

If only one digital line is intended to be used, then one of the "Dig-Out" lines ("Out-0" to "Out-15") can be selected as output channel. In the *Segments* section of the Pulse Generator the state of the digital line can be set either to "1" (active = TTL level), "-1" (inactive = 0 V), or to "0" (don't change, use the state that is set in the I/O control dialog).

The *Dig-Out word* output allows to switch several digital lines simultaneously. What channels are switched is defined in the *Segments* section by the corresponding bit value. If *Voltage* in the *Segments* section is set to "7", then the first three digital lines are turned on: $1 * 10^2 + 1 * 10^1 + 1 * 10^0 = 4 + 2 + 1 = 7$.

Note: For such trigger channels it is suggested to switch from Common Timing to Separate Timing. With this setting, the duration and number of segments is completely independent from the first (main) channel.



Leak: If selected, leak templates are output (see chapter `P/n Leak Correction`, 10.15 on page 127). This option allows a rapid deselecting of leaks without changing the leak settings. By activating this option the Leak Traces storing option is set by default to "Store Average".


***Note:** The leak template options are hidden in the **Pulse Generator** as long as this flag is not activated.*



10.10.2 AD input channel settings

AD: The AD-channel for input.

- AD-0...2: AD channel.
- E-mon: Voltage monitor of the amplifier.
- I-mon: Current monitor of the amplifier.
- Mux: Multi-use channel.
- off: No input.
- Virtual: Virtual input. A virtual *Trace* is used to generate a *Trace* where the data come from a source different from an ADC, e.g., external data loaded via the option *Import Trace*, or data which generate virtual *Traces*, such as the `LockIn`.
- Dig-in (word): Digital input.
- Dig-0...15: Digital input channels.
- Chirp Traces: Storage of particular *Spectroscopy* values, such as e.g. *Admittance*, *Phase*



A screenshot of a software interface showing a dropdown menu for AD input channel settings. The menu is open, displaying a list of options. The 'Virtual' option is highlighted with a blue checkmark in a small box to its left. The options listed are: AD-0, AD-1, AD-2, AD-3, AD-4, E-mon, I-mon, Mux, AD-8, AD-9, AD-10, AD-11, AD-12, AD-13, AD-14, AD-15, off, Virtual, Dig-in (word), Dig-0, Dig-1, Dig-2, Dig-3, Dig-4, Dig-5, Dig-6, Dig-7, Dig-8, Dig-9, Dig-10, Dig-11, Dig-12, Dig-13, Dig-14, Dig-15, Chirp_Avg, Chirp_Phase, Chirp_Admit(Y), Chirp_Real(Y), Chirp_Imag(Y), Chirp_Real(Z), Chirp_Imag(Z), and Chirp_Imp|Z|.



- **LockIn Traces:** Storage of particular *LockIn* values, such as e.g. C_m , G_m , G_s
- **Imaging/Photometry Traces:** Storage of particular *Imaging* or *Photometry* values (e.g. mean fluorescence, calcium concentration, fluorescence ratio.)

LockIn_Avg
LockIn_Phase
LockIn_Real(Y)
LockIn_Imag(Y)
LockIn_Admit(Y)
LockIn_Real(Z)
LockIn_Imag(Z)
LockIn_Imp Z
Photo_W1
Photo_W2
Photo_W3
Photo_R
Photo_Ca

Unit: Unit for input channels. Default is "A" for current (in voltage clamp mode). Use "V" if a plain AD channel is sampled or any other applicable unit if virtual *Traces* are used or the channel is scaled.

Link: Number of the stimulus channel that is associated with this input. This setting is important for later analysis of the *Trace*. It links the stimulus segments to the recorded *Trace*.

Compr.: Enter the number of points for the compression (compression factor) and choose the compression method from the list. The compression factor should be chosen such that the compression interval (or range) covers the complete period of interest. The number of data points divided by the number of compressed data points yields the compression factor.

***Note:** Normal Traces are acquired and can be compressed. Virtual Traces are generated.*

- **Compression Mode:**
 - single sample: Takes one point per compression interval.
 - mean: Averages all points of a compression interval.
- **2-byte integer / 4-byte real:** Data type for storage.
- **Set Skip:** With the *Set Skip* function, a specified number of points of the compression interval can be skipped before the compression starts. In a compression interval, you can only skip data points at the beginning of the interval.
- **Set Offset:** With this function, the start of the compression interval can be influenced. A negative offset moves the start of the compression interval to the left and a positive offset to the right. A negative offset essentially adds dummy data points on the negative time axis. Since compression starts at the first recorded point, the compression range is moved to the left (towards smaller times) with respect to the data points.

Compression Mode:
☒ single sample
☐ mean
☒ 2-byte integer
☐ 4-byte real
Set Skip (Skip = 0)
Set Offset (Offset = 0)
Digital Filter
Set Defaults
Build Instructions:
empty

***Note:** For an example describing the use of *Skip* and *Offset* compression functions please refer to the chapter *High Speed Fluorescence Measurements in the PATCHMASTER Tutorial*.*

- **Digital Filter:** If selected, the raw data are filtered digitally before compression. The filter factor is given by the compression factor. Example: If a *Trace* is acquired with at sampling rate of 5 kHz and the compression factor is 10, then the digital filter is set to 500 Hz. With a compression factor of 5 the filter bandwidth is increased to 1 kHz.



- **Set Defaults:** Sets all settings back to default.

Build Instructions: Allows computing a virtual *Trace* from other *Traces* and inputs. The instruction string starts with the math operation. The formula is terminated by a semicolon (;). Then a list with additional instructions all separated by a comma (,) can follow.

- **Mathematical Operations:**
 - **An:** Anodic integral of *Trace* with index *n*
 - **Bn:** *Trace* buffer with index *n*
 - **Cn:** Cathodic integral of *Trace* with index *n*
 - **Dn:** Differential of the *Trace* with index *n*
 - **In:** Integral of the *Trace* with index *n*
 - **M:** Mean of a *Trace*. It must be followed by a bracket enclosing 3 numbers, separated by comma:
 1. The *Trace* count of the *Trace* from which to compute the mean
 2. The time offset (in [s]) from the beginning of the **stored** *Trace* data.
 3. The time span (in [s]) over which to compute the mean. Example: "M(2,0.0,0.005)"
 - **Tn:** *Trace* with index *n*
 - **Vn:** Value with index *n*
 - **Zn:** Zero data of *Trace* with index *n*
 - **Operators:** +, -, *, /
 - **Brackets:** ()
- **Other Assignments:**
 - **CC-Switch:** Switches the recording mode to Current Clamp mode at the beginning of the specified segment. CC-Gain needs to be in 0.1 pA/mV or 1 pA/mV range. Example: ";CC-Switch=2" = Switch to Current Clamp Mode at the beginning of segment 2.
 - **Name="String"** (separated from the math operation by a semicolon ';'). "String" is just a placeholder for any name.
 - **No LockIn:** Suppressing LockIn computation
 - **Tcount:** *Trace* count with a value between 1 and 16
 - **S:** Stands for "Shift". It must be followed by a bracket enclosing 2 numbers, separated by comma:
 1. The source *Trace* index
 2. The number of samples by which to shift the source *Trace* Example: "S(2,50)" = *Trace* with index "2" shifted by 50 samples to the beginning of the *Trace*. Negative numbers shift the data towards the end of the trace. The undefined data will be set to invalid (NaN). Invalid data points are not displayed.
 - **Slave=[number]:** Copying the *Trace* data from the "slave" POTMASTER instance via the batch communication protocol. The equal sign must immediately follow "slave", e.g. "Slave=2". The given number is the *Trace* count of the *Trace* to get. In the given example data from the source *Trace* with *Trace* count = '2' is read.
 - **StimSwp():** Only *Sweeps* defined in the list will output the respective segment amplitudes. The other *Sweeps* will use the amplitude of segment '1' for all segments. Values are to be separated by commas, and *Sweep* ranges can be defined by a hyphen, e.g.: "StimSwp(1,3-5,7-9)".
 - **VC-Switch:** Switches the recording mode to Voltage Clamp mode at the beginning of the specified segment. Current Gain needs to be in medium gain range. Example: ";VC-Switch=2" = Switch to Voltage Clamp Mode at the beginning of segment 2.



- $Yrange=real$ (separated by a colon ':'). The "Yrange" sets the scaling of the Oscilloscope display. The value of "Yrange" corresponds to the total length of the Y-axis. $1/10$ of "Yrange" corresponds to the scale bar displayed in the Oscilloscope window.
- $Yrange([N],[factor])$. Sets the "Y-range" of a Trace proportional to the value of another Trace. "N" is the Trace index of the source Trace and "factor" is the factor to multiply.

Note: The total length of the Build Instruction is limited to 128 characters.

Example 1: $V1*T1;Name="m1", Yrange=1E-9$

Trace 1 is multiplied with Value 1. The name of the new virtual Trace is "m1" and the total range of the Y-axis is $1e-9$.

Example 2: $V1*T1+(T6-T7);Name="Baseline",Yrange=1E-6$

Trace 1 is multiplied with Value 1 and the difference between Trace 6 and Trace 7 is added. The name of the new virtual Trace is "Baseline" and the total range of the Y-axis is $1e-6$.

Example 3: $I1; Name="Charge", TCount=5, Yrange=1E-5$

Trace 1 is integrated. If Trace 1 is a current Trace we get the charge. The new virtual Trace gets the name "Charge", the total range of the Y-axis is $1e-5$ (i.e. $10 \mu C$). The Trace # is set to "5".



No other operator priority is considered besides the brackets.

The Build Instruction can also be used to give a particular Trace a name (label). This is convenient for having the name of the Trace displayed in the Replay window.

It is also possible to specify an equation as a build instruction similar to the equation handling in the Analysis (see chapter 13.3.2.7 on page 187), the Buffer menu (see chapter 4.6 on page 33) or the Calculator dialog (see chapter 4.3 on page 22). It is important to put the equation in double-quotes and start at the first character.

Example 4: $"t[1]-t[2]"$ or $"\sin(y)"$.

Points: Number of resulting points for this channel.

Store: If checked, data acquired for this channel will be stored.

Note: In case Store is not activated but the file is opened with write permission, the last Sweep will be stored temporarily. This Sweep is marked with a "#" in its label. It will be deleted when the next Sweep is acquired or a new Group is created or the file is closed.

Zero: Number of the segment on which base the Zero Offset will be calculated. Usually the Zero Offset is calculated from the second half of that segment to avoid contamination by tails of capacitive transients. In order to specify a custom range for the Zero Offset calculation please specify a StartSeg and a StartTime >0. Then the Zero Offset will be calculated between StartTime and end of the segment.

Note: When the segment number is set to "0", no Zero Offset subtraction will be calculated.

Note: Subtract Zero in the Display menu is kept "On" even if the conditions for Zero Subtraction are not met. If one wants to see the subtracted Trace offline, one has to switch to another Trace in the Replay Tree and then back to the just acquired Trace. A simple "Wipe" or "Redraw" of the displayed data is not sufficient.

Leak: Defines how the leak responses are stored.

- No Leak: No leak correction applied.



- Store Avg.: Stores the leak responses as average.
- Store Each: Stores every leak response separately.
- Store None: Leak correction is applied but the leak response is not stored.

Note: Leak correction will only be performed and stored when leak correction is activated (see chapter DA output channel settings, 10.10.1 on page 114)

Note: The main data are always stored with leak correction performed. I.e., to replay non-corrected data, the leak responses have to be stored.

Note: The leak traces storing method is set by default to "Store Average".

10.11 Segments

10.11.1 Segment Classes

Segments	1	2	3	4	5	6
Segment Class	Constant	Ramp	Ramp	Ramp	Constant	
Potential [V]	val 0.000	val 500.0m	val -200.0m	val 0.000	val ---	val ---
Duration [s]	val 100.0μ	val 5.000	val 7.000	val 2.000	val ---	val ---
E-incr. Mode	Increase	Increase	Increase	Increase	Increase	Increase
E-fact./incr. [V]	1.00 0.000	1.00 0.000	1.00 0.000	1.00 0.000	---	---
t-incr. Mode	Increase	Increase	Increase	Increase	Increase	Increase
Scan-fact./incr.	1.00 0.000	1.00 0.000	1.00 0.000	1.00 0.000	---	---
Scan Rate [V/s]	---	100.0m	-100.0m	100.0m	---	---

A pulse pattern consists of an arbitrary number of segments. Segments are shown as a horizontally scrolling matrix; clicking on the arrows does scrolling.

Store: On the top of the segment column there is a Store field with a selection list determining whether or not the data of that segment is to be stored to disk.

The following options are available:

- Not Stored: The data for that segment are sampled but removed from the Trace before storage. A non-stored segment must *not* be a *StartSegment* or a *Relevant Y-segment*. For a description on the usage of non-stored segments please read Non-Stored segments in the Pulse Generator in the PATCHMASTER Tutorial.
- Stored: The data are sampled and stored.
- First Sweep: These segments are output only with the first Sweep of the Series but are not stored.
- Last Sweep: These segments are output only with the last Sweep of the Series but are not stored.

Note: There are some important consequences when using P/n leak correction (see chapter P/n Leak Correction, 10.15 on page 127). Non-stored segments are not used for the P/n template. Thus, for proper leak correction it is recommended to insert a constant segment of the same value as the non-stored segment before and after that segment. The duration of these inserted segments should be sufficient to settle capacitive transients. Since P/n signals are generated for these inserted constant segments, P/n leak correction will work.



Segment Class:


Segments can be the following:

Segment Class	Constant	Ramp	Sine	Square	Chirp	Continuous
---------------	----------	------	------	--------	-------	------------

- Constant: Segment of constant amplitude.
- Ramp: Segment with ramp from the amplitude of the previous segment to the amplitude of this segment.
- Continuous: Identifier for continuous data acquisition.

Note: Only the last segment can be of that class.

- Sine: Segment with sine characteristics. Amplitude and cycle duration are defined in *Sine Wave* parameters (see chapter 10.9.1 on page 104).
- Square: Wave segment with rectangular characteristics. Amplitude and cycle duration are defined in *Square Wave* parameters (see chapter 10.9.2 on page 107).
- Chirp: Wave segment with increasing sine wave frequency. Wave characteristics are defined in *Chirp Wave* parameters (see chapter 10.9.3 on page 109).



The list entries *Insert*, *Duplicate*, and *Delete* are used to create or remove segments.

- Insert: Inserts a constant segment (default duration = 0) at the actual location (moves the selected segment to the right).
- Duplicate: Creates a copy of the selected segment and inserts it at the actual location.
- Delete: Deletes the selected segment at once.
- Duplicate...: Creates multiple copies of a number of segments. You have to enter how many segments you want to copy. Segments are counted starting with the selected segment. How often specifies the number of copies from these segments. They will be inserted at the actual location.
- Delete...: Deletes a specified number of segments at once.

Potential: The *Potential* of a segment is either a numeric value (*val*, in mV), a *E-init* value at time of sequence execution, or a value from the *PGF parameters* list (*p1-p10*). The *Potential* field may turn into "Current (A)" in *Galvanostatic* mode, into "Wavelength (nm)" when the segments are used for photometry stimulation or into "Amplitude" when the *Use for DA scaling* is used. The valid amplitude range is a function of the given DA output voltage range of the interface and the appropriate stimulus scaling factors or the wavelength to voltage conversion formalism of the *Photometry Extension*.)

Duration: Duration of a segment, entered in seconds (s). The value may be adjusted by dragging the mouse or typing the number. For segment duration, *PGF parameters* can also be used. In this case, all data are taken in seconds. Make sure that the duration of segments are even multiples of the *Sample Interval*. A warning is given if they are not, in case *Checking* is activated.

val / E-init / p1...10: Allows setting the value.

- val: Standard value.
- E-init: Keeps the potential at *E-initial* during a potentiostatic measurement. If the channel is used for photometry, the resting wavelength as specified in the *Photometry* window is set.
- p1...10: Sets the value to the value given in the corresponding *PGF parameter*. For a description of usage of the *PGF parameters* please read *PGF Parameters*, chapter 10.16 on page 130.



10.11.2 Scan Rates

The *Scan Rate* options are only available if *Use Scan Rates* is activated (see chapter 10.6 on page 103).

- Scan-fact./incr.: If a value for the increment is entered the field *t-incr. Mode* gets active.
- Scan Rates [V/s]: The user can enter a defined value into the field *Scan Rates* whereby the duration of the segment will be calculated or vice versa. The *Scan Rate* is calculated as Voltage/Duration.

10.11.3 Increment Modes

The various increment modes are flexible tools to create sequences with varying durations or amplitudes. Please note that first the amplitudes or durations in logical order are calculated as specified by the parameters *Potential/Duration*, *E-fact./t-fact* and *E-incr./t-incr.*. Then the logical order is converted to the physical output by applying the selected *Increment Mode*.


Increment Mode: This determines the order of incrementing the segment voltage and/or duration. The options are as follows (the numbers give an example for a *Series* with 6 *Sweeps*):

☒	Increase	☒	Increase
☐	Decrease	☐	Decrease
☐	Interleave +	☐	Interleave +
☐	Interleave -	☐	Interleave -
☐	Alternate	☐	Alternate
☐	Toggle	☐	Toggle
☒	t * Factor	☒	V * Factor
☒	dt * Factor	☐	dV * Factor

- Increase: first *Sweep* comes first (e.g. 1, 2, 3, 4, 5, 6)
- Decrease: last *Sweep* comes first (e.g. 6, 5, 4, 3, 2, 1)
- Interleave +: ascending interleaved (e.g. 1, 3, 2, 5, 4, 6)
- Interleave -: descending interleaved (e.g. 6, 4, 5, 2, 3, 1)

Example: If you apply a *Series* of 6 *Sweeps* at an increment of 10 mV starting at -40 mV the logical sequence will be: -40, -30, -20, -10, 0 and +10 mV. With the mode *Interleave -* activated the pulses will be output in the following (physical) order: +10, -10, 0, -30, -20 and -40 mV.

- Alternate: first, last, second, penultimate... (e.g. 1, 6, 2, 5, 3, 4)
- Toggle: Starts with the first *Sweep*, second the voltage increment is added and third the voltage segment is subtracted. Then twice the voltage segment is added. ... E.g. using 6 *Sweeps* with a *E-incr.* of 1 mV and the first *Sweep* steps to 10 mV. The output voltages are as follows: 10, 11, 9, 12, 8, 13 mV.



10.11.3.1 Logarithmic increment modes

V * Factor: In mode *V * Factor* the logarithmic increment is based on the voltage of the first *Sweep*. Therefore the voltage cannot be zero. The segment's voltage of the *i*th *Sweep* is then calculated as:

$$V_i = \text{Voltage} * V\text{Factor}^{i-1} + (i - 1) * dV\text{incr}$$

In mode *V * Factor* the increment may be zero. Let *Voltage* be 10 mV, $\Delta V\text{Incr.} = 0$ mV, and *VFactor* = 2 then the *Series* 10, 20, 40, 80 mV, ... is obtained.

dV * Factor: In mode *dV * Factor* the logarithmic increment is based on the linear increment step. Therefore the voltage increment cannot be zero. The segment's voltage of the *i*th Sweep is then calculated as:

for V-factor = 1:

$$V_i = \text{Voltage} + (i - 1) * dV\text{-incr}$$

for V-factor $\neq 1$:

$$V_1 = \text{Voltage} \quad V_{i,i>1} = \text{Voltage} + dV\text{-incr} * V\text{-Factor}^{i-2}$$


In mode *dV * Factor* the first segment may be zero and is then logarithmically incremented. Let *Voltage* = 0 mV, *dV-Incr* = 1 mV, and *V-Factor* = 2 then the Series 0, 1, 2, 4 mV, ... is obtained.

In some kind of experiments it might be useful to toggle between two fixed potentials. To achieve this, you have to:

- Define your starting "Voltage".
- Select "Increase" and "V*Factor".
- Select "V-fact." to "-1".

There are three cases for setting the voltage increment:

1. "0": E.g. +100 mV, -100 mV, +100 mV, -100 mV,...
2. "positive" (e.g. +20 mV): E.g. +100 mV, +80 mV, +100 mV, +80 mV,...
3. "negative" (e.g. -20 mV): E.g. +100 mV, +120 mV, +100 mV, +120 mV,...



In analogy to the logarithmic amplitude increments the duration increments are calculated as follows:

t * Factor: In mode *t * Factor* the logarithmic increment is based on the duration of the first Sweep. The segment's duration of the *i*th Sweep is then calculated as:

$$t_i = \text{Duration} * t\text{Factor}^{i-1} + (i - 1) * dt\text{Incr}$$

In mode *t * Factor* the increment may be zero. Let *Duration* be 10 ms, *dt-Incr.* = 0 ms, and *t-Factor* = 2 then the Series 10, 20, 40, 80 ms, ... is obtained.

dt * Factor: In mode *dt * Factor* the logarithmic increment is based on the linear increment step. Therefore the duration increment cannot be zero. The segment's duration of the *i*th Sweep is then calculated as:

for t-Factor = 1:


$$t_i = \text{Duration} + (i - 1) * dt\text{-incr}$$

for t-Factor $\neq 1$: $t_1 = \text{Duration}$

$$t_{i,i>1} = \text{Duration} + dt\text{-incr} * t\text{-Factor}^{i-2}$$

In mode $dt * Factor$ the duration of the first segment may be zero and is then logarithmical incremented. Let $Duration = 0$ ms, $dt-Incr = 1$ ms, and $t-Factor = 2$ then the Series 0, 1, 2, 4 ms, ... is obtained.

10.12 Miscellaneous



Common Timing / Separate Timing: If *Common Timing* is selected, the segments of all channels are timed like the first one. If *Separate Timing* is selected, all channels are timed separately, i.e. segment boundaries may be different in different channels.

Break Condition: For each acquired *Trace* of the AD input that is longer than 1 second a *Break* condition can be specified. The first condition that is *TRUE* terminated the acquisition of the *Sweep*.

- No Break: No break condition for that *Trace*.
- Break if >: Terminates the acquisition if the acquired values become larger than the value entered in the field to the right.
- Break if <: Terminates the acquisition if the acquired values become smaller than the value entered in the field to the right.
- Break if abs >: Terminates the acquisition when an absolute sample value of that trace exceeds the given threshold.
- Break on "next": Terminates the acquisition of the *Sweep* if the user presses the *Next* button in the Control window. The acquisition then continues with the next *Sweep*.

Note: The delay between the break condition = TRUE and leaving the acquisition routine can be up to 50 ms.

Recording Mode: Stimulation sequences can be restricted to *Potentiostatic* or *Galvanostatic* modes. The following options are available:

- Any Mode: Allows the stimulus to be executed in both modes, e.g. for photometry.
- Potentiostatic: Restricts the protocol to Potentiostatic mode. Stimulus amplitudes are entered in (V), *Holding* corresponds to *E-initial*. If the stimulus is output at a DA channel, the stimulus amplitude is scaled with the *Stim. Scale* factor specified in the Configuration.
- Galvanostatic: Restricts the protocol to Galvanostatic mode. Stimulus amplitudes are entered in (A), *Holding* corresponds to *I-initial*. If the stimulus is output at a DA channel, the stimulus amplitude is scaled with the *Stim. Scale* factor specified in the Configuration.




Filter Factor: The *Filter Factor* is implemented for the EPC 9 and EPC 10 amplifiers. It is used to define the automatic filter setting relative to the sample rate (activated by *Auto Filter* in the Configuration window).

Example: for *Sample Interval* = 250 μ s (4 kHz) and *Filter Factor* = 4, a filter cutoff frequency closest to 1 kHz (= 4 kHz / 4) will be selected. The suggested filter frequency is shown in parentheses.

Analysis: Enter an *Analysis Method* that should be executed automatically after the data acquisition. Note that *Automatic Stimulus Control* has to be activated in the Analysis window for this feature to become effective.

Relevant Segments: The *Rel X-seg* (relevant X-segment) specifies a segment of interest that is mainly used as X-axis reference for later analysis. The *Rel Y-seg* (relevant Y-segment) specifies the segment where the analysis is performed (e.g., determination of peak current). For the measurement of an h-infinity curve, for example, the *Relevant X-Segment* would be the conditioning segment of variable voltage, while the *Relevant Y-Segment* would be the test segment, where the peak current is determined.

10.13 Stimulus Template Preview



After each editing operation, the stimulus template is refreshed to reflect the changes made. Sweep other than the first one are shown as dashed lines. If part of the pulse pattern exceeds the DA limits, the forbidden voltage region is indicated in the picture by shading. The corresponding warning box "Segment with too large voltage encountered" only appears once. Segments drawn in red color refer to the *relevant Y-segment*. The type of graph scaling is chosen in Display → Labeling → PGF-Editor Grid.

Preview Mode: Select from the list what shall be displayed in the preview:


- Draw: Active Channel, all Sweeps: Displays the settings of the active channel for all Sweeps. Recommended if you use increment modes in your settings.
- Draw: Active Channel, 1. Sweep: Displays the settings of the active channel for the first Sweep.
- Draw: Active Channel, Last Sweep: Displays the settings of the active channel for the last Sweep.
- Draw: All Channels, all Sweeps: Displays the settings of all channels for all Sweeps. Recommended if you use *Increment* modes in your settings.
- Draw: All Channels, 1. Sweep: Displays the settings of all channels for the first Sweep.
- Draw: All Channels, Last Sweep: Displays the settings of all channels for the last Sweep.
- Draw no cartoon: Disables the display of the stimulus template.

Delay: Displays the hardware delay time for the DA output and the AD input with respect to the first DA output channel.

Delay: DA AD




10.14 E-initial



E-initial (V): This control displays the presently selected initial potential. It is **only** used for the sequence cartoon as reference and can be changed without actually affecting the initial potential in the Potentiostat window (*E-initial*). When the PGF template of a stored file is reviewed, *E-initial* is the holding voltage of that particular experiment.

Set Last Seg. Amplitude: At the end of the *Sweep*, *E-initial* is set to the amplitude of the last segment, e.g., 0.5 V, if this was the potential of the last segment. Normally, the potential is reset to *E-initial*.

10.15 P/n Leak Correction



PGF Leak Pulse settings: *No of Leaks*: 4, *Leak Delay*: 100 ms, *Leak Size*: 0.250, *Leak Hold*: -120 mV. The *V-Holding* is set to -70 mV.

Leak min/max: The minimum and maximum voltage which will be applied during the leak pulses is shown here. The calculation of the minimum and maximum values of the leak pulses is:

$$L(i) = (V(i) - V_h) * size + L_h \quad (10.1)$$

i=pulse number, V(i)=voltage of segment i, L(i)=voltage of leak segment, V_h=holding potential, L_h=leak holding, size=leak size;

Example: i=2, V(2)=-50 mV, L(2)=voltage of leak segment, V_h=-70 mV, L_h=-120 mV, size=0.25;

$$L(2) = (-50mV + 70mV) * 0.25 - 120mV = \underline{-115mV} \quad (10.2)$$

No of Leaks: Arbitrary numbers of so-called "P/n pulses" are supported.


Leak Delay: This is a waiting time between the leak pulses and the main pulse. In between leak pulses no extra waiting time is supported. P/n pulses precede the test pulse if a negative *Leak Delay* is supplied or they follow it if *Leak Delay* ≥ 0.

Leak Size: Determines the amplitude of the leak pulses relative to the main pulse.

Leak Hold: Specifies the potential from which leak pulses are generated.

If executing a PGF sequence with such *P/n Leak Correction* settings the leak pulses will look like this:





Classical P/4 protocol with 4 leak pulses scaled down to a fourth of the original pulse amplitude. In the classical P/n protocol the *Leak Size* is always equal to 1 divided by the *No of Leaks*. This is not necessary with the POTMASTER software: the leak pulses are summed and scaled with a factor of $1/(\text{No of Leaks} * \text{Leak Size})$.


Note: You can improve the signal-to-noise ratio if you increase *Leak Size* and *No of Leaks*, but be sure that the leak pulses do not reach the activation level of the channels you want to study.

Note: The scaling of leak Traces depends on the scaling of the parent Trace (Trace to which the leak Trace belong). Therefore you have to select the respective parent Trace in the *Oscilloscope* when you want to scale the leak Traces.

Software/Hardware Leak: In case an iTEV 90 amplifier is used the user can select between Software P/n and Hardware P/n function.

Leak Alternate: If *Leak Alternate* is on, leak pulses of alternating polarity are generated.

The following figure illustrates a pulse template.




Alt. Leak Average: This procedure was intended to be used with the *Average* function provided in the *Conrol Window* (see chapter 8.2 on page 92). The *Average* function acquires a given number of averages acquired for one *Sweep*. Only the average is stored to disk. *Alt. Leak Average* provides the option to generate leak pulses of alternating polarity while averaging. That means, if the number of averages in the *Protocol Editor* is set to "4", then the polarity of the leak pulses for the 2nd and 4th average is inverted. Averaging of the leak pulses thereby eliminates slow capacitive currents arising from the jump from holding to *Leak Hold*.


Note: An even number of Sweeps to average is required to give correct results. The entire leak template of every other Sweep is then inverted.

Wait:

- wait = abs. hold: Keeps the cell potential at the chosen absolute holding potential during wait periods.




- wait = rel. hold: Keeps the cell potential at the chosen relative holding potential during wait periods.



- wait = abs. leak hold: Keeps the cell potential at the chosen absolute leak holding potential during wait periods.





10.16 PGF Parameters

p1	p2	p3	p4	p5	p6	p7	p8	p9	p10
0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000

Use the *PGF parameters* as variables in the segment settings for *Voltage* or *Duration*. This allows you to change multiple settings in one or more sequences with changing only one parameter. POTMASTER provides 10 *PGF parameters* which will be stored with the Pulse Generator file. Instead of *p1* to *p10* you can assign use specific names to make the parameter input more intuitive.

Example: You want to increase the duration of three segments during a loop in a protocol. Enter the start duration in field "p1", set the duration of all segments to "p1" and then increase "p1" during loop execution.

For a description of usage of the PGF parameters please we refer to the chapter *Global Variables in PATCHMASTER* in the PATCHMASTER Tutorial.


10.17 Error Handling

If POTMASTER encounters an unreasonable value in the Pulse Generator window, the user is requested to change the corresponding parameter before proceeding. In case of multiple errors, one cannot exit the dialog until all parameters are set correctly. If it seems impossible to solve the error situation and you consistently get error messages, you can load a valid PGF file from disk.

Values in the Pulse Generator controls are rounded to the exact values as displayed. This prevents unexpected results caused by rounding problems, such as the *Sample Interval* being an odd number.


11 Protocol Editor Window

In the Protocol Editor window you can assemble complex experimental arrangements by combining PGF-templates with other operations (e.g. breaks, IF-THEN loops, setting changes). This window is the heart of the POTMASTER software concerning the automation of experiments.



With protocols, very complex sequences can be edited. During protocol execution no opening/closing of windows is necessary, thus the execution is pretty fast. Furthermore, in protocols it is possible to use conditional events, depending on e.g., analog or digital input, analysis results or key commands.

11.1 Protocol Handling



Protocol Pool: The first row displays a section of the pool of available protocols. It is a paging bar in units of six protocols. Two arrows at each side allow scrolling through the available protocols (the innermost arrows move in increments of one page, i.e., six protocols; the outermost arrows move to the start/end of the sequence list). A protocol is selected by clicking on it or by entering the protocol number in the blue field on the right side next to the arrows.

After modification of the existing pool of protocols, the entire Protocol Editor file (*.pro) should be saved to disk (see below). It can be saved under any name. The default name is DefProt.pro. If a custom *.pro file should be loaded into the Protocol Editor as a default, the new name of the *.pro file has to be specified in the Configuration window as *Protocol File* (see chapter 5.5 on page 51) and the *Configuration File* has to be saved.

LOAD : Loads the pool of available protocols (*.pro file).

MERGE : Loads protocols from another (.pro file) into the currently opened file. In case a protocol name is not unique, you are prompted to either skip this protocol, mark it invalid, or give it a new name.

SAVE : Saves the pool of available protocols (*.pro file).

Examp11 : The name of the present protocol is displayed in the blue field. To edit this name, click into the field. End editing either by pressing RETURN or by clicking with the mouse into another area of the window.

NEW : Creates a new protocol in the pool.

COPY : Creates a copy of the active protocol and inserts it at the end of the protocol pool.

MOVE : Moves the present protocol to another position in the protocol pool.

DELETE : Removes the present protocol from the pool.

Record Macros : Activates the macro recording. In the following the command lines for all activated events are recorded and inserted into the active protocol. Press again on *Record Macros* to finish the macro recording.

Relative Value : Interprets the subsequent *Vhold* potential as a relative value ("True"). It is a macro command and has to be activated via the *Record Macro* button. It has to be repeated prior to each *Vhold* macro command.


STEP : Steps to the next line in the event list and executes the event.

TO END : Runs the protocol through to its end.

Write : Writes the events to the Notebook window during execution of a protocol. The *Write* button should also be activated if you want to log the execution of single PGFs in the Notebook.

11.2 Event Handling

In the event area, you can select events for the actual event list that should be performed during the experiment.



Events: Here the position of the present event in relation to the total number of events is displayed.

Insert Before / Insert After: Here you can choose a new event from the drop-down list. It will be positioned either above or below the present entry.

Duplicate: Duplicates the present event. The duplicate will automatically be inserted above the duplicated event.

Delete: Deletes the present event from the list without further notice.

Move: Allows you to move the present event in the event list to another position by entering a new position number.

***Note:** Multi-line events such as Repeat loops or conditional statements can not be moved.*



***Note:** Since protocol files are stored in text format, editing of protocols with standard text editors is also possible.*

11.3 Recurring Functions

These functions and information fields exist in every event area.

Skip : Omits this step in the protocol execution. This option is included to allow the user a quick disabling of events without losing all settings.

Delay : Sets a delay time in seconds. The total time of the event is subtracted from the entered delay time value. E.g. an *Acquire Series* event needs in total 8 s for the execution and a *Delay* of 10 s is specified for this event. This would result in a delay (waiting time of the event) of 2 s (Delay time - Event execution time). The delay is executed after the event.

Repeat Status	Repeat at line 2: 2 of 5
IF etc. Result	Auto C-slow = 1.0000 < 1.0000 => FALSE

Repeat Status: Displays the status of loops. The following example shows that the repeat step happens at line 2 of the protocol and that there have been 2 repeats in a 5 cycle loop.

IF etc. Result: Displays the last IF condition that has been triggered.



11.4 Events - Overview

The available events are categorized into eight groups. Not all of the events are explained in detail here. Please follow the given links to other chapters.

Protocol Sequence	Data/Display
Break	Analysis
Chain Protocol	Digital Filters
Clear Key	Display Properties
Goto	Export
Goto_Mark	File Operation
IF...Then	Replay
Elself...Then	Trace Buffer
Else	Value/Parameter
Launch	PGF Parameters
Macro Command	Set Solutions
Repeat	Set Value
Switch Window	Messages
Wait	Annotation
Acquisition	Beep
Acquire Each Sweep	Write Icon Value
Acquire Properties	Extensions
Acquire Series	LockIn
Set Sweep Label	EIProScan
Test Pulse	Photometry/Imaging
Hardware	DocuLUX
Amplifier	E-Chem Methods
Serial Output	Cyclic Voltammetry
Set DAC	Impedance Spectroscopy
Set Digital Bit	Open Circuit Potential
Set Digital Word	Square Wave Voltammetry
Set Solution Changer	Optical Spectroscopy



11.4.1 Protocol Sequence

11.4.1.1 BREAK

Use to terminate a repeat loop or the protocol.

The dialog box is titled "BREAK". It contains a checkbox labeled "Skip" which is unchecked. Below this, there is a label "Type" followed by a button labeled "End Protocol".

Type:

- End Protocol: This kind of *BREAK* stops the protocol execution completely at this point.
- End Repeat: This kind of *BREAK* ends the execution of a repeat loop (e.g. *Acquire Each Sweep...*) and jumps out of it to the following events. If this *BREAK* is used outside of a loop, the execution jumps to the end of the protocol.
- End If: This kind of *BREAK* ends the execution of an IF-THEN statement and jumps out of it to the following events. If this *BREAK* is used outside of a loop, the execution jumps to the end of the protocol.

11.4.1.2 Chain Protocol

Use this event to link one protocol to another protocol.

The dialog box is titled "Chain Protocol". It contains a checkbox labeled "Skip" which is unchecked. To the right of the "Skip" checkbox is a "Delay" field with a value of "0.00 s". Below these, there is a "Protocol Name" label followed by a text input field. At the bottom, there is a checkbox labeled "Return" which is unchecked, and two buttons: "Edit Protocol" (highlighted in yellow) and "Select Protocol:".

Protocol Name: The name of the protocol is displayed in this field after one protocol is select in the *Select Protocol* selection list.

Return: If this option is not checked the protocol sequence will stop after executing the selected protocol. If the *Return* option is enabled the protocol sequence will come back to the main protocol and will be continued.

Edit Protocol: Opens the protocol named in the field *Protocol Name*.

Select Protocol: Select a protocol to define the link. After a selection is made the name of the protocol is displayed in the field *Protocol Name*.

11.4.1.3 Clear Key


Normally a keystroke is stored during the execution of a protocol until a new key is pressed. In case the key buffer has to be cleared, the *Clear Key* event can be used. This is typically used when a protocol should react once on a single key stroke (e.g. execute an acquisition). In case the key would not be cleared the protocol would react repeatedly until another key is entered.

The dialog box is titled "Clear Key". It contains a checkbox labeled "Skip" which is unchecked.



11.4.1.4 GOTO

Use to jump to a *GOTO_Mark*.




The GOTO event configuration interface consists of a title bar labeled 'GOTO'. Below the title bar, there is a checkbox labeled 'Skip' and a 'Delay' field set to '0.00 s'. A 'Mark' label is followed by a text input field.

Mark: Enter the name of the mark that denotes the destination of the GOTO jump (see below).

11.4.1.5 GOTO_MARK

Creates a *GOTO_Mark*.




The GOTO_MARK event configuration interface consists of a title bar labeled 'GOTO_MARK'. Below the title bar, there is a checkbox labeled 'Skip' and a 'Mark' label followed by a text input field.

Mark: Enter the name of the mark for a possible destination of a GOTO jump.

11.4.1.6 IF...THEN

Allows the definition of conditional events. Note that the list of available channel variables depends upon the chosen source.



The IF...THEN event configuration interface consists of a title bar labeled 'IF ... THEN'. Below the title bar, there is a checkbox labeled 'Skip' and a 'Delay' field set to '0.00 s'. The interface is divided into two main sections: 'Left Source' and 'Right Source'. The 'Left Source' section contains a text input field with 'ADC' and a dropdown menu with 'AD-0'. The 'Right Source' section contains a text input field with 'Number' and a dropdown menu with '0.0000'. A comparison operator '>' is located between the two source sections.

Left Source:



- ADC: AD channel to be read
 - Digital In: Digital Input
 - Parameters: Choose from a list of parameters, e.g., *Imon*, *C-slow*...
 - Analysis: Choose from the list of available functions (1...16)
 - Value: Use one of the *Values 1...16* that can be defined in the event *Set Value*. In addition, results from the *Analysis* can be stored in these "Values" (see below). Therefore, the "Values" constitute an important interface between the *Protocol Editor* and the *Analysis*.
 - Key: Input a key command
 - Icon Value: Reads the value of an icon.
 - Repeat Count: Enter a number that will be compared with the actual *Repeat Count*. If the *Repeat Count* is... then...
 - Sweep Count: Enter a number that will be compared with the actual *Sweep Count*. If the *Sweep Count* is... then...
 - Last Sweep: If this condition is true
 - Sweep Aborted: In case a *Sweep* has been aborted with a *Break* command or condition in the *Pulse Generator* sequence.
 - Selection Failed: If the defined condition for a selection fails the loop can be interrupted.
 - Break Alert: The "IF Break Alert" event pops up a dialog with a title (can be specified by the user) and the two buttons *Break* and *Continue*. A click on *Continue* proceeds with the first event after the *IF... THEN* loop. *Break* executes the events within the *IF... THEN* loop.
- Note: The Title strings must not include double quotes ("), since these double quotes are used as separators for splitting the event text.*
- Batch Comm.: Checks if the last received batch control string contains a user-defined phrase (see *Controlling PATCHMASTER* in the *PATCHMASTER* Tutorial).
 - Serial Comm.: Checks if the last received string of the serial port communication contains a user-defined phrase (see chapter 5.4.9 on page 50).
 - Series Name: Checks if the last name of a *Series* contains the user-defined phrase.
 - Overflow: Checks if an FIFO overflow occurred ("True") on the AD/DA board.
 - True / False: Choose if the IF loop shall be used (True) or not (False).

✓	ADC
	Digital In
	Parameter
	Analysis
	Value
	Key
	Icon Value
	Repeat Count
	Sweep Count
	Last Sweep
	Sweep Aborted
	Selection Failed
	Break Alert
	Batch Comm.
	Serial Comm.
	Series Name
	Overflow
	TRUE
	FALSE

Right Source:

- ADC: AD channel to be read
- Digital In: Dig-0...15. The operators for the Digital Input set the voltage.
- Parameters: Choose from a list of parameters, e.g., *Imon*, *C-slow*... Additionally, the *Timer* parameter allows to time processes, e.g., reset a timer with the *Display Properties* event and measure time lapse since this event.
- Analysis: Function 1...16
- Value: One of the global variables (*Value-1* to *Value-n*) can be specified.
- Number: User-defined values are entered manually.
- True / False: Choose if the IF loop shall be used (True) or not (False).

☒	ADC
☐	Digital In
☐	Parameter
☐	Analysis
☐	Value
☐	Number
☐	TRUE
☐	FALSE

If Operators: Select the appropriate logical operator for the *IF... THEN* operation.

- >: Greater than
- <: Smaller than
- ABS >: Absolute value greater than
- ABS <: Absolute value smaller than
- =: Equal to
- <>: Between the ranges of
- MOD: A Modulo operation; it is an integer operation. The modulo operator will return TRUE, if the operation parameter value of *MOD* equals zero.
- invalid: Condition of a channel, value etc. has to be *invalid*.

☒	>
☐	<
☐	ABS >
☐	ABS <
☐	=
☐	<>
☐	MOD
☐	invalid

11.4.1.7 ELSEIF... THEN

Allows a secondary conditional event, e.g. a check after an IF-THEN step. Works identically to IF-THEN, see above.

ELSIF ... THEN		☐ Skip
Left Source		Right Source
ADC	>	Number
AD-0		0.0000

11.4.1.8 ELSE

Final step of the IF condition.

ELSE	☐ Skip



11.4.1.9 Launch

The *Launch* event allows to start another application. The path to the application is defined relative to POTMASTER's home path.

- **Application:** Enter path where the application is located.
- **Arguments:** Enter command line arguments if necessary.

11.4.1.10 Macro Command

Use to execute an individual *Macro Command*. Many controls that can not be accessed via special events can be controlled using *Macro Command*.

A listing of all *Macro Command* can be plotted to the Notebook using the function *List All Macro Items* from the Help menu (see chapter 4.10 on page 39.)

Alternatively, you can view the *Macro Command* for individual icons via opening the Dialog Control window (see chapter 2.2 on page 6.)

Command: Enter the name of the *Macro Command* you wish to include, e.g., "E Zap".

11.4.1.11 REPEAT

Inserts a repeat loop. There are two types of repeat loops.

- **Repeat Counts:** Repeats for a given number of repeats.
- **Repeat Indefinite:** Repeats in an indefinite loop.

Duration: Time between repeats.

Increment: Increments duration.

Note: REPEAT events cannot be moved in the event list!



11.4.1.12 Switch Window

Brings the selected window in front.

The 'Switch Window' dialog box contains a 'Skip' checkbox, a 'Delay' field set to '0.00 s', and a drop-down menu currently showing 'Switch to Control Window'.

In the drop-down menu you can select the a POTMASTER window you want to switch to:

A list of POTMASTER windows with a checkmark next to 'Switch to Control Window'.

- Switch to Control Window
- Switch to Configuration
- Switch to Oscilloscope
- Switch to Potentiostat
- Switch to Replay
- Switch to Protocol Editor
- Switch to Pulse Generator
- Switch to Analysis
- Switch to Analysis Window 1
- Switch to Analysis Window 2
- Switch to Parameters
- Switch to I/O Control
- Switch to Photometry/Imaging
- Switch to Zero Potential (E-zero)
- Switch to LockIn
- Switch to Spectroscopy
- Switch to Solution Base
- Switch to Solution Changer
- Switch to Markers
- Switch to Protocol Methods
- Switch to Calculator
- Switch to Notebook

11.4.1.13 WAIT

The 'Wait' dialog box contains a 'Skip' checkbox, a 'Wait Type' field set to 'Absolute', and a 'Duration' field set to '0.00 s'.

Wait type:

- Absolute: Waits exactly for the given delay time after the previous event.
- Relative: Waits for the given delay time after the actual end of the previous event (e.g., plus additional delay times).
- Resume Icon: Waits until the *Resume* button is clicked (i.e. sets the *WAIT* button).
- Key: Waits until the specified key is entered. You can either enter a key or mark the *Any Key* option.
- Comment Alert: Will display a standard alert window where the user has to enter a new comment text. This comment will be used as long as no other is given.



11.4.2 Acquisition

11.4.2.1 Acquire Series

Used to execute an entire PGF sequence.

Edit PGF Template: Opens the PGF template of the selected sequence in the Pulse Generator window.

Sequence: Indicates the name of the selected PGF.

Label: Enter a *Series* label, e.g. "Drug A".

Comment: Enter a comment to the *Series*. The entry is stored with the *Series* and therefore available in the *Comment* field for the *Series* in the *Parameter* window.

Averages: A PGF sequence can be repeated several times. Only the average is stored to disk. Please enter the number of averages acquired for one *Sweep*. If the box in front of the "Averages" label is not checked, then this setting is ignored.

Break Keys: Define a key to break the acquisition of the running *Series*.

11.4.2.2 Acquire Each Sweep

This event allows executing other protocol events between the acquisition of *Sweeps*. E.g. use the *Amplifier* event to execute an *AutoCSlow* update before acquisition of each *Sweep* in a *Series*. A repeat loop is inserted with an *Acquire Each Sweep* event inside.

Duration: Defines the total cycle time for one *Sweep*.

Increment: Sets a defined time increment for the *Sweeps*.

For a description of the *Acquire Each Sweep* event please refer to the *Acquire Series* event above.



11.4.2.3 Acquire Properties

Update R-membrane and I-pipette: Compute and update the calculated *I-pipette* and *R-membrane* after every acquired *Sweep*. Enable the checkbox *Set It* in front of this option if you really want to execute it.

Wipe Display at Start: Determines whether the display is cleared at the start of a protocol or not. Enable the checkbox *Set It* in front of this option if you really want to execute it.

Write Events to Notebook: Writes the events to the *Notebook* window. This function is identical to the *Write* button. Enable the checkbox *Set It* in front of this option if you really want to execute it.

Set Min. Wait Time: Sets the time POTMASTER reserves to wait for the correct time to start when executing individual repeat loops such as acquisition of a *Series of Sweeps* or an *Repeat* loop in the *Protocol Editor*. Enable the checkbox *Set It* in front of this option if you really want to execute it. The *Min. Wait Time* can also be set in the *Configuration* window (see chapter 5.3.3 on page 44).

Note: A long Min. Wait Time will make the program to react very slowly. A very short Min. Wait Time might reduce the accurate timing of the starting of events. The timing within an acquisition; e.g sampling rate, is not affected by this parameter. Sampling within a Sweep is always accurate.

11.4.2.4 Set Sweep Label

Allows to put a label on the *Sweep* to be acquired next.

Sweep Label: Enter the name of the label into this entry field.

Sweep Label Target:

- Set in Next Sweep: The *Sweep Label* is acquired in the next *Sweep*.
- Set in Selected Sweep: The *Sweep Label* is acquired already in the selected *Sweep*.

The *Sweep Label* can be edited after acquisition in the *Replay* window by pressing the *Label* button.

11.4.3 Hardware

11.4.3.1 Amplifier

This event is for controlling the main amplifier settings.



☐ Set Amplifier Mode Standby
☐ OCP Set
☐ E-initial 0.0000 V
☐ Auto Range Once
 Auto Filter: keep

Set Amplifier Mode: Change the *Amplifier Mode* of the Potentiostat/Galvanostat.

- Standby: The amplifier is set to the Standby mode
- OCP: The amplifier is set to the OCP mode.
- Potentiostatic: The amplifier is set to the Potentiostatic mode.
- Galvanostatic: The amplifier is set to the Galvanostatic mode.

OCP Set: The *OCP* is measured and the value is transferred to *E-init*.

E-initial: Sets the desired initial value of the cell potential in the potentiostatic mode.

Auto Range Once: The *Auto Range* in the Potentiostat window is executed once to search for the best *Gain* value.

Auto Filter: This option automatically sets a hardware filter according to the chosen sample interval.

- keep: The status of the *Auto Filter* option is unaffected.
- ON: The *Auto Filter* option is enabled.
- OFF: The *Auto Filter* option is disabled.

Note: Further reference can be found in chapter 7 on page 71.

11.4.3.2 Serial Output

Use to send a string to the opened serial interface. For configuration of the *Serial Output* see Configuration Window, chapter 5.4.9 on page 50.

Serial Output ☐ Skip Delay 0.00 s
 String
 [Empty text input field]

String: Enter a string that will be sent to a device via the *Serial Output*.

Examples:

GO	The command "GO" is send.
GO\r	The command "GO" + <Carriage Return> is send.
GO\l	The command "GO" + <Line Feed> is send.
GO\032	The command "GO" + ASCII Code 32 (<Space>) is send.

11.4.3.3 Set DAC

Set DAC ☐ Skip Delay 0.00 s
 DA-0 = 0.000 V
☐ Use Dac Scaling



Channel: Select a channel for analog output.

Operation: See 11.4.5.3 on page 151.

Voltage: Enter a voltage that will be output via the given channel.

Use Dac Scaling: If this option is active then the given value is multiplied by the DAC scaling factor specified in the Configuration dialog.

Note: The Set DAC event does not consider any scaling applied by equation or lookup table. Only entered scale factors are considered if the Use Dac Scaling is active.

11.4.3.4 Set Digital Bit

Channel: Select a channel for digital output.

low: If selected, 0 Volt will be sent via the channel. If not selected, 5 Volt (standard TTL setting) will be sent.

11.4.3.5 Set Digital Word

Digital Word: The entered number will be automatically translated into the digital bits.

Digital Bits: Allows selecting the digital bits individually.

11.4.3.6 Set Solution Changer

Solution: Select a *Solution* (1-16) at the connected *Perfusion System*.

Note: The output channels and their aligned solutions have to be defined first in the Solution Changer dialog (chapter 19 on page 225).

Set Marker: If selected, a *Marker* will be set when the *Solution* is changed by the *Solution Changer* event. The label of the *Marker* can be entered in the entry field behind.



11.4.4 Data/Display

11.4.4.1 Analysis

Analysis Mode:

- Analysis: Keep: Preserves the current settings of the Analysis.
- Analysis: Use Given Method: Sets the Analysis mode to *Use Selected Method* and selects the *Analysis Method* specified in the field Name.
- Analysis: Use Selected Method: Sets the Analysis mode to *Use Selected Method* and uses the active *Analysis Method*.
- Analysis: Auto Stim Control – Pool: Sets the Analysis mode to *Automatic Stimulus Control: Pool* and uses the *Analysis Method* specified in the PGF sequence.
- Analysis: Auto Stim Control – Assigned: Sets the Analysis mode to *Automatic Stimulus Control: Assigned* and uses the *Analysis Method* which was assigned to the PGF sequence.
- Analysis: No Analysis: The Analysis is deactivated.

Name: Enter the *Analysis Method* you wish to start.

Edit: Opens the Analysis window with the given method for direct editing.

Wipe: Can be used to wipe individual graphs or all graphs at once.

- No Wipe
- Wipe All Graphs
- Wipe Graph 1...16

Set Graph Pos.: Activate the checkbox in front to apply the *Graph Positions* to the selected *Analysis Method*.


11.4.4.2 Digital Filters

Defines the cut-off frequencies of digital filters in Hz for *Trace 1...16* (1 – 16) and *Buffer 1...4* (B1 – B4). The checkbox is used to mark which of the filters are to be set.



11.4.4.3 Display Properties

Use to control settings of the `Display` menu and the Oscilloscope window.



Wipe Osci: Deletes all displayed data in the Oscilloscope window.

Wipe Analysis: Deletes all displayed data in the Online windows.

Reset Timer: Resets the timer in the Oscilloscope window. These options are also described in section 4.5 on page 29.

Overlay Series and Overlay Sweeps:

- keep: Keep the state that was set before this event.
- ON: Overlay on.
- OFF: Overlay off.

Subtract Leaks:

- keep: Keep the state that was set before this event.
- ON: Subtract leak signals on.
- OFF: Subtract leak signals off.

Subtract Zero:

- keep: Keep the state that was set before this event.
- ON: Subtract zero offset on.
- OFF: Subtract zero offset off.

Show Leaks:

- keep: Keep the state that was set before this event.
- ON: Show leak signals on.
- OFF: Show leak signals off.

Show Tree Info:

- keep: Keep the state that was set before this event.
- ON: Show Tree Info in the Notebook - on.



- OFF: Show Tree Info in the Notebook - off.

Show Traces 1...9, 0...6: Changes the show flag of all 16 *Traces*. This function can be used to show only a subset of acquired *Traces* in the Oscilloscope window (see chapter 4.5 on page 29). Note that you have to mark *Show Traces* for the changes to take effect.

11.4.4.4 Export

Exports the acquired data according to the export settings made in the `Replay` menu. Unless you activate the checkbox *Full Sweep* only the displayed data is exported (see chapter 4.4 on page 23).

Filename: Define a name for exported file according to the auto filename components described in the Configuration (see chapter 5.5.2 on page 52).

Target: Declares the *Target Specifier* in the tree. 4 comma separated integers are allowed, specifying the index of *Group*, *Series*, *Sweep* and *Trace* (Gr_Se_Sw_Tr). "+" and "-" define relative index from the present selection. An index of "0" defines the present selection.

Examples:

- "0": Exports the present *Group*.
- "0,0": Exports the present *Series*.
- "0,0,0,0": Exports the present *Trace*.
- "0,-1": Exports the previous *Series* of the present *Group*.
- "1,4": Exports the 4th *Series* of the 1st *Group*.
- "1,1,4,1": Exports the 1st *Trace* of the 4th *Sweep* of the 1st *Series* of the 1st *Group*.

Overwrite: Enables to overwrite already existing output files.

Full Sweep: The full *Sweep* is exported independent of what is shown in the Oscilloscope.

11.4.4.5 File Operation

Store Data:



- **keep:** Keeps the status that was set before this event (either ON or OFF of the *Store* button).
- **ON:** Acquired data will be saved to disk. This corresponds to the active *Store* button.
- **OFF:** Data will not be saved to disk. This corresponds to the deactivated *Store* button.

Save after Break:

- **keep:** Keeps the status that was set before this event (either ON or OFF of the *Save after Break* parameter in the Configuration window).
- **ON:** Acquired *Sweep* data will be stored if the *Sweep* was terminated with the *Break* command.
- **OFF:** *Sweep* data will not be saved to disk if the *Sweep* was terminated with the *Break* command.

Close File: The data file will be closed.

New Group, New Experiment, New File: A new group, experiment or file can be created. The ability to be able to create a new data file is very convenient when the auto file name generation is used.


Label: It is active when a new group or a new experiment is created and it specifies the group label.

Update File: Updates and saves the file.

File Status: Prints the file status to the Notebook (see chapter 4.1 on page 17).

11.4.4.6 Replay

Data of the opened file can be replayed.



Select the target level of data which should be replayed:

- **Nothing:** Nothing is selected.
- **Root:** Select the *Root*.
- **Group:** Select the *Group*.
- **Series:** Select the *Series*.
- **Sweep:** Select the *Sweep*.
- **Trace:** Select the *Trace*.

The options refer to the target level selection (see above).

- **Select Next:** Selects the next target (no replay).
- **Select Present:** Selects the present target (no replay).
- **Select Last:** Selects the last target (no replay).



- **Replay Next:** Selects the next target and replays the data.
- **Replay Present:** Selects the present target and replays the data.
- **Replay Last:** Selects the last target and replays the data.

More options can be selected via the checkboxes:

Marked Only: Only marked data in the Replay tree will be affected.

Wipe Osci: Deletes all displayed data in the Oscilloscope window.

Wipe Online: Deletes all displayed data in the Online window.

11.4.4.7 Trace Buffer

Trace Buffer Source: Define the data source you want to operate with.

Operation: Define the operation which should be applied to data before storing it in a *Buffer*.

- no operation
- add to
- subtract from
- accumulate in
- deaccumulate from

Target: Select the target *Buffer* (*Buffer 1...4*).

Replace: Replaces the selected *Trace* with the *Buffer*

Clear Buffer: Clears the target *Buffer* before storing the data.

Scaling:

- no operation
- **Scale:** Scales the *Trace Buffer Source* (*Trace* is selected) or *Buffer* (*Buffer* is selected) with a specified *Factor* and *Offset*.

- **Equation:** Calculates the scaling of the *Buffer* via an equation. For details on the equation syntax please refer to *Equation Syntax*, chapter 21.1 on page 232.

Update Display: Updates the display of the *Buffer* in the Oscilloscope window.


Update Analysis: Updates the display of the *Buffer* in the *Analysis* graphs.



11.4.5 Value/Parameter

11.4.5.1 PGF Parameters

PGF parameters can be set or modified during the protocol to automatically adapt the acquisition sequences.



The image shows a software dialog box titled "PGF Parameters". It contains a checkbox labeled "Skip" which is unchecked. To its right is a "Delay" label followed by a text input field containing "0.00 s". Below these, there is a text input field containing "PgParameter-4", followed by an equals sign "=" in a separate box, and then another text input field containing "0.0000".

PGF Parameters 1...10: Specify a PGF parameter to be set.

Operation: To modify a PGF parameter the following functions can be used.

- Functions working on manually entered *Values*:

- =
- increase by
- decrease by
- multiply by
- divide by


- Functions working on global *Values*:

- = value
- inc by value
- dec by value
- mul by value
- div by value

Value: Manually entered value or selection of a global *Value* (see PGF Parameters, chapter 10.16 on page 130).

Note: Please keep in mind that the stimulus amplitude is given relative to 1 V or 1 nA in voltage clamp and current clamp mode, respectively. I.e., to obtain a stimulus amplitude of 100 pA in current clamp mode you have to enter a value of "100m" ($= 0.1 * 1 \text{ nA}$).

11.4.5.2 Set Solutions



The image shows a software dialog box titled "Solutions". It contains a checkbox labeled "Skip" which is unchecked. To its right is a "Delay" label followed by a text input field containing "0.00 s". Below these, there are three rows of controls:

- A checkbox labeled "Set Internal" followed by a text input field containing "0".
- A checkbox labeled "Set External" followed by a text input field containing "0".
- A checkbox labeled "Set Marker" followed by an empty text input field.

Set Internal: User-defined index number for the internal solution. Activate the checkbox to apply the setting.


Set External: User-defined index number for the external solution. Activate the checkbox to apply the setting.

Set Marker: Set a solution marker when the solution is changed. In the entry field a label for the marker can be defined. Activate the checkbox to apply the setting. See also *Solutions*, chapter 18 on page 221.



11.4.5.3 Set Value

Allows you to freely define and store values that can be used in an *IF... THEN* event such that the *If* event can make decisions based on these stored values. The values can also be set via "batch control", thus allow triggering an *IF* decision by an external input, see *Controlling Patchmaster in the PATCHMASTER Tutorial*. In addition, results from the *Analysis* can be stored in these values (see *Analysis Functions*, chapter 13.3 on page 181). The values, therefore, constitute an important interface between the *Protocol Editor* and the *Analysis*.




The 'Set Value' dialog box contains the following elements:

- A checkbox labeled 'Skip'.
- A 'Delay' field with a value of '0.00 s'.
- A row of three buttons: 'Value-1', '=', and '0.0000'.
- A 'don't copy' button below the row.

Value 1...16, User 1-2: Refers to the global *Values* and the *User Parameters*. *User Parameters* can be defined in the *Configuration* window, see *I/O Parameters*, chapter 5.7.2 on page 59.


Operation: Set value to...

- manually entered values:
 - =
 - increase by
 - multiply by
 - divide by
 - modulo
 - keep: no change to the value
- an *Item*:
 - = item
 - inc by item
 - dec by item
 - mul by item
 - div by item
- *Parameters*:
 - = param
 - inc by param
 - dec by param
 - mul by param
 - div by param
- Global *Values*:
 - = value
 - inc by value
 - dec by value
 - mul by value
 - div by value
- *Analysis* results:
 - = online



A dropdown menu with a checkmark icon next to the '=' option. The menu items are:

- =
- increase by
- multiply by
- modulo
- keep
- = item
- inc by item
- dec by item
- mul by item
- div by item
- = param
- inc by param
- dec by param
- mul by param
- div by param



A dropdown menu showing a list of options:

- = value
- inc by value
- dec by value
- mul by value
- div by value
- = analysis
- inc by analysis
- dec by analysis
- mul by analysis
- div by analysis
- = equation
- = input
- = batch comm. value



- inc by online
 - dec by online
 - mul by online
 - div by online
- *equation* results: = equation. An equation can be entered to compute a value. For details on the equation syntax please refer to *Equation Syntax*, chapter 21.1 on page 232.
 - = input: A dialog opens for user input of a new numeric value.
 - = batch comm. value: A value received via the batch communication interface is assigned to the respective value.

Value:

- in case of constant value: Enter manually the constant value.
- in case of *Item* operation: Enter the item ID, e.g. "E CFastTot" to read the value of a program control (here compensated *C-fast* value). For a description how to get info on the control IDs please refer to *Modifying Dialogs and Controls*, chapter 2.2.3 on page 10.
- in case of *Parameter* operation: Choose a parameter from the selection.
- in case of *Value* operation: Choose one of the 16 global *Values*.
- in case of *Analysis* operation: Choose one of 16 *Analysis* results.

Copy:

- don't copy: No further operation.
- copy to item: Copies the value to a specified item.
- add to item: Adds the value to a specified item.
- multiply to item: Multiplies specified item with the value.

11.4.6 Messages

11.4.6.1 Annotation

Use this function to add comments to your protocol and improve the readability.


A screenshot of a software dialog box titled "Annotation". It features a light gray header bar with the title. Below the header, the word "Annotation" is displayed in a smaller font. At the bottom of the dialog, there is a text input field with a blue border and a light blue background, currently empty.

Text can be entered that will be shown in the event list.



11.4.6.2 Beep

Use to execute acoustic alerts.




The Beep event configuration dialog has a title bar 'Beep'. It contains a checkbox labeled 'Skip' which is unchecked, and a 'Delay' field set to '0.00 s'.

An acoustic signal will be output. This only works if your computer is properly equipped for sound (e.g. sound card, loudspeakers...).

11.4.6.3 Write Icon Value

Outputs the value of the specified icon into the Notebook like this:

[Title] [ItemID]: [value]



The Write Icon Value event configuration dialog has a title bar 'Write Icon Value'. It contains a checkbox labeled 'Skip' which is unchecked, and a 'Delay' field set to '0.00 s'. Below these are two text input fields labeled 'Title' and 'Item ID'. At the bottom, there are two checkboxes: 'Notebook' (checked) and 'Alertbox' (unchecked), followed by a checkbox labeled 'LF' (checked). To the right of these is a text field labeled 'Icon Text'.

Title: Title for the *Item ID*, leads to an improved readability of the output but is not mandatory.

Item ID: Unique identifier of the icon. Can be inspected by opening the Icon configuration dialog (see *Modifying Dialogs and Controls*, chapter 2.2 on page 6) – it is given in the title of that window.

Notebook: Writes to Notebook.

Alertbox: Opens an *Alertbox* containing the text.

LF: When LF is on, a linefeed signal is appended.

Icon Text / Icon Value:


- **Icon Text:** Writes the text from a field. For example from the icon *E Gain* would return "1.0 mV/pA".
- **Icon Value:** Writes the value of a field. For example from the icon *E Gain* would return "8.0", meaning the 8th item of the *Gain* list corresponding to "1.0 mV/pA".

11.4.7 Extensions

11.4.7.1 ElProScan

Note that the `ElProScan` Extension has to be active to allow the editing of the following events (see *Configuration*, chapter 5.4.8 on page 49).

The *ElProScan* event enables users to modify the motor and scan parameters.



The ElProScan event configuration dialog has a title bar. It contains a checkbox labeled 'Skip' which is unchecked, and a 'Delay' field set to '0.00 s'. Below these is a section titled 'Absolute Position' containing four checkboxes and their corresponding values: 'Set X-Position' (0.000 µm), 'Set Y-Position' (0.000 µm), 'Set Z-Position' (0.000 µm), and 'Set Piezo-Position' (0.000 µm). At the bottom of this section is a checkbox labeled 'No Scan'.

Position:

- **Absolute Position:** The motors can be moved to an absolute position by entering the destination position
- **Relative Position:** The motors can be moved by a certain distance.

X-Position: Enter a value into the entry field to define a *X-Position*. Enable the checkbox in front to activate the entered value.

Y-Position: Enter a value into the entry field to define a *Y-Position*. Enable the checkbox in front to activate the entered value.

Z-Position: Enter a value into the entry field to define a *Z-Position*. Enable the checkbox in front to activate the entered value.

Piezo-Position: Enter a value into the entry field to define the *Piezo-Position*. Enable the checkbox in front to activate the entered value.

Scan Mode:

- **No Scan:** No scan will be performed.
- **Tip-Down:** Moves the electrode towards the substrate in Z-direction.
- **Pull Up:** retracts the electrode from the substrate in Z-direction.
- **DepthScan:** A two dimensional scan in the X-Z plane can be performed.
- **2D-Scan:** A line scan along one of the three axis can be performed.
- **3D-Scan:** A two dimensional scan in the X-Y plane with correction of the Z-axis height to compensate for a double tilted substrate can be performed.
- **Matrix Scan:** A *Series* will be executed at each scan point of a two dimensional plane.
- **Template Scan:** Like a *Matrix Scan* but using loaded template.
- **Const-Dist 3D:** Like a *3D-Scan* but using a constant distance.
- **Const-Dist Matrix:** Like a *Matrix-Scan* but using a defined constant distance.
- **SF-Spectra Scan:** Performs a *Shear-Force* scan to determine the test frequency.
- **Fast 3D Scan:** Can only be used if X and Y piezo motors are present. A two dimensional scan in the X-Y plane with correction of the Z-axis height to compensate for a double tilted substrate can be performed. Exclusive use of piezo motors in all directions.
- **Fast Matrix Scan:** Can only be used if X and Y piezo motors are present. A *Series* will be executed at each scan point of a two dimensional plane. Exclusive use of piezo motors in all directions.
- **Fast C. D. 3D Scan:** Like a *Fast 3D Scan* but using a constant distance mode from the ElProScan Menu.
- **Fast C. D. Matrix Scan:** Like a *Fast Matrix Scan* but using a constant distance mode from the ElProScan Menu.
- **Surface Tracked 3D Scan:** Like a *3D Scan* but keeping a constant distance to the substrate



11.4.7.2 Photometry/Imaging

Note that the `Photometry` Extension has to be active to allow the editing of the following events (see `Configuration`, chapter 5.4.7 on page 48 and `Photometry Extension`, chapter 22 on page 237).

Photometry

☐ Skip Delay

☐ New Resting Wavelength

☐ Set Resting Wavelength

☐ Set New Wavelength

☐ Set New Speed

☐ Open Shutter

☐ Close Shutter

New Resting Wavelength: Sets the variable *Resting Wavelength* to the given value.

Set Resting Wavelength: Sets outputs *Resting Wavelength*. Note that this function is called "Close shutter" for those systems that have a shutter. In this case, the option *New Resting Wavelength* disappears.

Set New Wavelength: Sets a specified wavelength.

Set New Speed: Sets a new speed setting of the filter wheel.

Open Shutter / Close Shutter: Note that this function is called *New Resting Wavelength* for those systems without shutter (see above).

*Note: The amount of available options for the Photometry event depend on the type of photometry device selected in the **Configuration** window.*

11.4.8 Echem-Methods

The *Echem-Methods* of the Protocol Editor are explained in detail in chapter 12 on page 157.





12 Protocol Editor Window - Electrochemical Methods


The electrochemical methods are a special feature of POTMASTER. They differ from other operations in the Protocol Editor because they automatically create a transient *Pulse Generator Sequence* and an *Analysis Method* upon execution of the protocol. Transient means that they will not be saved with other changes made in the Pulse Generator Window and Analysis Window, but instead created or overwritten when executing the protocol. Therefore, if changes should be made by the user to both the *Pulse Generator Sequence* and *Analysis Method* a copy with a new name of the created *Pulse Generator Sequence* and *Analysis Method* has to be made.

The idea of these *E-Chem Methods* for standard electrochemical techniques is that the user does not have to create a complex *Pulse Generator Sequence* with several segments and an *Analysis Method* from scratch, but instead a standardized method can be executed immediately where only measurement parameters have to be entered in the Protocol Editor.


In the following, the process of executing an *E-Chem Method* is explained for the example of the Cyclic Voltammetry Method.

1. Create a new *Protocol* in the Protocol Editor Window and give it a meaningful name. We chose "CV-Method".
2. The operation E-Chem Methods – Cyclic Voltammetry is added to the protocol. By clicking on it in the protocol list on the left (black background when selected), a window appears on the right side where the parameters for the experiment are defined. Insert the desired parameters and close the Protocol Editor Window.







3. In the Control Window you can now see and execute the *Protocol* "CV-Method".




4. When executing the *Protocol* a *PGF* called "CV" is created and appears in the Control Window and an *Analysis Method* is created.





5. You can open the Pulse Generator Window to view the newly created *Sequence* called "CV".



6. You can open the Analysis Window to view the newly created *Analysis Method* called "CV".



7. If you want to make changes to this *Pulse Generator Sequence* or this *Analysis Method*, you have to make a copy with a different name. The automatically created *Pulse Generator Sequence* "CV" and *Analysis Method* "CV" will be overwritten upon repeated execution of all *Protocols* using the Cyclic Voltammetry Method. Also, the *Pulse Generator Sequence* "CV" and *Analysis Method* "CV" will not be saved with the other *Pulse Generator Sequences* and *Analysis Methods*.




12.1 Cyclic Voltammetry (CV)

Cyclic Voltammetry (CV) is one of the most commonly used electrochemical methods. It is also called *Triangular Wave Method*, because a triangular potential time curve is applied at the working electrode by a function generator in potentiostatic mode. A redox system of interest is investigated at the working electrode by recording the current which results from the triangular stimulation. The speed of potential change during the anodic and cathodic ramps of the triangular wave is characterized by the *Scan Rate*. Cyclic Voltammetry is used as steady state method (slow *Scan Rates* <1 mV/s) and also transient method (fast *Scan Rates* >10 V/s). The method is used for analytical applications of common electrochemistry, applications for battery research, investigation of fuel cells, characterization of electrocatalysis, electrode characterization in electrolysis cells, investigation of new electrode materials, characterization of sensors, corrosion investigation, investigation of chemical following reactions and electron transfer reactions.

A cycle, i.e. in the convention of POTMASTER one whole *Sweep*, starts with the first reversal potential (Reversal Pot. 1). From here a linear *Sweep* with a constant scan rate reaches the second reversal potential (Reversal Pot. 2). This is the first *Segment* of the PGF Template (also defined as *Relevant Segment*). From the second reversal potential (Reversal Pot. 2) we apply a potential ramp with constant scan rate back to the first reversal potential (Reversal Pot. 1). This is the second *Segment* of the PGF Template.

The cycle is enclosed by a start phase (*Segments* of type *First Sweep*) and an end phase (*Segments* of type *Last Sweep*), which are not shown and stored. The start phase runs a linear *Sweep* from the start potential to the first reversal potential. The end phase runs a linear *Sweep* from the first reversal to the second reversal and then back to the start/end potential. This makes the entire CV stimulus symmetrical with respect to oxidation and reduction processes.



If the numbers of cycles is larger than one, additional cycles (see red and black ramps in the figure above) are executed in the *Gap Free Mode*. Note that the green parts are executed only at the beginning and at the end of the recording.

12.1.1 CV Parameters

Cyclic Voltammetry			
☐ Skip	Delay 0.00 s		
Start/End Potential		0.000 V	
☐ Use E-init	☐ Keep at End	☐ Auto Ranging	
Reversal Pot. 1	-0.100 V	Increment	0.000 V
Reversal Pot. 2	0.400 V	Increment	0.000 V
Scan Rate	20.0 mV/s	Increment	0.00 V/s
Number of Cycles	3	☐ + 1 end cycle	
Anodic Peak: between	-0.100 V	and	0.100 V
Cathodic Peak: between	0.100 V	and	-0.100 V
Baseline: Zero		00:02:40	

Start/End Potential: Potential at which the start phase begins and the end phase stops. At the start of the CV the potential steps from the initial potential to the *Start/End Potential*.

Use E-init: If checked, the initial potential is used as *Start/End Potential*.

Keep at End: If checked, the *Start/End Potential* is set as new initial potential.

Auto Ranging: If checked, the complete CV method will be executed multiple times until it has reached the optimal current range. Only the last CV will be stored.

Reversal Pot. 1: By definition, the CV should start with the more negative reversal potential.

Increment: If multiple cycles are executed the first reversal potential can be altered by a fixed increment from *Sweep to Sweep*.

Reversal Pot. 2: By definition, the second reversal potential should be more positive than the first one.

Increment: If multiple cycles are executed the second reversal potential can be altered by a fixed increment from *Sweep to Sweep*.

Scan Rate: Anodic and cathodic *Sweeps* use the same *Scan Rate* (see also 10.6 on page 103).

Increment: If multiple cycles are executed the scan rate can be altered by a fixed increment from *Sweep to Sweep*.

Number of Cycles: If the number of cycles is larger than one, multiple *Sweeps* are executed in *Gap Free Mode*.

+1 end cycle: This option ensures a symmetric amount of anodic and cathodic peaks. Therefore the last cycle will be prolonged until the last cycle is completed.

Anodic Peak: The *Anodic Peak* will be analyzed as *Maximum* in the segment scanning from *Reversal Pot. 1* to *Reversal Pot. 2*. Scanning towards positive potentials is required. The voltage range for analyzing the maximum current can be defined.

Cathodic Peak: The *Cathodic Peak* will be analyzed as *Minimum* in the segment scanning from *Reversal Pot. 2* back to *Reversal Pot. 1*. Scanning towards negative potentials is required. The voltage range for analyzing the minimum current can be defined.


Baseline: For the peak analysis you can choose if the *Baseline* should be zero or a line drawn through the intersections between current and cursor lines.

Total Length: This display field shows the total length of the CV protocol (1 *Sweep*).




12.1.2 Acquiring a Cyclic Voltammogram

You start the electrochemical method by executing the protocol, which contains the method.




In the Oscilloscope Window the recording is shown versus the time. The method creates a standard *Analysis Method* "CV", which plots the cyclic voltammogram in *Graph 1* and analyzes anodic and cathodic current peaks and the corresponding peak potentials. The results are plotted into the Notebook.

The raw data recorded by the Cyclic Voltammetry method are stored in the data tree as *Series* with multiple *Sweeps*. Each *Sweep* containing the current and the voltage *Trace*.



The parameters of the method, which were used during the acquisition of the data, can be viewed in the *Acquisition Method* window. The window can be opened either via selecting *Show Protocol Method* in the *Replay* menu or via selecting *Protocol Methods* in the *Windows* menu.




The protocol event itself can be copied in the currently active protocol or the corresponding *Analysis Method* copied to the currently opened *Analysis pool*.

12.1.3 Customizing the Analysis Method

If you are acquiring multiple CVs at different *Scan Rates*, you might want to extend the standard CV analysis by a plot of *I-Peak* versus square root of the *Scan Rate*.

Replay the *Series* that contains the CV data. Then open the *Analysis* window and select the *Analysis Method* "CV". Press *Copy* to create another method (default name CV-1), where you can make your changes. Remember: The transient *Analysis Method* "CV" will be overwritten upon repeated execution of all *Protocols* using the *Cyclic Voltammetry Method* and not be saved.



Now, you have to define a new *Analysis Function*. In the *Analysis Function* section in the lower part of the *Analysis* window press *New* and select the function *sqrt* from the *Math* group. You can assign a name to the function and have to set the index of the *Analysis Function* of the *Scan Rate* (here number 3) as input for the square root function.

Analysis Functions

8 CV-max 9 CV-min 10 **sqr(scan rate)** 11

Function Type

Name

Input

Output

☒ Notebook

If you are replaying the data again, the square root of the *Scan Rate* will be written in the Notebook as additional result. Now we want to define a graph that plots *I-peak* versus $\sqrt{\text{scan rate}}$. In the lower section of the Analysis window switch to *Graph 2* and turn it on. Then, select $\sqrt{\text{scan rate}}$ as 'X' and *I-Peak_a* as 'Y'. Make sure that the little box above is checked.

Analysis Graphs

☒ Graph 1 **Graph 2** Graph 3

Scale Axis ☒ Overlay

	Min	Max	Scale
X	200.m	450.m	Auto Swp.
Y	4.00n	6.00n	Auto Ser.

☐ Share X-axis

Graph Entries


	1	2	3	4
X	☒ sqr(scan r	☐ I-mon	☐ I-mon	☐ I-mon
Y	☐ I-Peak_a	☐ I-mon	☐ I-mon	☐ I-mon

Graph Positions

Graph in Window 1 ☒ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐ ☐

Finally, you have to define in which Analysis window the *Graph 2* should be shown. On the bottom of the Analysis window, in the section *Graph Positions* make a check in the second box of the upper row. This means, that the *Graph 2* is shown in Analysis Window 1.


After another replay of the *Series* you should see the second graph which plots *I-peak* versus square root of the *Scan Rate* in the Analysis Window 1.



12.1.4 Customizing the CV Acquisition Sequence

Upon execution of a *Protocol* which contains the *E-Chem Method Cyclic Voltammetry* a transient *Pulse Generator Sequence* is created. This *Pulse Generator Sequence* will be overwritten by the next *E-Chem Method Cyclic Voltammetry*-containing *Protocol* and also not saved.

If you want to keep a specific configuration of the CV in your pgf pool, you should copy the *Pulse Generator Sequence* "CV" and rename it. The default name of the new sequence is CV-1. This sequence will be saved as usual.



You can now also customize your CV in a way that would not be possible using the parameter window of the *Echem Method Cyclic Voltammetry* in the Protocol Window (see page 158). The screenshot above shows a cyclic voltammogram where the upper potential is held for 10 s before starting the reverse scan. This was easily achieved by inserting a new Segment of the Class "Constant" into the existing segments. Unnumerous changes can be made to a common CV in this way.

12.2 Electrochemical Impedance Spectroscopy (EIS)

Electrochemical Impedance Spectroscopy is a characterization method where the impedance of a system depending on the frequency of an applied AC voltage is determined. Because the measured impedance is the sum of all components in the system, the data are analyzed using an equivalent circuit. Electrochemical Impedance Spectroscopy is commonly used to investigate electron transfer processes and diffusion of electrochemical reactions or the capacitance of electrode materials. Short-time behavior and fast processes correspond to high frequencies whereas long-time behavior and slow processes correspond to low frequencies. Interesting for electrochemists are commonly three processes which have different characteristic frequencies:

- double layer capacitance: 4.1 kHz (fast)
- faradaic reactions: 310 Hz (medium)
- diffusion of molecules: 0.4 Hz (slow)

Therefore a reasonable frequency range for studying these processes lies between some mHz and some kHz.

Before using the EIS Method you should make sure that the amplifier is calibrated for EIS. Please refer to section Calibrated EIS in the Amplifier menu 4.9 on page 36.

12.2.1 Parameters for EIS

The parameters for the measurement can be adjusted in the Protocol Editor Window by clicking on the *E-Chem Method - Impedance Spectroscopy* in the list of operations.

Impedance Spectroscopy ☐ Skip Delay 0.00 s

Start Frequency 10.0 kHz

End Frequency 1.00 mHz

Freq. per Decade Max. 5

Sine Amplitude 0.010 V

Min. Cycle Count 3

Cycles to Skip 2

☒ Auto Ranging ☒ Use Calib. Table

Total Length 29 -> 03:10:37

Start Frequency: Enter the first frequency of the impedance spectrum. It is advisable to go from high frequencies to low frequencies in order to be able to review the results quickly. The acquisition of data points for very low frequencies will take a long time.

End Frequency: Enter the last frequency of the impedance spectrum.

Freq. per Decade: You can set a resolution of the impedance spectrum by defining the number of frequencies which are measured per frequency decade. Note that a large number of frequencies increases the measurement time dramatically, especially when the spectrum spans into a frequency range below 1 Hz.

Sine Amplitude: Defines the voltage amplitude of the applied sine wave.

Min. Cycle Count: The number of sine cycles is adapted that a measurement time of 1 second is filled. If the measurement time per frequency is larger than 1 second, the minimum number of cycles are executed.

Cycles to Skip: Defines the number of sine cycles that are executed but not analyzed. Commonly, the first cycle should be skipped by writing a 1 into the box.

Auto Ranging: If checked the current *Gain* will be adapted that the signal is between the *Min* and *Max* range given by the *Auto Ranging* controls of the potentiostat.


Use Calib. Table: The EIS calibration table contains frequency dependent attenuation and phase shift data for the potentiostat. If checked, the recordings will be corrected with respect to the calibration data.

Total Length: Shows the estimated total time required for the spectrum recording.

The raw data recorded by the *Electrochemical Impedance Spectroscopy* method is stored in the data tree as *Group* with one *Series* per frequency. Each *Series* containing one *Sweep* with the current and voltage *Traces*, and a set of 7 *LockIn Traces*.

12.2.2 Acquiring an Electrochemical Impedance Spectrum

You start the electrochemical method by executing the protocol, which contains the method.




In the Oscilloscope Window the recording at each frequency is shown versus the time. The method creates a standard *Analysis Method* "EIS", which plots

- the Nyquist Plot of the impedance: the imaginary part of the impedance is plotted vs. its real part
- the Bode Plot of the impedance: the impedance is plotted vs. the frequency, both axis have a logarithmic scale
- the Bode Plot of the phase: the phase is plotted vs. the frequency, the frequency has a logarithmic scale


- the Nyquist Plot of the admittance: the imaginary part of the admittance is plotted vs. its real part
- the Bode Plot of the admittance: the impedance is plotted vs. the frequency, both axis have a logarithmic scale

The impedance format emphasizes the values at low frequency, which typically are of greatest importance for electrochemical systems that are influenced by mass transfer and reaction kinetics. Therefore, Nyquist and Bode plots working with impedance values are used.

The raw data recorded by the Impedance Spectroscopy method are stored in the data tree as *Series* with multiple *Sweeps* containing all the *Traces*.




The parameters of the method, which were used during the acquisition of the data, can be viewed in the **Acquisition Method** window. The window can be opened either via selecting **Show Protocol Method** in the **Replay** menu or via selecting **Protocol Methods** in the **Windows** menu.



12.2.3 Customizing the Analysis Method

If you want to make changes to the *Analysis Method*, replay the *Series* that contains the EIS data. Then open the *Analysis* window and select the *Analysis Method* "EIS". Press *Copy* to create another method (default name *EIS-1*), where you can make your changes. Remember: The transient *Analysis Method* "EIS" will be overwritten upon repeated execution of all *Protocols* using the *Cyclic Voltammetry Method* and not be saved.

For example, if you are studying faradaic reactions, only plots using the impedance data are of importance. The graphs for admittance plots can be deactivated.




12.3 Open Circuit Potential (OCP)

The Open Cell Potential (OCP) is measured between the working electrode and the reference electrode in order to get information about the electrochemical equilibrium. The OCP during corrosion measurements is known as corrosion potential. The condition of electrochemical equilibrium is current zero in respect to the sum of the cathodic and anodic currents. But this *current less* measurement of a potential is not possible in practice. The current which flows through the instrumentation during the potential measurement has to be kept as small as possible. This is realized by a very high impedance input of the reference electrode (100 G Ω) and the counter electrode is switched off during the OCP measurement. We measure the potential between two half-cells. The first half-cell is formed by the working electrode and the electrolyte, the second half-cell is formed by the reference electrode and the electrolyte.

12.3.1 OCP Parameters

The parameters for the measurement can be adjusted in the Protocol Editor Window by clicking on the *E-Chem Method - Open Circuit Potential* in the list of operations on the left.




Duration: Enter the duration for the OCP measurement.

Set OCP: If this option is checked, the measured OCP will be set as initial potential (*E-init/V-membr.*). The *Cell Mode* is not changed.

12.3.2 Acquiring the OCP


You start the electrochemical method by executing the protocol, which contains the method.



The Data tree for the OCP is very simply and only shows the recorded potential E-mon as the Trace.




The parameters of the method, which were used during the acquisition of the data, can be viewed in the Acquisition Method window. The window can be opened either via selecting **Show Protocol Method** in the **Replay** menu or via selecting **Protocol Methods** in the **Windows** menu.



12.4 Square Wave Voltammetry (SWV)

Square Wave Voltammetry belongs to the pulse voltammetric techniques. In this technique, a linear potential ramp is superimposed by a square wave. The advantage of the pulse voltammetric techniques is their ability to discriminate against charging currents (capacitive currents). Therefore, they exhibit a higher sensitivity for oxidation or reduction currents (faradaic currents) than conventional DC voltammetry. Square Wave Voltammetry yields peaks for faradaic currents instead of the sigmoidal waveform obtained with DC voltammetry which results in improved resolution for multiple analyte systems and more convenient quantification. Square wave voltammetry has received growing attention as a voltammetric technique for routine quantitative analyses. The acquisition of a square wave voltammetry is very fast using scan rates of 1 V/s and higher. A typical experiment can be performed in a matter of seconds.



12.4.1 SWV Parameters

The parameters for the measurement can be adjusted in the Protocol Editor Window by clicking on the *E-Chem Method - Square Wave Voltammetry* in the list of operations on the left.

Start Potential	-0.100 V	☐ Rel. to E-init
End Potential	0.100 V	☐ Keep at End
Amplitude	0.010 V	Scan Rate 1.00 V/s
Tau	10.0 ms	Frequency 100. Hz
		Pot. Increment 0.010 V
I sampled from	80 %	to 100 %
Keep Traces	☐ I-mon ☒ I-forward ☒ I-diff	
	☐ E-mon ☒ I-reverse ☒ E-mean	
Number of Cycles	1	☐ Alternate Scan Direction
1 Peak, Given Bounds	Cursor-1	-0.100 V 0.100 V
Total Time	200. ms	

Start Potential: Potential at which the Sweep of the SWV begins. At the start of the SWV the potential steps from the initial potential to the *Start Potential*.

Rel. to E-init: If checked the potentials will be applied relative to the E-Init.

End Potential: Potential at which the Sweep of the SWV ends.

Keep at End: If checked, the end potential will be set as *E-Init/V-membr.* after the Sweep has ended.

Amplitude: Voltage amplitude of a square half-wave.

Scan Rate: Overall Scan Rate of the potential increase.

Tau: Duration of a single square wave cycle.

Frequency: Frequency of the square wave, equals the inverse of *Tau*.

Pot. Increment: Potential increment from one square wave cycle to the next.

I sampled from: The current will be sampled at the end of each half cycle. The data in-between the cursor positions will be averaged. The result from the first half cycle is *I-forward* and from the second half-cycle *I-reverse*.


Keep Traces: The user can define which of the result *Traces* should be stored.

- I-mon: Raw data trace of the current recording.
- E-mon: Raw data trace of the voltage recording.
- I-forward: Result trace containing the analyzed current from the first half-cycle.
- I-reverse: Result trace containing the analyzed current from the second half-cycle.
- I-diff: Result trace containing *I-forward* - *I-reverse*.
- E-mean: Result trace containing the mean potentials each square wave cycle.

Number of Cycles: If the number of cycles is larger than one, multiple *Sweeps* are executed in *Gap Free Mode*.

Alternate Scan Direction: If checked the scan direction will be alternated when more than one *Sweep* is recorded. E.g. the second *Sweep* starts at the end of the first *Sweep* and scans back to the *Start Potential*.

Peaks: The standard analysis allows online peak analysis. The peak of *I-diff* and the corresponding potential will be analyzed.




- Compute No Peak: Peak analysis is turned off.
- 1 Peak, Auto Bounds: One peak will be analyzed. The algorithm will use the derivative of *I-diff* in order to find the main peak. A noise free recording is required for this *Analysis Method*. In order to reduce the noise you might apply a digital filter to the *I-diff Trace*. The digital filter is set in the *Oscilloscope* window.
- 1 Peak, Given Bounds: One peak will be analyzed. You can set the voltage range in which the peak should be searched.
- 2 Peaks, Given Bounds: Two peaks will be analyzed. You can set the voltage ranges in which the peaks should be searched.

2 Peaks, Given Bounds	Cursor-1	-0.100 V	0.100 V
	Cursor-2	0.099 V	0.100 V


Total Time: Displays the duration of the complete Square Wave Voltammogram.





12.4.2 Acquiring a Square Wave Voltammogram

You start the electrochemical method by executing the protocol, which contains the method.



The method creates a standard *Analysis Method* "SWV", which plots the square wave voltammogram in *Graph 1* in the form of I-diff vs. E-mean and analyzes the current peak and the corresponding peak potential. The results are plotted into the Notebook.

The raw data recorded by the *Square Wave Voltammetry* method are stored in the data tree as *Series* with multiple *Sweeps*. Each *Sweep* containing the recorded *Traces*.



The parameters of the method, which were used during the acquisition of the data, can be viewed in the Acquisition Method window. The window can be opened either via selecting **Show Protocol Method** in the **Replay** menu or via selecting **Protocol Methods** in the **Windows** menu.

12.5 Optical Spectroscopy

The Optical Spectroscopy method can be used to study the local light absorption properties of materials with high spatial resolution by scanning the sample in the light path. The illumination and detection area are in the range of a few micrometers. It can be employed to study photoelectrocatalysts and semiconductors, where also band gaps can be determined. The spectroscopy experiments are complimentary to and can be combined with other spatially-resolved experiments which can be performed with the EIProScan. The method allows the recording of absorbance spectra between 200 nm and 1200 nm using the designated spectrometer.

Please contact us to get further information about technical requirements to use this application.

The Protocol Editor Method *Optical Spectroscopy* dialog needs certain parameters for the measurement:

Integration Time: The integration time defines the time of collecting data at the spectrometer detector. Typical integration time are between 10 ms and 1 s depending on the light intensity. The intensity of the signal will increase with increasing integration time. For a good signal quality (signal-to-noise ratio) a value of 10k should be achieved. At 80k the limit of the detector is reached and the integration time needs to be reduced.

Averages: A signal which is measured at the certain integration time can be averaged. The number of averages is used to further improve the signal-to noise ratio. Increasing this value does not increase the signal intensity. Integration time and Averages are both used to achieve a good signal quality.

Light Sync Out: This box can be checked if the illumination source should be switched off between recording spectra.

Dark Ref. Spectrum: A dark reference (with illumination switched off) has to be recorded before the spectrum acquisition. In this step the dark reference was loaded into the buffer to correct subsequent spectra. The default is Buffer 1 and has to be selected to retrieve the right data.

Reference Spectrum: A reference spectrum at a blank spot has to be recorded before the spectrum acquisition. In this step the reference spectrum was loaded into the buffer to correct subsequent spectra. The default is Buffer 2 and has to be selected to retrieve the right data.


Threshold: A threshold can be defined to cut data with very small intensity. Due to the detector, the spectra are automatically recorded between 200 and 1200 nm. When the light source has a smaller spectral range, intensities



can be close to zero in the specific areas. This leads to not-meaningful calculation of absorbance. The threshold can be set to cut these areas from the spectrum.



13 Analysis Window



The Analysis allows you to immediately calculate and display data based on the acquired *Traces*, thus giving you a fast overview of your results.

POTMASTER will automatically plot the analysis to Analysis Windows 1 or 2 after or during execution of a *Series* (based on the settings made in the various controls inside this window).

Nomenclature: The Analysis is structured as follows. An Analysis file with the extension `*.onl` holds a set of *Analysis Methods*. Each method is an assembly of *Analysis Functions* and *Analysis Graphs*. Based on the type of incoming data, the user can specify a number of *Analysis Functions*. These functions produce analysis results which may be displayed in the Notebook and/or shown in an *Online Graph* inside either Analysis Window 1 or Analysis Window 2. Elements of a graph are *Graph Entries*, i.e. couples of analysis results to be used as X- and Y-references. Up to 4 *Graph Entries* can fit into one graph and multiple graphs can fit into one Analysis Window.

Thus, based on the hierarchy of the components involved in constructing an *Analysis Method*, *Analysis Functions* must first be defined because only then the respective analysis results can be placed as *Graph Entries* in graphs and windows.

13.1 Stimulus Control

Stimulus Control: There are four possible options for controlling the particular *Analysis Method* of choice:

Auto Stim Control: Pool

- Use Selected Method: Uses the active (selected) *Analysis Method*.
- Auto Stim Control: Pool: Uses the *Analysis Method* stored in and loaded from the Analysis file, e.g. 'DefAnal.onl'. With this kind of Analysis one defines one *Analysis Method* for all *Series* until the *Analysis Method* is changed.
- Auto Stim Control: Assigned: Uses the *Analysis Method* stored in and loaded from the data file bundle. With this kind of Analysis one creates a permanent link of an *Analysis Method* to a type of *Series*.
- No Analysis: No Analysis will be performed.

13.2 Analysis Methods



An *Analysis Method* is defined by all the information stored in the areas of *Analysis Graphs* and *Analysis Functions*. POTMASTER provides a pool of such *Analysis Methods* and such pools of methods are stored on disk as Analysis files (*.onl).

1...4: Scrolling bar of available *Analysis Methods* within this pool. The active (selected) *Analysis Method* is highlighted.

Load: Loads a pool of *Analysis Methods*.

Merge: Adds *Analysis Methods* from an Analysis file (*.onl) to the already loaded pool of methods.

Save: Saves the actual pool of *Analysis Methods*.


Analysis Method Name: Text field in which the name of the current *Analysis Methods* can be edited – "IV" in this example.

New: Creates a new *Analysis Method*.

Copy: Creates a copy of the active *Analysis Method* and inserts it at the end of the Analysis pool.

Move: Changes the positioning of a particular *Analysis Method* within the scroll bar.

Delete: Deletes the active *Analysis Method*.



If *Auto Stim Control: Assigned* is selected the three buttons *Load*, *Merge* and *Save* are replaced by:

Duplicate: Makes a copy of the currently selected *Analysis Method*.

Assign: Assigns the active *Analysis Method* to the selected *Series* in the data tree. The corresponding PGF template is automatically linked to this *Analysis Method*.

To Pool: Copies the active *Analysis Method* to the standard analysis pool (*Auto Stim Control: Pool*).

13.3 Analysis Functions

In this section of the *Analysis* dialog, *Analysis Functions* can be added to the *Analysis Method* and parameters of already existing *Analysis Functions* can be modified.

New *Analysis Functions* can be added by clicking on an empty button in the list at the top of the section. For changing an already existing *Analysis Function* just click on the *Change* button (see below).

List: Writes a list with the name of all *Analysis Functions* into the *Notebook* window.

 1...4  : Scrolling bar of an arbitrary number of available *Analysis Functions*. The buttons contain the *Name* of the defined function.

Function Type: Here the *Function Type* of the presently activated *Analysis Function* is displayed, e.g. "Amplitude".

Name: Here you can rename the selected *Analysis Function Type*.

Change: Click on this button to open the *Function Type* dialog for changing the *Analysis Function*.

New: Creates a new *Analysis Function*. You can also create a function by clicking on an empty *Function* button.

Delete: Deletes the selected *Analysis Function*.

Move: Changes the position of a particular *Analysis Function* within the scroll bar.

13.3.1 Analysis Function Settings

The *Analysis* provides four sections for customizing the selected *Analysis Function*. Here, we describe which settings are in general available. The availability depends of course on the selected *Analysis Function* and may vary from function to function.

Input

Input	
Use Given Value	0.00
Y-Trace	Trace 1
X-Trace	Compute Theoretical Amplitude

Value: Either use "Use Given Value" entered into the field beneath the function or "Get Value-1...16" which takes the value stored in *Value-1...16*.

Y-Trace: Select which *Trace* (1...16) or *Buffer* (1...4) shall be analyzed using the *Analysis Functions*. The *Trace* index usually corresponds with the channel number in the PGF sequence (if the *Trace* number is not set to another value in the build instruction of the *Trace* (see 10.10.2 on page 119)).



X-Trace: The voltage can either be calculated from a theoretical stimulus (*Compute theoretical Amplitude*), e.g. when there is no voltage *Trace* available, or it can be taken from a recorded *Trace* (*Amplitude from Trace 1...16, Amplitude from Buffer 1...4*).

Operation: Refer to the result of an already existing *Analysis Method*. In this example the *Math* operation 'a + b' refers to the *Analysis Method* index '3' and '4'.

Cursors

X-, Y-seg. Offset: The X- and Y-Segment offsets are set relative to the *Relevant X- and Y-Segments* as defined in the Pulse Generator window.


Cursor Bounds: Within the relevant segment, a left bound and a right bound (in %) determine the actual time period for analysis. Pressing the button lets pop-up the cursors in the Oscilloscope window. When moving the cursors in the Oscilloscope window, the cursors position is automatically transferred to the fields to the right of the button. The cursor limits can be outside the limits of the selected segment, i.e., they can be smaller than 0 % and greater than 100 %.

Cursor Type: Sets the definition of the cursor type. The following options are available:

- Cursors relative to Segment
- Cursors relative to Trace
- Cursors: Copy from Function-1...16: The active function gets the cursor bounds and cursor type from the selected function.

Adjust to Position of Function: Specify an existing *Analysis Function* to set the cursor bounds relatively to the found position of a peak or threshold. E.g. if one function analyzes the *Time to Maximum* and the second function analyzes the *Tau*, one can use *Adjust to Position of Function* in the second function to set the cursor settings in respect to the found *Time to Maximum* results. This means every time the *Time to Maximum* value changes the cursors settings (here: 1-15%) for the calculation of *Tau* will be adapted.





Note: Be aware that in this example the cursor range for the Tau function gets shifted to the right by Time to Maximum because the new 0% value is now the value of Time to Maximum.

Output

Output

☒ Notebook

Not Stored in Values

Result Unit

Notebook: If selected, the *Analysis Result* of this function will be written into the Notebook window.

Not Stored in Values / Store in Value-1...16: *Analysis Results* can be stored in one of the *Values-1...16*. They can then be retrieved and used as a *Constant* math function. In addition, they can be used in the Protocol Editor (see *Set Value*, 11.4.5.3 on page 151). The *Values-1...16* are used to exchange information between different *Analysis Methods* and between the Analysis and the Protocol Editor.

Result Unit: If e.g. a *Math* operation was used one can define via this entry box the unit of the output value.

Operation Options

Operation Options

☐ Fit

Baseline: Zero

Fit: If selected, the peak value is determined by a polynomial fit around the first numeric estimate of the peak. With this function, the influence of noise can be reduced in the determination of *Peak Amplitudes*.






The fit function must not be used when non uniformly continuous changes are expected around the peak value. E.g. if you want to analyze the peak value of a saw-tooth signal, the Fit function easily lead to erroneous results, because of the quick jump from the peak level to base line at the end of the saw-tooth.

Baseline: Define the *Baseline* for the selected *Analysis Function*. There are three options available:

- **Baseline: Zero:** Calculates the integral between the two intersection points of the cursors and the *Trace*, starting at the zero line.
- **Baseline: Cursor Intersections:** Calculates the integral, starting at a virtual baseline between the two intersections points of the cursors and the *Trace*.
- **Auto Bounds:** Makes a derivation of the *Trace* and sets the cursors to the minimum and maximum value of the derivation. This option can be helpful to find the proper peak if some other global maximum (or minimum) exists.




Crossings: Set the direction of the *Crossing*. There are three options available:

- **Positive and Negative Crossing:** Every threshold crossing is detected.
- **Positive Crossing:** A crossing is only detected if the signal runs through the threshold from negative to positive values.
- **Negative Crossing:** A crossing is only detected if the signal runs through the threshold from positive to negative values.


Average Points: Enter the number of points to be used for a running average to apply before extracting the voltage. The function applies to the source *Trace* as well as to the result *Trace*.

Ramps only: The *Analysis Function* does only work on *Ramp* segments if the checkbox is selected.



Equation: Enter an equation into the text field or modify an already existing one by clicking on the *Edit* button.

Histogram Bins: Enter the number of bins for the histogram.



13.3.2 Analysis Function Types

When selecting or changing an *Analysis Function* the *Function Type* dialog pops up. Here, you can select an *Analysis Function*. The functions are organized in several groups.

The 'Function Type' dialog box displays a grid of analysis function categories and their respective options. The categories are: Timing, Measurements, LockIn, Math, Trace, Power Spectra, Stim. Properties, Trace Param., Sweep Param., Histogram, and EIProScan. Each category has a list of functions with radio buttons for selection. The 'Timing' category includes Sweep Count, Analysis Index, Time, Timer Time, Series Time, and Real Time. The 'Measurements' category includes Extremum, Maximum, Minimum, Extr. Amplitude, Min. Amplitude, Max. Amplitude, Time to Extremum, Time to Maximum, Time to Minimum, Time to Threshold, Threshold Ampl., Thres. Crossings, Mean, Variance, Integral, Anodic Q, Cathodic Q, Zero Crossing, Slope, Intercept, Tau, and Y(x): y at pos = x. The 'LockIn' category includes LockIn_Phase, LockIn_Freq, LockIn_Ampl, Trace Param., OCP, Electrode Area, Solution, Solution Value, Zero Offset, Sweep Param., Temperature, Digital-In, and User_1 through User_10. The 'Math' category includes Equation, Constant, a + b, a - b, a * b, a / b, a in b, abs, log, sqrt, arctan, 1/a, 1/log, 1/sqrt, and 1/arctan. The 'Trace' category includes Trace, Equation, Q = Integral, 1 / (trace), 1 / (Q), ln (trace), ln (Q), log (trace), log (Q), dt = Differential, Trace x-axis (time), and Stimulus. The 'Power Spectra' category includes log(Frequency), Density, Histogram, Histogram Ampl, Histogram Bins, EIProScan, X-position, Y-position, Z-position, Matrix-X, Matrix-Y, 3D_X, and 3D_Y. The 'Stim. Properties' category includes Amplitude, Duration, Rel. Seg. Time, Abs. Seg. Time, and Scan Rate. The 'Trace Param.' category includes OCP, Electrode Area, Solution, Solution Value, Zero Offset, Sweep Param., Temperature, Digital-In, and User_1 through User_10. The 'Sweep Param.' category includes Temperature, Digital-In, and User_1 through User_10. The 'Histogram' category includes Histogram, Histogram Ampl, Histogram Bins, EIProScan, X-position, Y-position, Z-position, Matrix-X, Matrix-Y, 3D_X, and 3D_Y. The 'EIProScan' category includes X-position, Y-position, Z-position, Matrix-X, Matrix-Y, 3D_X, and 3D_Y. The dialog box has 'Cancel' and 'Done' buttons at the bottom right.

13.3.2.1 Timing

Sweep Count: Index of the *Sweep* within a *Series*.

Analysis Index: Index of the *Analysis* during the experiment. The *Analysis Index* is initialized with a *Wipe* command. The maximum number of *Analysis Sweeps* that can be handled, can be set in the *Configuration* window (see chapter 5.6.2 on page 54).

Time: Time starting with the relative "first sweep" analyzed (or created) since the last *Wipe Analysis* command.

Timer Time: Time of the *Timer* in the *Oscilloscope* at the start of *Sweep* acquisition.

Series Time: Time in respect to the *Sweep* acquisition of the corresponding *Series*.

Real Time: Time elapsed since midnight at the start of *Sweep* acquisition.

13.3.2.2 Stim. Properties

These are functions that are usually used as an X-reference in a graph. They are based on the *Relevant X-segment*, i.e. they are also sensitive to the X-segment offset.

Amplitude: Amplitude of the *Relevant X-segment* (usually "Voltage").

Duration: Duration of the *Relevant X-segment*. This may be useful if the length of the *Relevant X-segment* is changing during sweep acquisition.

Rel. Seg. Time: The *Relative Segment Time* function returns the start time of the *Relevant X-segment* with respect to the first stored point of the stimulus. The duration of non-stored segments is not taken into account.

Abs. Seg. Time: The *Absolute Segment Time* function returns the start time of the *Relevant X-segment* with respect to the first point of the stimulus. In opposite to *Relative Segment Time* the duration of non-stored segments is taken into account.

Scan Rate: Scan rate of the *Relevant X-segment*. Can be used e.g. to analyze the slope of a ramp.

Note: Do not mix up the Scan Rate with the Slope analysis function! While the Slope is calculated from the recorded Trace, the Scan Rate is taken from the stimulus.

Stim. Frequency: Returns the set *Stimulation Frequency* of a corresponding *Sine Wave* or *Square Wave* segment.

13.3.2.3 Measurements

These are functions that are usually used as an Y-reference in a graph. They are based on the relevant Y-segment, i.e. they are also sensitive to the Y-segment offset.

Extremum: Extreme value of data in the cursor region; either a maximum or a minimum.

Maximum: Maximum value of data in the cursor region.

Minimum: Minimum value of data in the cursor region.

Extr. Amplitude: Calculates the voltage applied at the position of the peak current (either minimum or maximum).

Min. Amplitude: Calculates the voltage applied at the position of the minimum peak current.

Max. Amplitude: Calculates the voltage applied at the position of the maximum peak current.

Time to Extremum: Time from the beginning of the *Trace* or the relevant segment to the extreme value in the cursor range.

Time to Maximum: Time from the beginning of the *Trace* or the relevant segment to the maximum value in the cursor range.

Time to Minimum: Time from the beginning of the *Trace* or the relevant segment to the minimum value in the cursor range.

Time to Threshold: Time from the beginning of the *Trace* or the relevant segment to the specified threshold value in the cursor range.

Threshold Ampl.: This function returns the corresponding X-axis value (often called amplitude) of the threshold crossing point. Use this function e.g. if you want to know the voltage value for a specific (recorded) current value.

Thres. Crossings: Counts the number of threshold crossings between the two cursors.

Mean: Mean value of the data in the cursor region.

Variance: Variance value of data in the cursor region (square of standard deviation).

Integral: Integral value of data in the cursor region.

Anodic Q: Integral of positive current in the cursor range.

Cathodic Q: Integral of negative current in the cursor range.

Zero Crossing: This function searches for the zero crossing of the target *Trace* and computes the corresponding value of a second *Trace* (i.e. voltage).

Slope: Calculates the slope (b) in the defined region, using a linear regression ($y = a + bx$).

Intercept: Calculates the intercept (a) in the defined region, using a linear regression ($y = a + bx$).

tau: Calculates the time constant (τ) from an exponential curve. Computes τ in two steps:

1. semi-logarithmic regression of x vs $\ln(y)$



2. $\text{abs}(1/\text{slope})$

Y(x) : y at pos = x: Calculates the Y-value at a defined X-position (e.g. time, frequency...). The X-position value can be entered into the field beneath the function.

13.3.2.4 Lock-In

The following parameters can be retrieved from the `LockIn` Extension. Please note that the *LockIn* returns a mean value of all data of the *Trace*. In case a mean value in the cursor range is desired, then please acquire/store the parameter as individual *Trace* and use the *Mean* function of the *Measurements*.

LockIn_Phase: Returns the phase setting of the software *LockIn*.

LockIn_Freq: Returns the frequency of the sinusoidal wave used for *LockIn* measurements.

LockIn_Ampl: Returns the amplitude of the sinusoidal wave used for *LockIn* measurements.

13.3.2.5 Trace Parameters

The following parameters can be retrieved for analysis from the parameter set that is stored with the data (see *Parameter* window, 14 on page 197).

Trace Count: Gives back the index number of the *Trace* which is set in *Trace #*. If no *Trace* with the index number exists a "NAN" result will be reported.

OCP: The value of the Open Cell Potential.

Electrode Area: The value of the Electrode Area.

Solution: Index number of the solution.

Solution Value: Relevant concentration value of the solution.

Zero Offset: Returns the calculated *Zero Offset* value.

13.3.2.6 Sweep Parameters

The following parameters can be retrieved for analysis from the parameter set that is stored with the data (see *Parameter* window, 14 on page 197).

Temperature: Temperature (from a recording device).


Digital-In: State of the Digital-In port of the PG 300/600 potentiostat.

User_1 - User_10: Parameter as set by the user in the *I/O Control* section of the *Configuration* window, 5 on page 41.

13.3.2.7 Math

Math supplies various calculations on one or two function results. By combining several *Math* functions, complex expressions can be generated.

The following control shows: "Result 1" "- Operand" "Result 2".



Equation: An equation can be entered to compute a value. For details on the equation syntax please refer to the section *Equation Syntax*, 21.1 on page 232.



Constant: Sets a constant value if *Use Constant Value* is selected. The constant value may be determined either by entering a value in the input field beneath the function name or by using a value (see also function $Y(x)$: y at $pos = x$ above.)

a + b: Calculates the sum of two Analysis results.

Note: Depending on the Math function one or two results of the Analysis are available for the calculation.

a - b: Calculates the difference of two Analysis results.

a * b: Calculates the product of two Analysis results.

a / b: Calculates the quotient of two Analysis results.

a in b: Calculates the amount of one Analysis result that is present within another Analysis result.

abs: Calculates the absolute value of an Analysis result.

log: Calculates the logarithmic value of an Analysis result.

sqrt: Calculates the square root of an Analysis result.

arctan: Calculates the arc tangent of an Analysis result.

1 / a: Calculates the reciprocal value of an Analysis result.

1 / log: Calculates the reciprocal logarithmic value of an Analysis result.

1 / sqrt: Calculates the reciprocal square root of an Analysis result.

1 / arctan: Calculates the reciprocal arc tangent of an Analysis result.

13.3.2.8 Trace

This analysis is applied to a selected *Trace*. The result is again a *Trace*.

Trace: Sets a *Trace* as an Analysis result.

Equation: An equation can be entered to compute a *Trace*. For details on the equation syntax please refer to the section *Equation Syntax*, 21.1 on page 232.

Q=integral: Calculates the integral over time of the *Trace*.

1 / (trace): Calculates the reciprocal of a *Trace*.

1 / (Q): Calculates the reciprocal of the integrated *Trace*.

ln (trace): Calculates the natural logarithm of the *Trace*.

ln (Q): Calculates the natural logarithm of the integrated *Trace*.

log (trace): Calculates the logarithm of the *Trace*.

log (Q): Calculates the logarithm of the integrated *Trace*.

dt = Differential: Calculates the differential over time of the recorded *Trace*.

Trace x-axis (time): Returns the *Trace* time (first data point of *Trace* has time = 0). You may use this as X-reference when plotting a *Trace* versus time in the Online Window.

Stimulus: The stimulus that is linked to the *Trace* is returned.



13.3.2.9 Power Spectra

For calculation of the power spectrum. Useful for measurement of the noise performance. Please note, that at least 1024 data points are required for compiling such a power spectrum.

log(Frequency): Returns the $\log(\text{Frequency})$ for a *Power Spectrum*. It should be used as X-axis in a *Power Spectrum* display. E.g. a value of '2' means 100 Hz, '3' means 1000 Hz.

Density: Returns the $\log(\text{Distribution})$ of a *Power Spectrum*. The distribution is also called spectral density. It should be used as Y-axis in a *Power Spectrum* display. The displayed values are the linear exponents, e.g. '-25' means $1 * 10^{-25}$. The unit is in A^2/Hz .

13.3.2.10 Histogram

The histogram functions can be used e.g. for a fast online detection of different amplitude levels.

Histogram Ampl: Returns the *Histogram Amplitude* distribution. You can select the *Trace* from which the amplitude should be calculated.

Histogram Bins: Defines the number of *Histogram Bins* for the analysis of *Histogram Amp*. The number of bins is entered manually in the *Function Type* window:

The width of a bin is computed as:

$$\text{bin width} = \text{Y-range} / \text{number of bins}$$

Note: A fixed Y-range is used to analyze all Traces of one Series. In Fixed mode it is the Y-range of the first analyzed Trace, in Auto mode it is the largest Y-range of all analyzed Traces of the parent Series.

13.3.2.11 ElProScan

The functions listed in this section are only available when the ElProScan Extension is activated in the Configuration window (Hardware pane).

X-position: Returns the X-axis position of the step motor in respect to the defined position of origin.

Y-position: Returns the Y-axis position of the step motor in respect to the defined position of origin.

Z-position: Returns the Z-axis position of the step motor in respect to the defined position of origin.

Matrix-X: Returns an index number of the X-position of the selected *Series*.


Matrix-Y: Returns an index number of the Y-position of the selected *Series*.

3D_X: Enter the index number of the *Analysis Function* to get the 3D X-position.

3D_Y: Enter the index number of the *Analysis Function* to get the 3D Y-position.



13.4 Analysis Graphs



Here, the properties of the graph display for the Analysis results are defined.

- **Graph 1**, **Graph 2**, **Graph 3** - : Scrolling bar of *Analysis Graphs* (currently to a maximum of 12 graphs). All information shown in the lower part of this panel is only valid for the highlighted graph.

Graph 1 (Graph Display Mode): A graph can either be turned *On* or *Off*.

- On: Displayed in Analysis Window 1 or Analysis Window 2.
- Off: Not displayed.

Save to Tree: Saves the active *Graph Entries* of the selected *Analysis Graphs* to the end of the data tree as a new *Series*.

Copy: Allows to automatically copy the settings (functions, scaling etc.) of a graph into the selected graph, rather than having to enter each entry manually. Enter the index number of the graph from which you want to copy the settings.

Print: Prints the Analysis results as displayed.

Redraw: Redraws the Analysis results displayed in the Analysis Windows, e.g. after changing the axis scaling.

13.4.1 Scale Axis

Overlay: The default setting in the Analysis is to wipe the data before a new Series is started. If you mark the *Overlay* option, the graph plot in the On-line window will not be cleared when performing a new analysis (new Series acquisition).

Scale Axis		☐ Overlay	☒ No Wrap
	Min	Max	Scale
X	-100.m	100.m	Auto Ser.
Y	-200.p	2.00n	Auto Ser.
Modify Axis			☐ Share X-axis

Note: This setting is overruled by the setting Wipe Display at Start and Wipe at Start (see 11.4.2.3 on page 142 and 5.6.2 on page 54).

Wrap: Selecting *Wrap* leads to a wrapping of the data at the end of the display, e.g. the display will start again at the left edge.

- No Wrap: No adaption of the X-axis scaling when the amount of data points exceeds the X-axis scaling.
- Wrap + wipe: Wiping of the online graph after the wrapping around. This behavior is typically used during Chart Recording (see PATCHMASTER Tutorial).
- Wrap, no wipe: No wiping of the online graph after wrapping around. Old data points are overlayed by newer data points when the Series exceeds the length of the graphical display.

Note: The Wrap option is limited to Fixed Scaling of the X-axis and Y-axis and the X-axis mode has to be linear.

If both *Overlay* and *Wrap* are selected, the data will wrap around at the end of the graph without clearing the graph.

More complex and custom display behavior can be controlled from a protocol, e.g. wiping individual graphs only, by using the Analysis protocol event (see 11.4.4.1 on page 145).

In case *Trace* data are displayed in the graphs, the options *Overlay* and *OvrlSwp* are available.

Scale Axis		☒ Overlay	☒ OvrlSwp
	Min	Max	Scale
X	-3.40m	120.	Auto Swp.
Y	-877.n	40.5μ	Auto Swp.
Modify Axis			☐ Share X-axis

- No selection: If neither *Overlay* nor *OvrlSwp* is selected the online graph is wiped after every Sweep acquisition.
- Overlay: General overlay, see also description above. Overlays Sweeps of different Series. Compare to *Overl. Ser* in the Oscilloscope window.
- OvrlSwp: *Overlay Sweep* function overlays all Sweeps within a Series. Compare to *Overl. Swp* in the Oscilloscope window.

Note: The autoscale functions Auto Ser. and Auto Swp. are incompatible with the OvrlSwp function. Only the last Sweep will be drawn after re-scaling the graph.

X, Y: Here you can enter the analysis display settings.

- Min: Minimum of axis range.
- Max: Maximum of axis range.
- Scale: There are several possible scaling options available once you click on the *Scale* field.





- Fixed: Analysis results can be shown in the graph immediately after acquisition or replay of a Sweep, if the scaling is known from the beginning. Therefore, a *Fixed Scaling* option is provided and the X-, Y-scaling of the analysis graph can be specified.
- Auto after a Series: After all Sweeps of a Series have been acquired or replayed, the maximal and minimal values of the currently selected abscissa and ordinate values are determined and used to scale the graph.
- Auto after each Sweep: After every Sweep, the maximal and minimal values of the currently selected abscissa and ordinate values are determined and used to scale the graph.
- Fixed with last Min/Max: Uses the maximum/minimum from the last graph drawing to copy into the Min/Max fields. They can now be used as scaling factors when the mode is switched to Fixed.

Modify Axis: Clicking on this button opens the Scale Properties window.

- Unit: The units of the first graph entry are shown on the graph display.
- Zero-line: The zero-line is depicted in the online graph.
- Position: Y(X)-min, Y(X)-zero, Y(X)-max). Determines the position of the axis either at the minimum, at zero or at maximum of the other axis, respectively.
- Mode: Defines how the respective axis is drawn: *linear*, *log*, *inverse*, *sqrt*, *square*.
- Tics: Sets the number of tics plotted on the axis. A value of zero suppresses the display of axis tics.
- Grid, Factor: A grid will be displayed in the given Analysis Graph. Enter a Factor for the grid density. If Factor is set to "1" then one line per Tic is drawn. A factor <1 increases the grid density.

The Scale Properties dialog box has two columns for X-axis and Y-axis settings. The X-axis settings are: Unit (checked), Zero-line (checked), Position (Y-min), Mode (linear), Tics (7), Grid, Factor (checked, 1.00), Include Zero (unchecked), Nice Values (checked), Centered (unchecked), Labels (checked), Header (IV-Analysis), Normalize (unchecked), and Sort (unchecked). The Y-axis settings are: Unit (checked), Zero-line (checked), Position (X-min), Mode (linear), Tics (5), Grid, Factor (checked, 1.00), Include Zero (unchecked), Nice Values (unchecked), Centered (checked), Labels (checked), Header (IV-Analysis), Normalize (unchecked), and Sort (unchecked). There are Cancel and Done buttons at the bottom.



- Include Zero: The Y- and X-axis always contains "0" at their origin independent of wrapping events.
- Nice Values: If selected, the axis limits will be rounded to the next possible "nice" value (e.g. 0, 1, 2, 5...).

- **Centered:** If selected, the first data point will be set in the center of the graph. This allows the monitoring of ongoing experiments. The displayed scaling defined by the values entered in the *Fixed Scaling* is not considered as an absolute value range but defines a relative one.

Note: This option is only available if Fixed Scaling is on and if Nice Values is off.

- **Labels:** If selected, the name of the *Analysis Function* is used as label for X- and/or Y-axis
- **Header:** Text for the graph header can be entered.
- **Normalize:** Select the type of normalization of the data:
 - None
 - $y' = y / \max$
 - $y' = (y - \min) / (\max - \min)$
- **Sort:** If selected, the data will be sort the data according to ascending X-values.

Note: The Sort flag should be set for summary data and non-ordered, non-continuous Traces before performing the Save to Tree operation.


Share X-axis: If *Share X-axis* is checked, then the same X-axis parameters (e.g. "Min", "Max", "Scale") are used for all graphs



13.4.2 Graph Entries


1...4: Activate up to four graph entries from this list by checking the box.

Axis: Now choose from the list of available analysis result variables for the X- and the Y-axis. An example could look like that shown on the right in the diagram above.




The following variables are shown in the figure beneath:

- Max_1: Maximum value between two cursor of *Trace 1*.
- Timer: Time of the *Timer*.
- C-slow_1: C-slow value of *Trace 1*.
- Int_1: Integral value between two cursors of *Trace 1*.
- 5...16: Empty item.
- Select by index...: Select the variables by their index number. This is necessary when more than 16 analysis results are defined.



Clicking on  opens the Marker Properties window:



Type: Choose between the available symbols (point, plus, star, diamond, cross, square).

Size: Size of the symbols.

Connect: If selected, the symbols will be connected with a line.

Set Color: Choose a color for the symbols and the lines.

Same Color as Trace: Use the *Trace* color for the symbols and the lines. The *Set Color* setting is ignored.



13.5 Graph Positions

Graph Positions		1	2	3	4	5	6	7	8	9	0	1	2
Graphs in Window 1:		☒	☒	☐	☐	☐	☐	☐	☐	☐	☐	☐	☐
Graphs in Window 2:		☐	☐	☒	☒	☐	☐	☐	☐	☐	☐	☐	☐


Graph Positions: In this section of the Analysis dialog, one can specify if and to what Analysis window (1 or 2) a graph is plotted. The currently selected graph (in the *Graph Entries* section) is indicated by red color.

Note: If a graph is not checked in the Graph Positions section, then it is not displayed, even if it is turned on in the Graph Entries section.





14 Parameters Window



Parameters are additional values describing the experimental and measurement conditions. The *Parameters* are required to interpret the data correctly. In POTMASTER we distinguish different types of *Parameters* depending to which level of the data structure they relate.


Note: If possible POTMASTER will use internal values to set the Parameters. E.g. the Amplifier Mode will be automatically read from the amplifier and stored with the Parameters. For some Parameters it might be necessary to assign a source (e.g. AD channel or manual input) in the I/O Control section of the Configuration window. For other amplifiers than EPC 10 or EPC 9 most of the external Parameters have to be entered manually via the I/O Control window.

In the *Parameter* window either the *Parameters* that are saved with already acquired data, or the *Parameters* of the last acquisition (stored or not stored) are displayed.



14.1 Parameter Tabs

14.1.1 Root Items




Level 0 of the data tree. The *Parameters* of the *Root* level are:

- Date: Date at the time of acquisition.
- Entries: Number of *Groups* in the file.
- Program Version: Version number and date of POTMASTER.
- Text: A descriptive text can be entered and stored if the data file is opened in the *Open Modify...* mode.



14.1.2 Group Items




Level 1 of the data tree. The *Parameters* of the *Group* level are:

- Header: Index number of the *Group* (e.g. GROUP_1).
- Label: Text identifier; default label is the experiment number.
- Count: Number of *Group* entries.
- Entries: Number of *Series* in the *Group*.
- Exp. Number: Number of the experiment. The *Exp. Number* can be incremented by creating a new *Experiment*. An *Experiment* might contain a collection of different *Groups* with the same *Exp. Number*.
- Text: A descriptive text can be entered and stored if the data file is opened in the *Open Modify...* mode.
- MatrixWidth: Width of the acquired image in the *Matrix Scan* mode (only for data acquired with the ELPROSCAN).
- MatrixHeight: Height of the acquired image in the *Matrix Scan* mode (only for data acquired with the ELPROSCAN).



14.1.3 Series Items




Level 2 of the data tree. The *Parameters* of the *Series* level are:

- Header: Index number of the *Group* and the *Series* (e.g. SERIES_1_4).
- Label: Text identifier; default label is the name of the *Series*.
- Count: Index number of the *Series*.
- Entries: Number of *Series* in the *Group*.
- Number of Sweeps: Original number of *Sweeps* in the *Series*. The number of *Entries* gets lower after deleting individual *Sweeps*.
- User Name: Name of the user which was entered in the Configuration (5.3.5 on page 45).
- Date: Date at the time of acquisition.
- Time: Time of the acquisition.
- Timer: Timer time of the acquisition.
- User Param 1/2: User-defined external values (see Configuration, 5.7.2 on page 60).
- Aux3...6: Auxiliary *Parameter* fields only used when certain extension are switched on.
- Comment: Text identifier (see Control window, 8.1 on page 92).
- Method: Name of the used method (see E-Chem Methods, 12 on page 157)



14.1.4 Sweep Items



Level 3 of the data tree. The *Parameters* of the *Sweep* level are:

- Header: Index number of the *Group*, the *Series* and the *Sweep* (e.g. SWEEP_1_4_1).
- Label: Text identifier; label of the *Sweep* (not the index number).
- Count: Index number of the *Sweep*.
- Entries: Number of *Traces* in the *Sweep*.
- Stim. Count: Total number of *Series* in the data tree.
- Time: Time of the acquisition.
- Timer: Timer time of the acquisition.
- Marker-1...4: Position of the *Sweep Marker* (see *Markers*, 20 on page 227).
- Temperature: Temperature value from an external device (see *Configuration*, 5.7.2 on page 60).
- User 1/2: User-defined external values (see *Configuration*, 5.7.2 on page 60).
- Digital In: Status of the input bits of the digital channels (see *I/O Control*, 16 on page 213).
- Digital Out: Status of the output bits of the digital channels (see *I/O Control*, 16 on page 213).



14.1.5 Trace Items

Marked Items	Root Items	Group Items	Series Items	Sweep Items	Trace Items	Amplifier Items	EIProScan Items
☒ TRACE	☐ Titles	☐ LF	☐ Flag Group 1	☐ Flag Group 2	☐ Flag Group 3	☐ Flag Group 4	
☒ Header	TRACE 1 1 3 1	☐	Copy Flags from	Gr.1 Gr.2 Gr.3 Gr.4	Info to Notebook	Gr.1	
☒ Label	I-mon	☐	Default Flags	Target to Notebook	Target to File		
☐ Count	1	☐	Clear all Flags	Marked to Notebook	Marked to File		
☐ Data Points	641	☐					
☐ File Offset	5384	☐					
☐ Average Count	1	☐	☐ Y-range	1.024 A	☐	☐ X-position	0.000 m
☐ Amplifier Mode	Potentiostatic	☐	☐ Y-min	-20.00mA	☐	☐ Y-position	0.000 m
☐ Amplifier	Disk	☐	☐ Y-max	60.00mA	☐	☐ Z-position	0.000 m
☒ Data Kind	Adc6-Imon	☐	☐ Zero Offset	0.000 A	☐		
☐ Data Format	2-byte integer	☐	☐ Bandwidth	300.0 kHz	☐		
☐ Data Factor	31.25μ	☐	☐ DA Channel	1	☐		
☐ Time Offset	0.000 s	☐	☐ AD Channel	1	☐		
☐ Y-unit	A	☐	☐ Holding	0.000 V	☐		
☐ X-unit	s	☐	☐ OCP	0.000 V	☐		
☐ X-interval	5.000μs	☐	☐ Electrode Area	1.000 cm2	☐		
☐ X-start	0.000 s	☐	☐ Solution		☐		
☐ Y-Offset	0.000 A	☐	☐ Solution Value	0.000 M	☐		

Level 4 of the data tree. The *Parameters* of the *Trace* level are:


- Header: Index number of the *Group*, the *Series*, the *Sweep* and the *Trace* (e.g. TRACE_1_1_1_2).
- Label: Text identifier; label of the *Trace*.
- Count: Index number of the *Trace*.
- Data Points: Number of stored data points.
- File Offset: Offset of the data in the according memory block (minimum number is 256 which is the size of the header).
- Average Count: Number of acquired *Traces* for the calculation of the stored average *Trace*.
- Amplifier Mode: Indicates the operation mode of the device, e.g. "Potentiostatic" or "Galvanostatic" mode.
- Amplifier: Indicates the used working electrode, e.g. "Disk" or "Ring".
- Data Kind: AD channel index number and kind of acquired data (e.g. Adc6-Imon).
- Data Format: Format of the data (either 2-byte integer, 4-byte integer, 4-byte real or 8-byte real).
- Data Factor: Scaling factor for the conversion of interface values into e.g. current or voltage values.
- Time Offset: Time delay of DA or AD channels (see *Delay*, 10.13 on page 126).
- Y-unit: Unit of the Y-axis.
- X-unit: Unit of the X-axis.
- X-interval: *Sample Interval* (see *Timing*, 10.6 on page 102).
- X-start: Indicates any delay of the X-axis, mostly a time delay when using a *Start-Segment* (see *Timing*, 10.6 on page 102). But it may also indicate other delays when X-axis is not time but e.g. frequency.
- Y-Offset: Y-axis offset in the Oscilloscope.
- Y-range: Maximal range of the Y-axis (one direction only). This depends on the *Gain* settings.
- Y-min: Minimum Y-value of the *Trace* (e.g. current or voltage).

- Y-max: Maximum Y-value of the *Trace* (e.g. current or voltage).
- Zero Offset: Calculated *Zero Offset* (see AD input channel settings, 10.10.2 on page 120).
- Bandwidth: Filter frequency of the acquired *Trace*.
- DA Channel: DA channel number (see Channel Settings for DA Output and AD Input, 10.10 on page 114).
- AD Channel: AD channel number (see Channel Settings for DA Output and AD Input, 10.10 on page 114).
- Holding: Holding potential/current value.
- OCP: Value of the Open Cell Potential.
- Electrode Area: Size of the *Electrode Area*.
- Solution: Name of the used solution.
- Solution Value: Value entered in the selected solution
- X-position: X-position value in relation to the origin.
- Y-position: Y-position value in relation to the origin.
- Z-position: Z-position value in relation to the origin.



14.1.6 Amplifier Items

Potentiostat/Galvanostat state record. Describes the potentiostat/galvanostat settings at the beginning of the *Series*. Available for fully software controllable amplifiers like PG 310 or PG 340.




The parameters of the *Amplifier* level are:

- **List with Traces:** If enabled all flagged *Parameters* will be exported either to the Notebook via "Target to Notebook", "Info to Notebook" or to a file via "Target to File" (see *Flagging*, 14.2 on page 208).
- **Amplifier:** Type of potentiostat/galvanostat used for data acquisition (e.g. "PG310_USB").
- **Serial Number:** Serial number of the device and revision state.
- **Calibration:** Indicates if the calibration file were loaded.
- **Recording Mode:** Indicates the Cell Connection, e.g. "Cell", "Standby" or "OCP" (see *Cell Connection*, 7.1.1 on page 72).
- **Potentiostat Mode:** Indicates the operation mode of the device, e.g. "Potentiostatic" or "Galvanostatic mode" (see *Amplifier Mode*, 7.1.2 on page 73).
- **E-initial:** *E-Initial* value of the cell potential.
- **Gain:** *Current Range* setting.
- **External Stimulus:** Scaling factor when the *External Stimulation Input* was used (see *External Input*, 7.1.4 on page 76).
- **Electrode Mode:** Indicates if the *3 or 4 Electrode Mode* was used (see *Electrode Mode*, 7.1.4 on page 76).
- **Notch Filters:** Indicates if the notch filter was used (see *Notch Filter*, 7.1.2 on page 73).
- **Voltage Bandwidth:** *Voltage Filter* setting (see *Voltage Filter*, 7.1.2 on page 74).
- **Current Bandwidth:** *Current Filter* setting (see *Current Filter*, 7.1.2 on page 74).
- **Control Bandwidth:** *Control Amplifier Bandwidth* setting (see *Control Amplifier Bandwidth*, 7.1.2 on page 74).
- **Current Filter:** Type of used *Current Filter*, e.g. "Bessel" (see *Current Filter*, 7.1.2 on page 74).
- **Stimulus Filter:** *Stimulus Filter* setting (see *Stimulus Filter*, 7.1.2 on page 74).
- **E-inital Range:** *E-initial range* setting (see *E-initial range*, 7.1.4 on page 77).
- **IR-Compensation:** *IR-Compensation* value (see *IR-Comp.*, 7.1.4 on page 75).

- Electrode Area: *Electrode Area* value (see *Electrode Area*, 7.2.6 on page 84).
- Ref. Electrode: Type of used reference electrode, e.g. "SCM" or "Calomel" (see *Zero Potential (E-zero)* window, 17 on page 217).
- E-zero: Value of *Manual Input* for *E-zero* (see *Zero Potential (E-zero)* window, 17 on page 217).
- OCP Potential: Value of the OCP (see *Cell Connection*, 7.1.1 on page 72).
- Anodic Charge: Value of the *anodic charge* (see *Potential/Current settings*, 7.2 on page 79).
- Cathodic Charge: Value of the *cathodic charge* (see *Potential/Current settings*, 7.2 on page 79).
- LockIn Phase: Phase of the LockIn (see *Software LockIn Extension*, 24 on page 249).
- LockIn Attenuation: Attenuation value of the LockIn (see *Software LockIn Extension*, 24 on page 249).
- LockIn Mode: "Sine+DC", "PiecewiseLinear" or "On Cell" (see *Software LockIn Extension*, 24 on page 249).



14.1.7 ElProScan Items



The *Parameters* of the *ElProScan* level are:

- **List with Series:** If enabled all flagged *Parameters* will be exported either to the Notebook via "Target to Notebook", "Info to Notebook" or to a file via "Target to File" (see *Flagging*, 14.2 on page 208).
- **Scan Mode:** Indicates the used *Scan Mode*, e.g. "TipDown", "2D Scan" or "3D Scan" (see *ElProScan* menu → *Scan* dialog).
- **Scan Alternate:** Indicates if the *Alternate* for a 3D- or Matrix Scan was used (see *ElProScan* menu → *Scan* dialog).
- **Scan Axis:** Indicates the scanned axis, e.g. "XX", "Piezo", or "SF Frequency" (see *ElProScan* menu → *Motors* dialog).
- **Scan Speed:** Speed of the performed scan (see *ElProScan* menu → *Scan* dialog).
- **X-origin:** Absolute value (calibrated) of the X-axis.
- **Y-origin:** Absolute value (calibrated) of the Y-axis.
- **Z-origin:** Absolute value (calibrated) of the Z-axis.
- **X-start:** Starting value on the X-axis in relation to the origin.
- **Y-start:** Starting value on the Y-axis in relation to the origin.
- **Z-start:** Starting value on the Z-axis in relation to the origin.
- **X slope:** Compensating parameter for the X-axis e.g. if the sample and the electrode are lopsided to each other (see *ElProScan* menu → *Scan* dialog).
- **Y slope:** Compensating parameter for the Y-axis e.g. if the sample and the electrode are lopsided to each other (see *ElProScan* menu → *Scan* dialog).
- **CD Amplitude:** Indicates the used Shear force amplitude in C.D. Mode (see *ElProScan* menu → *Constant Distance Modes* dialog).
- **CD Frequency:** Indicates the used Shear force frequency in C.D. Mode (see *ElProScan* menu → *Constant Distance Modes* dialog).

- CD Oper. Mode: Indicates the used *constant distance mode*, e.g. "Shear Force" or "Surface Tracking"(see ElProScan menu → Constant Distance Modes dialog).
- CD Reg. Input: Indicates the used *Regulation Parameters* for the *Constant Distance Mode*, e.g. "SF Amplitude" or "Imon-2" (see ElProScan menu → Constant Distance Modes dialog).
- CD Distance: Indicates the used constant distance (see ElProScan: Scan → Const. Dist. scan modes dialogs → Distance)
- CD Slope Dist: Indicates the slope distance for the used *constant distance mode* (see ElProScan menu → Constant Distance Modes dialog).
- Collision Prev.: Indicates if the *Tip Collision Prevention* for a 3D- or Matrix Scan was used (see ElProScan menu → Scan dialog).
- PullUp Distance: Indicates the used Pull Up Distance in a Constant Distance experiment (see ElProScan menu → Constant Distance Modes dialog)
- Stop Level: Defined deviation of the measured current size in relation to the initially measured current size (see ElProScan menu → Scan dialog).
- Scan Interval: Indicates the set *Line Wait* (resting time before each line scan) (see ElProScan menu → Scan dialog).
- Min. Increment: Indicates the set minimum displacement size (template resolution).
- Matrix X-index: Indicates the X-index number in the defined grid.
- Matrix Y-index: Indicates the Y-index number in the defined grid.
- HPGL Scaling: Scaling factor of the used plotter (HPGL) file (see ElProScan menu → Settings dialog).
- Amplitude: Stimulation amplitude of the SHEARFORCE (see ElProScan menu → ShearForce dialog).
- Frequency: Stimulation frequency of the SHEARFORCE (see ElProScan menu → ShearForce dialog).
- Break After Delta: *Break Criteria* of the *Constant Distance Mode*, either in "V", "A" or "o" (see ElProScan menu → ShearForce dialog).
- 3D Y Increment: Line scan distance of a 3D Scan.
- Piezo Step Size: Defined step size of the Piezo (see ElProScan menu → ShearForce dialog).
- Measure Interval: Total time available for the electrode approximation and the measurement.
- Calib. Interval: Measuring interval during a *Constant Distance Scan* (see ElProScan menu → ShearForce dialog).
- Calib. Distance: Total distance for the measuring intervals during a *Constant Distance Scan* (see ElProScan menu → ShearForce dialog).



14.1.8 Marked Items

All marked (ticked checkboxes) *Parameters* will be displayed in this tab sorted according to their origin (*Root, Group, Series, Sweep, Trace or Amplifier*).

Parameter Options:

☐ **SERIES** This checkbox is available in every *Parameter* tab, if there is no tick set all *Parameters* of this tab will not be displayed in the *Marked Items* tab independent how many *Parameter* checkboxes are ticked.

☒ **SERIES** A highlighted red background color indicates that the display level of the tree is identical to the display tab of the *Parameters*. Here, a *Series* in the *Replay* tree and the *Series Parameters* have been displayed.

☒ **Titles** If selected, the *Parameter* names will be exported together with their corresponding values. Otherwise, only the values will be exported.

☒ **LF** If selected, values and/or *Parameter* names will be separated with a linefeed when exported.

14.2 Flagging

Via the checkboxes (flag options) you can select information that you want to export to the *Notebook* window or to a file (see section before).

Flag Groups 1...4: Four different flag groups can be defined, each containing the information for all checkboxes in this window.

By default, *Flag Group 1* contains the following flags:

- Group: Header, Exp. Number.
- Series: Header, Label, User Name, Date, Time.
- Sweep: Header, Label, Time, Timer.
- Trace: Header, Label, Data Kind.
- Amplifier State: Amplifier, Recording Mode, V-Pipette, Gain, Filter 2, C-fast, C-slow.

Copy Flags from: Copy flag definitions from one group to another. First click on the *Flag Group* that you want information copied into, then on the copy group button of the group whose flags you want to copy to the active group.

Info to Notebook: The contents of the *Parameters* of the selected *Flag Group* are written to the Notebook window when the data is replayed and when `Replay` → *Show Tree Info* is on.

Default Flags: Set the *Flag Group 1* back to the default setting.

Clear all Flags: Reset all flags, i.e. no selected *Parameters*.

Target to Notebook: Copy all flagged information of the actual target in the *Replay* window to the Notebook.

Marked to Notebook: Copy all flagged information of the marked target in the *Replay* window to the Notebook.

Target to File: Copy all flagged information of the actual target in the *Replay* window to a file. You will be asked for a file name.


Marked to File: Copy all flagged information of marked targets in the *Replay* window to a file. You will be asked for a file name.





15 Notebook Window

The Notebook window is used to display messages and warnings of the program, such as error messages, analysis results, and information about displayed data. In many of the POTMASTER windows (e.g. Parameters, Configuration, Analysis), there are checkboxes or options that can be selected to target settings and results to be written into the Notebook window. The content of the Notebook can be stored on disk; its maximal size has to be specified in the drop-down menu **Notebook** → **Set Length**. To keep a better log of an experiment, the names of opened files and of executed series are written to the Notebook window.



When the window is activated (by clicking with the mouse pointer into it), the text editing functions are activated and applicable in the Notebook window. Thus, the Notebook window is an editor window of the memory-resident text file `*.txt`. Therefore, one can modify text in the Notebook window just as in any other text file. This option can be used to add further information to the text file, or to get rid of messages that should not be stored to the disk file.

The applicable menu commands are described in the **Edit** and **Notebook** menus (see **Menus**, 4.2 on page 21 and 4.7 on page 34).

***Note:** The **Cut**, **Copy** and **Paste** commands copy to and from the clipboard.*





16 I/O Control Window

The I/O Control window can have the following sections:

Digital In/Out, DA-Channels, AD-Channels, Values, Serial Out, and Parameters. Which sections and items are shown in the I/O Control window can be selected in the I/O Parameters section of the Configuration window (see 5.7 on page 56).

The idea of the I/O Control window is that the settings in the Configuration window ideally should be edited only during major software setups, whereas the I/O Control window should be used for *Parameter* editing during data acquisition. Hence, only those *Parameter* should be displayed in the I/O Control window that are not set internally and are subject to change during the experiment. These *Parameters* do not have an internal source. That means, are not defined by EPC 10 or read from a specified AD-channel. E.g. some amplifier settings such as recording bandwidth can be entered when a non-telegraphing amplifier is used.

The screenshot shows the 'I/O Control' window with the following sections:

- Dig-In**: A row of 16 checkboxes, with indices 0 through 15 displayed below.
- Dig-Out**: A row of 16 checkboxes, with indices 0 through 15 displayed below. Below this is a 'Clear' button, a blue field containing '0', and a checked 'Single Bit' checkbox.
- DA-Channels**: Four blue fields, each containing '0.000 V'.
- AD-Channels**: Two rows of four blue fields, each containing '0.000 V'.
- Value-1** through **Value-4**: Four blue fields, each containing '0.000'.
- Comm-1**: A 'Send' button, a blue field, and a 'Configure' button.
- Solution**: A blue field containing '1'.
- 1: Normal Frog Ringer**: A blue field.
- Temperature**: A blue field containing '20.00 C'.
- Curr. Bandwidth**: A blue field containing '66.75 Hz'.
- OCP**: A blue field containing '0.000 V'.
- Electrode Area**: A blue field containing '1.000 cm2'.

16.1 Dig-In / Out

Dig-In/Out: Control field of the digital channels. The digital input bits are only displayed. The digital output bits can be set by selecting or deselecting the checkboxes.

Clear: Clear all settings for the digital output, e.g., deselect all bits. The number in the blue field represents the output bits. It can be used to set the bits. Clear actions are written to the Notebook.

Single Bit: If selected, only one bit can be set for *Dig-Out*. The outputted digital value is set to the highest bit.

This close-up shows the 'Dig-In' and 'Dig-Out' sections. 'Dig-In' has 16 unchecked checkboxes. 'Dig-Out' has 16 unchecked checkboxes. Below 'Dig-Out' is a 'Clear' button, a blue field with '0', and a checked 'Single Bit' checkbox.



16.2 DA / AD-Channels

DA/AD-Channels: This section allows you to set the output of the DA-channels. The currently sampled voltages at the AD-channels are only displayed.

***Note:** When using these DA-channels to control external devices from the I/O Control window, make sure that the same DA-channels are not used by other parts of POTMASTER, i.e. as a stimulation channel in the Pulse Generator.*

DA-Channels			
0.000 V	0.000 V	0.000 V	0.001 V
AD-Channels			
-0.000 V	-0.001 V	-0.001 V	0.001 V
0.000 V	0.000 V	-0.003 V	0.012 V

16.3 Values

Values (1-32) are global variables which can be used for calculations in the Protocol Editor and in the Analysis. In addition, they are used to exchange information between these two program modules. In the I/O Control dialog the *Values* are displayed and they can be manually changed. The number of displayed *Values* (4, 8, 12, 16, 20, 24, 28 or 32) can be specified in the Configuration dialog (see 5.7.3 on page 60). In addition, the names of the *Values* are displayed above the numeric *Values*. The name can be set in the Configuration dialog.

Value-1	Value-2	Value-3	Value-4
0.000	0.000	0.000	0.000

16.4 Serial Output

Serial Port: A string can be sent (*Send* button) to a device via the serial port (RS-232) of the computer. The *Configure* button opens the Serial Port Configuration window (see *Serial Out*, 5.4.9 on page 50). An example is shown in the Protocol Editor, 11.4.3.2 on page 143.

Comm-1	Configure
Send	

***Note:** This section is only visible, if the serial communication is turned on in the Configuration dialog, 5.4.9 on page 50.*

16.5 Solutions

Here, it is possible either to monitor the presently active solution or to manually change the solution index. The solution information provided here will be stored in the data file of POTMASTER.

Solution	1
1: Normal Frog Ringer	

***Note:** When changing the solution in the I/O Control window there will be no active setting of external solutions. If you want to do that please use the Solution Changer event in the Protocol Editor (chapter 11.4.3.6 on page 144) and the Solution Changer dialog (chapter 19 on page 225).*

16.6 Parameters

Various *Parameters* can be selected in the Configuration window to be listed in the I/O Control window for display or editing. E.g. if a telegraphing amplifier is used the bandwidth that is read via an AD channel can be displayed.

The individual *Parameters* are described in the context of the Configuration window, 5.7 on page 56.

Temperature	20.00 C
Curr. Bandwidth	66.75 Hz
OCP	0.000 V
Electrode Area	1.000 cm ²





17 Zero Potential (E-zero)

A reference electrode provides a constant potential which should not vary during an experiment. In many cases it is necessary to relate the potentials which have been recorded with your own reference electrode, to a potential of a special reference electrode in the literature or a user defined reference potential. The *Zero Potential* option allows the selection of a desired reference potential (=Zero Potential) which is used to correct recorded potential traces.

Clicking on the **Zero Potential** entry in the **Windows** menu will open a dialog window. In this window a reference potential can be specified either by manual input or by a list of reference potentials. Selecting potentials from the reference list allows additionally to relate these potentials to the Normal Hydrogen Electrode (pH 0) or a Relative Hydrogen Electrode of a different pH value between pH 0...pH 16. Recorded potential traces will be corrected in respect to the selected *Zero Potential*. The value of this *Zero Potential* is visible in the blue box beside *Manual Input*. Recorded potential traces will be corrected against this value by clicking *Done*.

Note: Only the blue box beside Manual Input displays the Zero Potential which is used for a correction of recorded potential traces. Please do not forget to select the options E vs. HE and Use present value in the Pg 300 in order to see the applied potential corrections.

Manual Input: Activating this option allows typing of a user-defined reference potential in mV, for instance the formal redox potential of Ferrocen/Ferrocinium under special experimental conditions. Clicking *Done* applies the *Zero Potential* to recorded potential traces. The short name of the actually tagged reference system will be stored as a comment.

Zero Potential (E-zero)

☒ Manual Input 0.0 mV ☐ Add to Comment

☐ Reference Potential

☐ Normal H-Electrode to ... (pH = 0)

☒ Relative H-Electrode to ... pH: 7.00 at 298 K

☒ Hg / Hg ₂ Cl ₂ / KCl (sat.)	241 mV	SCE
☐ Hg / Hg ₂ Cl ₂ / KCl (1.0 M)	280 mV	NCE
☐ Hg / Hg ₂ Cl ₂ / KCl (0.1 M)	334 mV	Calomel (0.1 M)
☐ Ag / AgCl / KCl (sat.)	197 mV	silver chloride (sat.)
☐ Ag / AgCl / KCl (1.0 M)	236 mV	silver chloride (1.0 M)
☐ Ag / AgCl / KCl (0.1 M)	290 mV	silver chloride (0.1 M)
☐ Hg / Hg ₂ SO ₄ / K ₂ SO ₄ (sat.)	640 mV	SSE (sat.)
☐ Hg / Hg ₂ SO ₄ / K ₂ SO ₄ (0.5 M)	680 mV	mercurous sulphate (0.5 M)
☐ Hg / HgO / NaOH (1.0 M)	140 mV	mercurous oxide (1.0 M)
☐ Hg / HgO / NaOH (0.1 M)	165 mV	mercurous oxide (0.1 M)

☐ User defined 0 mV user defined name

Undo Cancel Done



Reference Potential vs. Normal Hydrogen Electrode: Some of commonly used reference potentials which are related to the Normal Hydrogen Electrode (= NHE) can be selected by tagging the radio button beside the desired reference system of a list of reference potentials. The potential values of the list of reference potentials are the values related to the Normal Hydrogen Electrode at 298 K and 1013 hPa. In the last line of the potential list a user defined option enables typing a user defined potential and short name that will be stored as a comment. The user defined potential here is also related to the Normal Hydrogen Electrode!

Zero Potential (E-zero)

☐ Manual Input mV ☐ Add to Comment

☒ Reference Potential

☒ Normal H-Electrode to ... (pH = 0)

☐ Relative H-Electrode to ... pH: at 298 K

☐ Hg / Hg ₂ Cl ₂ / KCl (sat.)	241 mV	SCE
☐ Hg / Hg ₂ Cl ₂ / KCl (1.0 M)	280 mV	NCE
☐ Hg / Hg ₂ Cl ₂ / KCl (0.1 M)	334 mV	Calomel (0.1 M)
☐ Ag / AgCl / KCl (sat.)	197 mV	silver chloride (sat.)
☐ Ag / AgCl / KCl (1.0 M)	236 mV	silver chloride (1.0 M)
☒ Ag / AgCl / KCl (0.1 M)	290 mV	silver chloride (0.1 M)
☐ Hg / Hg ₂ SO ₄ / K ₂ SO ₄ (sat.)	640 mV	SSE (sat.)
☐ Hg / Hg ₂ SO ₄ / K ₂ SO ₄ (0.5 M)	680 mV	mercurous sulphate (0.5 M)
☐ Hg / HgO / NaOH (1.0 M)	140 mV	mercurous oxide (1.0 M)
☐ Hg / HgO / NaOH (0.1 M)	165 mV	mercurous oxide (0.1 M)
☐ User defined	0 mV	user defined name

Undo Cancel Done

A selected potential trace will be corrected versus the Zero Potential value in this field

List of reference potentials, the potentials are related to the potential of the Normal Hydrogen Electrode

Reference potential vs. Relative Hydrogen Electrode: The potentials of the potential list can be related to a Relative Hydrogen Electrode. The pH differs from pH 0 (activity 1 of the H₃O⁺-ions) and the potentials of the Relative Hydrogen Electrode is calculated in dependency of the pH value assuming a potential difference of 59.2 mV per pH at 298 K and 1013 hPa. The pH can be varied between 0 and 16, the calculated potentials of the Relative pH-Electrode are listed in the following table.

Table 17.1: pH dependence of the potential of the Relative Hydrogen Electrode at 298 K (1013 hPa)

pH	Potential [mV]
0	0,0
1	59,2
2	118,4
3	177,6
4	236,8
5	296,0
6	355,2
7	414,4
8	473,6
9	532,8
10	592,0
11	651,2
12	710,4
13	769,6
14	828,8
15	888,0
16	947,2

Selecting now a reference potential of the reference potential list, the potential of the Relative H-Electrode is added to the selected reference potential.

Zero Potential (E-zero)

☐ Manual Input mV ☐ Add to Comment

☒ Reference Potential

☐ Normal H-Electrode to ... (pH = 0)

☒ Relative H-Electrode to ... pH: at 298 K

A selected potential trace will be corrected versus:
 $290 \text{ mV} + 7 \cdot 59.2 \text{ mV} = 704.4 \text{ mV}$

☐ Hg / Hg ₂ Cl ₂ / KCl (sat.)	241 mV	SCE
☐ Hg / Hg ₂ Cl ₂ / KCl (1.0 M)	280 mV	NCE
☐ Hg / Hg ₂ Cl ₂ / KCl (0.1 M)	334 mV	Calomel (0.1 M)
☐ Ag / AgCl / KCl (sat.)	197 mV	silver chloride (sat.)
☐ Ag / AgCl / KCl (1.0 M)	236 mV	silver chloride (1.0 M)
☒ Ag / AgCl / KCl (0.1 M)	290 mV	silver chloride (0.1 M)
☐ Hg / Hg ₂ SO ₄ / K ₂ SO ₄ (sat.)	640 mV	SSE (sat.)
☐ Hg / Hg ₂ SO ₄ / K ₂ SO ₄ (0.5 M)	680 mV	mercurous sulphate (0.5 M)
☐ Hg / HgO / NaOH (1.0 M)	140 mV	mercurous oxide (1.0 M)
☐ Hg / HgO / NaOH (0.1 M)	165 mV	mercurous oxide (0.1 M)
☐ User defined	0 mV	user defined name

Undo Cancel Done

List of reference potentials, the potentials are related to the potential of the Normal Hydrogen Electrode

Done: Applies the *Zero Potential* to a selected potential trace and closes the window *Zero Potential*. The value of the *Zero Potential* will remain in the window and all potential traces are corrected versus it until a new change is done.

Undo: Sets the window to the default settings, no corrections of the recorded potential traces are done.

Cancel: Closes the window **Zero Potential**, no corrections are applied to selected potential traces.



18 Solution Base

In most electrophysiological experiments bathing solutions are changed during the experiment. Thus, it is of great importance to keep track, which solutions were on both sides of the membrane.

POTMASTER provides two ways to handle solution changes during your experiments:

1. A simple one is to use the entry *Solution* in the I/O Control window. To activate this option please go to the Configuration window and active the *Show Solution* checkbox in the I/O Control tab.




You can set the entry manually or via the event *Set Solutions* in the Protocol Editor (11.4.5.2 on page 150). The numbers can be the entry of an external list of solutions, or the index of a specific solution in a POTMASTER Solution Data Base (see 18.1). In case you are using a multiple amplifier and you selected *Show Amplifier Solutions* in the Configuration (5.7.2 on page 60) you have to specify the used solutions for every amplifier. The solution indices are stored together with the recorded data and can be reviewed in the Parameters window (14.1.5 on page 202).

2. A more advanced method is to create a *.sol file. Such a data base is a *Tree* of solution entries ordered by the given identifying index.

This way the indices stored with the data are converted to solutions from the *Solution Data Base*. Here, it is also possible to use the event *Set Solutions* in the Protocol Editor (11.4.5.2 on page 150).

18.1 Activating the Solution Base

First, turn on the *Solution Base* in the Configuration window (5.5.2 on page 52). If no *Solution Data Base* has been loaded before and stored with the current configuration, a dialog box shows up asking you if you want to open an existing data base or creating a new data base.



During the experiment, solutions can be set either via the I/O Control window (16.5 on page 214) or via the *Set Solutions* event in the Protocol Editor (11.4.5.2 on page 150).



18.2 Using Solution Indices

The indices can be given to solutions in any arbitrary way but it certainly is of advantage if one sticks to some consistent concept in order to be able to identify solutions by their index easily. Since the solution indices are numbers between 0 and 2'147'483'648 (many more than you will find bottles in the laboratory), there is plenty of freedom to organize them.

Here is an example: Usually one has several standard solutions, which are frequently used and modified slightly for various experiments. One could assign numbers divisible by 1000 or 10000 to them. Then one has 999 or 9999 possibilities for modifications of this solution, respectively. An example would be that certain concentrations of toxin are added to the standard solution 1000 yielding numbers 1001 through 1099. Another toxin could occupy the numbers 1100 through 1199, etc..

***Note:** The indices given for internal and external solutions are pointing into the same Solution Data Base. It is therefore a good idea to index internal and external solutions such that they are consistent and easily identifiable. For example, you could use odd thousands (or ten thousands) for external solutions and even thousands (or ten thousands) for internal solutions.*

18.3 Solution Data Base

Solution Base

1 Normal Frog Ringer

Numeric Name: Free Ca 1.80m

pH: 7.20 NaOH Osmol.: 285.m

Index	Ingredient	Conc.
1	NaCl	140. mM
2	KCl	2.50 mM
3	CaCl ₂	1.80 mM
4	Hepes	10.0 mM

1 entry

Create Entry

Duplic. Entry

Delete Entry

Next Entry

Last Entry

Export Label

Export Listing

SAVE

UNDO

DONE

To open the Solution Data Base you have to select Solution Base in the Windows menu of POTMASTER.

Solution Index: Index number for the solution. By clicking into the field you are able to enter a defined index number directly which might be helpful if you have a large Solution Data Base.

Name: Name of the solution.

Numeric Name: An editable field that may hold a feature of the solution that is not easily determined from its ingredients (e.g., free calcium concentration, etc.).

Numeric Name Value: An editable field containing a number or value (Osmolarity, pH) which is in relation to the name entered in the field Numeric Name.

pH: Holds the value of the pH of the solution and the substance used to adjust it.

Osmol.: Holds the value of osmolarity of the solution.

Ingredient / Conc.: Each ingredient is defined by its index number and its concentration in the solution.

Insert, Append, Delete: By clicking into an index number of the solution ingredients options for modifying the solution content get available:

- **Insert:** Creates a new entry line for an ingredient before the selected index number.
- **Append:** Creates a new entry line for an ingredient at the end of the ingredient list.
- **Delete:** Deletes the selected ingredient.

Entries: Number of entries in the solution data base.

Create / Duplic. / Delete Entry: Used to generate, copy, and remove solutions in the data base.

Next / Last Entry: This option moves through the data base by selecting the next or last available solution.

Export Label: This is used to output labels of the shown solution:

- **ASCII File:** Exports the solution information as an ASCII text file (*.label).
- **Printer:** Two labels with two different size containing the solution information are created.
- **Igor Text:** Exports the solution information as IGOR text (*.label).
- **Notebook:** Exports the solution information to the Notebook.

Export Listing: This is used to output a list of the entire solution file:

- **ASCII File:** Exports the index number and the name of the solutions as an ASCII text file (*.list).
- **Printer:** Prints the index number and the name of the solutions.
- **Notebook:** Lists the index number and the name of the solutions in the Notebook.

SAVE: Saves the file to disk.

UNDO: Reverts the edited solution to its original form.

DONE: Exits the dialog.






19 Solution Changer

The Solution Changer dialog can be accessed via the `Windows` menu of POTMASTER. The Solution Changer dialog has two main purposes:

1. Manage analog or digital outputs of your perfusion system
2. Align external or internal solutions to the perfusion system

Thus, using the Solution Changer helps to keep track with the solutions used during an perfusion experiment. The solution information is stored in the data file of POTMASTER. To look up this information use the `Parameters` dialog.


Any external perfusion system with analog or digital inputs can be used together with the Solution Changer and e.g. an EPC 10 USB amplifier.



The Solution Changer dialog has two main tabs. In the *Control* tab one sets the active perfusion channel manually ("Set"). Further the solution index and the solution name from the `Solution Base` is displayed here. By clicking into the selection field it is possible to change the solution index/solution name.

Note: The solution selection does only work if a `Solution Base` is defined and loaded.





In the *Configure* tab you can define up to 16 solutions used in the perfusion experiment. Further you can define if the used solutions is an internal or an external solution. In the last column one defines the output channel e.g. analog or digital. If an analog output is selected you will have to define the voltage amplitude necessary to activate the perfusion channel.


Note: Please do not forget to define an output channel for the "Off" position.

DA-0
DA-1
DA-2
Stim-DA
DA-4
DA-5
DA-6
DA-7
✓ off
Dig-out (word)
Dig-0
Dig-1
Dig-2
Dig-3
Dig-4
Dig-5
Dig-6
Dig-7
Dig-8
Dig-9
Dig-10
Dig-11
Dig-12
Dig-13
Dig-14
Dig-15
Clear all



20 Markers

The Markers dialog can be opened by selecting **Markers** from the **Windows** menu.



With this dialog, *Markers* can be set at any time during a running experiment even between *Sweep* acquisitions. A *Marker* is drawn as a blue vertical line in the Oscilloscope.

20.1 New Marker

In this section one can add *Markers* and define the name of the label.


Add Marker: Adds a new *Marker* to the data file. The *Marker Text* and the target *Sweep/Trace* information is stored as well.

Text: Enter a label for the *Marker* you want to add.


Trace: A *Marker* can be linked either to a *Sweep* or a *Trace* which can be specified by this selection menu.

If the *Marker* is defined for a *Sweep* it is labeled "[i] Text". The i in square brackets indicates the number of the *Marker* in the list (index number) and "Text" shows the label of the *Marker*.





If the *Marker* is defined for a *Trace* it is labeled "[i] i: Text". The i in square brackets indicates the number of the *Marker* in the list (index number) whereas the index number thereafter indicates the *Trace* index number. "Text", again, shows the entered label for the *Marker*.



Time: The time of a new *Marker* is the actual time if the field is empty. It can be the given time in the time format "hh:mm:ss.mmm" e.g. if you add a *Marker* offline. Optionally the time format may be preceded by a date "yyyy/mm/dd".

From Sweep: Gets the time information of the *Sweep* selected in the Replay tree.

20.2 Edit Marker

Marker Index: The *Marker* index e.g. "1 of 7" indicates the index number of the selected *Marker* in the *Marker* list.

Show Sweep: Selects the according *Sweep* in the Replay tree and replays it in the Oscilloscope.

Show Marker: Shows the appropriate *Marker* information in the *Marker* list if the selected *Series* in the data tree owns a *Marker*.

Delete: Deletes the selected *Marker*.


Text: Modifies the label of the set *Marker*.

Trace: Modifies the link (*Trace* or *Sweep*) of the set *Marker*.

Time: Displays the time information of the selected *Marker*.

: The two arrow buttons enable to scroll through the list of added *Markers*.

Write to Notebook: Writes the time information of acquired *Series* and set *Markers* into the Notebook.




Note: In the *Protocol Editor* event "Set Solutions" it is possible to add *Solution Markers*. The *Markers* set there are also included into the *Markers* list.






21 Calculator Window and Equations

Equations are a very flexible way to perform complex calculations within POTMASTER. Equations can be used at various places, such as *Analysis*, *Trace Buffer*, or for *DA-Scaling*. In addition, a separate **Calculator** window allows to use the equations syntax as replacement of the calculator provided by the operating system.



In case an equation has to be edited in the *Analysis*, *Protocol Editor* or *Trace Buffer*, then the following dialog can be opened by pressing on the *Edit* button.



Command Line: Enter the equation string into this field.

Results: The result of the equation will be shown here.

Equation: You can enter an equation name which will be loaded into the *Command Line* after pressing RETURN. Alternatively if an equation was selected by the option *Select Equation* the equation name will be displayed in this field.

Select Equation: After loading an equation file you can select already stored equations.

Save Equation: For saving an equation enter a name in the *Equation* field and press the *Save Equation* button. When the entered name does not already exist in the equation file you will be asked if you want to save it as a new entry.

? : Prints a help text on the equation syntax to the Notebook.

LOAD: Brings up a file selector to open an equation file (e.g. *Equation.txt*).

SAVE: Saves the modifications made in the equation file.

Cancel: Closes the dialog without accepting the equation.

Done: Accepts the new equation string and closes the dialog.



21.1 Equation Syntax

Normal math syntax is to be used. All 3 types of brackets ('('and')', '['and']', '{'and'}') are equivalent and can be freely used and nested. The opening and closing brackets must match.

Math functions:

- +, -, *, /, ^ (power), % (remainder)
- & (and), nor, or, xor: bit operation on integer values
- >, <, =: boolean operations: true=1, false=0
- round: rounds towards nearest value using the banking rules (-2.5 → -3.0)
- trunc: rounds towards the smaller amount (-2.5 → -2.0)
- ceil: rounds towards the positive (-2.5 → -2.0)
- floor: rounds towards the negative (-2.5 → -3.0)
- exp, sqrt, ln, log
- sin, cos, tan
- arcsin/asin, arccos/acos, arctan/atan
- abs, rad, deg
- random: returns random number between 0.0 and 1.0
- f[i], i = online function index = 1...n
- p[i], i = amplifier parameter index = 1...n
 - p[0] → Icell (Direct Current Monitor value)
 - p[1] → Ecell (Cell Potential Monitor value)
 - p[2] → Overload (Overload status)
 - p[3] → OCP (Open Cell Potential value)
 - p[4] → ElectrArea (Electrode Area value)
 - p[5] → Temp (Temperature value)
 - p[6] → AnodicQ (Anodic Charge value)
 - p[7] → CathodicQ (Cathodic Charge value)
 - p[8] → User1 (User-defined parameter 1)
 - p[9] → User2 (User-defined parameter 2)
 - p[10] → Timer (Time of the Timer)
 - p[11] → Time (Time)
 - p[12] → Xpos (ElProScan specific parameter)
 - p[13] → Ypos (ElProScan specific parameter)
 - p[14] → Zpos (ElProScan specific parameter)
 - p[15] → PiezoPos (ElProScan specific parameter)
- v[i], i = value index = 1...n
- icon[icon descriptor], e.g. 'icon[E VHold]'



Constants:

- e, pi

Special case:

'@' as the first character will replace the given label by the formula in the equation file with that label. E.g. '@first 1+2' results in '1+2'.

Comments:

Anything after ';' is considered a comment.

21.2 Equations - Where to Use

21.2.1 Analysis

Trace Param.	Math	Trace	Spectra
☐ Trace Count	☒ Equation	☐ Trace	☐ Frequency
☐ C-slow	☐ Y(x): y at pos = x	☐ Equation	☐ Distribution
☐ R-series	☐ Constant	☐ Q = Integral	
☐ Rs-value	☐ a + b	☐ 1 / (trace)	Histogram
☐ Leak Comp.	☐ a - b	☐ 1 / (Q)	☐ Histogram Ampl
☐ M-conductance	☐ a * b	☐ ln (trace)	☐ Histogram Bins
☐ Cell Potential	☐ a / b	☐ ln (Q)	
☐ Seal Resistance	☐ a in b	☐ log (trace)	
☐ Pip. Pressure	☐ abs	☐ log (Q)	
☐ Int. Solution	☐ log	☐ dt = Differential	
☐ Int. Sol. Value	☐ sqrt	☐ Trace x-axis (time)	
☐ Ext. Solution	☐ arctan	☐ Stimulus	
☐ Ext. Sol. Value	☐ 1/a		
	☐ 1/log		
Sweep Param.	☐ 1/sqrt		
☐ User_1	☐ 1/arctan		
☐ User_2			
☐ Temperature			
☐ Digital-In			

Equation:

Name:

The *Equation* functions in the *Math* and *Trace* section of the Analysis Function dialog provide the possibility to assign the result of an equation to a function. The advantage of the equation is to perform multiple steps of a calculation with just one function. In an equation of the Analysis function the results of other Analysis functions can be easily combined.

Math - Equation

The following additional identifiers can be used in the *Math - Equation* function:

- f[i], i = online function index = 1...n
- p[i], i = amplifier parameter index = 1...n
- v[i], i = value index = 1...n
- icon[icon descriptor], e.g. 'icon[E VHold]'

Example - working with the icon descriptor: A parameter of the Parameters window can be easily plotted in the Analysis graph.

Trace Param.	Math	Trace	Spectra
☐ Trace Count	☒ Equation	☐ Trace	☐ Frequency
☐ C-slow	☐ Y(x): y at pos = x	☐ Equation	☐ Distribution
☐ R-series	☐ Constant	☐ Q = Integral	
☐ Rs-value	☐ a + b	☐ 1 / (trace)	Histogram
☐ Leak Comp.	☐ a - b	☐ 1 / (Q)	☐ Histogram Ampl
☐ M-conductance	☐ a * b	☐ ln (trace)	☐ Histogram Bins
☐ Cell Potential	☐ a / b	☐ ln (Q)	
☐ Seal Resistance	☐ a in b	☐ log (trace)	
☐ Pip. Pressure	☐ abs	☐ log (Q)	
☐ Int. Solution	☐ log	☐ dt = Differential	
☐ Int. Sol. Value	☐ sqrt	☐ Trace x-axis (time)	
☐ Ext. Solution	☐ arctan	☐ Stimulus	
☐ Ext. Sol. Value	☐ 1/a		
	☐ 1/log		
Sweep Param.	☐ 1/sqrt		
☐ User_1	☐ 1/arctan		
☐ User_2			
☐ Temperature			
☐ Digital-In			

Equation:

Name:

As *equation* string you choose the icon descriptor of the *C-slow* parameter from the Parameters window. Please note that the Parameters window has to stay open to access the value of this icon. However, you may minimize the window. The advantage of using the *C-slow* parameter from the Parameters window instead of the Amplifier window is that the *Parameter* is available also during offline replay and analysis of the acquired data.

Analysis Functions

4 Mean end 5 Time 6 "C-slow" 7

X, Y-seg. Offset:

Cursor Bounds (%):

Trace #: ☒ Notebook

not stored in Value

Trace - Equation

The Analysis Function Trace - Equation provides the possibility to scale a Trace before plotting it in the online graph. The advantage of the *equation* is to perform multiple steps of a calculation with just one function. In an *equation* of the Analysis function the results of other Analysis functions as well as other trace data of the selected trace can be easily combined.


In addition to the identifiers of *Math - Equation* the following identifiers can be used:

- y: The trace y-value, i.e. "amplitude"
- x: The trace x-value, i.e. "x-axis value = x-start + index * x-increment"
- Q: The integral up to the respective index, i.e. " $(y[i] + y[i-1]) / 2 * dx$ "
- dt: The differential between present and preceding y-value, i.e. " $(y[i] - y[i-1] / dx)$ "
- i: The index of the trace data point
- tb[i]: The y-value of equivalent sample in buffer i, i = buffer index, 1...4
- tr[i]: The y-value of equivalent sample in trace i, 1...n. One can use the trace-ID (e.g. "I-mon") to specify a trace instead of the numeric trace-count "i". The text must be identical to the text in the trace-count menu.

21.2.2 Trace Buffer

In case you would like to permanently modify (rescale) acquired data or at least let the rescaled data then run through the **Analysis** then you can use the *equation* functionality of the *Trace Buffer* menu or protocol event.

An example: As *Trace 3* we acquire the frequency dependence of the *Admittance* (Gain). This is an output *Trace* of the *Spectroscopy Extension*. The X-axis of the *Admittance Trace* is scaled in frequency units. The *Admittance* itself is scaled in linear units. For some applications it is useful to rescale the *Admittance* in dB units.



The protocol event above shows the following: We take *Trace 3* as source, clear the destination buffer, add the source to the buffer-3, apply the equation (here we calculate $20 \cdot \log(y)$, y = values in the destination buffer after performing the operation), and then put the result back into *Trace 3*. Now we can run *Trace 3* through the **Analysis** again and apply the analysis function on an *Admittance* (Gain) *Trace*, which is now scaled in decibel. E.g. we can apply the *Time to Threshold* function with a threshold of -3 (dB) to search for the cutoff frequency.

The operands are identical to the ones used in *Trace - Equations* (see 21.2.1 on the preceding page).


21.2.3 DA Scaling

Special Operands:


- **ampl** = The value for output or amplitude as defined by the PGF-stimulus. Used for equations in *DA-Scaling* only.

In the following we will describe how to output two consecutive pulses with exponentially increasing voltages.


We first enter in the **I/O Control** pane of the **Configuration** window for the *DA Scaling* an equation of a single exponential (5.7.1 on page 57). As *Parameters* we use the values 1 to 3. As time or x-value we use the *Ampl Parameter* of the stimulus.



In a second step, we define the appropriate segment pattern for the DA 0 channel and select *Use DA Scaling*. In segments of amplitude 0 a constant voltage is put out. In case we want to generate the exponentially increase, we set the segment type to *Ramp* and choose for amplitude and duration the same value (1 Volt corresponds to 1 Second). In case the amplitude has to be reset to zero, we insert a constant segment with duration 0 and amplitude 0.



We now set the *Values 1 to 3* as desired via the I/O Control window and execute the sequence. The result is shown in the Oscilloscope window. By changing the values 1 to 3 the stimulus can be easily adjusted.



21.2.4 Values

The protocol event *Set Value* provides the possibility to assign the result of an equation to a value. The advantage of the equation is to perform multiple steps of a calculation with just one protocol event. In an equation of the *Set Value* event results of the *Analysis* can be easily combined. Therefore the definitions are identical to the ones in *Math - Equation* (see 21.2.1 on page 233).

22 Photometry Extension

This chapter provides a brief overview of the settings for each of the Photometry Extensions mentioned above. A complete tutorial/exercise for photometry measurements using POTMASTER is provided in the PATCHMASTER Tutorial.

After starting POTMASTER, open the Configuration window and select the appropriate Photometry Extension in the Hardware section of the dialog. The options are:

- TILL Photonics / Cairn OptoScan
- DG-4 / DG-5
- Lambda-10
- PTI DeltaRAM

Some controls are independent from the photometry hardware. These controls are described in this section. Hardware specific options will be described below.

R-max	1.0000E+00	R-min	0.0000E+00	Kd * Sf	1.0000E+00
Background1	0.0000E+00	Background2	0.0000E+00	Background3	0.0000E+00
Dead Time	2.00 ms	Compute Traces of Marked Targets			

R-max, R-min: Minimum and maximum ratio (F1/F2).

Kd*Sf: Effective Kd (dissociation constant) of the dye.

Note: These values are required for calculating the free calcium concentration (according to the Grynkiewicz formalism). The calcium concentration can be stored as a Trace by use of the "Photo_Ca" input channel in the Pulse Generator dialog.

Background 1...3: Background fluorescence for the three wavelengths F1, F2, F3.

*Note: The parameters R-max, R-min, Kd*Sf and Background 1...3 are stored as "Series Parameters" with the raw data file and are listed in the Parameters dialog (see figure below).*

☐ Photo1 Backgroui	0.000	☐
☐ Photo2 Backgroui	0.000	☐
☐ Photo3 Backgroui	0.000	☐
☐ PhotoCa Kd	1.000	☐
☐ PhotoCa Rmin	0.000	☐
☐ PhotoCa Rmax	1.000	☐

Dead Time: Time to allow for settling of the commanded wavelength before starting the image exposure.

Compute Traces of Marked Targets: This function allows to recompute stored "Photo" Traces (Photo_W1, Photo_W2, Photo_W3, Photo_R, and Photo_Ca) with modified fluorescent background values.



22.1 Photometry Configuration TILL Photonics/Cairn OptoScan

This Photometry selection can either be used with the TILL Photonics Polychrome or the Cairn OptoScan monochromator.

Photometry: TILL Photonics / Cairn OptoScan

R-max	1.00E+00	R-min	0.00E+00	Kd * Sf	1.00E+00
Background1	0.00E+00	Background2	0.00E+00	Background3	0.00E+00
Dead Time	2.00 ms	Compute Traces of Marked Targets			

Undo **Set Defaults**

Emit-1 AD-1 **0.00V**
 Emit-2 off **0.00V**

Resting Wavelength **280 nm** **Set** Excit. DA-1 **-9.95V 280 nm**

No Shutter

Hide **Show Calibration**

Wavelength Calibration **Polychrom I - IV**

from **-10.00V** to **10.00V** **Compute Factors**

wavelength	lower	volt	wavelength	upper	volt
380 nm	Scan	-4.50V	505 nm	Scan	2.50V

Wavelength to Volts

Slope	Offset	Angle β
7.2500E-02	1.8200E+00	30°

Undo: Resets all changes made since the last opening of the Photometry window.

Defaults: Resets all changes to the default entries.

Resting Wavelength: Wavelength to be set during resting periods.

Set: Only after pressing the Set button the wavelength entered in the *Resting Wavelength* field is output.

Shutter Selection:

- No Shutter: No shutter available.
- Shutter Control: Shutter can be closed/opened via a button.
- Digital-out 7: Shutter control via digital output.

Emit-1, Emit-2: Select the AD channels for sampling the photometry responses. *Emit-1* is always active whereas *Emit-2* can be turned off. It may be used for dual emission fluorescence experiments.

Excit.: Select the DA channels for the voltage output. This function is used to control the monochromator. Once it is calibrated, the output can be defined either in Volts (left) or in nm (right).

22.1.1 Wavelength Calibration

Hide		Show Calibration	
Wavelength Calibration		from	-10.00V to 10.00V
		Compute Factors	
wavelength - lower -	volt	wavelength - upper -	volt
380 nm	Scan -4.50V	505 nm	Scan 2.50V

In this area the monochromator is calibrated, i.e. the relationship between output voltage and the wavelength is defined. To do so, voltage ramps ranging between the values specified by *from...to* are output when the *Scan* buttons are pressed. This is done for two bandpass filters with known transmission maxima. The corresponding wavelengths are entered in the *wavelength* fields. After both scan operations, i.e. when the voltages corresponding to the peak transmission have been determined successfully, the calibration constants for the control of the monochromator are calculated upon *Compute Factors* is pressed.

from - to: Scan range for voltage ramps. For a single bandpass filter usually the full range of ± 10 V can be scanned. However, when a single multi-band filter is used, the scan range has to be limited in order to detect only the peak of interest.

Compute Factors: Compute calibration constants based on the determined relationships of wavelength and voltage.

wavelength: Enter here the maximum of the bandpass filter (in nm).

lower - Scan: Do the scan for the low-wavelength calibration filter.

Volt: This field shows the voltage corresponding to the peak transmission.

upper - Scan: Do the scan for the high-wavelength calibration filter.

Volt: This field shows the voltage corresponding to the peak transmission.

Wavelength to Volts: *Slope*, *Offset* and *Angle* are internal calibration constants used to control the monochromator. These values are the result of the calibration procedure.

Wavelength to Volts		
Slope	Offset	Angle β
7.2511E-02	1.8235E+00	30°

The conversion between voltages and wavelength is done as follows:

$$F(\lambda) = \arcsin(\lambda/1666.67 * \cos(\pi/360^\circ * \beta)) \quad (22.1)$$

$$V = (2\pi * F(\lambda) - Offset)/Slope \quad (22.2)$$

$$\lambda = (1666.67 * \cos(\pi/360^\circ * \beta) * \sin((Slope * V + Offset)/2\pi)) \quad (22.3)$$

$$Slope = 2\pi * (F(\lambda_1) - F(\lambda_2))/(V_1 - V_2) \quad (22.4)$$

$$Offset = 2\pi * F(\lambda_1) - Slope * V_1 \quad (22.5)$$

with

- β = typically 30 degree; depends on your hardware. Please refer to the manual of the monochromator.
- λ = requested wavelength in [nm]; λ_1 = calibration wavelength 1.
- V = DA voltage in [V]; V_1 = DA voltage of calibration wavelength 1.
- arguments of trigonometry functions in [RAD].

Here an example: We calibrate our monochrometer at the wavelength 380 nm and 505 nm. With the 380 nm and 505 nm filters we measure an intensity maximum at -4.5 V and 2.5 V, respectively. Now, we calculate $F(\lambda)$ for both wavelength: $F(\lambda_1) = 0.2383$, $F(\lambda_2) = 0.3191$. Then the *Slope* is 0.0725 and the *Offset* is 1.8235.

22.1.1.1 Calibrating the different TILL monochrometers

Polychrome I and Polychrome II: Two bandpass filters are provided with the Polychrome I and II. Enter the peak wavelength in the *Wavelength* fields of the *Wavelength Calibration* section, then put the first filter in place and press the left *Scan* button. The voltage of the peak intensity will be automatically detected and shown in the *Volt* field. Now, put the second filter in place and press the right *Scan* button. Finally, press *Compute Factors* to calculate all calibration parameters.

Polychrome IV: The Polychrome IV comes with a triple-band filter. When using this filter the scan range to detect the first or the last peak of the three has to be limited. Therefore, you should limit the scan range for the first peak from -10 V to -2 V and for the third peak from +2 V to +10 V. We omit the center peak in this calibration. Please enter the peak wavelengths in the corresponding *Wavelength* fields and the scan range before performing each scan in the *from – to* fields. Then, perform the two scans and finally, press *Compute Factors* to calculate all calibration parameters.

Polychrome V: The Polychrome V features an auto-calibration. Please ask T.I.L.L. Photonics for a pair of calibration values. E.g. which analog control voltages refer to the wavelength 380 nm and 505 nm. Then, please enter the wavelength and voltages in the *Wavelength* and *Volts* fields and press *Compute Factors* to calculate all calibration parameters.

22.2 Photometry Configuration DG-4/DG-5

Undo: Resets all changes made since the last opening of the Photometry window.

Defaults: Resets all changes to the default entries.

Resting Wavelength: Wavelength to be set during resting periods.

Set: Only after pressing the *Set* button the wavelength entered in the *Resting Wavelength* field is output.

Emit-1, Emit-2: Select the AD channels for sampling the photometry responses. *Emit-1* is always active whereas *Emit-2* can be turned off. It may be used for dual emission fluorescence experiments.

Position: Enter a number for the filter wheel position of the DG-4/DG-5 between 0 and 15.

Shutter Selection:

- No Shutter: No shutter available.

- Shutter Control: Shutter can be closed/opened via a button.
- Digital-out 7: Shutter control via digital output.

Shutter Open/Close: Toggles between an open and a closed shutter.

Note: To control the DG-4/DG-5 via the POTMASTER software you need a special cable connecting the Digital Outputs of the interface and the DG-4/DG-5. Please contact HEKA (eimeaorder-scmth@harvardbioscience.com) if you need such a cable (Order Number: 895102).

22.2.0.1 PGF Setting of DG-4/DG-5

When using the DG-4/DG-5 you need to define the *Filter Position* in your PGF sequence. Therefore you have to set the DA output channel to *Dig Out (word)* and use for *wavelength*. Then you can enter the *Filter Position* directly in the segments (Wavelength) of the PGF sequence.

Pulse Generator File: DefPgf

Full View | Condensed View | Cartoon View

9 Time Expand | 10 LTP | 11 LockIn | 12 TestSeries | 13 DG4/DG5 | 14

Pool: LOAD MERGE SAVE Name: DG4/DG5 LIST COPY MOVE UNDO DELETE

Interactive Mode | Gap Free Mode

Photometry Wave

Timing: No wait before 1. Sweep | Not Triggered

No of Sweeps: 1 | Use Durations

Sweep Interval: 0.00 s | StartSeg: 0

Sample Interval: 50.0 µs (20.0kHz) | StartTime: 0.00

Checking | EXECUTE

Sweep Length: 1.000 s | 20000 pts

Channel Length: 1.000 s | 40060 bytes

1.000 s | 20000 pts

2	DA	Unit	Stimulus -> DA	Leak	AD	Unit	Link	Compr.	Points	Store	Zero	Leak
Ch-1	Stim-DA	V	StimScale	☐	Imon2	A	1	1	20000	☒	0	No Leak
Ch-2	Dig-out (wo	nm	Wavelength	☐	Photo W1	A	2	4000	5	☒	0	No Leak
Ch-3	off	V	absolute voltage	☐	Photo W2	A	2	4000	5	☒	0	No Leak
Ch-4	off	V	absolute voltage	☐	Photo R	A	2	4000	5	☒	0	No Leak

Segments | 1 | Stored | 2 | Stored | 3 | Stored | 4 | Stored | 5 | Stored | 6 | St

Segment Class	Constant	Constant	Constant	Constant	Constant	Constant
Wavelength	val 2	val 4	val 2	val 4	val 2	val 4
Duration [ms]	val 100.00	val 100.00	val 100.00	val 100.00	val 100.00	val 100.00
W-fact./incr.	Increase	Increase	Increase	Increase	Increase	Increase
t-incr. Mode	1.00 0	1.00 0	1.00 0	1.00 0	1.00 0	1.00 0
t-fact./incr. [ms]	1.00 0.00	1.00 0.00	1.00 0.00	1.00 0.00	1.00 0.00	1.00 0.00

Draw: Active Channel, all Sweeps | Delay: DA 0.00 s | AD 55.0 ms

100.mnm | 100.ms

Leak Pulses

No of Leaks: 0

Leak Delay: 10.0 ms | Leak Alternate

Leak Size: 1.000 | Alt Leak Average


Leak Hold: --- | wait = abs. hold

Set Last Seg. Amplitude

p1	p2	p3	p4	p5	p6	p7	p8	p9	p10
100.00m	0.0000	100.00m	90.000m	10.000m	45.000m	0.0000	0.0000	0.0000	0.0000

Traces 4

22.3 Photometry Configuration Lambda-10



Undo: Resets all changes made since the last opening of the Photometry window.

Defaults: Resets all changes to the default entries.

Resting Wavelength: Wavelength to be set during resting periods.

Set: Only after pressing the Set button the wavelength entered in the *Resting Wavelength* field is output.

Emit-1, Emit-2: Select the AD channels for sampling the photometry responses. *Emit-1* is always active whereas *Emit-2* can be turned off. It may be used for dual emission fluorescence experiments.

Position: Enter a number for the filter wheel position between 0 and 15.

Speed: Enter a number between 0 and 7.

Shutter Selection:


- No Shutter: No shutter available.
- Shutter Control: Shutter can be closed/opened via a button.
- Digital-out 7: Shutter control via digital output.

Shutter Open/Close: Toggles between an open and a closed shutter.

Note: To control the Lambda-10 via the POTMASTER software you need a special cable connecting the Digital Outputs of the interface and the Lambda-10. Please contact HEKA (eimeaorder-scm@harvardbioscience.com) if you need such a cable (Order Number: 895102).

22.3.0.1 PGF Setting of Lambda-10

When using the Lambda-10 you need to define the *Filter Position* in your PGF sequence. Therefore you have to set the DA output channel to *Dig Out (word)* and *use for wavelength*. Then you can enter the *Filter Position* directly in the segments (Wavelength) of the PGF sequence.



22.3.1 Lambda 10-3

Before using the Lambda 10-3 controller for controlling the shutter and the filter wheel via the POTMASTER software you need to set the Lambda 10-3 controller into “Parallel Port” mode. By default the Lambda 10-3 controller is configured for “USB” mode. To set it into the “Parallel Port” mode you need to press the following keys at the controller:

- Switch on the device
- Press “Local”
- Press “Mode”
- Press “7 - Default”
- Press “1 - Set the new default”
- Press “4 - Comm Port”
- Press “3 - Parallel Port”
- Press “0 - Set no other defaults”

Once this prerequisite is fulfilled there are two reasonable ways to control the shutter via the Lambda 10-3:


1. Using the digital output channel “Dig-7” controlled in the Photometry dialog $\Rightarrow$ TTL Mode
2. Using the digital bit (word) commands either in the Pulse Generator, Protocol Editor or I/O Control Window $\Rightarrow$ Conditional Mode.

TTL Mode

To control the shutter via the “Dig-7” digital port you need to enable “TTL” mode at the Lambda 10-3 controller:

- Switch on the device
- Press “Local”
- Press “Mode”
- Press “3 - TTL”
- Press “Enable TTL”

If the “TTL” mode at the Lambda 10-3 is disabled it is not possible to use the Photometry extension to control the shutter.




Now, one can use the *Shutter* button in the Photometry Dialog or the digital output channel “Dig-7” in the I/O Control Window or elsewhere to control the shutter.

Conditional Mode

To control the shutter in the *Conditional Mode* you need to disable “TTL” mode at the Lambda 10-3 controller:


- Switch on the device
- Press “Local”
- Press “Mode”
- Press “3 - TTL”
- Press “Disable TTL”

One can activate the “Conditional Mode” of the Lambda 10-3 e.g. in the I/O Control Window and enter the bit “299” (deselect *Single Bit* first). This means the shutter closes automatically when the filter wheel is moved and opens again when the filter wheel movement has finished.



To switch off the “Conditional Mode” you need to send either bit “298” or bit “300”. Bit “298” disables the “Conditional Mode” but leaves the shutter open, bit “300” disables the “Conditional Mode” but closes the shutter.

22.4 Photometry Configuration pti DeltaRAM



Undo: Resets all changes made since the last opening of the Photometry window.

Defaults: Resets all changes to the default entries.

Resting Wavelength: Wavelength to be set during resting periods.

Set: Only after pressing the *Set* button the wavelength entered in the *Resting Wavelength* field is output.

Emit-1, Emit-2: Select the AD channels for sampling the photometry responses. *Emit-1* is always active whereas *Emit-2* can be turned off. It may be used for dual emission fluorescence experiments.

Excit.: Select the DA channels for the voltage output. This function is used to control the monochromator. Once it is calibrated, the output can be defined either in [V] (left) or in [nm] (right).

Shutter Selection:

- No Shutter: No shutter available.
- Shutter Control: Shutter can be closed/opened via buttons.
- Digital-out 7: Shutter control via digital output.




Wavelength to Volt *Volts/nm* and *Wavelength at 0 V* are internal calibration parameters, which must be given as described in the specifications of your DeltaRAM system. The PTI system has a linear relationship between the applied voltage and the wavelength of the excitation light.



23 SmartLUX (Imaging Extension)

SmartLUX allows to perform simultaneous image acquisition and electrophysiological recordings. The timing of a fluorescence excitation light source, the camera and the electrophysiological command signals are synchronized by the Pulse Generator of POTMASTER. Fluorescence mean values from regions of interest (ROI), which can be marked by the user in the image, are plotted together with e.g. the current Trace in the Oscilloscope window and stored together with the other data in POTMASTER's data tree. When stepping through the data tree, SmartLUX shows the corresponding image. In case more images have been acquired during a single Sweep, the individual images are shown when stepping through the data points of the fluorescence Trace using the scan function of the Oscilloscope.



SmartLUX consists of an Imaging window in POTMASTER, an additional Control window and an Image window. SmartLUX supports the following hardware:

- Fluorescence Excitation light sources which can be controlled in the same way as:
 - Polychrome and Oligochrome from TILL Photonics
 - DG-4/DG-5 wavelength switcher from Sutter Instrument Company
 - Lambda-10 filter wheels from Sutter Instrument Company
 - DeltaRAM monochromator from pti

-
- Cameras supporting the so-called "Bulb Trigger Mode" from Andor, Jenoptic, QImaging, Photometrics and Hamamatsu.

Image files are stored as 16-bit tiff stacks and can be opened by other programs such as ImageJ, IGOR Pro etc....

For further reading we refer to the `SmartLUX` manual.




24 Software LockIn Extension

24.1 LockIn Configuration Window

For the activation and the configuration of the *LockIn Extension* you have to enable the software lock-in function in the Configuration window (see 5.4.3 on page 47). After switching on the *LockIn*, the LockIn Configuration window opens, which is later on available in the Window menu.



In the POTMASTER manual only the main user functions are explained. If you are interested in a detailed description how the LockIn Extension works or what might be the best approach to determine membrane capacitances in different recording modes we refer to the PATCHMASTER Tutorial: "Capacitances Measurements using the LockIn Extension".



LockIn Mode: Set a *LockIn* Mode. You have these possibilities:

- Off: No *LockIn* mode is set.
- On: The *LockIn* mode is turned on.
- Phase Range: 0..180: Limits the phase to a range of 0° to 180°.
- Phase Range: 0..360: Allows to use the full phase range.

Phase Shift: The user can specify a *Phase Shift* to be applied to the residual admittance before the equivalent circuit parameters are calculated. This *Phase Shift* is only applied to the actual current that is measured and not the component of admittance which is "nulled out" by *C-slow* compensation.



This feature can be used, for example, to "tweak" the phase determined by *Measured* or *Calculated Calibration*. Another example where this would be useful is if data are recorded with a set of "critical parameters" which differs from those that were used when a *Measured Calibration* was performed (this is prevented by the software if an EPC 9 or EPC 10 is used, but can occur for other amplifiers). The phase difference can be determined by performing a new *Measured Calibration* and can be specified as the *Phase Shift* for replay of the recorded data.

When *Manual Calibration* is used, the *Phase Shift* sets the absolute phase of the software lock-in.

Note: Generally the Phase Shift should be left at 0.0°.

Attenuation: It is an amplitude scaling factor for the current signal that can be used to correct the transfer function of the complete system. This attenuation factor is applied to the residual admittance before the equivalent circuit parameters are calculated.

Note: Generally the Attenuation should be left at 1.0°.

When *Manual Calibration* is used the *Attenuation* should be set to the value estimated by the *Measured Calibration*.

Generally, the *Attenuation* might be set to a value unequal to 1 when the transfer function of the system is not ideal, i.e. the sine wave frequency is higher than the current filter frequency or the sine wave frequency approaches the overall bandwidth of the amplifier.

Parent Trace: This option is only available if *Manual Calibration* is selected. Select the corresponding *Trace* (e.g. current or voltage *Trace*) for the calibration, either the *Linked Trace* assigned in the PGF (*Link*) or any other *Trace* (*Trace 1...16*).

Write to Notebook: Enabling this option will print out the *LockIn* results to the Notebook window.

Note: The Write to Notebook option does only work if you acquire at least one LockIn Trace in your PGF.

Points to Average: A global number of points to average can be entered.

Offline Computation - Trace to create: These sections of the *LockIn* window are used for calculating *LockIn* data from pre-recorded current *Traces* (for further details see *The LockIn Traces*, 24.2).

Compute LockIn of Marked Targets: Computes the *LockIn Traces* selected in *Offline Computation - Trace to create* of the marked targets in the in the *Replay* tree.

Default Y-ranges: The user has to define *Default Y-ranges* for the resulting *LockIn Traces*. POTMASTER does not know the range of the forthcoming data because this depends on cell size. Thus, the user has to provide a default range (e.g. 0 to 40 pF for the C_m *Trace* would be a reasonable estimate).

SetDefaults: Resets the *Default Y-ranges* to the default values.

Skip: A global number of cycles to skip can be entered.



The global setting Skip, when activated, overrules the setting made in the Wave Parameters of the PGF.

24.2 The LockIn Traces

The *LockIn Extension* provides several output *Traces* which can be selected in the *AD-channel* section of the Pulse Generator (10.10.2 on page 117).

The following *Traces* which are independent from the *LockIn* mode are available:



LockIn_Phase: The phase angle between stimulus and measured signal.

$$Phase = \frac{180}{\pi} * \arctan abs(B/A)$$

LockIn_Real(Y): The real part of the admittance.

$$\begin{aligned} \text{Voltage Clamp: } Real(Y) &= A \text{ (Siemens).} \\ \text{Current Clamp: } Real(Y) &= \frac{A}{A^2+B^2} \text{ (Ohm).} \end{aligned}$$

LockIn_Imag(Y): The imaginary part of the admittance.

$$\begin{aligned} \text{Voltage Clamp: } Imag(Y) &= B \text{ (Siemens).} \\ \text{Current Clamp: } Imag(Y) &= \frac{B}{A^2+B^2} \text{ (Ohm).} \end{aligned}$$

LockIn_Admit(Y): The admittance (Siemens).

$$\begin{aligned} \text{Voltage Clamp: } Admit(Y) &= \sqrt{A^2 + B^2}. \\ \text{Current Clamp: } Admit(Y) &= \frac{1}{\sqrt{A^2+B^2}}. \end{aligned}$$

LockIn_Real(Z): The real part of the impedance.

$$\begin{aligned} \text{Voltage Clamp: } Real(Z) &= \frac{A}{A^2+B^2} \text{ (Siemens)} \\ \text{Current Clamp: } Real(Z) &= A \text{ (Ohm).} \end{aligned}$$

LockIn_Imag(Z): The imaginary part of the impedance.

$$\begin{aligned} \text{Voltage Clamp: } Imag(Z) &= \frac{B}{A^2+B^2} \text{ (Siemens)} \\ \text{Current Clamp: } Imag(Z) &= B \text{ (Ohm).} \end{aligned}$$

LockIn_Imp|Z|: The impedance (Ohm).

$$\begin{aligned} \text{Voltage Clamp: } Imp(Z) &= \frac{1}{\sqrt{A^2+B^2}}. \\ \text{Current Clamp: } Imp(Z) &= \sqrt{A^2 + B^2}. \end{aligned}$$

Abbreviations:

- A = Real part of the admittance.
- B = Imaginary part of the admittance.
- I_{dc} = DC current.
- E_{rev} = Reversal potential.
- V_h = Stimulus voltage.
- f_{sine} = Stimulation frequency.

A detailed description of all terms can be found in the literature (see reference list of the PATCHMASTER Tutorial: "Capacitances Measurements using the LockIn Extension").




24.3 Capacitance Measurements - Step by Step



The following section was made for PATCHMASTER and HEKA EPC 9 or 10 amplifier. If you are using CHARTMASTER or POTMASTER several modifications may be necessary!

Here, we make our first capacitance measurement using a model circuit and the `LockIn Extension`. We used an EPC 10 Single for this tutorial, but most settings are identical for EPC 10 Double, Triple or Quadro amplifiers and for EPC 9 Single, Double and Triple amplifiers. We will explicitly mention the particular amplifiers, where it is required.

24.3.1 Activating the LockIn Extension



By default, the `LockIn Extension` of POTMASTER is inactive. To activate the extension, we have to open the `Hardware` tab of the `Configuration` window and turn on the `LockIn`. This step only has to be done the first time one uses `LockIn`. The setting is stored in the POTMASTER configuration file `Potmaster.set`.

24.3.2 LockIn Configuration

Immediately after switching on the `LockIn Extension`, the `LockIn Configuration` window comes up. Please set the `LockIn` mode to `Sine + DC`. This is the recommended mode, if you are using an EPC 10 or EPC 9 amplifier. At the moment, no other settings need to be made. We will come back to the `LockIn Configuration` window later.



If you want LockIn calculations to be written to the Notebook, Write to Notebook should be checked and at least one LockIn trace has to be selected as an AD input channel in the PGF.

LockIn Configuration

LockIn Mode: **Sine + DC**

Phase Range: 0 .. 180
☐ Phase Range: 0 .. 360

Calibration Mode: **Calculated**

Phase Shift: **0.0°**

Attenuation: **1.000**

Parent Trace: **Linked Trace**

PL-Phase: **0.0°** **Compute**

Calib. Sequence: **IV**

Perform Measured Calibration

☐ Write to Notebook

Points to Average: **Off**

Offline Computation - Traces to create:

☐ Real(Y) ☐ Real(Z) ☒ CM ☐ DC
☐ Imag(Y) ☐ Imag(Z) ☐ GM ☐ CV
☐ Admit(Y) ☐ Imp(Z) ☐ GS ☐ GP
☐ Phase ☐ Sine Average

Compute LockIn of Marked Targets

Compute CV + GP from Real + Imag. Trace

Default Y-ranges: **Set Defaults**

Real(Y)	200.0n	Real(Z)	1.000
Imag(Y)	200.0n	Imag(Z)	1.000
Admit(Y)	200.0 nS	Imp(Z)	1.000 Ohm
Phase	180.0°		
CM	40.00 pF	DC	4.000 nS
GM	4.000 nS	CV	40.00 pF
GS	400.0 nS	GP	400.0 nS
☐ V-rev	0.000 V	☐ Skip	0

Note: The Default Y-ranges section can be used to set reasonable display ranges according to the expected values.

24.3.3 Creating a PGF sequence

The next step is to prepare a PGF sequence, that can be used for C_m measurements. Please open the Pulse Generator window. We could either modify an existing sequence or create a new one from the scratch. Let us create a new sequence this time. A click on an empty button in the sequence pool will create a new sequence and you are prompted to enter a name for the new sequence. We name the sequence we want to create "LockIn".

POTMASTER automatically creates a new and very simple protocol: it just consists of one *Constant* segment with a duration of 10 ms. 10 ms is quite short, therefore we will increase the duration to 100 ms. Furthermore, we have to change the *Segment Class* to *Sine*, since at least one sine wave segment is required for the *LockIn*. Of course, one can also use other segment classes (Constant, Ramp, Square) in the same sequence, for example to define a depolarizing voltage step at the beginning of the *Sweep*. But *LockIn* data can only be calculated from *Sine* segments.



Pulse Generator File: DefPgf_v9

Full View Condensed View Cartoon View

1 LockIn 2 3 4 5 6

Pool LOAD MERGE SAVE Name LockIn LIST COPY MOVE UNDO DELETE

Interactive Mode Gap Free Mode

Timing No wait before 1. Sweep Not Triggered

No of Sweeps 1 Use Durations

Sweep Interval 0.00 s StartSeg 0

Sample Interval 50.0 μ s (20.0kHz) StartTime 0.00

Checking EXECUTE

Sweep Length Total 100.0 ms 2000 pts

Channel Length Stored 100.0 ms 4000 bytes

Stimulus 100.0 ms 2000 pts

1	DA	Unit	Stimulus -> DA	Leak	AD	Unit	Link	Compr.	Points	Store	Zero	Leak	
Ch-1	Stim-DA	V	StimScale	☐	Imon2	A	1	1	C	2000	☒	0	No Leak
---	off		absolute voltage	☐	off				C	---	☐	---	No Leak
---	off		absolute voltage	☐	off				C	---	☐	---	No Leak
---	off		absolute voltage	☐	off				C	---	☐	---	No Leak

Segments ☒ Store 1 ☐ Store 2 ☐ Store 3 ☐ Store 4 ☐ Store 5 ☐ Store 6

Segment Class	Sine	Constant							
Voltage [mV]	holdV-memb	valu	---	valu	---	valu	---	valu	---
Duration [ms]	val 100.00	val	---	val	---	val	---	val	---
V-incr. Mode	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase
V-fact./incr. [mV]	1.00 0	---	---	---	---	---	---	---	---
t-incr. Mode	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase	Increase
t-fact./incr. [ms]	1.00 0.00	---	---	---	---	---	---	---	---

Common Timing

Voltage Clamp

Filter Factor

Analysis: Edit

Rel X-seg 1

Rel Y-seg 1

Draw: Active Channel, all Sweeps Delay: DA 0.00 s AD 0.00 s

V-membrane [mV] (display)

0 ☐ Set Last Seg. Amplitude

Leak Pulses

No of Leaks 0

Leak Delay 10.0 ms

Leak Size 0.250

Leak Hold [mV] ---

Leak Alternate

Alt. Leak Average

wait = abs. hold

p1	p2	p3	p4	p5	p6	p7	p8	p9	p10
0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000

Traces 1

The *Sample Interval* is set to 50 μ s. This corresponds to a sampling rate of 20 kHz. Please keep that sampling rate in mind. We will need that information later on. For the calculation of the *LockIn* parameters, we need to know the whole-cell conductance. The conductance can only be measured, if the holding potential is not "0". Therefore, you should not forget to enter a holding potential for the *Sine* segment. One can either enter a fixed value, for example "-70 mV" or set *Voltage* to "hold". If *Voltage* is set to "hold", the holding potential from the Amplifier window is used.

If a *Sine* segment exists in the PGF sequence an additional button (**Sine Wave**) appears in the dialog. This button can be used to modify the parameters (frequency, amplitude,...) of the sine wave segments in a protocol (see *Sine Wave Parameters*, 10.9.1 on page 104).

In this dialog, you can decide, if the *Sine* segments in the sequence should be used for *LockIn* measurements (Use as LockIn Sinewave) or just as a simple sine wave stimulus (Use as Simple Sinewave: common/separate frequency). Of course, we want to use it as *LockIn* sine wave.

Sinewave Parameters

Use as LockIn Sinewave

Peak Ampl. [mV]	10.0	value
Requested Freq.	800.0 Hz	
Actual Frequency	800.0 Hz	
Points / Cycle	25	
Cycles to Skip	0	Checking
Cycles to Average	1	Cancel
Total Cycles	80	Done
V-reversal [mV]	0.0	

Now, we increase the *Requested Frequency*. A frequency of 800 Hz is quite good for most approaches. All segment durations must be integral multiples of the sine wave period. Therefore the *Actual Frequency* might be slightly different from the *Requested Frequency*. An appropriate value is calculated by the software and cannot be changed.

Please note, that the *Sample Interval* in the PGF sequence is adopted, upon changing the setting for *Requested Frequency* or *Points/Cycle*. It is intuitive that for higher frequencies higher sampling rates are required. If you want to reduce the sampling rate, you can decrease the number of *Points/Cycle*. We will do that and set *Points/Cycle* to "25".



The number of *Points/Cycle* should never be below 10, otherwise the accuracy of the C_m measurement is dramatically reduced.

The setting of 10 mV for the sine wave amplitude is fine and a *V-reversal* of 0 mV is also correct for our model circuit. The *V-reversal* setting is actually not very critical if G_m is low. A value of zero is often used in the common situation where G_m is low and the actual reversal potential is unknown. Now, we can go back to the PGF sequence.

Since we decided to use the *Sine* segment in the sequence for *LockIn* measurements, the entry *Stimulus* → *DA* in the PGF sequence has automatically been changed from "StimScale" to "StimScale, LockIn" upon closing the Sinewave Parameters dialog. This *LockIn* flag is important. Without that flag, no C_m data can be calculated.

1	DA	Unit	Stimulus -> DA	Leak
Ch-1	Stim-DA	V	StimScale, LockIn	☐
---	off		absolute voltage	☐
---	off		absolute voltage	☐
---	off		absolute voltage	☐

The *AD input* is set to "Imon-2", the current input of your amplifier. This is undoubtedly a good idea, however, we need an additional *Trace* to store the C_m value. A click in an empty AD field opens a long list of available inputs. What we need is a *LockIn_CM* Trace. You will see, that the compression factor is automatically set to "25" (the value, we have set for *Points/Cycle* in the Sinewave Parameters dialog). That means, 25 data points are in one cycle and one C_m value is calculated from each cycle, therefore you will get 80 C_m data points in that 100 ms Sweep.

AD	Unit	Link	Compr.	Points	Store	Zero	Leak
Imon2	A	1	1	C	2000	☒	No Leak
LockIn CM	F	1	25	C	80	☒	No Leak
off		---	---	C	---	☐	No Leak
off		---	---	C	---	☐	No Leak

Please note, that *Link* is set to "1". This tells POTMASTER to use the first *Trace* (the current *Trace*) for calculating the *LockIn* information. This point is not crucial as long as you have only one current *Trace* and one *LockIn Trace*,



however, it is important if you should have more than one current *Trace* (e.g. if you are using a Double or Triple amplifier).

Store is active for both *Traces*. It would also be possible to store only the *LockIn* information and to discard the current data. However, in most cases, it is reasonable to store both *Traces*.

24.3.4 Amplifier Settings

Now it is time for the experiment. Switch the model circuit into the "10 M" position to simulate a 10 M Ω pipette that is open to the bath solution.


In the Amplifier window you can correct pipette offset potentials by adjusting the V_0 value or you can alternatively click on the *Auto V_0* button to let POTMASTER do this correction automatically for you. The same is done by calling the protocol "SETUP".

Now simulate a pipette sealed to the membrane by switching the model circuit into the middle position. Make an automatic fast capacitance cancellation by clicking on the *Auto C-fast* or the "SEAL" protocol button.


To break into the cell, set the switch of the model circuit to the "0.5 G" position. Make an automatic slow capacitance cancellation by clicking on the *Auto C-slow* or "WHOLE-CELL" protocol button. With the *V-membrane* control change the pipette holding potential to "-70 mV". Now we are ready to run the PGF sequence we defined before.

24.3.5 Running the PGF sequence

In POTMASTER, PGF sequences can be run from the Pulse Generator directly (*Execute* button) or from the Control window. We click on the *LockIn* button to start the sequence.



The results are displayed in the Oscilloscope. As defined in the PGF sequence, you can see the current *Trace* (black) and a second *Trace* (red) with the *LockIn* data.




Note: Change the display labeling to Grid + Values in order to display the axis scaling.

By default, all data points are connected by lines. If you want to see the individual data points, you can modify the *Trace Properties* in the *Display* menu. The MC-10 model cell circuit in the "0.5 G" position has a capacitance of 20-22 pF. The calculated C_m values are well within this range which can be seen in the *Notebook* window (only if *Write to Notebook* is activated in the *LockIn* Configuration).

24.3.6 Analysis

In the *Oscilloscope*, the *LockIn* information is displayed with a high time resolution. Even fast membrane capacitance changes can be observed in that way.

However, in some experiments, the expected C_m changes are in the range of several seconds or minutes. In that case, we should use the *Analysis* to monitor the changes in membrane capacitance. For that purpose, it is necessary to modify the *LockIn* PGF sequence. We will increase the *No of Sweeps* from 1 to 100. That is all we have to do for the moment.




Click on an empty button to define a new *Analysis Method*. We name the new analysis "LockIn". We decline the question if we want to copy the settings from an previous method. Again we will start from the scratch. Further, you have to do the following modifications:

- **Graph Positions:** Activate checkbox "1" in the row "Graphs in Window 1". This enables the display of the *Graph 1* (which will be defined below) in the Analysis Window 1. Enable "Use Selected Method" to make sure that the now specified *Analysis Method* is used.
- **Analysis Functions:** Click on an empty field and the *Function Type* dialog will open. First, select *Sweep Count* for the X-axis analysis then click again on an empty button field and select *LockIn_CM* for Y-axis analysis. Please take care that *Trace #* for the proper C_m values has to be *Trace 2*.
- **Analysis Graphs:** Switch on the *Graph 1*.
- **Graph Entries:** Activate the checkbox "1" and select *SweepC* (Sweep Count) for X-axis and *LI_CM_2* (CM for trace 2) for Y-axis.
- **Scale Axis:** Select *Auto Swp.* for X-axis and *Fixed* for Y-axis. 10 to 30 pF (10 p / 30 p) are sufficient for our model cell.
- **Analysis Window:** If not visible, go to the *Windows* menu and activate Analysis Window 1.

Function Type						
Timing	Measurements	LockIn	Trace Param.	Math	Trace	Power Spectra
☒ Sweep Count	☐ Extremum	☒ LockIn_CM	☐ Trace Count	☐ Equation	☐ Trace	☐ log(Frequency)
☐ Analysis Index	☐ Maximum	☐ LockIn_GM	☐ C-slow	☐ Y(x): y at pos = x	☐ Equation	☐ Density
☐ Time	☐ Minimum	☐ LockIn_GS	☐ R-series	☐ Constant	☐ Q = Integral	
☐ Timer Time	☐ Extr. Amplitude	☐ LockIn_Phase	☐ Rs-value	☐ a + b	☐ 1 / (trace)	Histogram
☐ Series Time	☐ Min. Amplitude	☐ LockIn_Freq	☐ Leak Comp.	☐ a - b	☐ 1 / (Q)	☐ Histogram Ampl
☐ Real Time	☐ Max. Amplitude		☐ M-conductance	☐ a * b	☐ ln (trace)	☐ Histogram Bins
	☐ Time to Extremum	AP Analysis	☐ Cell Potential	☐ a / b	☐ ln (Q)	
Stim. Properties	☐ Time to Maximum	☐ Baseline	☐ Seal Resistance	☐ a in b	☐ log (trace)	
☐ Amplitude	☐ Time to Minimum	☐ AP Amplitude	☐ Int. Solution	☐ abs	☐ log (Q)	
☐ Duration	☐ Time to Threshold	☐ Time to AP Ampl	☐ Int. Sol. Value	☐ log	☐ dt = Differential	
☐ Rel. Seg. Time	☐ Threshold Ampl.	☐ Repol Ampl	☐ Ext. Solution	☐ sqrt	☐ Trace x-axis (time)	
☐ Abs. Seg. Time	☐ Thres. Crossings	☐ Time to Repol Ampl	☐ Ext. Sol. Value	☐ arctan	☐ Stimulus	
☐ Scan Rate	☐ Mean	☐ Rise Time	Sweep Param.	☐ 1/a		
	☐ Variance	☐ Up Slope	☐ Temperature	☐ 1/log		
	☐ Integral	☐ Rise Time Delay	☐ Pip. Pressure	☐ 1/sqrt		
	☐ Anodic Q	☐ Decay Time	☐ Digital-In	☐ 1/arctan		
	☐ Cathodic Q	☐ Down Slope	☐ User_1			
	☐ Reversal	☐ Decay Time Delay	☐ User_2			
	☐ Slope	☐ Decay Tau				
	☐ Intercept					
	☐ Tau					

If we now run the *LockIn* PGF sequence, one *LockIn* data point per *Sweep* is plotted in the Analysis window. This single point is the mean value from all C_m values within a *Sweep*.





25 Spectroscopy Extension

25.1 Introduction

The `Spectroscopy Extension` allows to measure the transfer functions of a system. The complex transfer function can be presented as two real traces: *Magnitude* (or *Gain*) and *Phase*. E.g. if we send a sine wave stimulus of amplitude 1 into a system and we measure as response a sine wave of the same frequency but amplitude of 0.5, then we would get a *Magnitude* or *Gain* of 0.5.

In case the system converts a voltage stimulus to a current signal, we can interpret the *Magnitude* (or *Gain*) as *Admittance* (*Y*) or *Impedance* (*Z*).

The `POTMASTER LockIn Extension` performs this type of analysis at a given single frequency. In order to obtain a complete spectrum, repetition of the measurement at different frequencies is required. This ends up in a very elaborate procedure. The `Spectroscopy Extension`, however, uses a stimulus containing a set of multiple frequencies, allowing to calculate a complete spectrum from a single measurement.

25.1.1 Chirp Wave Forms

The Pulse Generator of POTMASTER provides three different chirp wave forms:

Linear Chirp:

In a *Linear Chirp*, the frequency $f(t)$ varies linearly with time:

$$f(t) = f_0 + kt \quad (25.1)$$

where f_0 is the starting frequency (at time $t = 0$), and k is the rate of frequency increase or chirp rate. The corresponding time-domain function for a sinusoidal *Linear Chirp* is:

$$x(t) = \sin[2\pi(f_0 + \frac{k}{2}t)t] \quad (25.2)$$

Exponential Chirp:

In an *Exponential Chirp*, the frequency of the signal varies exponentially as a function of time.

$$f(t) = f_0 k^t \quad (25.3)$$

where f_0 is the starting frequency (at $t = 0$), and k is the rate of exponential increase in frequency. Unlike the *Linear Chirp*, which has a constant chirp rate, an *Exponential Chirp* has an exponentially increasing chirp rate. The corresponding time-domain function for a sinusoidal *Exponential Chirp* is:

$$x(t) = \sin[\frac{2\pi f_0}{\ln(k)}(k^t - 1)] \quad (25.4)$$




Spectroscopy Chirp:

A chirp wave form created by inverse *Fourier Transform* from a spectrum with frequencies increasing with t^2 and constant amplitude. The `Spectroscopy` Extension uses this stimulus wave form as input only.

25.1.2 Transfer Function Ratio

The transfer function (H) of two signal processing units in series is given by:

$$H(A + B) = H(A) \otimes H(B) \quad (25.5)$$



In case one is interested the transfer function of an individual signal processing unit, the "ratio" of the transfer function can be calculated:


$$H(B) = \frac{H(A + B)}{H(A)} \quad (25.6)$$

The `Spectroscopy` Extension provides a so-called *Correction Mode* to calculate the *Transfer Function Ratio* of the recorded response (nominator) and the reference response (denominator).


In some application a reference element (e.g. calibrated resistor) is used for a calibration measurement of the system. In this case the value of the reference element has to be taken into account when calculating the transfer function of the measuring system without reference element. We offer the option to use the measured resistance or a given resistance for this correction.

25.2 Activating the Spectroscopy Extension

The `Spectroscopy` Extension is activated in the Configuration window on the tab *Hardware*.



Once the extension is turned *On*, the Spectroscopy window can be opened via the Windows menu list in POTMASTER.



Correction Mode: It is possible to correct the spectra during online and offline calculation with respect to another chirp response (see Transfer Function Ratio, 25.1.2 on the facing page).

- In case a simultaneous recording of the reference chirp is possible:
Under certain circumstances it might be possible to record the reference *Trace* simultaneously with the chirp response. In this case you can select the respective *Trace* which contains the reference chirp response in the *Correction Mode*.
- In case the reference chirp is recorded in advance:
The reference chirp has to be copied in one of the four *Trace Buffers* and in the *Correction Mode* the respective *Buffer* has to be selected.

If a correction *Trace* is selected, the admittance of the recorded response is divided by the normalized magnitude of the reference. In addition, the phase of the reference is subtracted from the phase of the recorded response.

Input Source: Defines the data type from which the spectra are calculated. In case a *Correction Mode* is used, the type of input data for the acquired data and the reference have to be the same.

Reference Element Correction: Must be used in case the measurement system has been calibration with respect to a reference element (e.g. a 10 M Ω resistor), which is removed and replaced by the cell under study. Two options are available:

1. Measured Resistance: The resistance of the reference element is measured at the lowest frequency.
2. Given Resistance: A given resistance can be used to normalize the calibration measurement.

In case *Reference Element Correction* is not activated the unscaled raw data of the reference transfer function are used. This method usually applies only if it is possible to measure from two different locations in the signal processing path.

Log Compression: Often the results of the chirp analysis are plotted on a logarithmic frequency scale. In this case the higher frequencies contain much more measuring points than the lower frequencies. For better display and post analysis (e.g. fitting of the spectra) a equidistant scaling on the logarithmic frequency axis is of advantage.



Therefore we offer the option of a logarithmic compression. The *Compression Factor* (usually a value between 1 and 2) defines the strength of the compression. When logarithmic compression is used the frequency axis is scaled in units of logarithm to basis 10.

Offline Computation - Traces to create: You can mark by using the checkboxes which *Traces* (*Admit(Y)*, *Real(Y)*, *Real(Z)*, *Phase*, *Imag(Y)*, *-Imag(Z)*, *Average*, *Imp(Z)*) should be created during re-calculation of the spectra.

Parent AD-Trace: Select the corresponding *Trace* for computing spectroscopy, either *All* or any other *Trace* (Trace 1...16).

Compute Spectroscopy of Marked Targets: Recalculates and generates the selected *Traces* (select and mark the target (*Sweeps* or *Series*) first).

Default Ranges: For the various spectroscopy *Traces* the default Y-ranges for display in the Oscilloscope window are defined. E.g. a Y-range of 1 S (Siemens) defines that at scaling of 1 and offset of 0, the Oscilloscope shows a range from -1 S to +1 S.

25.2.1 The Spectroscopy Traces

Chirp_Avg: Average of the chirp response. In case multiple *Chirp* segments are used in one *Sweep*, the *Chirp_Avg* contains the average of all *Chirp* segments.

Chirp_Phase: The phase angle between stimulus and response versus frequency. The default unit is degree(°). This is a frequency based *Trace* (X-unit is Hertz).

Chirp_Admit(Y): Contains the magnitude of the spectrum normalized to the stimulus amplitude. In case of a recorded current response this parameter can be interpreted as admittance. The default unit is Siemens (S). This is a frequency based *Trace* (X-unit is Hertz).

Chirp_Real(Y): The real part of the admittance. This is a derived frequency based *Trace* (X-unit is Hertz), calculated as follows:

$$Real(Y) = \frac{Admittance}{\sqrt{1 + \tan(Phase)^2}} \quad (25.7)$$

The default unit is Siemens (S).

Chirp_Imag(Y): The imaginary part of the admittance. This is a derived frequency based *Trace* (X-unit is Hertz), calculated as follows:

$$Imag(Y) = \tan(Phase) * Real(Y) \quad (25.8)$$

The default unit is Siemens (S).

Chirp_Real(Z): The real part of the impedance. This is a derived frequency based *Trace* (X-unit is Hertz), calculated as follows:

$$Real(Z) = \frac{Real(Y)}{Real(Y)^2 + Imag(Y)^2} \quad (25.9)$$

The default unit is Ohm (Ω).

Chirp_Imag(Z): The imaginary part of the impedance. This is a derived frequency based *Trace* (X-unit is Hertz), calculated as follows:

$$-Imag(Z) = \frac{Imag(Y)}{Real(Y)^2 + Imag(Y)^2} \quad (25.10)$$

The default unit is Ohm (Ω).



Chirp_Imp|Z|: The impedance. This is a derived frequency based *Trace* (X-unit is Hertz), calculated as follows:

$$Imp|Z| = \frac{1}{\sqrt{Real(Y)^2 + Imag(Y)^2}} \quad (25.11)$$

The default unit is Ohm (Ω).

25.3 Setting up a Spectroscopy Acquisition



The following section was made for PATCHMASTER and HEKA EPC 9 or 10 amplifier. If you are using CHARTMASTER or POTMASTER several modifications may be necessary!

Create a new PGF sequence with one *Chirp* segment as stimulus. When at least one segment contains a chirp wave form then a button *Chirp Wave* appears in the top right section of the *Pulse Generator* window.

25.3.1 The Chirp Wave Dialog

A detailed description of the available chirp wave parameters can be found in *Chirp Wave Parameters*, 10.9.3 on page 109.


25.3.2 Parametrization of the Chirp Stimuli

Possible Frequency Range: The maximal possible frequency in a chirp wave is half the sampling frequency (*Min. Points / Cycle* = 2).

- **Linear and Exponential Chirp:** The *Start* and *End Frequency* can be directly entered in the *Chirpwave Parameters* dialog. According to *Start* and *End Frequency* and given segment duration the rate of frequency increase (*k*) and the *Min. Points / Cycle* are calculated.

- **Spectroscopy Chirp:** *Start Frequency* and *End Frequency* can not be entered directly. Both parameters are defined by the *Sampling Frequency*, *Segment Duration*, and the *Min. Points / Cycle*. The *End Frequency* is given by *Sampling Frequency* divided by *Min. Points / Cycle*. The *Start Frequency* then depends on the duration of the segment. The longer the segment the smaller the starting frequency of the chirp.





Useful Frequency Range: In the preceding section we have discussed the possible frequency range in a chirp stimulus. Correct analysis of the chirp response, however, has to obey some additional boundary conditions. The most important rule is to obey the *Sampling Theorem* or *Nyquist Theorem*! That means that the response has to be low pass filtered to prevent aliasing artifacts. The filtering can be either done by the system under study itself or by additional filters. When using filters with steep cut-off (e.g. 8-Pole filter) unbiased analysis can be extended to about 80 % of the possible frequency range.

25.3.3 Adding Chirp Analysis Traces

Once you have configured the chirp stimulus you can add multiple chirp analysis *Trace* to the AD selection in the Pulse Generator.

Dig-15
Chirp_Avg
Chirp_Phase
Chirp_Admit(Y)
Chirp_Real(Y)
Chirp_Imag(Y)
Chirp_Real(Z)
Chirp_Imag(Z)
Chirp_Imp Z
LockIn_Avg

25.3.4 Minimizing Swing-In Effects


In order to minimize swing-in effects, two methods can be used:

- 1. Use at least two Chirp Segments: We duplicate the *Chirp* segment and set the *StartSeg* to "2". In case of very long chirp stimuli this method is very time consuming.


***Note:** When you have more than one Chirp segment, please deactivate the checking routines during editing the parameters of the segments. Once you have finished the edition, turn the Checking on again.*

In the following screenshot you see a sample Pulse Generator sequence for a chirp acquisition.






2. Use of a Pre-Chirp Segment: In order to reduce the time for swing in to a minimum, a *Constant* segment of fixed length can be defined before the *Chirp* segment. When the option *Pre-Chirp* is activated in the Chirpwave Parameters dialog, then this segment is automatically filled with appropriate swing-in stimulus data. Please note that the *Pre-Chirp* segment should not be recorded (use *Start Segment "2"*) to ensure accurate chirp analysis.



In the following screenshot the swing in signal of a *Pre-Chirp* segment is shown. The first 5 ms of the stimulus are the *Pre-Chirp* segment which provides an optimal swing-in of the chirp segment.




25.4 Display of Spectra in the Oscilloscope Window


The chirp analysis *Traces* are stored in the Replay window.

1	Imon-1	1
	Vmon-1	2
	Phase	10
	Admit(Y)	11
	Real(Y)	12
	Imag(Y)	13
	Real(Z)	14
	-Imag(Z)	15
	ImpZ	16

Their X-unit is Hertz and the number of points differs from the raw data *Trace*. Since raw data *Trace* and spectra have different X-axis scaling it is difficult to show them at the same time on appropriate scaling in the Oscilloscope, which supports only one X-scaling. By default you will see the raw data *Trace* in the Oscilloscope, since the Oscilloscope takes its X-scaling from the first *Trace* in the *Sweep*.



In order to show also the chirp analysis *Traces* in the Oscilloscope, please select *Fixed X-width* from the *Display* menu and enter as *X-max* (panel to the right in the Oscilloscope window) the maximum frequency in the spectrum (e.g. 12k, what stands for 12 kHz).



Trace 10 (orange) and Trace 11 (green) show the phase and the admittance, respectively.

26 Appendix I: File Overview

26.1 File Types

Figure 26.1: The different file types of PATCHMASTER, POTMASTER AND CHARTMASTER

File Type	Extension	Default File	Note
Amplifier file (only generated when multiple amplifiers are active (EPC 10 USB Double, Triple and Quadro))	*.amp		Chapter 27.1 on page 273
Data file (includes the measured data)	*.dat		
Dialog setting for the dialog of windows (if no dialog file exists, the software will start with its own default values.)	*.dia		Chapter 2.3 on page 10; (*.dia files are always named after the window that is saved, e.g., Oscilloscope.dia.)
EPC files for amplifier-dependent calibration parameters; SCALE for amplifier calibration; CFAST for Cfast calibration. Starting with EPC 10 amplifiers of revision "N" the scale file is stored on the EEPROM of the amplifier.	*.epc	SCALE-XXXXXX.epc	
Key file for key commands (includes all key commands)	*.key	Potmaster.key	Chapter 3 on page 13
Macro file for macros; not supported anymore starting with version 2.40	*.mac		Chapter 11 on page 131
Markers file for all marker information	*.mrk		Chapter 20 on page 227
Method file for Protocol Editor Methods	*.mth		Chapter 12 on page 157
Online file for Analysis settings (can include more than one <i>Analysis Method</i>)	*.onl	DefAnal.onl	Chapter 13 on page 179
PG3 files for amplifier-dependent calibration parameters SCALE for potentiostat calibration	*.pg3	SCALE-XXXXXX.pg3	Only for Potentiostats (PG Series)
Stimulus file for the Pulse Generator settings	*.pgf	DefPgf.pgf	Chapter 10 on page 99
Protocol file for a protocol pool (can include more than one protocol)	*.pro	DefProt.pro	Chapter 11 on page 131
Equation file for storing equation strings	*.txt	Equation.txt	Chapter 21 on page 231
Acquisition parameters file	*.pul		

Continues on next page



File Type	Extension	Default File	Note
Settings file	*.set	Potmaster.set	
Solution file for storing the entries of the Solution Database	*.sol	DefSolutionBase.sol	Chapter 18 on page 221
Filename template file	*.tpl		Chapter Using a Recorded Waveform as Stimulus in the PATCHMASTER Tutorial
Notebook file	*.txt		Chapter 4.7 on page 34



27 Appendix II: Data Format

In this chapter we describe the general structure of the files generated by POTMASTER.

27.1 Data Files

POTMASTER generates up to 8 files, when a data file is created:

1. The *Raw Data File* (*.dat).
2. The *Marker File* (*.mrk).
3. The *Method File* (*.mth).
4. The *Analysis File* (*.onl).
5. The *Stimulation Sequence File* (*.pgf).
6. The *Acquisition Parameters File* (*.pul).
7. The *Solution File* (*.sol).
8. The *Notebook File* (*.txt).

If *Make Bundle Files* is checked in the Configuration window (5.5.2 on page 52), then all files will be merged into one single bundle file with the file extension *.dat.

Except for the *Raw Data File* and the *Notebook File*, all other files have a "Tree" structure. The entire trees are kept in memory, whereas the raw data *Traces* are always loaded from disk, when needed.

27.1.1 Raw Data File

This *Raw Data File* has an optional header, followed by a continuous data stream. Each data point is a 16-bit signed integer or 32-bit IEEE real as defined in the respective PGF template. When a *Sweep* is stored, POTMASTER stores the various *Traces* (if available) sequentially as defined in the respective PGF template. *Traces* with leak pulses are stored leak subtracted. The leak *Traces* are normally stored after the *Traces* themselves.

Very long traces with "continuous" segments may be stored as interleaved blocks. The size of such a block is given in the "InterleaveSize" field (in bytes) of the *Trace* record, the distance to the next block in the "InterleaveSkip" field (in bytes).

The structure of the *Raw Data File* (*.dat) is defined by the file `DataFile_v9.txt`.



27.1.2 Markers File

It contains the *Marker* records. The structure of the *Markers File* (*.mrk) is defined by the file `MarkerFile_v9.txt`. It has a tree structure:

Table 27.1: Markers File Description

Record	Description
Root	Version information
Marker	Description of one marker event

27.1.3 Protocol Methods File

It contains possibly used protocol method records. The structure of the *Protocol Methods File* (*.mth) is defined by the file `MethodFile_v9.txt`.

27.1.4 Analysis Methods File

It contains the assigned analysis method for every *Series*. The structure of the *Analysis Methods File* (*.onl) is defined by the file `AnalysisFile_v9.txt`. It has a tree structure:

Table 27.2: Analysis Methods File Description

Record	Description
Root	Version information
Method	Description of one <i>Analysis Method</i>
Function	Description of one <i>Analysis Function</i>

27.1.5 Stimulation Template File

Stores the stimulation protocol. The structure of the *Stimulation File* (*.pgf) is defined by the file `StimFile_v9.txt`. It has a tree structure:

Table 27.3: Stimulus Template File Description

Record	Description
Root	Version information
Stimulation	Description of an ensemble of pulse patterns; e.g., I-V curve
Channel	Combines the definition for one output (DAC) and one input (ADC) <i>Trace</i>
Segment	Individual segment of a pulse pattern

Stimulation Files can be loaded into the *Pulse Generator*. In fact, the *Pulse Generator Files* for the stimulation protocols used during the experiments have the same data structure as the PGF files, which belong to the recorded data. In this way it is possible to exactly repeat an experiment by using a copy of a PGF file as *Pulse Generator File*.



27.1.6 Acquisition Parameters File

Stores parameters, such as e.g. *Gain*, *Capacitance*.... The pointer to the data stored in the *Raw Data File* is also contained in this file. The structure of the *Acquisition Parameters File* (*.pul) is defined by the file `PulsedFile_v9.txt`. It has a tree structure:

Table 27.4: Acquisition Parameters File Description

Record	Description
Root	Version information
Group	Larger section of an experiment; e.g., cell or patch
Series	Description of an ensemble of <i>Sweeps</i>
Sweep	Description of a <i>Sweep</i> , i.e. one collection of <i>Traces</i>
Trace	Description of an individual data <i>Trace</i>

A graphical template of the `Pulsed File (Tree)` is shown in the **Replay** window. It contains information necessary to reconstruct the experimental conditions as the data were recorded.

27.1.7 Solutions File

The file is only generated, when the option *Solution Base* is activated. It contains the solution record for every *Series*. The structure of the *Solutions File* (*.sol) is defined by the file `SolutionsFile_v9.txt`. It has a tree structure:

Table 27.5: Solutions File Description

Record	Description
Root	Version information
Solution	Description of one solution
Chemical	Description of one chemical compound

27.1.8 Notebook File


The *Notebook File* is a standard ASCII text file with line breaks.

27.2 The Tree Format

The idea of POTMASTER is to order the data of an experiment in "Trees". The trunk of the tree (*Root*) is the main descriptor of a data file (it could, for example, correspond to one cell or the entire experiments of one day). The next level is the *Group*. This level can be defined by the user to identify data that belong together. An example would be to open a new *Group* for each patch. The *Group* may contain several families of records. Such a family (e.g., records of a current-voltage relationship) is called *Series*. The individual records of a family are called *Sweeps*. Finally, each *Sweep* may be composed of *Traces*. A copy of this data tree is accessible to the user throughout the experiment (so one has an overview of what was recorded, and one can immediately edit the entries (e.g., discard bad records)).



The following is a description of the *Tree* format. An example tree with 5 levels can be diagrammed as follows:



The above tree has five levels:

- 0 : Root
- 1 : Group
- 2 : Series
- 3 : Sweep
- 4 : Trace

The format of a tree stored to a file is as follows:

1. Magic number: 054726565 (hex) = "Tree"
2. Number of levels
3. Level sizes, one per level
4. Tree records, top down, left-to-right. Each record has the format:
 - Record contents
 - Number of children

All of the values (except the record contents) are INT32 values, i.e., 32-bit (4 bytes) values.

Note: Check the record sizes in the file headers. The record sizes may differ from what you are expecting, e.g., because the file has been created by an older program version which used fewer fields than it is currently using, or a newer version with additional fields. **You must use the record sizes stored in the files themselves.**

***Note:** The "Magic2 number will be 065657254 (hex) = "eerT", if the file was written on an operating system with opposite byte ordering (e.g. written under Mac OS, read under MS Windows). In that case, appropriate byte swapping has to be performed (highest → lowest, etc.), when the data is read in.*

A detailed description of the POTMASTER data file format for programmers is available for download from our FTP server: <ftp://server.hekahome.de/pub/FileFormat/Patchmasterv9/>.

27.3 File Template

27.3.1 Filename

In POTMASTER you can use so-called template files for stimulation.

You have the following options how to use the file templates:

Option 1: One template per DA channel common to all Sweeps of a Series

In this case, the name of the template file would be:

[stimulus name]_[channel number].tpl

E.g., if the stimulus name is "IV", then POTMASTER looks for the *Template File* `IV_1.tpl` to be used as *Template File* for the first DA channel of all Sweeps.

These *Template Files* must be in a sub-folder of the folder with the pgf files. The sub-folder must be named identically to the stimulus, e.g., "IV".

Option 2: A different template per DA channel and Sweep

In this case, the name of the *Template File* would be:

[stimulus name]_[sweep index]_[channel number].tpl

E.g., if the stimulus name is "IV", then POTMASTER looks for the *Template File* `IV_1_1.tpl` to be used as *Template File* for the first DA channel of the first Sweep, `IV_1_2.tpl` for the first DA channel of the second Sweep and so on.

These *Template Files* must be in a sub-folder of the folder with the pgf files. The sub-folder must be named identically to the stimulus, e.g., "IV".

27.3.2 Data Format of the File Template

27.3.2.1 Potentiostatic Stimulation or plain DA output

The file should contain one voltage value per stimulus point. POTMASTER fills missing samples with the holding value, if *Apply StimScale* is active, or alternatively with zero. Excessive samples are ignored.

The voltage must be a "short" (4 byte), binary IEEE-floating point format number. All values must be in Volt, i.e., if a voltage of -80 mV has to be output, the required value is -0.080. The total number of samples should be equal to the total number of input samples.



Index

Analysis

- Analysis Graphs
 - Scrolling Bar, 194

- Analysis Methods
 - Scrolling Bar, 184

Input

- X-Trace, 186
- Y-Trace, 185

Analysis Window, 183

Analysis Functions, 185

- $1/(Q)$, 192
- $1/(\text{trace})$, 192
- $1/a$, 192
- $1/\text{atan}$, 192
- $1/\log$, 192
- $1/\text{sqrt}$, 192
- 3D_X, 193
- 3D_Y, 193
- a in b, 192
- $a*b$, 192
- $a+b$, 192
- $a-b$, 192
- a/b , 192
- abs, 192
- Absolute Segment Time, 190
- Amplitude, 189
- Analysis Function Types, 189
- Analysis Index, 189
- Anodic Charge, 190
- atan, 192
- Cathodic Charge, 190
- Change, 185
- Constant, 192
- Cursors, 186
- Delete, 185
- Differential, 192
- Digital-In, 191
- Distribution, 193
- Duration, 189
- Electrode Area, 191
- ElProScan, 193
- Equation, 191, 192
- Extremum, 190
- Extremum Amplitude, 190

- Frequency, 193

- Function Type, 185

- Histogram, 193

- Histogram Amplitude, 193

- Histogram Bins, 193

- Input, 185

- Integral, 190

- Intercept, 190

- List, 185

- $\ln/(Q)$, 192

- $\ln/(\text{trace})$, 192

- LockIn, 191

- LockIn_Amplitude, 191

- LockIn_Frequency, 191

- LockIn_Phase, 191

- log, 192

- $\log/(Q)$, 192

- $\log/(\text{trace})$, 192

- Math, 191

- Matrix-X, 193

- Matrix-Y, 193

- Maximum, 190

- Maximum Amplitude, 190

- Mean, 190

- Measurements, 190

- Minimum, 190

- Minimum Amplitude, 190

- Move, 185

- Name, 185

- New, 185

- OCP, 191

- Operation Options, 187

- Output, 187

- Power Spectra, 193

- $Q=\text{integral}$, 192

- Real Time, 189

- Relative Segment Time, 189

- Reversal, 190

- Scan Rate, 190

- Scrolling Bar, 185

- Series Time, 189

- Slope, 190

- Solution, 191

- Solution Value, 191



- sqrt, 192
- Stim. Frequency, 190
- Stimulation Properties, 189
- Stimulus, 192
- Sweep Parameters, 191
- Tau, 190
- Temperature, 191
- Threshold Amplitude, 190
- Threshold Crossings, 190
- Time, 189
- Time to Extremum, 190
- Time to Maximum, 190
- Time to Minimum, 190
- Time to Threshold, 190
- Timer Time, 189
- Timing, 189
- Trace, 192
- Trace Count, 191
- Trace Parameters, 191
- Trace Time, 192
- User_1,2, 191
- Variance, 190
- X-position, 193
- y at pos = x, 191
- Y-position, 193
- Z-position, 193
- Zero Offset, 191
- Analysis Functions
 - Sweep Count, 189
- Analysis Graphs, 194
 - Auto after a Series, 196
 - Auto after each Sweep, 196
 - Centered, 197
 - Copy, 194
 - Fixed, 196
 - Fixed with last Min/Max, 196
 - Grid,Factor, 196
 - Header, 197
 - Include Zero, 196
 - Labels, 197
 - Max, 195
 - Min, 195
 - Modify Axis, 196
 - Nice Values, 196
 - Normalize, 197
 - On/Off, 194
 - Overlay, 195
 - Position, 196
 - Print, 194
 - Redraw, 194
 - Save to Tree, 194
 - Scale, 195
 - Scale Axis, 195
 - Share X-axis, 197
 - Sort, 197
 - Tics/Direction, 196
 - Units, 196
 - Wrap, 195
 - X, Y, 195
 - Zero Line, 196
- Analysis Methods, 184
 - Assign, 185
 - Automatic Stimulus Control: Assigned, 184
 - Automatic Stimulus Control: Pool, 184
 - Copy, 184
 - Delete, 184
 - Duplicate, 185
 - Load, 184
 - Merge, 184
 - Move, 184
 - Name, 184
 - New, 184
 - No Analysis, 184
 - Save, 184
 - To Pool, 185
 - Use Selected Method, 184
- Cursors
 - Adjust to Position of Function, 186
 - Cursor Bounds, 186
 - Cursor Type, 186
 - X-, Y-seg. Offset, 186
- Graph Entries, 198
 - Axis, 198
 - Connect, 198
 - Same Color as Trace, 198
 - Set Color, 198
 - Size, 198
 - Type, 198
- Graph Positions, 199
- Input
 - Operation, 186
 - Value, 185
- Nomenclature, 183
- Operation Options
 - Average Points, 188
 - Baseline, 188
 - Crossings, 188
 - Equation, 188
 - Fit, 187
 - Histogram Bins, 188
 - Ramps Only, 188
- Output
 - Notebook, 187
 - Result Unit, 187
 - Store in Value, 187
- Stimulus Control, 184



Appendix I: File Overview, 275

File Types, 275

- *.amp, 275
- *.dat, 275
- *.dia, 275
- *.epc, 275
- *.key, 275
- *.mac, 275
- *.mrk, 275
- *.mth, 275
- *.onl, 275
- *.pg3, 275
- *.pgf, 275
- *.pro, 275
- *.pul, 275
- *.set, 275
- *.sol, 275
- *.tpl, 275
- *.txt, 275

Appendix II: Data Format, 277

Data Files, 277

- *.dat, 277
- *.mrk, 277
- *.mth, 277
- *.onl, 277
- *.pgf, 277
- *.pul, 277
- *.sol, 277
- *.txt, 277

Acquisition Parameters File, 279

Analysis Methods File, 278

Markers File, 278

Notebook File, 279

Protocol Methods File, 278

Raw Data File, 277

Solutions File, 279

Stimulation Template File, 278

File Template, 281

Tree Format, 279

Data Format of the File Template, 281

Filename, 281

Calculator Window, 235

Command Line, 235

Equation, 235

Equation Syntax, 236

Comments, 237

Constants, 237

Math Equation, 237, 238

Math functions, 236

Special Case, 237

Equations

Analysis, 237

DA Scaling, 239

Trace Buffer, 239

Values, 240

Equations - Where to Use, 237

Load, 235

Question Mark, 235

Results, 235

Save, 235

Save Equation, 235

Select Equation, 235

Configuration

Hardware

DocuLUX, 52

ElProScan, 53

Photometry, 52

Configuration Window, 45

Batch Communication

Connections, 49

Disable, 48

Enable as Receiver, 49

Enable as Sender, 49

Logging, 49

Polling, 49

Synchronize Files, 49

Display, 58

3D Graph, 59

Fonts and Colors, 58

Grid Digits, 59

Hide Charge Controls, 59

Hide Current Density Controls, 59

Hide PGF-Bar in Control Window, 59

Hide Protocol-Bar in Control Window, 59

Notebook, 58

Notebook Digits, 58

Show Amplifier Protocols, 59

Show Fixed Control Protocols, 59

Show Leak Pulses in PGF Editor, 59

Show Options, 59

Time Digits, 58

Wipe at Start, 58

Files, 55

Ask for Data File, 56

Auto Filename, 56

Batch Path, 55

Data File, 55

Dialog Files, 55

Equations File, 55

Experiment Number, 57

Home path, 55

I-gain Lookup, 56

Make Bundle Files, 56

Maximum File Size Alert, 57

Miscellaneous Settings, 56



- Naming Folders and Files, 56
- Online File, 55
- PGF Pool File, 55
- Protocol File, 55
- Save after Break, 56
- Save Settings File, 56
- Solution Base, 55, 56
- Startup Protocol, 57
- Temporary Path, 55
- Verify Quit, 56
- General, 47
 - Batch Communication, 48
 - Log Compression, 48
 - Max Shown Traces and Values, 47
 - Max. Shown Traces, 47
 - Max. Shown Values, 47
 - Maximum Number of Stimulation Traces x Points, 48
 - Maximum Sample Points, 48
 - Memory Allocation, 47
 - Minimum Wait Time, 48
 - Spectrum Buffer Size, 48
 - User Name, 49
- Hardware, 49
 - ITC-1600, 50
 - ITC-16, 50
 - ITC-18, 50
 - LIH 1600, 50
 - LIH 8+8, 50
 - Amplifier and Digitizer Selection, 50
 - Amplifier Selection, 50
 - Board Type, 50
 - Current In, 51
 - Demo Mode, 51
 - Digitizer Selection, 50
 - E-initial Out, 51
 - Imaging, 52
 - LockIn, 51
 - Lookup Tables, 51
 - Output and Input Channels, 51
 - PG 284, 50
 - PG 285, 50
 - PG 300, 50
 - PG 300 USB, 50
 - PG 330 USB, 50
 - PG 400 Local, 50
 - PG 400 Remote, 50
 - Serial Out, 54
 - SmartLUX, 52
 - Spectroscopy, 52
 - Stim. Scale, 51
 - Telegraphing Amplifier, 50
 - Voltage In, 51
- I/O Control, 60
 - AD scale, 63
 - AD/DA Input Scaling, 61
 - AD/DA Output Scaling, 61
 - Curr. Bandwidth, 64
 - DA scale, 64
 - Default, 60
 - Electrode Area, 64
 - List, 60, 63
 - Name of Value, 64
 - Show AD Values, 64
 - Show AD-channels, 64
 - Show DA Controls, 64
 - Show Digital In, 64
 - Show Digital In/Out, 64
 - Show Digital Out, 64
 - Show Solutions, 64
 - Show Values, 64
 - Source, 60
 - Temperature, 64
 - User Parameters, 64
- Individualize POTMASTER, 46
- Miscellaneous, 68
 - Analysis Maximum Sweeps Redraw, 69
 - Analysis Mode, 68
 - Analysis: Compress Vectors, 69
 - Analysis: Include Headers, 69
 - Analysis: Sweeps | Results, 68
 - Analysis: Write Sweep Index, 69
 - Clear Comment, 69
 - Replay Scroll Rate, 69
- Save, 47
 - Default Settings, 47
 - Default Windows, 47
 - SAVE, 47
 - SAVE AS..., 47
- Trace Assignment, 65
 - AD-Channel, 65
 - Amplifier, 65
 - Digital, 66
 - ElProScan, 67
 - Imaging, 67
 - Leak, 66
 - LockIn, 66
 - Photometry, 67
 - S.W.Voltammetry, 67
 - Spectroscopy, 66
 - Trace Count, 65
 - Virtual, 66
- Configuration Windows
 - I/O Control
 - OCP, 64
- Control Window, 95



- Controlling Data Acquisition, 96
 - Average, 96
 - Break, 96
 - Next, 96
 - Resume, 96
 - Stop, 96
 - Store, 96
 - Wait, 96
- Fixed Control Protocols, 97
- Information, 95
 - Comment, 96
 - State, 95
 - Time, 96
 - Timer, 96
- Start PGF Sequence, 97
- Starting a Protocol, 97
- I/O Control Window, 217
 - AD-Channels, 218
 - DA-Channels, 218
 - Dig-In, 217
 - Dig-In/Out
 - Clear, 217
 - Dig-In/Out, 217
 - Single Bit, 217
 - Dig-Out, 217
 - Parameters, 218
 - Serial Output, 218
 - Serial Port, 218
 - Solutions, 218
 - Values, 218
- Imaging Extension, 251
- Introduction, 1
 - Closing Potmaster, 7
 - Disclaimer, 1
 - Installing Potmaster, 6
 - Key Conventions, 5
 - Naming Conventions, 4
 - EPC Type Amplifiers, 4
 - Syntax, 5
 - Windows versions, 5
 - Online Help Controls, 7
 - Online Help on Keys, 6
 - POTMASTER Concept, 1
 - Scope, 1
 - Starting Potmaster, 6
 - Support Hotline, 8
 - Supported System Software, 4
- Keys, 17
 - Key file, 18
 - Listing, 17
- Markers, 231
 - Edit Marker, 233
 - Delete, 233
 - Marker Index, 233
 - Scrolling Marker List, 233
 - Show Marker, 233
 - Show Sweep, 233
 - Text, 233
 - Time, 233
 - Trace, 233
 - Write to Notebook, 233
 - New Marker, 231
 - Add Marker, 231
 - From Sweep, 233
 - Text, 231
 - Time, 232
 - Trace, 231
- Menus, 21
 - Buffer Menu, 37
 - Accumulate Trace, 37
 - Add Trace, 37
 - Buffer 1...Buffer 4, 37
 - Clear, 37
 - Clear All, 37
 - Deaccumulate Trace, 38
 - Equation, 37
 - Export, 37
 - Import, 37
 - Parameter Handling, 38
 - Replace Target Trace, 38
 - Scale, 37
 - Show, 37
 - Subtract Trace, 37
 - Use Full Trace, 37
 - With Marked Target and Children, 38
 - Display Menu, 33
 - 3D-Mode, 34
 - Auto Show, 33
 - Auto Y-scale, 33
 - Buffer, 35
 - Center, 36
 - Dimmed Overlay, 33
 - Done, 36
 - Extend, 35
 - Fixed X-width, 33
 - Fixed Y-scale, 33
 - Freeze Zero Line, 34
 - Labeling, 34
 - Marker, 36
 - Overlay Pages, 33
 - Overlay Series, 33
 - Overlay Sweeps, 33
 - Overlay Traces, 33
 - Reference Series Correction, 36



- Repaint, 36
- Show, 36
- Show Tree Info, 34
- Show Zero Line, 33
- Subtract Reference Series, 33
- Subtract Trace Buffers Series, 33
- Subtract Zero Offset, 33
- Trace, 35
- Trace Buffer Correction, 36
- Trace Properties, 33, 35
- Use Trace Scaling, 36
- Vectors, 36
- Edit Menu, 25
 - Clear, 25
 - Copy, 25
 - Cut, 25
 - Find Same, 25
 - Find Selection, 25
 - Find..., 25
 - Paste, 25
 - Replace Same, 25
 - Replace..., 25
 - Select All, 25
 - Undo, 25
- File Menu, 21
 - Close, 22
 - File Status, 22
 - Merge..., 22
 - New Experiment, 22
 - New Group, 22
 - New..., 22
 - Open Modify..., 22
 - Open Read Only..., 22
 - Page Properties Notebook..., 23
 - Page Properties Traces..., 24
 - Page Setup..., 23
 - Quit, 24
 - Update File, 22
 - Write Tree File, 22
- Help Menu, 43
 - About POTMASTER, 44
 - Hide Keys, 44
 - List all Items, 44
 - List all Macro Items, 44
 - List Keys, 44
 - Save Keys, 44
 - Show Keys, 43
 - Show Tooltips, 43
- Notebook Menu, 38
 - Auto Store, 39
 - Clear, 38
 - Clear when Saved, 38
 - Font Size..., 38
 - Line Numbers, 38
 - Merge..., 38
 - Print..., 38
 - Save, 38
 - Save as..., 38
 - Scientific Notation, 39
 - Set Length..., 38
 - Zoom, 38
- PG 300/PG 600 Menu, 40
 - Calibrate PG300, 41
 - E vs HE, 40
 - E vs OCP, 40
 - E vs Reference, 40
 - EIS Calibration, 41
 - Initialize PG300, 41
 - Serial Number, 43
 - Update Delay, 40
 - Update DelayStimulus Ranging, 41
 - Use Value in Data File, 40
- Protocols Menu, 39
 - Append Macro File, 39
 - Execute, 39
 - Execute while Recording, 39
- Replay Menu, 27
 - ASCII, 28
 - ASCII-Text Format, 30
 - Copy PGF to PGF-Pool, 27
 - Delete, 27
 - Export, 28
 - Export Format, 28
 - Export Full Sweep, 28
 - Export Mode, 29
 - Export Trace, 30
 - Igor Pro, 28, 29
 - Import Trace, 30
 - MatLab, 28
 - PICT, 28
 - Print, 28
 - Printer, 28
 - Reference Series, 27
 - Show, 27
 - Show Method, 27
 - Show PGF Template, 27
 - Trace Time, 28
 - With marked Target and Children..., 30
 - WMF, 28
- Windows, 26
- Windows Menu
 - Analysis Window, 26
 - Calculator, 26
 - Close Front Window, 27
 - Configuration Window, 26
 - Control Window, 26



- Enable Icon Configuration, 27
- I/O Control Window, 26
- Imaging Window, 26
- LockIn Window, 26
- Markers, 26
- Notebook, 27
- Oscilloscope Window, 26
- Parameters Window, 26
- Photometry Window, 26
- Protocol Editor Window, 26
- Protocol Methods, 26
- Pulse Generator Window, 26
- Replay Window, 26
- Reset Front Window Position, 27
- Save Front Dialog, 27
- Solution Base, 26
- Spectroscopy, 26
- Notebook Window, 215
- Oscilloscope
 - Display
 - Cursor shapes, 71
 - Measurements on Traces
 - Measure, 72
 - Overlay Options
 - Overlay Series, 72
 - Overlay Sweep, 72
- Oscilloscope Window, 71
 - Display, 71
 - Auto Series, 73
 - Auto Sweep, 73
 - Digital Filter, 73
 - Display Scaling, 73
 - Lassoing, 71
 - Overlay Options, 72
 - Page, 73
 - Page Reset, 73
 - Reset, 73
 - Start / End Time, 73
 - Sweep Length Reset, 73
 - Trace Color, 73
 - Trace Selection, 73
 - Y-offset, 73
 - Y-Scale, 73
 - Y-Scale Centering, 73
 - Zooming, 71
 - Display Refreshing, 72
 - Freeze, 72
 - Repaint, 72
 - Wipe, 72
 - Measurements on Traces, 72
 - Scan, 72
 - Navigation, 72
- Statistics
 - Group/Series/Sweep, 72
- Parameters Window, 201
 - Flagging, 212
 - Clear all Flags, 213
 - Copy Flags from, 213
 - Default Flags, 213
 - Flag Groups 1...4, 212
 - Info to Notebook, 213
 - Marked to File, 213
 - Marked to Notebook, 213
 - Target to File, 213
 - Target to Notebook, 213
 - Parameter Tabs, 202
 - Amplifier Items, 208
 - ElProScan Items, 210
 - Group Items, 203
 - LF, 212
 - Marked Items, 212
 - Parameter Options, 212
 - Root Items, 202
 - Series Items, 204
 - Sweep Items, 205
 - Titles, 212
 - Trace Items, 206
- Photometry Extension, 241
 - Background 1...3, 241
 - Compute Traces of Marked Targets, 241
 - Dead Time, 241
 - DG-4/DG-5
 - Emit-1, 244
 - Emit-2, 244
 - Filter Position, 244
 - Open/Close, 245
 - PGF Setting, 245
 - Resting Wavelength, 244
 - Set, 244
 - Shutter Selection, 244
 - Kd*Sf, 241
 - Lambda 10-3, 247
 - Conditional Mode, 248
 - TTL Mode, 248
 - Lambda-10, 244, 246
 - Emit-1, 246
 - Emit-2, 246
 - Filter Position, 246
 - Open/Close, 246
 - PGF Setting, 246
 - Resting Wavelength, 246
 - Set, 246
 - Shutter Selection, 246
 - Speed, 246



- pti DeltaRAM, 249
 - Defaults, 242, 244, 246, 249
 - Emit-1, 249
 - Emit-2, 249
 - Excitation, 249
 - Resting Wavelength, 249
 - Set, 249
 - Shutter Selection, 249
 - Undo, 242, 244, 246, 249
 - Wavelength to Volt, 250
- R-max, 241
- R-min, 241
- TILL Photonics/Cairn OptoScan, 242
 - Calibrating TILL monochrometers, 244
 - Compute Factors, 243
 - Emit-1, 242
 - Emit-2, 242
 - Excitation, 242
 - from - to, 243
 - lower - Scan, 243
 - Polychrome I, 244
 - Polychrome II, 244
 - Polychrome IV, 244
 - Polychrome V, 244
 - Resting Wavelength, 242
 - Set, 242
 - Shutter Selection, 242
 - upper - Scan, 243
 - Volt, 243
 - wavelength, 243
 - Wavelength Calibration, 243
 - Wavelength to Volts, 243
- Potentiostat Window, 75
 - PG 300 - Main Functions, 75
 - Amplifier Mode, 77
 - Amplifier Mode Switch, 80
 - Amplifier Protocols, 81, 82
 - Auto Filter, 77
 - Auto Range, 80
 - Break+Standby, 77
 - Cell, 76
 - Cell Connection, 76
 - Control Amplifier Bandwidth, 78
 - Current Filter, 78
 - Current Range, 76
 - E-cell, 76
 - E-initial, 76
 - E-initial range, 81
 - Electrode Area, 82
 - Electrode Mode, 80
 - Ext. Input, 80
 - Filter Settings, 77
 - I-cell, 76
 - I-initial, 76
 - IR-Compensation, 79
 - Monitor, 75
 - Notch Filter, 77
 - OCP, 76
 - Open Cell Potential, 76
 - Reset, 80
 - Standby, 76
 - Stimulus Filter, 78
 - Tuning, 77
 - Voltage Filter, 78
 - PG 340 - Special Functions, 89
 - Additional Settings, 90
 - Electrode Conditions, 89
 - Electrode Mode, 90
 - Potential/Current Settings, 90
 - PG 600 - Main Functions, 83
 - Amplifier Protocols, 87
 - Auto Filter, 85
 - Auto Range, 86
 - Break + Standby, 85
 - Charge Controls, 87
 - Control Amplifier Bandwidth, 85
 - Current Density Control, 87
 - Current Range, 84
 - DC Subtraction, 85
 - E-cell, 83
 - E-initial, 84
 - Electrode Area, 88
 - Electrode Mode, 86
 - Filter Settings, 85
 - Galvanostatic Mode, 83
 - I-cell, 83
 - I-initial, 84
 - IR Compensation, 86
 - Monitor, 83
 - OCP, 83
 - Open cell potential, 83
 - Output Filter, 85
 - Potentiostatic Mode, 83
 - Reset, 86
 - Standby, 83
 - Stimulus Filter, 85
 - Tuning, 84
 - Voltage Range, 84
 - PG 618 - Special Functions, 91
 - Amplifier-1, 91
 - Amplifier-2, 92
 - Amplifier-2 - Additional Features, 93
 - Amplifier-2 - Filter Settings, 92
 - Electrode Conditions, 91
 - Protocol Editor - E-Chem Methods, 161
 - Cyclic Voltammetry, 165



- +1 end cycle, 166
- Anodic Peak, 166
- Auto Ranging, 166
- Baseline, 166
- Cathodic Peak, 166
- Customizing Acquisition Sequence, 170
- CV Parameters, 165
- Increment, 166
- Keep at End, 166
- Number of Cycles, 166
- Reversal Pot. 1, 166
- Reversal Pot. 2, 166
- Scan Rate, 166
- Start/End Potential, 165
- Total Length, 166
- Use E-init, 166
- Impedance Spectroscopy, 172
 - Acquisition, 173
 - Auto Ranging, 172
 - Customizing Analysis, 175
 - Cycles to Skip, 172
 - EIS Parameters, 172
 - End Frequency, 172
 - Freq. per Decade, 172
 - Min. Cycle Count, 172
 - Sine Amplitude, 172
 - Start Frequency, 172
 - Total Length, 173
 - Use Calibration Table, 173
- Open Circuit Potential, 176
 - Acquisition, 176
 - Duration, 176
 - Parameters, 176
 - Set OCP, 176
- Optical Spectroscopy, 181
- Square Wave Voltammetry, 178
 - Acquisition, 180
 - Alternate Scan Direction, 179
 - Amplitude, 178
 - End Potential, 178
 - Frequency, 178
 - I sampled from, 179
 - Keep at End, 178
 - Keep Traces, 179
 - Number of Cycles, 179
 - Parameters, 178
 - Peaks, 179
 - Pot. Increment, 179
 - Rel. to E-init, 178
 - Scan Rate, 178
 - Start Potential, 178
 - Tau, 178
 - Total Time, 179
- Protocol Editor - EChem Methods
 - Cyclic Voltammetry
 - Acquisition, 167
 - Analysis, 168
- Protocol Editor Window, 135
 - Acquire Each Sweep
 - Duration, 145
 - Increment, 145
 - Acquire Properties
 - Set Min. Wait Time, 146
 - Update R-membrane and I-pipette, 146
 - Wipe Display at Start, 146
 - Write Events to Notebook, 146
 - Acquire Series
 - Averages, 145
 - Break Keys, 145
 - Comment, 145
 - Edit PGF Template, 145
 - Label, 145
 - Sequence, 145
 - Amplifier
 - Auto Filter, 147
 - Auto Range Once, 147
 - OCP Set, 147
 - Set Amplifier Mode, 147
 - V-membrane, 147
 - Analysis
 - Analysis Mode, 149
 - Edit, 149
 - Name, 149
 - Set Graph Position, 149
 - Wipe, 149
- BREAK
 - Type, 139
- Chain Protocol
 - Edit Protocol, 139
 - Protocol Name, 139
 - Return, 139
 - Select Protocol, 139
- Display Properties
 - Overlay Series, 150
 - Overlay Sweep, 150
 - Reset Timer, 150
 - Show Leaks, 150
 - Show Traces, 151
 - Show Tree Info, 150
 - Subtract Leaks, 150
 - Subtract Zero, 150
 - Wipe Online, 150
 - Wipe Osci, 150
- ElProScan
 - Piezo-Position, 158
 - Position, 158



- Scan Mode, 158
- X-Position, 158
- Y-Position, 158
- Z-Position, 158
- Event Handling, 136
 - Delete, 136
 - Duplicate, 136
 - Events, 136
 - Insert After, 136
 - Insert Before, 136
 - Move, 136
- Events, 138
 - Acquire Each Sweep, 145
 - Acquire Properties, 146
 - Acquire Series, 145
 - Acquisition, 145
 - Amplifier, 146
 - Analysis, 149
 - Annotation, 156
 - Beep, 157
 - BREAK, 139
 - Chain Protocol, 139
 - Clear Key, 139
 - Data/Display, 149
 - Digital Filters, 149
 - Display Properties, 150
 - ElProScan, 157
 - ELSE, 142
 - ELSEIF... THEN, 142
 - Export, 151
 - Extensions, 157
 - File Operation, 151
 - GOTO, 140
 - GOTO_MARK, 140
 - Hardware, 146
 - IF... THEN, 140
 - Imaging, 159
 - Launch, 143
 - Macro Command, 143
 - Messages, 156
 - Methods, 159
 - PGF Parameters, 154
 - Photometry, 159
 - Protocol Sequence, 139
 - REPEAT, 143
 - Replay, 152
 - Serial Output, 147
 - Set DAC, 147
 - Set Digital Bit, 148
 - Set Digital Word, 148
 - Set Solution Changer, 148
 - Set Solutions, 154
 - Set Sweep Label, 146
 - Set Value, 155
 - Switch Window, 144
 - Trace Buffer, 153
 - Value/Parameter, 154
 - WAIT, 144
 - Write Icon Value, 157
- Export
 - Filename, 151
 - Full Sweep, 151
 - Overwrite, 151
 - Target, 151
- File Operation
 - Close File, 152
 - File Status, 152
 - Label, 152
 - New Experiment, 152
 - New File, 152
 - New Group, 152
 - Save after Break, 152
 - Store Data, 151
 - Update File, 152
- GOTO
 - Mark, 140
- GOTO_MARK
 - Mark, 140
- IF... THEN
 - Left Source, 140
 - Logical operators, 142
 - Right Source, 141
- Macro Command
 - Command, 143
- PGF Parameters
 - 1...10, 154
 - Operation, 154
 - Value, 154
- Photometry
 - New Resting Wavelength, 159
 - Open Shutter / Close Shutter, 159
 - Set New Speed, 159
 - Set New Wavelength, 159
 - Set Resting Wavelength, 159
- Protocol Handling, 135
 - Delete, 136
 - Load, 136
 - Merge, 136
 - Move, 136
 - Name, 136
 - New, 136
 - Protocol Pool, 136
 - Record Macros, 136
 - Relative Value, 136
 - Save, 136
 - Step, 136



- To End, 136
- Write, 136
- Recurring Functions, 137
 - Delay, 137
 - If etc. Result, 137
 - Repeat Status, 137
 - Skip, 137
- REPEAT
 - Duration, 143
 - Increment, 143
- Serial Output
 - String, 147
- Set DAC
 - Channel, 148
 - Operation, 148
 - Use Dac Scaling, 148
 - Voltage, 148
- Set Digital Bit
 - Channel, 148
 - low, 148
- Set Digital Word
 - Digital Bits, 148
 - Digital Word, 148
- Set Solution Changer
 - Set Marker, 148
 - Solution, 148
- Set Solutions
 - Set External, 154
 - Set Internal, 154
 - Set Marker, 154
- Set Sweep Label
 - Sweep Label, 146
 - Sweep Label Target, 146
- Set Value
 - Copy, 156
 - Operation, 155
 - Value, 156
 - Value 1...16, User 1 – 2, 155
- Trace Buffer
 - Clear Buffer, 153
 - Operation, 153
 - Replace, 153
 - Scaling, 153
 - Target, 153
 - Trace Buffer Source, 153
 - Update Analysis, 153
 - Update Display, 153
- WAIT
 - Wait type, 144
- Write Icon Value
 - Alertbox, 157
 - Icon Text / Icon Value, 157
 - Item ID, 157
 - LF, 157
 - Notebook, 157
 - Title, 157
- Pulse Generator
 - Check and Execute
 - Execute, 108
 - Segment Classes
 - Delete, 126
- Pulse Generator Window, 103
 - Acquisition Modes, 106
 - Gap Free Mode, 106
 - Interactive Mode, 106
- AD Input Channel
 - AD, 121
 - Build Instructions, 123
 - Chirp, 121
 - Compression Modes, 122
 - Defaults, 123
 - Dig-0...15, 121
 - Dig-in (word), 121
 - Digital Filter, 122
 - E-mon, 121
 - I-mon, 121
 - Imaging, 122
 - Leak, 124
 - Link, 122
 - LockIn, 122
 - Mux, 121
 - Off, 121
 - Photometry, 122
 - Set Offset, 122
 - Unit, 122
 - Virtual, 121
 - Zero Offset, 124
- Auto Sample, 106
- Channel Settings DA - Output/AD Input, 118
 - AD Input Channel, 121
 - Channels, 118
 - DA Output Channel, 118
 - Use of Digital Outputs, 119
- Check and Execute, 107
 - Checking, 108
 - Not Checking, 108
- Chirp Wave Parameters
 - Amplitude, 114
 - End Frequency, 114
 - Exponential Chirp, 114
 - Linear Chirp, 114
 - Points/Cycle, 114
 - Pre-Chirp, 115
 - Segment Points, 114, 115
 - Spectroscopy Chirp, 115
 - Start Frequency, 114



- Value/Parameter, 114
- DA Output Channel
 - Analog output, 118
 - Apply StimScale, 119
 - Dig-0...15, 118
 - Dig-out (word), 118
 - Leak, 120
 - load from file template, 119
 - Off, 118
 - relative to Einit, 119
 - Stim-out, 118
 - Stimulus - DA, 119
 - Unit, 119
 - use DA-Scaling, 119
 - use for Imaging, 119
 - use for LockIn, 119
 - use for Spectroscopy, 119
 - use for Wavelength, 119
- E-initial, 131
- Error Handling, 134
- Increment Modes
 - Alternate, 127
 - Decrease, 127
 - Increase, 127
 - Interleave +, 127
 - Interleave -, 127
 - Toggle, 127
- Leak Correction, 131
 - Alt. Leak Average, 132
 - Hardware Leak, 132
 - Leak Alternate, 132
 - Leak Delay, 131
 - Leak Hold, 131
 - Leak min/max, 131
 - Leak Size, 131
 - No of Leaks, 131
 - Software Leak, 132
 - Wait, 133
- Logarithmic Increment Modes
 - dt*Factor, 128
 - dV*Factor, 128
 - t*Factor, 128
 - V*Factor, 127
- Miscellaneous, 129
 - Analysis Method, 130
 - Break Condition, 129
 - Break if >, 129
 - Break if <, 129
 - Break if abs <, 129
 - Break on next, 129
 - Common Timing, 129
 - Filter Factor, 130
 - Galvanostatic, 129
 - No Break, 129
 - Potentiostatic, 129
 - Recording Mode, 129
 - Relevant Segments, 130
 - Separate Timing, 129
- PGF Parameters, 134
- Photometry Wave Parameters
 - Adapt to Maximal Sweep Length, 116
 - Expand, 116
 - Number of Cycles, 116
 - Number of Segments, 116
 - Trunc, 116
- Recording Mode
 - Any Mode, 129
- Segment Classes
 - Chirp, 126
 - Constant Segment, 126
 - Continuous, 126
 - Delete, 126
 - Duplicate, 126
 - Duration, 126
 - Insert, 126
 - Potential, 126
 - Ramp, 126
 - Segment Class, 126
 - Sine, 126
 - Square, 126
 - Store, 125
 - Value setting, 126
- Segments, 125
 - Increment Modes, 127
 - Logarithmic Increment Modes, 127
 - Scan Rates, 127
 - Segment Classes, 125
- Sequence Pool, 105
 - Copy, 105
 - Delete, 105
 - List, 105
 - Load, 105
 - Merge, 105
 - Move, 105
 - Name, 105
 - New, 105
 - Save, 105
- Sequence Selector, 105
 - Sequence Buttons, 105
- Sine Wave Parameters, 108
 - Actual Frequency, 109
 - Common Frequency, 109
 - Cycles to Average, 111
 - Cycles to Skip, 110
 - LockIn, 110
 - Peak Amplitude, 109



- Points / Cycle, 109
- Requested Frequency, 109
- Separate Frequency, 109
- Total Cycles, 111
- Use Theoretical Stimulus, 110
- Value/Parameter, 109
- Square Wave Parameters
 - Actual Frequency, 111
 - Amplitude Increment, 112
 - Common Frequency, 111
 - Cycles to Average, 113
 - Cycles to Skip, 113
 - LockIn, 112
 - Negative Amplitude, 112
 - Peak Amplitude, 111
 - Points/Cycle, 113
 - Positive Duration Factor, 112
 - Requested Frequency, 111
 - Separate Frequency, 112
 - Total Cycles, 113
 - Use Theoretical Stimulus, 113
 - Value/Parameter, 111
- Stimulus Template, 130
 - Delay, 130
 - Preview Mode, 130
- Sweep and Channel Length, 108
 - Channel Length, 108
 - Stimulus, 108
 - Stored Sweep Duration, 108
 - Sweep Duration, 108
 - Sweep Length, 108
- Timing, 106
 - Allow Continuous Redraw, 106
 - No Wait before 1. Sweep, 106
 - No. of Sweeps, 106
 - Not Triggered, 107
 - Sample Interval, 107
 - Start Segment, 107
 - Start Time, 107
 - Sweep Delay, 106
 - Sweep Interval, 106
 - Trigger Mode, 107
 - Trigger Series, 107
 - Trigger Sweep, Main only, 107
 - Trigger Sweeps, Main + Leaks, 107
 - Use Durations, 107
 - Use Scan Rates, 107
 - Wait before 1. Sweep, 106
- Views, 104
 - Cartoon View, 104
 - Condensed View, 104
 - Full View, 104
- Wave Parameters, 108
 - Chirp Wave Parameters, 113
 - Imaging Wave Parameters, 116
 - Photometry Wave Parameters, 115
 - Square Wave Parameters, 111
- Replay Window, 99
 - Main window functions, 99
 - Apply, 100
 - Arrows, 100
 - Interpreting Tree objects, 100
 - Label, 99
 - Maneuvering, 101
 - Mark, 100
 - Mark a Target and its Children, 101
 - Mark All, 100
 - Repeat Rate, 100
 - Selecting/Marking multiple objects, 101
 - Show, 100
 - Text, 99
 - Tracing, 100
 - Tree Handling, 100
 - Unmark, 100
 - Visible, 100
 - Window Scroll Arrows, 100
- Software LockIn Extension, 253
 - Capacitance Measurements, 256
 - Activating the LockIn Extension, 256
 - Amplifier Settings, 260
 - Analysis, 261
 - Creating a PGF sequence, 257
 - LockIn Configuration, 256
 - Running the PGF sequence, 260
 - LockIn Configuration
 - Points to Average, 254
 - LockIn Configuration Window, 253
 - Attenuation, 254
 - Compute LockIn of Marked Targets, 254
 - Default Y-ranges, 254
 - LockIn Mode, 253
 - Offline Computation - Trace to create, 254
 - Parent Trace, 254
 - Phase Shift, 253
 - SetDefaults, 254
 - Skip, 254
 - Write to Notebook, 254
 - LockIn Mode
 - On Cell, 253
 - Phase Range: 0..180, 253
 - Phase Range: 0..360, 253
 - Piecewise Linear, 253
 - Sine + DC, 253
 - LockIn Traces, 254
 - LockIn_Admit(Y), 255



- LockIn_Imag(Y), 255
- LockIn_Imag(Z), 255
- LockIn_Imp|Z|), 255
- LockIn_Phase, 255
- LockIn_Real(Y), 255
- LockIn_Real(Z), 255
- Solution Changer, 229
- Solutions, 225
 - Solution Base, 225
 - Solution Data Base, 226
 - Append, 227
 - Concentration, 227
 - Create Entry, 227
 - Delete, 227
 - Delete Entry, 227
 - Duplicate Entry, 227
 - Entries, 227
 - Export Label, 227
 - Export Listing, 227
 - Ingredient, 227
 - Insert, 227
 - Last Entry, 227
 - Name, 226
 - Next Entry, 227
 - Numeric Name, 226
 - Numeric Name Value, 226
 - Osmolarity, 226
 - pH, 226
 - Solution Index, 226
 - Solution Indices, 226
- Spectroscopy Extension, 265
 - Display of Spectra, 273
 - Introduction, 265
 - Chirp Wave Forms, 265
 - Exponential Chirp, 265
 - Linear Chirp, 265
 - Spectroscopy Chirp, 266
 - Transfer Function Ratio, 266
- Spectroscopy Acquisition, 269
 - Adding Chirp Analysis Traces, 270
 - Minimizing Swing-In Effects, 270
 - Parametrization of the Chirp Stimuli, 269
 - Possible Frequency Range, 269
 - The Chirp Wave Dialog, 269
 - Useful Frequency Range, 270
- Spectroscopy Extension, 266
 - Chirp_Admit(Y), 268
 - Chirp_Avg, 268
 - Chirp_Imag(Y), 268
 - Chirp_Imag(Z), 268
 - Chirp_Imp(Z), 269
 - Chirp_Phase, 268
 - Chirp_Real(Y), 268
 - Chirp_Real(Z), 268
 - Compute Spectroscopy of Marked Targets, 268
 - Correction Mode, 267
 - Default Ranges, 268
 - Input Source, 267
 - Log Compression, 267
 - Offline Computation - Traces to create, 268
 - Parent AD-Trace, 268
 - Reference Element Correction, 267
 - The Spectroscopy Traces, 268
- Supported
 - Potentiostats/Galvanostats, 4
- User Interface
 - Dialog Controls, 9
 - Background Color, 10
 - Boolean, 10
 - Button, 9
 - Checkbox, 9
 - Drag, 9
 - Edit Text, 9
 - Enter, 10
 - Framed Text, 10
 - List, 9
 - Number, 10
 - Radio Button, 9
 - Switch, 9
- User Interface, 9
 - Copy and Paste Functions, 16
 - Dialog Controls
 - Numerical Input, 10
 - SI Units, 10
 - String Buffer, 10
 - Modifying Dialogs and Controls, 10
 - Deleting Controls, 14
 - Dialog Control Window, 11
 - Hiding Controls, 13
 - Saving Dialogs and Controls, 14
 - Switching Windows, 15
- Zero Potential Window
 - Manual Input, 221
 - Normal Hydrogen Electrode, 222
 - Reference Potential, 222
 - Relative Hydrogen Electrode, 222
- Zero Potential Windows, 221



