



THERAPY GLOVES
INSTRUCTIONS FOR USE



EASY TO USE.



Scan QR code for instructional video.




Carefully pull on your gloves using the cuff at the base of product near wrist. **Note: To maximize longevity of the gloves do not pull on fingers of gloves to put on or take off.**

- 2 To turn on, simply hold button near top of cuff for 2 seconds until a blue light is illuminated.



- 3 Press button once to begin low frequency vibration. Press button again to toggle between desired vibration levels.



Low frequency

High frequency

Pulse



- 4 Apply vibration therapy for 25 – 45 minutes for maximum level of pain relief.



- 5 When finished, simply hold button down for two seconds to turn off.

- 6 Replenish battery by plugging dual USB charger into the charge port of each glove.

- 7 Use daily to maximize pain relief.

EASY TO MAINTAIN.

Recharges in minutes with dual USB charger. Simply plug one end into AC wall adapter or powered USB port on alternative device and other end into each glove. Lights will illuminate as unit charges. Blue light indicates full charge. Fully charged, gloves will operate up to 45 minutes on high frequency. Use only the charger provided with Gloves. Charger included is not interchangeable with other products.

EASY TO SERVICE.

Brownmed warrants this product to be free from manufacturing defects. If your Intellinetix Gloves stop working within one year from date of purchase, we will replace them.

EASY TO CARE FOR.

- Spot clean as needed.
- Lay flat and allow to air dry.
- Keep charger opening dry while cleaning.
- For adult use only.
- Do not dry clean.

Need a replacement or extra parts?

Customer service is happy to help you.

Please call us at **1-877-853-5518**.

Show us how **you** vibe.
#Intellinetix

Learn more at intellinetix.com



Executive:

101 Federal Street
29th Floor
Boston, MA 02110
P 857-317-3354
F 857-317-3355

Manufacturing:

1300 Lundberg Drive West
Spirit Lake, IA 51360
TF 800-843-4395
P 712-336-4395
F 712-336-2874

©2023 Brownmed- Intellinetix is a registered trademarks of Brownmed.



WARNING: Lithium-ion Batteries and/or products that contain Lithium-ion Batteries can expose you to chemicals including cobalt lithium nickel oxide and nickel, which are known to the State of California to cause cancer and birth defects or other reproductive harm. For more information, go to www.P65Warnings.ca.gov.