


Quick Start Guide

LINECHEK® II

Model 620L



SAFETY CHECKLIST




- S**urvey the test station. Make sure it is safe & orderly.
- A**lways keep unqualified/unauthorized personnel away from the test area.
- F**amiliarize yourself with safety protocols in the event of a problem.
- E**xercise caution and never touch products or connections during a test.
- T**rain operators. Never touch clips directly and always connect the return lead first.
- Y**ou should always know when a test is being performed.




WARNING: THIS GUIDE WAS CREATED FOR OPERATORS HAVING SOME FAMILIARITY WITH ELECTRICAL SAFETY TESTING. AN ELECTRICAL SAFETY TESTER PRODUCES VOLTAGES AND CURRENTS THAT CAN CAUSE HARMFUL OR FATAL ELECTRIC SHOCK. TO PREVENT ACCIDENTAL INJURY OR DEATH, THESE SAFETY PROCEDURES MUST BE STRICTLY OBSERVED WHEN HANDLING AND USING A TEST INSTRUMENT. CONTACT US AT INFO@ARISAFETY.COM FOR MORE INFO ON HOW TO GET TRAINED ON ELECTRICAL SAFETY TESTING.

FRONT PANEL CONTROLS




- 1 **RESET BUTTON:** Resets the instrument. If a failure condition occurs during a test, pressing this button will reset the system, shut off the alarm and clear the failure condition. The Reset button must be pressed before performing another test or changing any of the setup parameters. This button also serves as an abort signal to stop any test in progress.
- 2 **TEST BUTTON:** Starts a test.
- 3 **GRAPHIC LCD:** 320 X 240 Monographic LCD.
- 4 **SOFT KEYS:** Multifunction keys used to select screens and change parameters.
- 5 **NUMERIC DATA ENTRY:** Keys used to enter numeric data.
- 6 **UP, DOWN, LEFT, AND RIGHT ARROW KEYS:** Keys used to scroll the highlighted area or cursor, up and down, left and right. When more than five steps are programmed in a test file, the left and right arrow keys will page through the screens of steps. The screens where the paging function is available are: Setup Tests, Perform Tests, Results Summary, and Results.
- 7 **POWER SWITCH:** Rocker style power switch with international ON (I) and OFF (O) markings.
- 8 **EXIT KEY:** Key used to escape from parameter editing and return to prior screens.
- 9 **ENTER KEY:** Key used to finalize parameter entries. The ENTER key may also be used to scroll the highlighted area to different parameters in the parameter setting screens.
- 10 **LINE VOLTAGE INDICATOR:** This indicator flashes to warn the operator that line voltage is present at the voltage output terminal.

BACK PANEL CONTROLS




- 1 EXTERNAL MEASURING DEVICE:** Contains an external Measuring Device PCB that will enable during a Leakage Current test when “External” is selected from the “Meas. Device” soft key. The external MD allows the operator to configure a custom measuring device using either a simple resistive component or a complex two pole network.
- 2 INTERNAL MD MODULE:** Provides the operator access to the measuring device that is currently selected. The BNC connector may be used to monitor that voltage directly across the terminals of the MD.
- 3 SCANNER CONNECTOR:** For connection of optional external 8 channel Scanner.
- 4 REMOTE OUTPUT:** Optional connector used to interface the 620L LINECHECK with an APT AC Power Source for remote memory selection.
- 5 CALIBRATION BUTTON:** To put the instrument into the calibration mode push this button and turn on the power switch simultaneously.
- 6 REMOTE SIGNAL OUTPUT:** 9-Pin D-type subminiature female connector for monitoring PASS, FAIL, and PROCESSING output relay signals.
- 7 REMOTE SIGNAL INPUT:** 9-Pin D-type subminiature male connector for remote control of TEST, RESET, and REMOTE.

BACK PANEL CONTROLS



- 8 DUT POWER INPUT LINE:** Provides DUT input power (line connection). Only a single phase unbalanced power source may be used.
- 9 DUT POWER INPUT NEUTRAL:** Provides DUT input neutral power (neutral connection). Only a single phase unbalanced power source may be used.
- 10 DUT POWER OUTPUT:** Provides line power (line connection) to the DUT during Leakage Current and Functional Run testing.
- 11 DUT POWER OUTPUT:** Provides line power (neutral connection) to the DUT during Leakage Current and Functional Run testing.
- 12 PROBE HI:** Connector used to attach the Probe-HI test lead to the DUT. When used with an appropriate test lead, this terminal connects one side of the MD (measuring device) to the DUT during Enclosure or Applied Part Leakage Current tests. This terminal will be enabled during a Leakage Current test whenever Probe-HI has been selected.
- 13 PROBE LO:** Connector used to attach the Probe-LO test lead to the DUT. When used with an appropriate test lead, this terminal connects one side of the MD (measuring device) to the DUT during Applied Part Leakage Current tests. The terminal is always used in conjunction with the Probe-HI terminal.
- 14 GND:** Connector used to attach the adapter box Ground lead to the instrument.
- 15 CASE (Optional):** Connector used to attach the return lead to the DUT case or dead metal. Provides the return for the Ground Bond, Dielectric Withstand, and Insulation Resistance tests. During a Leakage Current test and Functional Run test this terminal is isolated from the test circuits.

BACK PANEL CONTROLS



- 16 SAFETY TESTER CONNECTIONS (Optional):** Connector used to interface the 620L to an Associated Research electrical safety tester.
- 17 CHASSIS GROUND (Earth) CONNECTION:** This terminal should be connected to a good earth ground before operation.
- 18 FUSE RECEPTACLE:** To change the fuse, unplug the power (mains) cord and turn the fuse receptacle counter-clockwise. The fuse compartment will be exposed. Replace the fuse with one of the proper rating.
- 19 BUS INTERFACE:** Standard connector for interconnection to the USB/RS-232 Bus interface. Optional IEEE-488 interface may be substituted for the USB/RS-232.
- 20 INPUT POWER RECEPTACLE:** Standard IEC 320 connector for connection to a standard NEMA style line power (mains) cord.
- 21 INPUT POWER SWITCH:** Sets the line voltage configuration of the instrument. In the left position it is set for 115 volt operation, in the right position it is set for 230 volt operation.


INSTRUMENT SETUP



WARNING: LOCATE A SUITABLE TESTING AREA WITH A THREE-PRONG, GROUNDED OUTLET. BE SURE THAT YOUR THREE-PRONG OUTLET HAS BEEN TESTED FOR PROPER WIRING. READ THE SAFETY CHECKLIST OF THIS GUIDE BEFORE STARTING TO TEST.

1. Choose the correct input line voltage on the rear panel of the instrument, either 115V AC or 230V AC.
2. Connect the line cord into the rear of the instrument and plug the male end of the cord into a grounded power source.
3. Plug the interlock connector into the Signal Input connector on the rear panel of the instrument. LINECHEK is equipped with a Remote Interlock feature, which utilizes a set of closed contacts to enable the instrument's output. In order to perform a test, the interlock connector must be plugged into the Signal Input connection on the rear panel of the instrument.
4. Turn the POWER switch to ON.

The initialization screen will appear. After three seconds the Perform Test Screen will appear as shown below.



CHANGING TEST SETTINGS

ADDING A NEW TEST

From the Main Menu screen, press the “Setup Tests” soft key.

On the Setup Tests screen, the test and step number will be highlighted.

Press the “Add” soft key and specify the type of test with the soft key options given (Run Test (optional feature), Leakage Current test).

The parameters for the chosen test type will appear. Use the soft keys to edit parameters on the right side of the screen (examples: neutral, reverse, ground).

Use the up and down arrows to scroll between and edit test parameters on the left side of the screen (examples: leakage-hi, leakage-lo, voltage-hi).

Use the numeric keypad to change a parameter value. Once the user has entered the desired parameter, press the ENTER key to save.

Press EXIT to go back to the Setup Tests screen to add another test or press EXIT again to go back to the Main Menu.

EDITING & DELETING TESTS

If the default settings do not apply to user specifications, the test parameters are fully adjustable. Common parameters that can be adjusted include leakage-hi, leakage-lo, voltage-hi, voltage-lo, delay time, and dwell time.

From the Main Menu screen, press the “Setup Tests” soft key.

Use the arrow keys (up, down) to highlight the test to be edited and hit the “Edit” soft key.

Select a parameter to edit by scrolling with the up and down arrow soft keys.

Use the numeric keypad to adjust the parameter value.

Once the user has entered the desired value, press the ENTER key to save.

Press EXIT twice to get back to the Main Menu screen.

Deleting a test step is also done from the Setup Tests menu. Highlight the test to be deleted and hit the “Delete” soft key.

Press ENTER and the step will be deleted. Pressing EXIT will abort the deletion of the test.


CHANGING TEST SETTINGS

SAVING AND LOADING TEST FILES

Multiple tests for a particular product or individual product can be saved in files for an easy recall of test conditions.

From the Main Menu screen, press the "Setup Tests" soft key.

Hit the "File" soft key from the Setup Tests window. The files currently on the system will be displayed.



Using the up and down arrow keys, highlight the number of the file slot where the test sequence is to be saved. Hit the "Save As" soft key and the Create File screen will appear.

Use the arrow keys (up, down, left, right) to select letters on the virtual keypad or use the numeric keypad to enter numbers for the file name. When finished, hit the ENTER key. The system will exit back to the "Setup Tests" screen with the current file now loaded.

To recall any test file, hit the "File" soft key from the Setup Tests screen.

Use the up and down arrow keys to select the file to be loaded. Hit the "Load" soft key and the test sequence associated with the file will automatically be loaded.

CONNECTIONS

DUT INPUT POWER CONNECTIONS

Remove the cover plate on the rear panel of the 620L to access the DUT Power Input Connections. These terminals provide the connections from an external power source to the 620L.

Once removed, use the black 10 AWG cables (included) to connect to an unbalanced, single-phase power source for applications up to and including 40 amps.

The 220 - 240V US style line power IS NOT suitable to connect to the DUT inputs. This style of power distribution is a balanced type with two HOT or LINE conductors.



BE SURE THAT THE HOT LEAD OF THE POWER SOURCE IS CONNECTED TO PIN 1, L (LINE) TERMINAL OF THE DUT POWER INPUT TERMINAL BLOCK AND THE RETURN OR LOW VOLTAGE LEAD IS CONNECTED TO PIN 2, N (NEUTRAL) TERMINAL OF THE DUT POWER INPUT TERMINAL BLOCK. DO NOT CONNECT A LINE CONDUCTOR TO THE N OR NEUTRAL TERMINAL OF THE DUT POWER INPUT. THIS CONDITION CAN BE VERY DANGEROUS TO THE OPERATOR.

If voltage is applied to pin 2, N (neutral) terminal of the DUT Power Input and you attempt to execute a Run Test or Leakage Current test, a warning message will appear in the displayed messages portion of the screen that says Neutral-V.

If you see this message, you will need to correct the voltage problem before the instrument will allow you to execute Run Tests or Leakage Current tests.

DUT OUTPUT POWER CONNECTIONS

Remove the cover plate on the rear panel of the 620L. These connections provide power to the DUT during a Leakage Current or Functional Run test.


Once removed, the set of black 10 AWG cables (included) may be used for the DUT Output connections for applications up to and including 40 amps.

CONNECTIONS

TEST LEAD CONNECTIONS

Connect one or more of the black, alligator clip-terminated cables to the Probe-HI and Probe-LO terminals on the rear panel of the 620L. These cables are used to measure leakage current during either a Leakage Current test or a Functional Run test.

Connect the red test lead (also terminated in an alligator clip) to the GND terminal on the rear panel of the 620L. This will provide the return connection for all leakage current measurements.




ADAPTER BOX CONNECTIONS

The adapter box may be used for products that are terminated in either a two prong or three prong line cord.


Connect the adapter box to the DUT Output L (line) and N (neutral) terminals, and the GND terminal.

The adaptor box will now provide power to the DUT during Leakage Current testing and Functional Run testing.



WARNING: THE ADAPTER BOX MAY ONLY BE USED FOR APPLICATIONS UNDER 20 AMPS. IF AN APPLICATION IN WHICH MORE THAN 20 AMPS OF INPUT CURRENT IS REQUIRED, PLEASE USE THE STAND-ALONE INPUT AND OUTPUT CABLES PROVIDED WITH THE INSTRUMENT.

TEST RESULTS



VIEWING TEST RESULTS

From the Perform Tests screen, press the “Results” soft key.

Use the up and down arrows to select the test to be reviewed or use the left and right arrows to scroll between memories.

Results are displayed on the right hand side of the screen including the test type, reason for failure and test parameter results.

Press the EXIT key to go back to the Perform Tests screen and hit EXIT again to go back to the Main Menu.

INTERCONNECTION

620L LINECHECK II INTERCONNECTION

In order to interconnect the LINECHECK II to the OMNIA 8204 Electrical Safety Compliance Analyzer, system settings must change on the OMNIA.

First activate the PLC remote on the OMNIA.

Choose SETUP SYSTEM from the Main Menu screen.


Select HARDWARE from the Setup System screen.

Use the + and - soft keys to turn the PLC Remote Setting to ON.

Press EXIT twice to return to the Main Menu Screen.

INTERCONNECTION

INTERCONNECTION OF 620L LINECHECK II TO OMNIA 8204 AND CONNECTION TO A DEVICE UNDER TEST



1. Plug the double sided interlocks (P/N 38142) into the signal inputs of both instruments.
2. Using the rear panel connections, connect the following cords:
 - a. P/N 38556 - Plug one end into the return and sense (-) jack on the OMNIA. Hook the other end into the return jack on the 620L.
 - b. P/N 38555 - Plug one end into the current and sense (+) jack on the OMNIA. Hook the other end into the current jack on the 620L.
 - c. P/N HS-8-12 - Plug one end into the HV port on the OMNIA. Plug the other end into the HV port on the 620L.
 - d. P/N 05030DT-21- Plug the red coded end into the 620L interlock connector. Plug the blue coded end into the OMNIA signal output.
 - e. P/N 05030DT-22 - Plug the red coded end into the 620L signal output. Plug the blue coded end into the OMNIA interlock connector.
 - f. P/N CBLHC40-10TL - Hook the black Case lead to the rear panel Case connection on the 620L. Then connect the clip end of the test lead to the chassis ground of the DUT.
 - g. P/N 38777 - Remove the DUT INPUT/ DUT OUTPUT plate cover from the 620L. Using a screwdriver, screw in the line and neutral connections of the 620L adapter box to the L and N posts of the DUT OUTPUT on the 620L. Next, hook the GND lug on the adapter box to the GND jack of the 620L.
3. Connect Line and Neutral from an unbalanced power supply into the L and N posts on the DUT INPUT of the 620L.
4. Plug the DUT line cord into the adapter box.



SCAN FOR
**VIDEO
LIBRARY**

Experts In Electrical Safety Compliance.®



For additional information about these and other key features of the 620L LINECHECK® II, please consult the full Operation and Service Manual or call us toll-free 1-800-858-TEST (8378) or +1-847-367-4077

©2022 Associated Research • arisafety.com

FOLLOW US!

