



HEADACHE BAND

INSTRUCTIONS FOR USE



EASY TO USE.

- 1 To turn on, simply hold button on side of the band for 2 seconds until a blue light is illuminated.



- 2 Press button once to begin low frequency vibration. Press button again to toggle between desired vibration levels.

Low frequency



High frequency



Pulse



- 3 Apply vibration therapy for 25 – 45 minutes for maximum level of pain relief.



- 4 When finished, simply hold button down for two seconds to turn off.

- 5 Replenish battery by plugging USB charger into the charge port.

- 6 Use as needed to maximize pain relief.

EASY TO MAINTAIN.

Recharges in minutes with USB charger. Simply plug one end into AC wall adapter or powered USB port on alternative device and other end into the band. Lights will illuminate as unit charges. Blue light indicates full charge. Fully charged, band will operate up to 45 minutes on high frequency.

EASY TO SERVICE.

Brownmed warrants this product to be free from manufacturing defects. If your Intellinetix Headache Band stops working within one year from date of purchase, we will replace it.

EASY TO CARE FOR.

- Spot clean as needed.
- Lay flat and allow to air dry.
- Keep charger opening dry while cleaning.
- For adult use only.
- Do not dry clean.

Need a replacement or extra parts?
Customer service is happy to help you.
Please call us at 1-877-853-5518.

Show us how **you** vibe.
#Intellinetix

Learn more at intellinetix.com



Executive:
101 Federal Street 29th Floor
Boston, MA 02110
P 857-317-3354
F 857-317-3355

Manufacturing:
1300 Lundberg Drive West
Spirit Lake, IA 51360
P 712-336-4395
F 712-336-2874



WARNING: Lithium-ion Batteries and/or products that contain Lithium-ion Batteries can expose you to chemicals including cobalt lithium nickel oxide and nickel, which are known to the State of California to cause cancer and birth defects or other reproductive harm. For more information, go to www.P65Warnings.ca.gov.